

Anexa 72 Spălător pentru microplăci, Model: Stat Fax 2600

Numar de inregistrare AMDM: DM000437987

Specificarea tehnică deplină solicitată, Standarde de referință	Specificarea tehnică deplină oferita, Standarde de referință
<p>Spălător pentru microplăci Cod 250420 Descriere Este fabricat pentru curățarea diferitor tipuri de microplăci, complet programabil asigurând soluții de diferite concentrații, aspirație, aspirație/combinare/dozare. Parametru Specificația Canale de la 8 până la 12 Capabile să spele toate tipurile de microplăci obligatoriu Dotat cu calibrator obligatoriu Afișare LCD Programabil obligatoriu Alarmă obligatoriu, indicați lista Control automat Numărul ciclurilor de spălare, diapazon minim 1...15 Alimentare 220-240 V, 50 Hz</p>	<p>Spălător pentru microplăci -brosura Cod 250420 Descriere Este fabricat pentru curățarea diferitor tipuri de microplăci, complet programabil asigurând soluții de diferite concentrații, aspirație -brosura 2 Parametru Specificația Canale : 8 - brosură Capabile să spele toate tipurile de microplăci da - brosură Dotat cu calibrator automat da - brosură Afișare LCD – brosură Programabil da – brosură Alarmă da – manual pag 37=>42, sunt enumerate tipurile de erori, care sunt la rândul lor sunt atenționate de alarma auditive. Control automat -brosura Numărul ciclurilor de spălare diapazon 1...20 manual pag 3 Alimentare 230 V, 50-60 Hz - brosură</p>



Stat Fax[®] 2600 - Microplate Washer

Features

Washes flat, round, and v-bottom plates and strips.

No user adjustments required, automatic calibration, alignment, and last row detection.

Large, non-volatile memory for up to 50 user-entered wash protocols; Six factory-programmed wash/rinse modes.

Programmable automatic rinse cycle.

Precision operation with constant monitoring of vacuum and pressure. For extra quiet operation, internal pump cycles on only when needed.

Ready to run; 3 bottles, aerosol shield, and 8-probe manifold included.

Options & Accessories

12-probe and 16-probe manifolds.

Extra wash, rinse, and waste bottles.

2600 Plus added features:
Dual wash bottle system for added versatility and convenience.
Individual row selection for partial plate washing.
Built-in plate mixing function.

Distributed by:



Awareness Technology, Inc. PO Drawer 1679, Palm City, FL 34991 (772) 283 6540, fax (772) 283 8020
web site: <http://www.awaretech.com>, email: info@awaretech.com

Cost Effective by Design[®]

Stat Fax[®] 2600

Microplate Washer Specifications

Fluid Performance (with 8-probe head)

Residual Volume:	<3 μ l.
Dispense accuracy:	<3% CV at 300 μ l.
Processing time (full plate):	55 seconds aspirate/dispense with single or double aspirate.
Liquid Contact Materials:	Stainless steel, Delrin, silicone, nylon, PVC, polypropylene, HDPE.

Electronic

Display:	Two-line LCD, alphanumeric, 24 characters per line.
Keyboard:	16-key, domed membrane switch, 4x4, enunciating.
Power Requirements:	115V or 230V AC, 50-60Hz (switch selectable).
Microprocessor:	Z80A or Z180.
Non-volatile Memory:	Battery supported RAM, holds approx. 50 wash protocols.

Software

Washing Programs:	Aspirate, dispense, mix, soak - up to 99 minutes 99 secs.
Special Menu Options:	<i>Auto alignment</i> - senses physical geometry of the head, plate, and carrier mechanism. <i>Standby</i> - disables pumps and releases vacuum and pressure. <i>Self Check</i> - performs internal performance tests and auto alignment. <i>Auxiliary</i> - selection of well type, select auto/manual well depth and dispense depth, constant plate cycle time, viewing of pre-programmed tests, configure auto-rinse, set the time and date.

Other

Bottles:	Wash and waste 2L, rinse 1L, plastic with volume sensor probes.
Enclosure:	Fire retardant ABS plastic cover with metal base.
Dimensions: Approx.	15.5 x 13.5 x 7.5in (39 x 34 x 19cm). Weight 19 lbs. (8.6kgs).
Certifications:	NRTL Listed, CE Mark.



2600 Plus model - added features

Bottles:	Wash-1, Wash-2, and Waste (2L), Rinse (1L), plastic with volume sensor probes.
Mix function:	Programmable mix time during soak cycle.
Row selection:	User may select individual rows to wash, skipping used wells.


Stat Fax[®] 2600

Microplate Washer

Owner's Manual

Table of Contents

1. Introduction	1	3. Additional Tips and	
1.1 Warnings and Precautions	1	Information	29
1.2 Intended Use	3	3.1 Cleaning	29
1.3 Description	3	3.1.1 Rinsing	29
1.4 General Safety Summary	4	3.1.2 Cleaning the exterior	29
1.4.1 To Avoid Fire or Personal		3.1.3 Cleaning the probe head ...	29
Injury:	4	3.1.4 Disinfecting the Washer ...	30
1.4.2 Safety Terms and Symbols:	4	3.2 Maintenance	32
1.5 Installation	6	3.2.1 Adding solution / Changing	
1.5.1 Unpacking	6	bottles	32
1.5.2 Preparation	8	3.2.2 Emptying the Waste Bottle	32
1.5.3 Electrical Setup and Safety		3.2.3 Changing Probe Heads	32
Information	8	3.2.4 Exhaust filter replacement	33
1.6 Checkout Procedure	13	3.2.5 Valve tubing replacement .	33
1.7 Getting Started	15	3.2.6 Dispense volume repeatability	
1.8 Specifications	18	check	33
2. Operating Procedures	19	3.2.7 Volume calibration	34
2.1 Direct Operation	19	3.2.8 Plate Carrier Alignment	35
2.2 Auxiliary Menu/ SETUP Mode	22	3.2.9 Dispense Verification	35
2.2.1 Select well type	22	3.2.10 Pressure Adjustment	36
2.2.2 Run constant time	22	3.2.11 Storage	36
2.2.3 View a program	22	3.3 Troubleshooting/ Error Messages .	37
2.2.4 Configure auto-rinse	23	3.3.1 Wash and Rinse System ...	37
2.2.5 Set Date and Time	23	3.3.2 Aspiration System	37
2.2.6 Set Auto Depth	24	3.3.3 Error Messages	38
2.2.7 Set Dispense Depth (Top			
Wash)	25		
2.3 Programmed Operation	26		
2.3.1 Overview	26		
2.3.2 Store	26		
2.3.3 Delete program	28		
2.3.4 Run a program	28		
2.3.5 List programs	28		
2.3.6 View program	28		
2.3.7 Factory programs	28		



1. Introduction

1.1 Warnings and Precautions



WARNING

FOR IN-VITRO DIAGNOSTIC USE

Some diagnostic assays utilize material which is potentially biohazardous. Always wear protective apparel and eye protection while using this instrument. Always operate the instrument with the aerosol shield lowered.

Probe tips are sharp and may cause bodily injury.

Do not place hands or fingers under the probes while the instrument is in operation. Always set the power switch to OFF (O) before changing the probe heads. Never touch the probes while the instrument is operating.

If the waste bottle is overturned during operation, immediately set the power switch to OFF (O). If this occurs, the washer may discharge a small amount of waste material from the waste bottle via a fitting on the bottom of the unit. This material should be treated as potentially biohazardous.

If the exhaust filter becomes wet due to an overturned waste bottle, it will be blocked. Continued use of a blocked filter may impair washer effectiveness and/or result in damage to the instrument.



CAUTION

The voltage select switch setting must match the local AC line voltage or permanent damage to the instrument may occur.

Solvents such as acetone or thinner will damage the instrument. Do not use solvents to clean the unit.

Avoid abrasive cleaners. The cover, keypad and display are liquid-resistant, but are easily scratched.

The wash and rinse bottles are pressurized during normal operation. Do not remove bottle caps or tubing connections while the bottles are pressurized. Press STNDBY or

set the power switch to OFF (O) before changing bottles or tubing connections.

Particulate matter in wash solutions may clog the probe head easily. See the section on cleaning the probe head for special instructions on removing particulate matter from clogged heads.

Please take time to read this manual carefully before using the instrument. For best results, familiarize yourself with the instrument and its capabilities before attempting any clinical diagnostic tests. Refer any questions to your instrument dealer.

Retain the original packing material for future use in the event that the instrument is placed in storage, shipped to another location, or returned for service.

IMPORTANT OPERATING PRECAUTIONS!

The quality of washing often effects the validity of test results. To assure adequate washing: perform periodic dispense volume repeatability checks as described in this manual, rinse the probes after use, handle and store the probes carefully to prevent damage, use the prime cycle before each wash, watch the instrument to see that each dispense probe is functioning properly. Be sure to run a sufficient number of controls in each assay. If controls are not within their acceptable limits, or if you suspect incomplete or non-uniform washing, disregard test results.

Do not operate the instrument if the pressure is unstable or if any probe is damaged.

1.2 Intended Use

FOR IN-VITRO DIAGNOSTIC USE

The Stat-Fax[®] 2600 Microplate Washer is designed for use in processing enzyme-linked immunosorbent assays (“ELISA” or “EIA”), including clinical diagnostic assays, requiring multistep washing, rinsing, and soaking. This general purpose instrument is intended to be used by laboratory professionals who are capable of selecting the appropriate features and options for each specific clinical application.

1.3 Description

The Stat-Fax[®] 2600 Microplate Washer is a microprocessor-controlled automatic microplate washer that accepts many types of standard microplates and micro strip trays. The instrument will accept microwells having flat-, round-, or V-bottom configurations. It automatically accommodates partially filled micro strip trays. Plates may be washed in the A to H direction (8 strips of 12 wells each) or in the 1 to 12 direction (12 strips of 8 wells each) depending upon the plate orientation in the carrier and the number of wash probes on the head. Instruments are available with user changeable wash heads. The 8-probe head is standard; 12-probe and 16-probe (8 x 2) heads are available as optional accessories.

The instrument is fully programmable to provide multistep combinations of dispense, aspirate, combined aspirate/dispense, and timed soak cycles. Each user-programmed test can have up to 20 steps, and the instrument is capable of storing approximately 50 separate tests in non-volatile memory. As a convenience feature, the instrument can be programmed to automatically perform a rinse operation at a particular time of day. The instrument may also be used in the direct mode whereby wash operations are selected via the built-in keypad. A feature allowing user-assigned well depth may be also be used, if preferred to the automatic well sensing feature. For volumes larger than the capacity of the wells, or to reduce bubbles, a top-wash feature is included so the wash solution may be aspirated off the top as it is being dispensed. Instructions for all of these operations can be found in this manual.

In addition to the 8-probe combination dispense/aspirate wash head, each instrument is supplied with wash, rinse, and waste bottles utilizing integral electronic level sensors, and an aerosol shield. A stainless steel bottle rack is available as an optional accessory.

1.4 General Safety Summary

Review the following safety precautions to avoid injury and prevent damage to this product or any products connected to it. To avoid potential hazards, use this product only as specified.

Only qualified personnel should perform service procedures.



1.4.1 To Avoid Fire or Personal Injury:

- **Use Proper Power Cord.** Use only the power cord specified for this product and certified for the country of use.
- **Ground the Product.** This product is grounded through the grounding conductor of the power cord. To avoid electric shock, the grounding conductor must be connected to earth ground.
- **Observe All Terminal Ratings.** To avoid fire or shock hazard, observe all ratings and markings on the product. Consult this manual for further ratings information before making connections to the product.
- **Do Not Operate Without Covers.** Do not operate this product with covers or panels removed.
- **Use Proper Fuse.** Use only the fuse type and rating specified for this product.
- **Avoid Exposed Circuitry.** Do not touch exposed connections and components when power is present.
- **Do Not Operate With Suspected Failures.** If you suspect there is damage to this product, have it inspected by qualified service personnel.
- **Provide Proper Ventilation.** Refer to the installation instructions for details on installing the product so it has proper ventilation.
- **Do Not Operate in Wet/Damp Conditions.**
- **Do Not Operate in an Explosive Atmosphere.**
- **Keep Product Surfaces Clean and Dry.**

1.4.2 Safety Terms and Symbols:

Terms in This Manual. These terms may appear in this manual:



Warning: Warning statements identify conditions or practices that could result in injury or loss of life.



Caution: Caution statements identify conditions or practices that could result in damage to this product or other property.

Terms on the Product. These terms may appear on the product:

DANGER indicates an injury hazard immediately accessible as you read the marking.

WARNING indicates an injury hazard not immediately accessible as you read the marking.

CAUTION indicates a hazard to property including the product.

Symbols on the Product. These symbols may appear on the product:



WARNING
Risk of Shock



Protective Ground
(Earth) Terminal



CAUTION
Refer to Manual

1.5 Installation

1.5.1 Unpacking

Carefully unpack the instrument and remove it from the plastic bag. Report any visible damage to your freight carrier at once.

NOTE: Retain the original packing material for future use in the event that the instrument is placed in storage, shipped to another location, or returned for service.

You should find the following items packed with the instrument. Please locate each item before proceeding. Study the list below and the diagram (Figure 1) on the next page to become familiar with the items. Remove all packing material and retain.

<u>Included items</u>	<u>Description</u>
Aerosol shield	clear acrylic hinged cover, mounted on the instrument
8-probe wash head	white plastic with 8 pairs of metal tubes (shown installed in Figure 1)
Power cable	heavy black straight cord
Plate carrier	large square metal plate with holes
Bottle cap assembly	3 bottle caps, with tubing, sensor leads, and tube keeper connected
Bottles	3 bottles, labeled for wash, rinse, and waste
Tool kit	plastic bag with small tools
Owner's Manual	this document


Contact your dealer immediately if anything is missing.

Optional Accessories:

12-probe head: white plastic with 12 pairs of stainless tubes

16-probe head: white plastic with 16 pairs of stainless tubes (two rows of 8)

Bottle rack: stainless steel rack for holding wash, rinse, and waste bottles



*Figure 1. Stat-Fax[®] 2600 Microplate Washer with included items
(note that microplate shown in Instrument picture is not included)*

1.5.2 Preparation

Complete this procedure to prepare the washer for operation.

1. Place the instrument on a flat working surface, at least 24 inches (61cm) deep, capable of safely supporting the weight of the instrument and filled bottles (approximately 25 lb., 11.4 kg). A clearance of at least 3 inches around the instrument is required to assure optimal ventilation. It is recommended that the instrument be operated within an ambient temperature range of 15-35°C and humidity between 10 and 85%.
2. If the probe head has been shipped separately, install it as discussed in the section entitled *Maintenance: Changing Probe Heads*.
3. Refer to Figure 2, Rear Panel Connections, and Figure 3, Parts and Controls, for the remainder of this procedure. Place the bottles on the work surface behind the instrument. The order of the bottles, left to right, viewed from the front of the instrument, is WASH, RINSE, WASTE.
4. Remove the bottle plumbing (caps and tubing) and leads from the protective plastic bag. Locate each cap to its corresponding bottle. Hold the cap steady with one hand while turning the bottle with the other. The tubing should exit toward the front, and should not be kinked, twisted, or strained. Tighten the bottle caps firmly. Do not place undue stress on the bottle cap tubing connections or sensor lead connectors.
5. Cardboard valve supports have been included with your instrument to prevent the tubing from sticking during transport and storage. These supports must be removed. To access, open the valve door (rear panel) using a thumbscrew. On the inside of the door you will see the two valves, and the cardboard supports holding the valves open. Remove the supports and retain for future storage and/or shipment. Reclose the valve door.
6. Connect the tubes to the rear panel fittings. The tubes and the fitting are color-coded. In addition, all rear panel connections are shown on a color graphic overlay affixed to the rear panel.
7. Plug the sensor leads into the jacks on the rear panel. The sensor leads are color-coded.
8. Locate the power switch on the rear panel. Check that the power switch is in the **Off (O)** position.

1.5.3 Electrical Setup and Safety Information

Note: See section 1.4 for additional Safety Information.

1. Assure Proper Grounding: The safety classification of this instrument is Class 1.



WARNING: To avoid the risk of electric shock, the third prong of the AC power plug must be grounded at the main socket-outlet and connected to conductive parts internal to the equipment.

The internal connection is provided for when manufactured by means of internal toothed solder lugs and stainless steel screws and nuts, or metal contacts, tooth lock washers, terminals and rivets. The location of the protective grounding conductor terminal is marked internally by the IEC 417 symbol 5019 (see section “Safety Terms and Symbols”). DO NOT loosen or remove these screws, rivets or contacts. Do not defeat the safety purpose of the grounding plug. It is the operator’s responsibility to ensure the mains supply is properly grounded at the power outlet. If there is doubt that it is not properly grounded, contact a qualified electrician.

2. Assure Proper Power Availability and Voltage Select Switch Setting: the power requirements can be found on the specifications page.



CAUTION: The voltage select switch setting must match the local AC line voltage or permanent damage to the instrument may occur. The voltage select switch must be set for the appropriate input voltage prior to powering up.

Locate the voltage select switch on the rear panel. This is a 2 position slide switch that will configure the instrument to accept either 230V or 115V input. Do not connect the instrument to the power supply before assuring the proper position of the line voltage selection switch.

When you can see the 230V label, the instrument is set for 230V input. If you plug the instrument into a 115V power supply while 230V is selected, the instrument will have insufficient operating power.

To select 115V input, insert a straight screw driver blade (or similar instrument) into the slot on the switch, and slide it into its alternate position. Upon sliding the switch, you will see the 115V label appear.



CAUTION: If the instrument is configured to accept 115V and you plug it in to a 230V power supply, the fuses will blow and permanent damage to the electronics may result.

3. Assure Proper Power Cord Selection: use only the power cord specified for this product and certified for the country of use.

For units for use at 110-120 V inside the US: Use a listed cord set consisting of a minimum 18 AWG, Type SVT or SJT three conductor cord, maximum 3 meter (9.8 feet) in length, rated 10 A, 125 V, with a parallel blade, grounding type attachment plug. The cord set provided by the manufacturer meets these requirements.

For units for use at 220-240 V inside the US: Use a listed cord as above, except rated 250 V, with a tandem blade, grounding type attachment plug.

Connect the supplied power cable to the rear of the instrument as shown. Plug the other end of the power cable into an AC outlet.

4. Install the plate carrier in the plate bed. Refer to Figure 3, Parts and Controls. Slide the two metal pins into the two slots at the back of the plate bed, and lower the plate carrier into the plate bed. Do not

use excessive force on the plate carrier.

5. Assure Clean Power Availability: the circuit used should be substantially free of large voltage transients (Kilovolt amp loads) such as large pumps, large centrifuges, refrigerators and freezers, air conditioners, large autoclaves, ovens, and dryers. The instrument may fail to operate normally if the power supply is interrupted. If this occurs, turn the instrument off for several seconds. When you turn the instrument back on, it will resume normal operation, but information not stored in non-volatile memory will be lost.
6. Fuses: the fuses are located internally in the instrument; there are two fuses, fusing both sides of the main power supply. Fuse failure is a very rare occurrence and should indicate malfunction of the equipment requiring service by qualified personnel.

The fuses used with this instrument are 1/2 Amp, T rating (slow blow) 250 V. Cartridge size is 3AG or size '0', dimensions 1/4 x 1-1/4" (6.3 x 32 mm). For continued protection against risk of fire, use the same fuse for either 115 or 230 V line voltage selection. Disconnect power cord from mains supply before replacing fuses.

Do not attempt to make repairs or adjustments to the circuitry. Do not install any non-specified replacement parts. Awareness Technology will supply all service and accessories. Consult your dealer to make arrangements. Use of a fuse of the improper rating may constitute a fire hazard.




Figure 2. Rear panel connections




Figure 3. Parts and Controls




Figure 4. Keypad Layout

1.6 Checkout Procedure

Visually confirm the following:

- Each tube connected to the corresponding color-coded fitting and all sensor leads are plugged in.
- All tubing fully seated on fittings.
- Plate carrier installed to plate bed.
- Wash bottle filled with wash solution
- Rinse bottle filled with deionized water or other rinse solution.
- Waste bottle empty.
- All bottle caps tight.
- Power cable plugged into rear of unit and into AC outlet.
- Power switch set to OFF (O).
- Correct wash head (8, 12, or 16 probes) is installed.

The instrument is now ready for power-up.

- Turn on the power switch at the left rear of the instrument. The display will show:

```
Stat-Fax 2600      :V
```

The letter at the right side of the top line indicates the software revision (“V” in the example shown).

The pump will continue to run until the pressure stabilizes and vacuum is checked. After a few seconds, the display will show the **main prompt**:

```
Press <YES> to run prog  
Va OFF Pr 5.0 09:42:00
```

The message for “Va” indicates whether the vacuum is off or on; the value shown for “Pr” indicates pressure in pounds per square inch. During operation, the pump will cycle on and off automatically to maintain pressure within range. When the instrument is not performing wash functions, the pump should not run continuously for more than 60 seconds nor cycle on and off more often than once in 60 seconds. If the pump runs or cycles on and off excessively, check for leaks in the system. Check that the bottle caps are tight and the tubing connections are correctly located and secure.

- Place a microplate in the plate carrier. With an 8- or 16-probe head, orient the plate so that well number A-1 is at the left rear corner. With a 12-probe head, orient the plate so that well number A-1 is at the left front corner. Be sure that the plate is securely seated in the grooves on the plate carrier.
- Press **ALIGN**. The instrument will lower the head to the reservoir, then return the probe to the up position. The instrument will then check the orientation of the plate and measure the depth of the plate wells. When the instrument has completed this process, the probe and plate will return to home positions.
- Press **PRIME**. The instrument will alternately dispense wash solution into the reservoir and aspirate, which serves to prime the wash solution dispense system.

- If the display shows:

```
Rinse Bottle is Low  xxx  
                    or  
Wash Bottle is Low  xxx
```

Press **NO** twice to return to the main prompt. Check that the sensor is plugged into the proper receptacle (see **SET UP** section). Press **STNDBY** to disable the pumps and release pressure. Wait at least 15 seconds before removing caps from bottles. Add the appropriate solutions to the bottles and replace the caps. Tighten the caps securely to prevent pressure loss. Press **YES** to exit **STNDBY** mode and return to the main prompt.

Note: The 3 digit number which appears after the message is a direct reading of the sensor parameter, and is used for troubleshooting.

- If the instrument produces results other than those described here, set the power switch to OFF (O). Go to the section entitled “Setup” and review all steps carefully. Repeat the Checkout procedure. If the instrument still performs erratically, contact your dealer for assistance.

1.7 Getting Started

This section gives you brief instructions so that you may begin working with the instrument right away. For a closer look at the instrument's capabilities, refer to the other sections of the manual. Follow the steps below to perform wash operations on a microplate. Refer to the assay package insert for information on number of washes required, wash solution formulae, and other details.

Before continuing, be sure the instrument is prepared for use by completing the procedures in the sections *1.5 - Installation* and *1.6 - Checkout Procedure*.

Note: You may cancel any operation at any time and return to the main prompt by pressing **NO** twice in succession.

1.7.1 Fill the wash and rinse bottles

- Press **STNDBY**. The instrument will release the pressure from the bottles. Wait at least 15 seconds after pressing **STNDBY** for the pressure to fully release.
- Unscrew the wash bottle cap and fill the bottle with wash solution. Replace the wash bottle cap. Repeat with rinse bottle and rinse solution. Tighten all bottle caps firmly.
- Place a microplate in the plate carrier. With an 8- or 16-probe head, orient the plate so that well number A-1 is at the left rear corner. With a 12-probe head, orient the plate so that well number A-1 is at the left front corner. Be sure that the plate is securely seated in the grooves on the plate carrier.
- Press **YES** to exit the **STNDBY** mode and resume normal operation.

1.7.2 Aspirate

- Press **ASP**. The display shows:

Aspirate	
Double Aspirate	Y/N

Note 1: If round-bottom or V-bottom wells are selected, the double aspirate option will not be displayed.

Note 2: The instrument is preset for automatic well-depth detection. If you prefer to set the well-depth manually, see section *2.2.6 - Set Auto Depth*

The symbol at the right of the top line indicates the flat-bottom well type. If your microplate contains other than flat-bottom wells, you must first select the type of well. Refer to the section *2.1 - Direct Operation* for more information.

Double aspirate means that, for each strip, the instrument will first aspirate, then reposition the head and aspirate again. This produces the exceptionally low residual volume. Double aspiration is highly recommended on the initial and final aspirates.

- Press **YES** to use double aspiration, or press **NO** to use single aspiration. The single aspiration leaves more residual fluid in the plate, but is much faster. The display shows:

```
Aspirate           
Insert plate  --> <YES>
```

- Press **YES** and the instrument aspirates the contents of the plate.
- When the instrument has completed the plate, it will return the plate to the right side of the plate bed. You may inspect the wells for residual wash solution at this time.

1.7.3 Dispense

- Press **PRIME**. The instrument will alternately dispense wash solution into the reservoir and aspirate, to prime the dispense system. Repeat.
- Press **DISP**. The display shows:

```
Enter desired vol(μL):
<YES> =                300
```

Note: The instrument is preset for automatic dispense-depth. If you wish to use the top wash feature, see section 2.2.7 - *Set Dispense Depth*

- Press **YES** to accept the displayed volume, or enter a new dispense volume and then press **YES**. Enter a volume between 25 μL and the total well volume (350 μL for flat-bottom well), or up to 500 μL if you are using the top wash feature.

```
Dispense           
Insert plate  --> <YES>
```

- Press **YES** to dispense wash solution into all wells. When the dispense is complete, the plate will return to the right side of the plate bed.

1.7.4 Aspirate/Dispense

- Press **PRIME**. The instrument will alternately dispense wash solution into the reservoir and aspirate, to prime the dispense system. Repeat.
- Press **A/D**. The display shows:

```
Aspirate           
Double Aspirate   Y/N
```

Note: If round-bottom or V-bottom wells are selected, the double aspirate option will not be displayed.

The symbol at the right of the top line indicates the flat-bottom well type. If your microplate contains other than flat-bottom wells, you must first select the type of well. Refer to the section “**Direct Operation**” for more information.

Double aspirate means that, for each strip, the instrument will first aspirate, then reposition the head and aspirate again. This produces the exceptionally low residual volume. Double aspiration is highly recommended on the initial and final aspirates.

- Press **YES** to use double aspiration, or press **NO** to use single aspiration. The single aspiration leaves more residual fluid in the plate, but is much faster. The display shows:

```
Enter desired vol (µL)
<YES> =          300
```

- Press **YES** to accept the displayed volume, or enter a new dispense volume and then press **YES**. Enter a volume between 25 µL and the total well volume (350µL for flat-bottom well), or up to 500µL if you are using the top wash feature.

```
Dispense          
Insert plate  --> <YES>
```

- Press **YES** to aspirate and dispense wash solution into all wells, with each row of wells being aspirated and dispensed before the washer proceeds to the next row. When the aspirate/dispense cycle is complete, the plate will return to the right side of the plate bed.

1.7.5 Repeat as needed

- Repeat the **ASP**, **DISP**, or **A/D** selections described above, as recommended in your assay literature. You may use a single aspirate for intermediate washes. However, on the final aspirate, be sure to select double aspiration to ensure low residual wash solution.

1.8 Specifications

Physical:

8/12 probe microplate transport accepts flat, round, or V-bottom wells, plates & strip trays
 8-probe aspirate and dispense head (12- or 16-probe heads optional)
 Stainless steel plate bed
 Wash, rinse, and waste bottles with electronic level sensing on all bottles
 Transparent acrylic aerosol shield
 Dimensions 13.5" W x 16" D x 7.5" H (34.3 cm W x 40.6 cm D x 19 cm H)
 Bottle assembly 13.5" W x 5.5" D x 10.5" H (34.3 cm W x 14 cm D x 26.7 cm H)
 Weight 22 lb. (10 kg)

Electronic:

Z80 microprocessor, 2.0 MHz clock
 8K bytes non-volatile RAM
 32K bytes EPROM
 24 character x 2 line liquid crystal display (LCD)
 4 x 4 keypad

Power Source:

Switch selectable power supply (115V or 230V indicated)
 Voltage Source 110-120V / 220-240V from 50 to 60 Hz, CAT II
 Power Consumption less than 70 Watts
 Fuse 1/2 A, T rating, 250 V
 All power cords must be approved for the country of use.

Environmental Conditions for safe operation:

indoor use
 altitude up to 2000 m
 temperature 5° C to 40° C *
 humidity 80% for temperatures up to 31° C decreasing linearly to 50% humidity at 40° C
 mains supply voltage fluctuations not to exceed $\pm 10\%$ of the nominal voltage

Certifications and Compliances:

Intertek Testing Services, Listing to Standard UL 3101-1 for electrical equipment for laboratory use.
 CE Certificate of Conformity. Conforms with the following standards: EN 50082-1, EN 55011,
 EN 61010-1 following the provisions of the 73/23/EEC and 89/336/EEC.

Performance:

Residual (double aspiration)	$\leq 3 \mu\text{l}$ per well
Dispense precision	3% C.V. across 96 wells
Dispense accuracy	Mean volume $\pm 3\%$ from reference
Number of stored wash programs	Approximately 50

* Although it may be safe to operate in these conditions, it may not be suitable for the performance of your tests. Check with your supplier.

Design and specifications are subject to change without notice.

2. Operating Procedures

2.1 Direct Operation

Direct operation allows you to control the instrument from the keypad. You can select aspirate and dispense operations, and you can set various stored parameters, such as the auto-rinse time and well type.

For convenience, you can store wash operations as a program and simply run the program when needed. Refer to section 2.3 - *Programmed Operation* for more information.

The direct operation selections are described below.

<u>Keypad</u>	<u>Description</u>
YES, NO	User input The YES and NO keys are used in response to displayed messages. The YES key usually serves to select an option or to enter a value. The NO key serves to bypass a selection or to clear the entered value. Pressing the NO key twice during any wash operation will end the wash operation and return to the main prompt.
0-9	Numeric keys The numeric keys are used to directly enter numeric data when required. The numeric keys are also labeled with the direct operations described below.
PRIME	Prime system with wash solution PRIME alternately dispenses wash solution into the reservoir and aspirates. This serves to prime the system with wash solution. Always PRIME after filling the wash bottle and after changing the head.
RINSE	Flush system with rinse solution RINSE flushes the system with rinse solution by dispensing rinse solution into the reservoir while aspirating. Always perform a RINSE at the end of a usage period to prevent wash solution from drying and crystallizing in the dispense tubes.
DISP	Dispense plate DISP dispenses a selected amount of wash solution. The display shows: <pre>Enter desired vol (µL) <YES> = 300</pre> <p>Press YES to accept the displayed volume, or enter a new dispense volume and then press YES. You must enter a volume between 25 µL and the total well volume (350 µL for flat-bottom well), or up to 500µL if you are using the top wash feature. The instrument automatically senses the orientation of the plate and detects the presence of individual strips, unless you have selected manual depth mode..</p> <p>When the display shows:</p> <pre>Dispense  Insert plate --> <YES></pre> <p>press YES to perform the operation. The instrument will show its progress across the plate graphically by displaying icons.</p>

ASP Aspirate plate
ASP aspirates the contents of the wells into the waste bottle. If the current well type is flat-bottom, the display shows:

```
Aspirate           
Double Aspirate   Y/N
```

A double aspiration aspirates the well, then repositions the tube and aspirates again. It is very effective at drying the well, but is somewhat slower than single aspiration. Press **YES** to double aspirate, or press **NO** to single aspirate. Note that double aspiration is only available with flat-bottom wells. The instrument automatically senses the orientation of the plate and detects the presence of individual strips, unless you have selected manual depth mode. The display shows:

```
Aspirate           
Insert plate   --> <YES>
```

Press **YES** to perform the operation. The instrument shows its progress across the plate graphically with well icons.

A/D Combined aspirate / dispense
A/D performs a combined aspirate and dispense in one pass, aspirating and dispensing each row before proceeding to the next. Operation is identical to **ASP** and **DISP**, described above.

ALIGN Align mechanism
ALIGN causes the instrument to sense the position of the head, plate, and carrier mechanisms. Always perform an **ALIGN** after changing heads, the well type, or the plate type.

STORE Store user programs
STORE allows you to create multistep programs in memory. See the section entitled “**Programmed Operation**” for more information.

DELETE Delete user programs
DELETE allows you to remove stored programs from memory. See the section entitled “**Programmed Operation**” for more information.

STNDBY Enter standby mode
STNDBY disables the pump and releases the pressure and vacuum on the system. This allows you to change bottles or heads, without running the pumps. Press **YES to continue operation.**

NUM Select number of strips
When the display shows “Insert plate --<YES>”, in response to an aspirate, dispense, or aspirate/dispense selection, you may press NUM to enter the number of strips installed in the holder. The instrument will not attempt to auto-detect the presence of strips if this feature is selected.

8/12 Selects method of plate detection
Press 8/12. The display shows:

```
Select Wash Head Mode
8 position head   Y/N
(if no is selected, then...)
12 position head  Y/N
(if no is selected, then...)
16 pos (double row)Y/N
```

At the beginning of each program or direct operation. the probe head goes directly to the first row of the selected type of plate. Be sure that the “8/12” key setting is properly selected when changing probe heads to avoid damaging the aspiration tubes.

ALT	Aspirate reservoir Pressing ALT aspirates any fluid remaining in the reservoir.
SELFCK	Self-Check SELFCK performs internal tests. The instrument will display error messages in the event that errors are detected.
AUX	Auxiliary menu/SETUP mode The AUX menu allows access to various stored parameters and options described in the next section.

2.2 Auxiliary Menu/ SETUP Mode

Press **AUX** to enter the SETUP Mode.


2.2.1 Select well type

The display shows:

```
Setup Mode
Select well type Y/N
```

Press **NO** to continue to the next selection, or press **YES** to select the well type. You may select from flat-, V-, or U-bottom wells. If you press **YES**, the display shows the current well type. Press **NO** to select different well types, then press **YES** to make the displayed well type the current well type.

The well icons are:



After you have selected a well type, the display shows:

```
Mech must be realigned
Insert plate --> <YES>
```

Install a plate with the new well type and press **YES**. The instrument will realign and return to the main prompt.

2.2.2 Run constant time

The display shows:

```
Setup Mode
Set Const Time optY/N
(if yes is selected, then...)
Run Constant Time Y/N
```

Press **NO** to continue to the next selection, or press **YES** to enable constant run times. The constant time option allows uniform wash times for all strips in a partially filled tray. The instrument operates as it would with a full strip tray, except that it detects the empty rows and does not dispense into them. The last step of the program aspirates or dispenses only the filled rows.

2.2.3 View a program

The display shows:

```
Setup Mode
View a program Y/N
```

Press **NO** to continue to the next selection, or press **YES** to view the steps of a previously stored

program. If you press **YES**, the display shows:

```
View a program
Ent prog number --> <YES>
```

Enter the program number and press **YES**. The name of the program, along with the programmed well type, is briefly displayed. The display then shows information similar to the following:

```
02-01 Dbl Asp/Disp 300µL
Press <YES> for next step
```

The top line of the display shows the program number, the step number, and a description of the operation for that step. Press **YES** to view the next program step. If there are no more program steps, or you press **NO**, the display reverts to the main prompt.

2.2.4 Configure auto-rinse

The display shows:

```
Setup Mode
Set Auto Rinse Y/N
```

Press **NO** to continue to the next selection or press **YES** to configure the auto-rinse feature. When auto-rinse is enabled, the instrument automatically performs a **RINSE** operation at a specified time each day, then goes into standby mode. If the washer is in operation at the specified time, the auto-rinse time will be automatically advanced one hour. Once the auto-rinse has been completed, the auto-rinse time will be reset to the previously specified time for the next day.

If you press **YES**, the display shows:

```
Setup Mode
Turn Auto Rinse ON Y/N
```

Press **YES** to enable auto-rinse or **NO** to disable. If you press **YES**, the display will show the currently programmed auto-rinse time:

```
Setup Mode
Auto Rinse at 12:00:00
```

Press **YES** to accept the displayed time, or enter the desired auto-rinse time using 24-hour format HH.MM.SS and press **YES**. Using 24-hour format, 1:00 PM is entered as 13:00:00.

2.2.5 Set Date and Time

The display shows:

```
Setup Mode
Set Date/Time Y/N
```

Press **NO** to return to the main prompt, or press **YES** and the display shows:

```
Setup Mode
Enter Date as MM.DD.YY
```

Enter the current date using MM.DD.YY as the format and press **YES**. The display shows:

```
Setup Mode
Enter Time as HH.MM.SS
```

Enter the correct time using 24-hour format HH.MM.SS and press **YES**. Using 24-hour format, 1:00 PM is entered as 13:00:00.

2.2.6 Set Auto Depth

The display shows:

```
Setup Mode
Set Auto Depth    Y/N
(if yes is selected, then...)
Set Auto Depth
0=Automatic  1=Manual
```

Press **0** to select Automatic well depth selection, or press **1** to select Manual well depth selection. Partial plate auto-detection is disabled when you are in Manual well depth mode, so you are automatically prompted to set your strip number when washing plates.

If you have not yet set your manual plate depth parameter, you should do so. The value is retained when you switch between manual and automatic depth selection, so you need only perform this procedure once. The default value is 40 “counts” from the probe home position.

To set your manual plate depth parameter, select test #216. The display shows:

```
Enter Well Depth Y/N
```

If you have set this value before and are either re-entering it because of a memory failure or a software change, you may press **YES** to enter the same value as before. Otherwise, install a plate in the plate bed and press **NO**. The instrument will position the probe head with the tips in the first row of wells. The display shows:

```
Set Well Depth: YES=Done
0=+1, .=+5, 1=-1, 2=-5
```

To make the probe go deeper, press the **0** or **.** key. To make the probe go higher, press the **1** or **2** key. Press **YES** when the probe tips are where you desire the depth to be set. Auto-detection for partial plates is disabled when in manual mode, so use **NUM** to select the number of strips being washed in a partial plate. If you set the depth so the probes are not touching the bottom of the wells, you may notice an increased amount of residual liquid after aspirating, and single aspiration may leave less residual liquid than double aspiration.

2.2.7 Set Dispense Depth (Top Wash)

The display shows:

```
Setup Mode
Set Disp Depth    Y/N
(if yes is selected, then...)
Set Disp Depth
0=Automatic  1=Manual
```

Press **0** to select Automatic dispense depth selection, or press **1** to select Manual dispense depth selection.

If you have not yet set your manual dispense depth parameter, you should do so. The value is retained when you switch between manual and automatic depth selection, so you need only perform this procedure once. The default value is 1 “count” from the probe home position.

To set your manual plate depth parameter, select test #217. The display shows:

```
Enter Disp Depth Y/N
```

If you have set this value before and are either re-entering it because of a memory failure or a software change, you may press **YES** to enter the same value as before. Otherwise, install a plate in the plate bed and press **NO**. The instrument will position the probe head with the tips in the first row of wells. The display shows:

```
Set Disp Depth: YES=Done
0=+1, .=+5, 1=-1, 2=-5
```

To make the probe go deeper, press the **0** or **.** key. To make the probe go higher, press the **1** or **2** key. Press **YES** when the probe tips are where you desire the depth to be set - it should be so the aspirate tubes are below the level of the top of the wells (see figure below).



Test #217 Disp Depth Suggested Setting

2.3 Programmed Operation

2.3.1 Overview

You can program the instrument to perform multistep dispense, aspirate, combined aspirate/dispense, and soak operations in any order. You can store up to 20 steps in each program, and you can store approximately 50 programs in the instrument's non-volatile (battery-backed) memory. You can give each program a descriptive name; however, programs are recalled by number. You can list the stored programs, and you can view the steps of each program.

2.3.2 Store

1. From the main prompt, press **STORE** to enter the programming mode. The display shows:

```
Store a new program
Use  Well type Y/N
```

Press **YES** to use the displayed well type, or press **NO** to select a different well type. If you press **NO**, the display shows the available well types. Press **NO** to display the next well type, or press **YES** to select the displayed well type and continue.

2. The display shows:

```
Sel action for step: 1
Aspirate plate      Y/N
```

Press **NO** to select a different action, or press **YES** to select the displayed action and continue. Programmable actions are described below. Some programmable actions require additional input as noted below. You must select an action for the first step.

3. Repeat step 2 as needed to build the program. The display shows the current step number and the available actions.
4. Select the action "FINISHED PROGRAMMING" to save the program. The program is now stored in memory and the display shows:

```
Saved as program      7
Run Constant Time    Y/N
```

Press **YES** to run the test using the constant run time feature described in the section 2.2 - *Auxiliary Menu*. Otherwise press **NO**.

5. The display shows:

```
Saved as program      7
Name the program      Y/N
```

Note the number of the program in the upper right corner of the display. You should keep a written record of the program number and the actions programmed. Press **YES** to name the program, or press **NO** to return to the main prompt. If you select **YES** to name the program, the display shows:

```
ABCDEFGHIJKLMN OPQRSTUVWXYZ
<4 6> YES=Let STORE=Done
```

This allows you to build a 12-character name on the bottom line by selecting letters from the top line. The numeric keys **4** and **6** move the cursor left and right through the character set. The display will scroll as needed to display more letters. Press **YES** to select the letter under the cursor and add it to the program name. Press **NO** to remove the last letter added. When you have finished adding characters to the program name, press **STORE** to save the program name, and return to the main prompt.

Description of Programmable Actions

Repeat Last Step

Select this option to repeat the last step you programmed. This is not available on the first step.

Aspirate

Aspirate the contents of all wells. If the selected well type is flat bottom, the display shows “DOUBLE ASPIRATION Y/N”. Press **YES** for double aspirate, or press **NO** for single aspirate. Double aspirate is recommended for the first and last aspirate in the wash program.

Dispense

Dispense wash solution into plate wells. The display shows:

```
Enter desired vol (µL):
```

and the most recently entered dispense volume. Press **YES** to accept the displayed value or enter a new dispense volume and press **YES**.

Aspirate & Dispense

This performs a combined aspirate and dispense in one pass. Operation is identical to **ASPIRATE** and **DISPENSE**, as described in the section 2.1 - *Direct Operation*.

Soak

Pause for a selected amount of time, then continue to the next step. Enter the minutes and seconds separated by a decimal point, then press **YES**.

Pause for keystroke

Pause until a key is pressed. This is useful for tests in which multiple washes are performed. When this step is encountered during the program, the program will pause and display

```
Program XX in progress
Press a key to continue
```

When a key is pressed, the program resumes at the next step.

Finished Programming

Save the steps you have entered so far as a numbered program.

2.3.3 Delete program

Press **DELETE** to remove a program from memory. Enter the program number and press **YES**. The display shows:

```
Delete USER PROGRAM Y/N (or test name)
```

Press **YES** to delete, or press **NO** twice to cancel and return to the main prompt.

2.3.4 Run a program

From the main prompt, press **YES** and then enter the number of the program you wish to run. Press **YES** to begin running the program.

2.3.5 List programs

Run program 99 to view a list of stored programs. You will be shown the name of each program; if you have declined to name a program, it will be shown as “User Program”. If there are no programs in memory, running program 99 will return to the main prompt.

2.3.6 View program

To view the steps of a program, press **AUX**. Press **NO** until the display shows:

```
View a program Y/N
```

Press **YES**. Enter the program number and press **YES**. The name of the program is displayed briefly, along with the programmed well type. The display then shows information similar to the following:

```
02-01 Dbl Asp/Disp 300µL
Press <YES> for nxt step
```

The top line of the display shows the program number, the step number, and a description of the operation for that step. Press **YES** to view the next program step. If there are no more steps, the display reverts to the main prompt.

2.3.7 Factory programs

Several factory programs are included at program number 1 through 6 to enable you to begin using the instrument in programmed mode. They are permanent and cannot be erased.

The factory programs are:

<u>No.</u>	<u>Name</u>	<u>Description</u>
1	3x@300 Const	Wash 3 times with 300 µL/constant time
2	4x@300 Const	Wash 4 times with 300 µL/constant time
3	5x@300 Const	Wash 5 times with 300 µL/constant time
4	3x@300 Wash	Wash 3 times with 300 µL
5	4x@300 Wash	Wash 4 times with 300 µL
6	5x@300 Wash	Wash 5 times with 300 µL

3. Additional Tips and Information

3.1 Cleaning

3.1.1 Rinsing

A **RINSE** should be performed when the instrument will not be used for an extended period, e.g. overnight, at end of shift, etc. Proper rinsing will help to prevent clogging of the dispense tubes (see below). A rinse should also be performed when the wash solution is changed, to prevent contamination. The auto-rinse feature may be programmed to automatically perform a **RINSE** at a predetermined time of day. It is recommended that the rinse bottle be filled with deionized water. If the instrument is inactive for approximately 15 minutes, it will go into an idle mode. This will lower the probe head into the reservoir and release the vacuum and pressure. To allow for cases where the probe heads have not been rinsed prior to inactivity, a small amount of liquid will be dispensed into the reservoir, in order to keep the probe heads moist (to avoid crystallization of wash solution).

3.1.2 Cleaning the exterior



CAUTION: Solvents such as acetone or thinner will damage the instrument! Use only water!

Avoid any abrasive cleaners. The keypad and display areas are liquid-resistant, but are easily scratched.

The exterior of the instrument as well as the stainless steel plate bed may be cleaned with a soft cloth using plain water. If needed, a mild all-purpose or non-abrasive cleaner may be used. A 1.5% solution of chlorine bleach or 70% isopropyl alcohol may be safely used as a disinfectant. Take special care not to spill excessive liquid into the plate bed area.

3.1.3 Cleaning the probe head

The supplied tool kit contains gauged wires, used to clean the aspirate and dispense tubes. The dispense tubes may become clogged due to suspended particulate matter in the wash solution. Always use fresh wash solution and inspect frequently. The dispense tubes as well as the aspirate tubes can clog due to crystallization of the wash solution. To prevent this, always perform a **RINSE** operation after use.

To clean the probe head, put it in **STANDBY** mode and remove the probe head (see the section *Maintenance: Changing Probe Heads*). Use the gauged wires for the first step of cleaning the head. When using the gauged wires to clean the dispense tubes, be especially gentle. Do not scrub the wire vigorously up and down or place any horizontal pressure against the tube. Excessive or indelicate use of the cleaning wire can cause abrasion of the inside of the dispense tubes and may adversely affect dispense precision. After using the gauged wires, remove the large plastic screws on the probe head: the lower screws are for the dispense probes and the upper screws are for the aspirate probes. Run deionized water through the openings to flush out any particulate matter loosened by use of the gauged wires. Replace the head as described in the section *Maintenance: Changing Probe Heads*.

3.1.4 Disinfecting the Washer

Use all standard laboratory safety precautions when working with known or suspected biological and chemical hazards. The following procedure is recommended by the manufacturer for disinfecting the microplate washer.

- Disinfecting the waste bottle

Enter the Standby mode and wait 15 seconds, or turn off the washer, before disconnecting any bottle.

Holding the cap of the waste bottle still, unscrew the waste bottle and discard the waste material in an appropriate manner (See section 3.2.2 - *Emptying the Waste Bottle*).

Prepare a fresh 10% solution of bleach containing Sodium Hypochlorite. Fill the bottle to one-third capacity with this solution. Cap tightly. Then swirl the solution in the bottle to rinse it. Discard the rinse solution appropriately and repeat this step. Rinse the bottle at least 3 times with fresh water to remove the bleach if desired. Alternately a small amount of undiluted bleach may be retained in the bottle.

Replace the waste bottle on its cap and either turn the instrument on, or press the YES key to resume vacuum.

Other bottles can be disinfected similarly, however, complete rinsing is required afterwards.

- Disinfecting the probes and tubing

Prepare a fresh 10% solution of bleach containing Sodium Hypochlorite. Alternately, you may use any of the standard disinfectants used for laboratory decontamination. These will not harm the washer as long as they are completely rinsed out immediately after use. Many disinfectants can cause drying of the tubing over time. This will lead to premature aging and more frequent need for tubing replacement. Therefore, disinfectant should be rinsed out immediately after use.

Put the instrument in the Standby mode. Wait at least 15 seconds before disconnecting the rinse bottle. Then replace the rinse with the disinfecting solution. Reassemble and operate the rinse cycle at least two times to flush out the tubing.

Put the instrument in the Standby mode again. Wait at least 15 seconds. Then replace the disinfectant with a bottle of fresh deionized water. Reassemble and operate the rinse cycle 3-5 times. This should clear the disinfectant out of the tubing.

- Disinfecting the instrument surfaces

Prepare a fresh 10% solution of bleach containing Sodium Hypochlorite. Alternately, you may use any of the standard disinfectants used for laboratory decontamination. To be sure that these will not harm the cosmetic finish of the instrument, experiment first with a small section at the back of the washer. It is highly recommended that the disinfected surfaces be wiped with fresh water and dried immediately after disinfecting. This will avoid leaving a chemical residue that may cause premature aging of the instrument cover. The front panel is manufactured from chemically resistant stainless steel, and is unlikely to be damaged by any standard cleaners.

Turn the instrument off.

Remove the probe head (which may also be disinfected and rinsed). See instructions for cleaning the head in this manual.

Remove the plate carrier (which may also be disinfected and rinsed).

Use a soft cloth to wipe the surfaces with disinfectant. Do not pour liquid on the instrument or into the plate bed area. Do not allow liquid to soak up under the keyboard overlay.

Next, wipe the surfaces with a cloth wetted with fresh water to remove all of the chemical residue. Dry the instrument.

70% isopropanol can be used to remove dirt and fingerprints from the keypad overlay. It may also be used to remove ink marks or other stains that remain after cleaning with water. This may be left to evaporate and will not harm the instrument surfaces.

3.2 Maintenance

3.2.1 Adding solution / Changing bottles

Press **STNDBY** to disable the pumps and release pressure and vacuum. Wait at least 15 seconds before removing caps from bottles. Transfer solution and replace caps. Tighten caps securely to prevent pressure loss. Press **YES** to exit **STNDBY** mode and return to the main prompt.

3.2.2 Emptying the Waste Bottle

If the display shows:

```
Empty the waste Bottle  
press <YES> to resume
```

If you have the optional bottle rack, loosen the velcro strap on the waste bottle. Carefully unscrew the bottle from the cap. Empty the waste bottle and replace the cap. Press **YES** to continue.



WARNING: Use adequate precautions when disposing of waste. Eye protection, gloves, and proper clothing are indicated. Waste material should be treated as potentially biohazardous.

3.2.3 Changing Probe Heads

The head may be changed at any time other than during a wash operation. To change the head, first set the power switch to OFF (O). Remove the plastic tubing connected to the fittings on the side of the head. Remove the two knurled thumbscrews holding the head to the probe arm. Place the new head beneath the probe arm and install the thumbscrews. Connect the plastic tubing to the fittings on the new head, observing the color code.

After changing the head (or reinstalling the original head after cleaning), you must realign the head with the plate carrier.

Note: Ensure that you first change the washer to the proper head type (8-, 12, or 16-probe) using the 8/12 key.

After selecting the head type, perform an **ALIGN**:

Place a microplate in the plate carrier. If you are using an 8- or 16-probe head, orient the plate so that well number A-1 is at the left rear corner. If using 12-probe head, orient the plate so that well number A-1 is at the left front corner. Be sure that the plate is securely seated in the grooves on the plate carrier. Press the **ALIGN** key and the instrument will sense the position of the new head in relation to the plate bed and the microplate.

Note: After a head change has been selected via the 8/12 key, the instrument will ask “Last plate was XX way; Is XX way head installed” during the ALIGN process. Press YES to confirm that the head has been properly changed and selected.

After the **ALIGN** is complete, select test #214 and perform the plate alignment discussed in section 3.2.8-*Plate Carrier Alignment*. Then, press **PRIME** to remove any trapped air from the newly installed head. Repeat the **PRIME** until a steady stream of fluid can be seen exiting the dispense tubes. To assure uniform washing, it is important to observe the proper functioning of each probe before washing a plate.

3.2.4 Exhaust filter replacement

The filter that removes contaminant particles from the exhaust of the vacuum pump should be replaced only as needed. If the “WAIT” message for the pump is taking longer to clear, the filter may be clogged and should be replaced. In the event that the filter gets wet due to a waste bottle spill, the filter will need replacement.



CAUTION: once the filter has become wet, it will be blocked. Continued use of a blocked filter may impair washer effectiveness and/or result in damage to the washer.

Contact your dealer for a replacement filter. When replacing the filter, note that the INLET side should be toward the bottles.

3.2.5 Valve tubing replacement

It is not recommended to replace any tubing while the washer is performing properly. The short lengths of 1/16" silicone tubing used in the internal valves may become clogged or worn with age. If this occurs, contact your dealer for replacement valve tubing. Follow the instructions included with the replacement valve tubing.

3.2.6 Dispense volume repeatability check

Dispense volume repeatability should be checked at least once every 6 months to assure that the probes are not clogged or damaged.

- Fill the wash bottle with a solution containing a wetting agent (such as a drop of liquid soap).
- Press **PRIME** to prime the instrument.
- Dispense 300 µL into each well of a clean dry microplate.
- Remove the plate and examine the wells visually to see that each appears evenly filled.

If any row of wells has lower volumes than the other rows, suspect a clogged probe. Using the tool kit provided, attempt to clean the dispense probe that fills the affected row. Repeat the test. If the wells do not appear to be filling evenly, consult your dealer for service, and do not use the washer in clinical testing until the problem is corrected.

3.2.7 Volume calibration

The calibration of the dispense volume can be checked using a 50 mL graduated cylinder. Volume calibration is also recommended every 6 months.

- From the main prompt, press **YES**. Enter 213 and press **YES**. The display will show:

```
PD:40   DD: 1   03/14/96
Pl:125  Pr: 5.0 Vol:27.0
```

The top line shows the plate depth parameter (used for manual depth), the dispense depth parameter, and the date of last parameter change. The bottom line shows the values for Plate Alignment, pressure, and volume. The value shown for volume in the example may be different than actual. Write down the value shown for volume.

- Press **YES** to return to the main prompt.
- From the main prompt, press **YES**. Enter 212 and press **YES**. The display shows:

```
Run a program
Ent calibration vol   Y/N
```

- Press **NO**. The display shows:

```
Run a program
Calibrate volumes    Y/N
```

- Press **YES**. The display shows:

```
Unplug dispense tube
Place in cyl -> <YES>
```

- Unplug the plastic dispense tube from the lower (red) fitting on the side of the head. Place the end of the tubing in a clean, dry 50 mL graduated cylinder. Position the tubing so that the tubing remains level with the head. Press **YES** and the unit will dispense a relatively large amount of fluid (but not more than 50mL). When this is complete, the display shows:

```
Enter measured vol (mL):
```

- Read the amount dispensed from the graduated cylinder. Enter the new value read from the graduated cylinder and press **YES**. If you do not want to change the value, press **NO** twice in succession to cancel and return to the main prompt. Record the volume each month. If the volume is trending lower each time, verify the pressure setting.
- Connect the plastic dispense tube to the fitting on the head. Be sure to **PRIME** before performing any dispense operation, to remove any trapped air from the head.

3.2.8 Plate Carrier Alignment

The plate carrier is aligned by means of a software adjustment value. The factory setting need not be changed. This information is included for verification only.

To enter the plate carrier alignment parameter from the calibration data label (located on the bottom of the instrument), select test #210. The display shows:

```
Plate Carrier alignment
```


Type the value shown on the calibration data label and press **YES**.

To realign the plate carrier, first run **ALIGN**, then select test #214. The display shows:

```
Plate Carrier alignment  
0=Lft, 1=Rt, YES=Accept
```

The figure in the margin shows the desired alignment. Press 0 (to move left) and 1 (to move right) to achieve the results pictured. Press **YES** when complete.

Aspirate tube must be as far to right as possible without touching the side of the well.



3.2.9 Dispense Verification

The following test may be run to ascertain proper performance characteristics. A scale capable of reading tenths of a gram (XX.X grams) and a standard 96 well microplate are necessary.

- Fill the wash bottle of the washer with deionized water containing a wetting agent (such as a drop of liquid soap). Press **PRIME** to prime the system, then repeat the priming.
- Place a clean, dry 96-well standard microplate on the scale and tare it to 0.0 grams.
- Install the plate into the washer.
- Press **DISP**. Enter 200 and press **YES**. The instrument will dispense 200 μ L of water into each well of the plate.
- When the dispense is complete, place the filled plate on the scale. The scale reading should be between 18.0 and 20.0 grams. If not, perform a Volume Calibration, then repeat.

If the washer does not pass this test, contact your dealer.

3.2.10 Pressure Adjustment

Some users may find they wish to adjust the pressure up or down in order to increase or decrease the wash force, or to reduce the forming of bubbles (in addition to the top wash feature, see *Section 2.2.7*). To change the pressure, first put the instrument in Standby mode by pressing the STNDBY key. Then, select test #201. The display shows:

Pressure (PSI) = ?

Enter the desired pressure amount in PSI. You will then need to recalibrate the volume setting, following the instructions in *Section 3.2.7*. Note that reducing the value below 2.5 may decrease your precision; run a dispense plate and verify that all wells are filled uniformly.

3.2.11 Storage

The instrument may be stored under the following environmental conditions:

Temperature 5 to 70°C
Humidity Max. 85% non-condensing.

Before storing or shipping the instrument, follow this procedure:

1. Fill the wash and rinse bottles with deionized water.
2. Perform a **PRIME** several times, then perform a **RINSE** several times to flush all solutions from the instrument.
3. Empty all bottles and replace.
4. Perform a **PRIME** to purge with air.
5. Perform a **RINSE** several times to purge with air. Repeat as needed until no more fluid is seen flowing out of the head or into the waste bottle.
6. Turn off the power switch and remove the power cord.
7. Disconnect all tubing and sensor leads from the rear panel.
8. Remove the plate carrier and dry the plate bed and reservoir thoroughly using lint-free absorbent towels. Replace the plate carrier. The instrument is now ready for storage.

Note: Always handle and store the wash head very carefully to avoid physical damage to the probes.

3.3 Troubleshooting/ Error Messages

Successful troubleshooting will be aided by an understanding of the basic elements of the system. The instrument consists of an interdependent pressure and vacuum system and position systems for the plate and wash head. These systems are regulated and positioned to the programmed levels under microprocessor control. Pinch valves control the dispense functions. One pinch valve controls the wash solution and one controls the rinse solution. A third valve directs output of the pump to pressurize the bottles and maintain pressure under microprocessor control. A fourth valve is used to relieve the pressure system during the idle period and in the standby mode. The first two of these valves and most plumbing can be accessed through the door on the rear of the unit. **Always turn the instrument off and disconnect the power cord before opening this door or disconnecting any reservoir tubing.** Motorized lead screws and optical sensors provide for precise plate and probe positioning.

3.3.1 Wash and Rinse System

The wash and rinse bottles are pressurized by the common pump to a level of 5 psi (default setting) during active periods. Note that the user may change the pressure setting (see section 3.2.10 - *Pressure Adjustment*). During standby, or after an inactive period of approximately 12 minutes, the pressure is relieved by means of the pressure relief valve. The displayed pressure on line 2 of the display will change from 5 ± 2 psi to OFF indicating a sleep period. Pressing any key will reactivate this system. In the sleep period, the probe head will lower to the tray and a small amount of liquid will be released in order to keep the probe heads moist and free from crystallized wash solution.

The wash and rinse bottles are connected to the pressure system by means of a tee fitting located at the bottle caps. The unfilled portions of these bottles are the pressure reservoir for dispensing. The outlet of each bottle is connected to a pinch valve mounted on the valve door. The exit side of these valves are again teed together and connected to the dispense chamber of the wash head. Control of the dispensed volume is achieved by precisely timed activation of these valves. Conductivity sensors in each bottle sense fluid levels.

3.3.2 Aspiration System

The aspiration system is based on a strong vacuum pump; the pump runs continuously when active. The pump draws the waste bottle down to this vacuum level through a fine hydrophobic filter, and is directly connected to the aspirate chamber of the wash head. A conductivity sensor connected to the bottle senses waste level. When a full condition is detected, the instrument goes into the standby mode, beeps to alert the operator, and relieves the pressure system. Prior to going into standby mode, the instrument will complete any wash program in progress. The alert occurs when there is sufficient remaining wash solution (or sufficient room in the waste bottle) to allow the completion of the wash program.

3.3.3 Error Messages

Below are described some of the observations and error messages that will help you detect trouble. The solution provided for each is a suggested cause and remedy. If you have a problem that the suggested actions fail to solve, do not use the washer for clinical testing. Consult your dealer to arrange for further assistance.

Problem: Incomplete aspiration.

Explanation/Solution:

- a.) Not using dummy wells in partial strips - place empty wells to complete rows.
- b.) Aspirate needle(s) clogged—use cleaning tool.
- c.) Tubing kinked—check tubing.

Problem: Incomplete dispense

Explanation/Solution:

- a.) Dispense needle(s) clogged—use cleaning tool.
- b.) Dispense tube kinked. Dispense valve stuck. Check for pinched tubing.
- c.) Check volume calibration.

Problem: Inaccurate dispense

Explanation/Solution:

- a.) Check dispense needles. Perform dispense repeatability and calibration procedures. Verify pressure.
- b.) Check volume calibration.

Problem: Excess bubbles affecting washing

Explanation/Solution: most washes are not affected by the amount of bubbles (caused by wash solution surfactant), but if your assay is sensitive or if you wish to decrease the level of bubbles in any case, use the top wash feature (see section 2.2.7).

Problem: Wash or Rinse fluid level not detected

Explanation/Solution: Check that sensors are plugged in to bottle caps and valve door. Clean/dry sensor connections and the cap at the terminal entry points.

Problem: “Empty Waste” message, bottle not full.

Explanation/Solution: Check that sensors are plugged in to bottle caps and valve door. Clean/dry sensor connections and the cap at the terminal entry points.

Problem: Vacuum pump is slow.

Explanation/Solution: Filter clogged, or Vacuum leak. Check tubing and fittings.

Problem: Pump runs during inactive periods

Explanation/Solution: Check system for leaks.

Problem: CAUTION: Probe lowering

Explanation/Solution: Head is about to go into sleep mode. Move your hands to avoid injury.

Problem: Dispense system has been rinsed. Prime first? Y/N

Explanation/Solution: This message occurs when a dispense, aspirate, or program is selected prior to the system being primed. Press YES to prime the system.

**Problem: Last plate was XX way
Is XX way head installed**

Explanation/Solution: After a head change has been selected via the 8/12 key, the instrument will ask this question the first time a wash function is attempted. Press YES to confirm that the head has been properly changed and selected.

Problem: * MECH JAM: PROBE***
(or ***MECH JAM: PLATE***)
CLEAR JAM THEN --> <YES>**

Explanation/Solution: Check for obstructions and clear them, then press YES. If you cannot clear the obstructions easily, in most situations, pressing YES will cancel the mode and return to the “Press <YES> to run prog” message.

Problem: Check Vacuum System

Explanation/Solution: Check tubing - may be kinked or clogged, or head may be clogged.

Problem: Check Pressure System

Explanation/Solution: Possible leak on the air or fluid side. Check for air leak or fluid leak.

Problem: Vac OverRng-Check Filter

Explanation/Solution: This will occur if the vacuum over-ranges, and is most likely due to a blocked vacuum filter. Replacement of the filter is needed.

Problem: ***CHECKSUM ERROR*******

Explanation/Solution: Stored parameters have been lost. The display shows:

```
Enter Plate Align param
```

Enter the plate alignment parameter from the calibration data label and press **YES**. The display shows:

```
Enter measured vol (mL):
```

Enter the calibration volume from the calibration data label on the bottom of the instrument (or the most recent value determined from periodic volume calibration) and press **YES**. If you are using other stored parameters, such as manual depth and dispense depth, or have modified the pressure setting, you should also reset these at this time.

Problem: Cannot delete Perm Prog

Explanation/Solution: The first six tests are permanent. This message occurs when you try to delete one.

Problem: MEMORY FULL

Explanation/Solution: Insufficient memory to store program. Delete previously stored programs to make more memory available.

Problem: MEMORY ERROR: Prog Ended

Explanation/Solution: Checksum error in recalling the program. Contact your dealer if you continue to have problems.