

Certificate of Analysis



Anode Buffer Container 3500 Series

Product No. 4393927
Lot No. 2402925
Date of Manufacture 09APR2024
Expiration Date 25JUN2025

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION Material is clear, colorless and free of particulates		
Appearance - Conformance		Pass
PH TEST Measured at 25.0°C ± 0.3°C		
pH Value	7.6 - 8.2	7.939
CONDUCTIVITY TEST Measured at 25.0°C ± 0.1°C		
Conductivity	1500 - 2000 µS/cm	1787 µS/cm

For Research Use Only. Not for use in diagnostic procedures.

**Manufactured in compliance with our ISO 13485 certified quality management system.
Site: Bedford, MA USA**

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

Janelle Him
Quality Assurance
Issued 10APR2024