

# STORZ

## KARL STORZ—ENDOSKOPE

**en Instructions for use**  
**POWER LED 300**

Digitally signed by Gherman Irina  
Date: 2026.07.09 09:36:16 EEST  
Reason: MoldSign Signature  
Location: Moldova

MOLDOVA EUROPEANĂ



05-2024

**Copyright ©**

All product illustrations, product descriptions, and texts are the intellectual property of  
KARL STORZ SE & Co. KG.

Their use and reproduction by third parties require the express approval of KARL STORZ SE & Co. KG.

All rights reserved.



## Table of contents

|                                                                    |           |
|--------------------------------------------------------------------|-----------|
| <b>1 General information .....</b>                                 | <b>5</b>  |
| 1.1 Read the instructions for use .....                            | 5         |
| 1.2 Read the instructions for use of compatible products .....     | 5         |
| 1.3 Scope .....                                                    | 5         |
| 1.4 General signs and symbols .....                                | 5         |
| 1.5 Description of warning messages .....                          | 6         |
| <b>2 Normal use .....</b>                                          | <b>7</b>  |
| 2.1 Intended use .....                                             | 7         |
| 2.2 Indications .....                                              | 7         |
| 2.3 Contraindications .....                                        | 7         |
| 2.4 Clinical benefits .....                                        | 7         |
| 2.5 Residual risks .....                                           | 7         |
| 2.6 Target user populations .....                                  | 7         |
| 2.7 Patient population .....                                       | 7         |
| 2.8 Intended conditions of use .....                               | 7         |
| <b>3 Safety and warning .....</b>                                  | <b>8</b>  |
| 3.1 Serious incidents .....                                        | 8         |
| 3.2 Correct handling and product testing .....                     | 8         |
| 3.3 Combination with other components .....                        | 8         |
| 3.4 Working in the field of vision .....                           | 9         |
| 3.5 Product not clean .....                                        | 9         |
| 3.6 Dangers from electrical current .....                          | 9         |
| 3.7 Damage due to ingress of liquid in electrical components ..... | 9         |
| 3.8 Risk of explosion and fire .....                               | 9         |
| 3.9 Electromagnetic interference .....                             | 10        |
| 3.10 Hot components .....                                          | 10        |
| 3.11 High light intensity .....                                    | 10        |
| 3.12 Functionality of the touch screen .....                       | 10        |
| 3.13 Failure of products .....                                     | 10        |
| <b>4 Product description .....</b>                                 | <b>12</b> |
| 4.1 Product overview .....                                         | 12        |
| 4.2 Possible combinations .....                                    | 13        |
| 4.3 Technical data .....                                           | 14        |
| 4.4 Symbols employed .....                                         | 14        |
| 4.4.1 Symbols on the packaging .....                               | 14        |
| 4.4.2 Symbols on the product .....                                 | 15        |
| 4.4.3 Symbols on the type plate .....                              | 16        |
| 4.4.4 Symbols on the user interface .....                          | 16        |
| 4.5 Ambient conditions .....                                       | 17        |
| <b>5 Preparation .....</b>                                         | <b>18</b> |
| 5.1 Unpacking the product .....                                    | 18        |
| 5.2 Inspecting products .....                                      | 18        |
| 5.3 Reprocessing the product .....                                 | 18        |
| 5.4 Setting up the product .....                                   | 18        |
| 5.5 Connecting the product .....                                   | 19        |

|           |                                                                                               |           |
|-----------|-----------------------------------------------------------------------------------------------|-----------|
| 5.6       | Connecting the light cable .....                                                              | 20        |
| 5.7       | Putting the product into operation .....                                                      | 22        |
| <b>6</b>  | <b>Application .....</b>                                                                      | <b>23</b> |
| 6.1       | User interface .....                                                                          | 23        |
| 6.2       | Switching the light on and off .....                                                          | 23        |
| 6.3       | Manual light adjustment .....                                                                 | 24        |
| 6.4       | Automatic light control .....                                                                 | 24        |
| 6.5       | Settings .....                                                                                | 24        |
| 6.5.1     | Service .....                                                                                 | 25        |
| 6.5.2     | Setting the language .....                                                                    | 25        |
| 6.5.3     | Event protocol .....                                                                          | 25        |
| 6.5.4     | Product information .....                                                                     | 26        |
| 6.5.5     | Audio settings .....                                                                          | 27        |
| 6.6       | Switching off the product .....                                                               | 27        |
| 6.7       | Information signals .....                                                                     | 27        |
| 6.7.1     | Visual information signal .....                                                               | 27        |
| 6.7.2     | Acoustic information signal .....                                                             | 28        |
| 6.7.3     | Availability signal .....                                                                     | 28        |
| <b>7</b>  | <b>Maintenance, servicing, repairs, and disposal .....</b>                                    | <b>29</b> |
| 7.1       | Maintaining the product .....                                                                 | 29        |
| 7.1.1     | Maintenance .....                                                                             | 29        |
| 7.2       | Changing a fuse .....                                                                         | 29        |
| 7.3       | Safety inspection in accordance with IEC 62353 .....                                          | 30        |
| 7.3.1     | Visual inspection .....                                                                       | 31        |
| 7.3.2     | Electric measurements .....                                                                   | 31        |
| 7.3.3     | Functional test .....                                                                         | 31        |
| 7.4       | Repairing the product .....                                                                   | 31        |
| 7.5       | Disposing of the product .....                                                                | 31        |
| <b>8</b>  | <b>Accessories and spare parts .....</b>                                                      | <b>32</b> |
| 8.1       | Accessories .....                                                                             | 32        |
| 8.2       | Spare parts .....                                                                             | 32        |
| <b>9</b>  | <b>Electromagnetic compatibility .....</b>                                                    | <b>33</b> |
| 9.1       | General notes on the operating environment .....                                              | 33        |
| 9.2       | Accessories, transducers and cables .....                                                     | 33        |
| 9.3       | Test-Tables .....                                                                             | 34        |
| 9.3.1     | Table 1 – Compliance level for immunity tests .....                                           | 34        |
| 9.3.2     | Table 2 – Test levels for proximity fields from RF wireless communications<br>equipment ..... | 36        |
| 9.3.3     | Table 3 – Emission class and group .....                                                      | 36        |
| <b>10</b> | <b>Errors and messages .....</b>                                                              | <b>38</b> |
| 10.1      | Troubleshooting .....                                                                         | 38        |
| <b>11</b> | <b>Overview of mitigating warnings .....</b>                                                  | <b>39</b> |
| <b>12</b> | <b>Subsidiaries .....</b>                                                                     | <b>40</b> |



# 1 General information

## 1.1 Read the instructions for use

If the instructions for use are not followed, patients, users, and third parties may be injured or the product may be damaged.

- ▶ Read the instructions for use carefully and follow all the safety notes and warnings.
- ▶ Read the reprocessing instructions carefully and follow all the safety notes and warnings. The reprocessing instructions can be downloaded from [www.karlstorz.com/ifu](http://www.karlstorz.com/ifu) by entering the item number.
- ▶ Keep the instructions for use and reprocessing instructions in a safe place.

## 1.2 Read the instructions for use of compatible products

If the instructions for use of compatible products are not followed, patients, users, and third parties may be injured or the product may be damaged.

- ▶ Read the instructions for use of the compatible products carefully and follow all the safety notes and warnings.
- ▶ Read the reprocessing instructions of the compatible products carefully and follow all the safety notes and warnings.

## 1.3 Scope

This instruction manual is valid for:

| Product name  | Article number |
|---------------|----------------|
| POWER LED 300 | TL300          |

## 1.4 General signs and symbols

The signs and symbols used in this document have the following meaning:

### Practical tip

-  This sign refers to useful and important information.

### Actions to be performed

Action to be carried out by several steps:

- ✓ Prerequisite that must be met before carrying out an action.
- 1. Step 1
  - ⇒ Interim result of an action
- 2. Step 2
  - ⇒ Result of a completed action

Actions in safety notes or in the case of a single step:

- ▶ Step 1

### Lists

- 1. Numbered list
  - Unnumbered list, 1st level

- Unnumbered list, 2nd level

## 1.5 Description of warning messages

To prevent any injury to persons or damage to property, the warnings and safety notes in the instructions for use must be observed. The warnings use the following levels of danger:

### **⚠ WARNING**

#### **WARNING**

Designates a possible imminent risk. If this is not avoided, it could lead to death or serious injuries.

### **⚠ CAUTION**

#### **CAUTION**

Designates a possible imminent risk. If this is not avoided, it could lead to minor injuries.

### **NOTICE**

#### **NOTICE**

Designates a possibly harmful situation. If this is not avoided, the products could be damaged.

## 2 Normal use

### 2.1 Intended use

#### White-light light sources

White-light light sources are intended for white light illumination in diagnostic and surgical interventions.

White-light sources do not have body contact and are meant for short-term use.

### 2.2 Indications

White-light light sources are suitable for generating white-light illumination for medical investigations and visualization during diagnostic and surgical interventions.

### 2.3 Contraindications

Light sources, light cables and adapters must not be used for ophthalmologic interventions.

Light sources, light cables and adapters are not used in body contact with the patient but provide light illumination for medical imaging.

Furthermore, there are no contraindications for the use of the light sources, light cables and adapters directly associated with the product.

### 2.4 Clinical benefits

Light sources, light cables and adapters are widely used during endoscopic diagnostic and therapeutic procedures.

### 2.5 Residual risks

No residual risks directly related to the product were identified.

### 2.6 Target user populations

The application of the products in question is carried out under the responsibility of a medical specialist.

### 2.7 Patient population

There are no restrictions in terms of patient groups for this product.

### 2.8 Intended conditions of use

The product may only be used in hospitals and doctors' offices in suitable ambient conditions.

| Condition             | Application                                                   |
|-----------------------|---------------------------------------------------------------|
| Frequency of use      | One or more times a day                                       |
| Length of use         | Several minutes to several hours a day                        |
| Place of installation | Positioning on a level, vibration-free surface                |
| Mobility              | Can be moved if placed on a cart.                             |
| Combination           | Can be used on the patient at the same time as other devices. |
| Control               | Can be controlled via the KARL STORZ SCB.                     |

### 3 Safety and warning

#### **⚠ WARNING**

##### **Danger due to non-observance of warnings and safety notes**

This chapter contains warnings and safety notes structured according to hazards and risks.

- ▶ Carefully read and observe all warnings and safety notes.
- ▶ Follow the instructions.

### 3.1 Serious incidents

A 'serious incident' includes incidents which, directly or indirectly, had, could have had or could have any of the following consequences:

- Death of a patient, user, or another person
- Temporary or permanent serious deterioration in the medical condition of a patient, user, or another person
- A serious threat to public health
- ▶ The manufacturer and appropriate authority must be notified of all serious incidents.

### 3.2 Correct handling and product testing

If the product is not handled correctly, patients, users, and third parties may be injured.

- ▶ Only persons with the necessary medical qualification and who are acquainted with the application of the product may work with it.
- ▶ Check that the product is suitable for the procedure prior to use.
- ▶ Check the product for the following properties, for example, before and after every use:
  - Functionality
  - Damage
  - Changes to the surface
  - In the case of several components: completeness and correct assembly
- ▶ Do not continue to use damaged products.
- ▶ Dispose of the product properly; see *Disposing of the product*.

### 3.3 Combination with other components

The use of unauthorized devices and components may result in injuries.

- ▶ Ensure that any additional devices connected to electrical medical devices comply with the relevant IEC or ISO standards.
- ▶ When connecting additional devices to medical electrical equipment, ensure that all configurations of the resulting system comply with the standardized requirements for systems (see IEC 60601-1). National laws and regulations take precedence over the aforementioned standards.
- ▶ Only combine the product with devices and components that the manufacturer has approved for combined use, see chapter *Possible combinations*.
- ▶ Only use devices and components that have standardized interfaces and do not breach the normal use of the product.
- ▶ Observe the instruction manuals and interface specifications of the devices and components used in combination.
- ▶ Do not make any modifications to the product.



### 3.4 Working in the field of vision

Using the product outside the field of vision can cause injury to tissue or can damage the product.

- ▶ Only use the product in the field of vision.

### 3.5 Product not clean

The product is not clean when delivered. The use of products that have not been cleaned poses a risk of infection for patients, users, and third parties.

- ▶ Reprocess the product in line with the reprocessing instructions before initial use and every subsequent use.

### 3.6 Dangers from electrical current

An improper power supply may cause an electric shock and injure patients, users, or third parties.

- ▶ Ensure that all electrical installations of the operation room in which the product is connected and used conform with the applicable IEC standards.
- ▶ Use either the power cord supplied by KARL STORZ or a power cord which has the same properties and which bears a national mark of conformity.
- ▶ The product may only be operated with the line voltage stated on the rating plate.
- ▶ Position the product appropriately so that the power cord can be unplugged at any time. The product is only voltage-free when the mains plug has been disconnected.
- ▶ Connect the product to a power supply with protective conductor.
- ▶ Ensure potential equalization according to the applicable national rules and regulations.
- ▶ To ensure reliable protective earth grounding, connect the product to a properly installed socket that is approved for use in the operation room. Routinely inspect the electrical plug and cord and do not use if the inspection reveals damage.

In the case of electrical products, individual components or the product itself may be live. Live parts can cause electric shocks in the event of contact and injure patients, users, and third parties.

- ▶ Do not open the product.
- ▶ Do not touch the output jacks of the product and the patient at the same time during use
- ▶ Have servicing carried out by KARL STORZ or a company authorized by KARL STORZ.
- ▶ Always pull out the mains plug before carrying out any cleaning and maintenance work.

If several products supplied with energy are used simultaneously, the patient leakage currents accumulate. These leakage currents can exceed the limit values and injure patients.

- ▶ The patient applied parts of the simultaneously used products must be type BF or type CF.

### 3.7 Damage due to ingress of liquid in electrical components

In the case of electrical products, individual components or the product itself may be live. Liquid ingress into an electrical product may result in a short circuit or an unintentional transfer of current. The product is damaged as a result and patients, users and third parties may be injured.


- ▶ Do not store liquids near the product or on the product.
- ▶ If liquid has entered the product, pull out the plug and allow the product to dry completely.

### 3.8 Risk of explosion and fire

The product can generate sparks, which cause combustible or flammable gases and liquids to ignite or explode. This may cause injuries to patients, users, and third parties.



- ▶ When using explosive narcotic gases: Operate the product outside of the hazard zone.



- ▶ Do not use the product in the presence of flammable anesthetics.
- ▶ The product must not be operated in oxygenated environments.
- ▶ Only connect or disconnect the power plug to or from the power supply outside explosive atmospheres.

### 3.9 Electromagnetic interference

Medical electrical devices are subject to special precautions regarding electromagnetic compatibility. If other devices (e.g., MRT, CT, diathermy, electrocautery, or RFID) emit electromagnetic radiation, the product's functionality may be impaired. High-frequency (HF) communication equipment can affect electrical medical devices and impair their performance.

- ▶ During installation and operation of the product, please take note of the information on electromagnetic compatibility, see chapter *Electromagnetic compatibility*.

### 3.10 Hot components

The high level of light intensity produced by the light source may cause the distal end, the light connections, and adjacent components to heat up. This can cause burns to patients, users, and third parties.

- ▶ Set the output of the adjustable light sources to a level that is just high enough to ensure optimal illumination of the operating area.
- ▶ Prevent the distal end, light connections, and adjacent components from coming into contact with tissue and operating room accessories.

### 3.11 High light intensity

The high level of light intensity produced by the light source may lead to permanent eye damage or blindness, and may cause tissue and items facing the light output to heat up.

- ▶ Do not look into the light output.
- ▶ Set the output of the adjustable light sources to a level that is just high enough to ensure optimal illumination of the operating area.

### 3.12 Functionality of the touch screen

If the functionality of the touch screen is limited, the product cannot be used correctly. Patients, users, and third parties may be injured.

- ▶ Do not use the product if the touch screen is defective.
- ▶ Do not tap the touch screen in several places at the same time.


### 3.13 Failure of products

The product may fail during use.

- ▶ Have a replacement product ready for each application or plan for an alternative surgical technique.

## 4 Product description

### 4.1 Product overview



*POWER LED 300 – Front view*

- 1 Power switch
- 2 TFT touch screen

- 3 Light outlet point



POWER LED 300 – Back view

- |   |                   |   |                   |
|---|-------------------|---|-------------------|
| 1 | Grounding plug    | 3 | Power fuse holder |
| 2 | Power cord socket | 4 | SCB connectors    |

## 4.2 Possible combinations

It is recommended that the suitability of the products for the intended procedure be checked prior to use. Please note that the described products in this medium may not be available in all countries due to different regulatory requirements.

The basic system consists of the following components:

- Power LED 300
- Camera control unit (CCU), IMAGE1 S system
- Light cable with endoscope or videoendoscope

### Compatibility with light cables

Series 495xx light cables are compatible with the standard adapter.

- i** Only KARL STORZ Hopkins telescopes are recommended in combination with KARL STORZ light cables.

### Compatibility with camera control units (CCU)

The light source can be connected to the KARL STORZ camera control unit (CCU).

The following camera control units (CCUs) can be combined:

- IMAGE1 S CONNECT (TC200)
- IMAGE1 S CONNECT II (TC201)
- IMAGE1 S H3-LINK (TC300)
- IMAGE1 S X-LINK (TC301)

- IMAGE1 S D3-LINK (TC302)
- IMAGE1 S 4U-LINK (TC304)

### 4.3 Technical data

| Description                                | Value              |
|--------------------------------------------|--------------------|
| Power supply (AC)                          | 100–240 V          |
| Operating frequency                        | 50/60 Hz           |
| Line fuse                                  | 2 x T 4.0 AH 250 V |
| Power consumption                          | 175 VA             |
| Electrical protection class                | I                  |
| Applied part type according to IEC 60601-1 | CF                 |
| Dimensions (L x H x W)                     | 313 x 155 x 305 mm |
| Weight                                     | 8.6 kg             |

### 4.4 Symbols employed

#### 4.4.1 Symbols on the packaging

| Symbol                                                                              | Meaning                                                |
|-------------------------------------------------------------------------------------|--------------------------------------------------------|
|  | Manufacturer                                           |
|  | Date of manufacture                                    |
|  | Medical device                                         |
|  | Article no.                                            |
|  | Serial number                                          |
|  | Number of products in the product packaging            |
|  | Unique Device Identifier                               |
|  | Consult the printed or electronic instructions for use |
|  | Fragile, handle with care                              |



| Symbol                                                                            | Meaning                                                                                                                                                                                                                                                                                                                                                 |
|-----------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  | Keep dry                                                                                                                                                                                                                                                                                                                                                |
|  | Temperature limit                                                                                                                                                                                                                                                                                                                                       |
|  | Humidity limit                                                                                                                                                                                                                                                                                                                                          |
|  | Air pressure limit                                                                                                                                                                                                                                                                                                                                      |
| <b>Rx only</b>                                                                    | Federal (USA) law restricts this device to sale by or on the order of a physician.                                                                                                                                                                                                                                                                      |
| <b>CE</b>                                                                         | <p>CE marking<br/>With this marking, the manufacturer declares the conformity of the product with the applicable EU directives. A code number after the CE mark indicates the responsible notified body.</p> <p>The EU directives relevant to the product can be found in the EU Declaration of Conformity, which can be requested from KARL STORZ.</p> |

#### 4.4.2 Symbols on the product


| Symbol                                                                              | Meaning                                                                                                                                                                                                                                                                                                                                                                              |
|-------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  | Follow the instructions for use.<br>The color may differ on the product. The symbol is black/white on the packaging label.                                                                                                                                                                                                                                                           |
|  | ON                                                                                                                                                                                                                                                                                                                                                                                   |
|  | OFF                                                                                                                                                                                                                                                                                                                                                                                  |
|  | <p>The potential equalization is responsible for equalizing the potentials of different metal parts that can be touched at the same time, or for reducing potential differences that could occur between the body, electromedical devices, and external live parts during use.</p> <p>The potential equalization complies with the requirements for a medical electrical system.</p> |
|  | Applied part of the type CF                                                                                                                                                                                                                                                                                                                                                          |
|  | <p>KARL STORZ SCB interface<br/>Remote control of functions and remote display of parameters</p>                                                                                                                                                                                                                                                                                     |

#### 4.4.3 Symbols on the type plate

| Symbol                                                                              | Meaning                                                                                                                                                                                                                                                                                                                                         |
|-------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|    | Manufacturer                                                                                                                                                                                                                                                                                                                                    |
|    | Date of manufacture                                                                                                                                                                                                                                                                                                                             |
|    | Medical device                                                                                                                                                                                                                                                                                                                                  |
|    | Federal (USA) law restricts this device to sale by or on the order of a physician.                                                                                                                                                                                                                                                              |
|    | Prevention of pollution by electronic devices                                                                                                                                                                                                                                                                                                   |
|    | Separate collection of electrical and electronic devices.<br>Do not dispose of in household refuse.                                                                                                                                                                                                                                             |
|  | CE marking<br>With this marking, the manufacturer declares the conformity of the product with the applicable EU directives. A code number after the CE mark indicates the responsible notified body.<br><br>The EU directives relevant to the product can be found in the EU Declaration of Conformity, which can be requested from KARL STORZ. |

#### 4.4.4 Symbols on the user interface

| Symbol                                                                              | Meaning        |
|-------------------------------------------------------------------------------------|----------------|
|  | Settings       |
|  | Standby        |
|  | Start/stop     |
|  | Increase value |
|  | Decrease value |

| Symbol                                                                              | Meaning              |
|-------------------------------------------------------------------------------------|----------------------|
|    | Confirm              |
|    | Cancel               |
|    | Remove               |
|    | Scroll back          |
|   | Scroll forward       |
|  | Audio signals on/off |

## 4.5 Ambient conditions

| Transport and storage conditions   |                                  |
|------------------------------------|----------------------------------|
| Temperature                        | -10°C ... +60°C (14°F ... 140°F) |
| Relative humidity (non-condensing) | 5 – 95 %                         |
| Air pressure                       | 500 – 1,080 hPa                  |

| Operating conditions               |                                |
|------------------------------------|--------------------------------|
| Temperature                        | 10°C ... 40°C (50°F ... 104°F) |
| Relative humidity (non-condensing) | 20–85%                         |
| Max. operating altitude            | 3,000 m                        |

## 5 Preparation

### 5.1 Unpacking the product

1. Carefully remove the product and accessories from the packaging.
2. Check the delivery for missing items and possible damage.
3. In the case of damage, hidden defects, and short deliveries, document their nature and extent and contact the manufacturer or supplier immediately.
4. Keep packaging for further transport.

### 5.2 Inspecting products

Inspect all products used for damage, e.g.:

- Damaged power cord
- Damaged glass plate
- Damaged housing
- Sharp edges, sharp corners

### 5.3 Reprocessing the product

- ▶ Reprocess the product in line with the reprocessing instructions before using it.

### 5.4 Setting up the product

#### **⚠ WARNING**

#### **Overheating! Risk of fire!**

Insufficient ventilation can cause an internal build-up of heat, resulting in a safety shut-down. If the product overheats, there is a risk of fire. Patients, users, and third parties may be injured.

- ▶ Ensure that there is sufficient air circulation.
- ▶ Keep air inlets and outlets free.

This product and connected components may only be used in medical rooms with electrical installations that conform to applicable national regulations. When the product is installed, the position of the user must be taken into account. When the product is being operated, the user stands within a viewing cone with an angle of view of  $\pm 45^\circ$  at a distance of approx. 30–70 cm from the front panel.

1. Set the product down on a horizontal, flat surface or a video cart.
2. Position the product so that it is in the user's field of vision.
3. Keep the product out of reach of patients.

## 5.5 Connecting the product

1. Connect the potential equalization cable.



2. Connect the power cord. Push the power plug fully into the power socket.



3. Connect the SCB cable.



4. Connect the other end of the SCB cable to a KARL STORZ control device or other SCB devices; see the instruction manual for the KARL STORZ SCB control NEO system.
5. To remove the SCB cable, pull on the plug (the SCB cable is equipped with a protection device to prevent it from being pulled out accidentally).



## 5.6 Connecting the light cable

### **⚠ WARNING**

#### **Hot light connections! Risk of burns!**

The high level of light intensity produced by the light source may cause the light connections and adjacent components to heat up. This can cause burns to users and third parties.

- ▶ Switch off the light source before changing the light cable.
- ▶ Allow the product to cool down before changing the light cable.

**i** The light outlet point is provided with an antiglare flap, which does not allow any direct outlet of light. Light only appears when the light cable is attached.

**i** We recommend using original KARL STORZ light cables, because light cables from other manufacturers may not be optimized for light transmission. The degree of protection against electric shock at the applied part as type CF according to IEC 60601-1 is guaranteed when using KARL STORZ light cables, see chapter *Accessories* [p. 32].

1. Insert the light cable into the light outlet point until it engages. Hold the light cable only by the handle, never by the cable.



2. Connect the light cable by twisting the knurled screw on the endoscope screw base through a quarter turn.





## 5.7 Putting the product into operation

1. Switch the product on with the power switch.



⇒ The following start screen appears and the self-test is performed:



- ⇒ After a successful self-test, the ready signal sounds and the light source is ready for use.
- ⇒ If the self-test fails, the product assumes the safe system state and an information signal sounds:

203: Fan initialization error




2. Switch the product on and off.
3. Check whether the product is correctly connected.

## 6 Application

### 6.1 User interface

The product is in manual standby mode after being switched on. The light switches off at a light intensity of 5%. The most recently used light intensity is displayed.



1 Adjust the brightness


3 Light on/off

2 Settings

 The **Settings** button appears only when the light is switched off.

### 6.2 Switching the light on and off

1. Tap on the **Start/Stop** button to switch the light on and off: orange = on, blue = off.




2. To adjust the brightness, see chapter *Manual light adjustment* [p. 24].

## 6.3 Manual light adjustment

The brightness can be adjusted when the light is switched on or off.

1. Tap the **Plus** button to increase the brightness.
  2. Tap the **Minus** button to decrease the brightness.
- ⇒ The set brightness is shown in %.
- ⇒ The orange triangles indicate the set value in relation to the possible setting range.




## 6.4 Automatic light control

If the product is connected to the IMAGE1 S camera control unit via the SCB connecting cable, the light is controlled automatically and optimally adjusted to each situation, see instructions for use IMAGE1 S.


## 6.5 Settings

1. Make sure that the light is switched off and the **Start/Stop** button turns blue.



2. Tap the **Menu** button.
- ⇒ The **Settings** screen appears with the following sub-menus:






### 6.5.1 Service

The service area is reserved for the service employees of KARL STORZ and is therefore password protected. The settings are described in the service manual.

### 6.5.2 Setting the language

1. In the **Settings** screen, tap the **Language** sub-menu.  
⇒ The **Language** screen appears.



2. Select the language and confirm with the **Checkmark**.  
⇒ The **Settings** screen appears.
3. Cancel the selection with the **Cross**.
4. Browse back through the screen with the **Arrow**.


### 6.5.3 Event protocol

Alarms and information reports are saved in the event protocol at the time of occurrence.

Each line contains the following event data:

- Date
- Time
- Info ID

1. In the **Settings** screen, tap the **Event protocol** sub-menu.  
⇒ The **Event protocol** screen appears.




2. Browse back through the screen with the **Arrow**.

## 6.5.4 Product information

Information on the product can be retrieved, e.g., serial number, software version, and operating hours.


1. In the **Settings** screen, tap the **Device Info** sub-menu.  
⇒ The **Device Info** screen appears.



2. Browse back through the screen with the **Arrow**.

### 6.5.5 Audio settings


1. Tap on the **Audio signal on** and **Audio signal off** buttons to switch the acoustic signals on or off.
  - ⇒ The function is active when the button turns orange.
  - ⇒ When the audio signal is switched off, a crossed-out bell appears in the screen header.



2. Confirm the screen with the **Checkmark** or cancel it with the **Cross**.

## 6.6 Switching off the product

1. Switch the product off with the power switch.



## 6.7 Information signals

Information signals are continuously output when they indicate the cause of an inoperable product. All other information signals are output as long as the signal conditions exist. To prevent confusion between the signals when the conditions exist only for a very short time, the signals are displayed for at least 5 s.

Signals of a higher priority overwrite signals of lower priority, or signals of lower priority are suppressed as long as signals of higher priority are present. In the event of multiple signal conditions with the same priority, the most recently detected condition will appear in the title line.


### 6.7.1 Visual information signal

The information signal is displayed with blue writing on a white background in the title line.

### 6.7.2 Acoustic information signal


The acoustic information signal is issued for as long as the signal conditions apply. However, at least one complete signal sequence sounds. Acoustic signals can be temporarily switched on or off in Settings and are always active when the product is restarted.

The pitch of the information signal is modulated at a frequency of 1.5 Hz by  $\pm 2$  Hz at a time. 5 different harmonics are generated and the sound sequence occurs once. Signals indicating the product's safety state are repeated and sound every 15 s.



### 6.7.3 Availability signal

The pitch of the availability signal is modulated with a frequency of 1.5 Hz by  $\pm 2$  Hz each time. 5 different harmonics are generated.



*Harmonics of the availability signal*

## 7 Maintenance, servicing, repairs, and disposal

### 7.1 Maintaining the product

#### **⚠ WARNING**

**Risk of injury due to product degradation!**

Patients, users and third parties may be injured as a result of product and accessory degradation.

- ▶ Shut down the product.
- ▶ Have the deficiencies repaired by persons authorized by KARL STORZ.

If they are not described in more detail here, maintenance activities may only be performed by KARL STORZ or by a company authorized by KARL STORZ.

#### 7.1.1 Maintenance

The following maintenance intervals are recommended:

| Interval | Activity    | To be performed by             |
|----------|-------------|--------------------------------|
| annually | Safety test | KARL STORZ service technicians |

- ▶ Check the device configuration after every software update.

### 7.2 Changing a fuse

- ✓ The product is switched off.
  - ✓ The power cord is disconnected from the product.
1. Remove the screw inserts on the line fuse holder with a screwdriver.



2. Remove the defective fuse.

3. Insert a new fuse. Only use fuses with the specified values; see chapter *Technical data* [p. 14].



4. Introduce the screw inserts into the line fuse holder again.
5. Connect to the power supply.



6. Switch on the product and test for proper operation.



## 7.3 Safety inspection in accordance with IEC 62353

### **⚠ WARNING**

#### **Risk of injury due to product degradation!**

Patients, users and third parties may be injured as a result of product and accessory degradation.

- ▶ Shut down the product.
- ▶ Have the deficiencies repaired by persons authorized by KARL STORZ.



Regardless of the national accident prevention regulations and testing intervals for medical devices, for this device safety checks must be performed as repeat inspections according to IEC 62353 and recorded by a qualified electrician at least once a year. Detailed specifications regarding the scope and execution of the safety inspection can be found in the service manual.

### **7.3.1 Visual inspection**

1. Check the product and accessories for any mechanical damage.
2. Check labels for readability.

### **7.3.2 Electric measurements**

 Limit values for electrical measurements can be found in the current IEC 62353.

1. Inspect the device safety fuses
2. Measure the protective ground resistance.
3. Measure the earth leakage current.
4. Measure the touch current.
5. Measure the patient leakage current.

### **7.3.3 Functional test**

1. Perform a functional test in line with the instructions for use.
2. Document the results of the safety inspection.

## **7.4 Repairing the product**

Repair work may only be performed by KARL STORZ or by a company authorized by KARL STORZ. The interventions described in this instruction manual are exempt from this rule.

- Please contact your local KARL STORZ subsidiary or authorized dealer (see the list of subsidiaries).

Contaminated devices may not be shipped. To prevent contact infections and airborne infections, products must first be decontaminated. KARL STORZ reserves the right to send back contaminated products.

## **7.5 Disposing of the product**

The product meets the requirements of the Directive on Waste Electrical and Electronic Equipment (WEEE).

Within the scope of application of this directive, KARL STORZ SE & Co. KG is responsible for the proper disposal of this product.

1. The product must be disposed of in accordance with the applicable national laws and regulations at a suitable collection point for the reprocessing of electrical and electronic equipment.
2. Contact KARL STORZ SE & Co. KG, a KARL STORZ branch or an authorized dealer to find out the address of the collection point in your area.

## 8 Accessories and spare parts

### 8.1 Accessories

| Article                            | Order no. |
|------------------------------------|-----------|
| All KARL STORZ light cables        | 495xx     |
| Three-way adaptor for light source | TL005     |
| SCB Connecting Cable, 100 cm       | 20090170  |
| Power cord, length 300 cm          | 400A      |
| Power cord, US version, 200 cm     | 400B      |

 Not all articles are available in all regions.

### 8.2 Spare parts

| Article                                                    | Order no. |
|------------------------------------------------------------|-----------|
| Mains fuse, 100 – 240 V, T 4.0 AH 250 V AC, IEC 127 format | 2027690   |

## 9 Electromagnetic compatibility

### 9.1 General notes on the operating environment

#### Special environment

The product is suitable for use in close proximity to an active HF electrosurgical device in professional healthcare facility environments. Professional healthcare facility includes physician offices, dental offices, limited care facilities, freestanding surgical centers, freestanding birth centers, multiple treatment facilities, hospitals (emergency rooms, patient rooms, intensive care, surgical rooms, outside the RF shielded room of an ME system for MRT).

 This product has been evaluated for compatibility with high-frequency surgical equipment.

 The emission characteristics of this product make it suitable for use in professional healthcare environment as well as residential environment (CISPR 11 Class B). This product offers adequate protection to radio communication service. In the rare event of interference to the radio communication service, the user might need to take mitigation measures, such as relocating or re-orienting equipment.

#### **WARNING**

##### **Electromagnetic interferences! Malfunction!**

Use of this product adjacent to or stacked with other equipment could result in improper operation.

- ▶ This situation should be avoided.
- ▶ If such use is necessary: Verify that this equipment and the other equipment are operating normally.



#### **CAUTION**

##### **MR unsafe!**

This product is MR unsafe.

- ▶ Keep the product away from the Magnetic Resonance Imaging (MRI) Scanner Room and mobile MRI scanner.

### 9.2 Accessories, transducers and cables

#### **WARNING**

##### **Reduced immunity! Malfunction!**

The use of an accessory, transducers and cables with the product other than those specified in this manual may result in increased emissions or decreased immunity.

- ▶ Preferably use the accessories specified in the manual.
- ▶ When using other than those specified in this manual, it becomes the responsibility of the user to determine compliance with IEC 60601-1-2.

**⚠ WARNING**
**Degradation of performance! Malfunction!**

Portable RF communications equipment (including peripherals such as antenna cables and external antennas) could result in degradation of the performance of the product.

- Use portable communications equipment no closer than 30 cm (12 inches) to any part of the product, including cables.

The following accessories and cables are defined for EMC compliance.

| Type       | Shielded | Maximum length | Contains ferrite | Use                               |
|------------|----------|----------------|------------------|-----------------------------------|
| PE         | No       | >3 m           | No               | Potential equalization            |
| Power cord | No       | 3 m            | No               | Connection of device to AC mains. |

## 9.3 Test-Tables

### 9.3.1 Table 1 – Compliance level for immunity tests

| Interference immunity tests                                                                    | EN/IEC 60601-1-2 test level                                                                                                                                                                                                                 | Compliance level                                                                                                                                                                                                                            | Electromagnetic environment – guidance                                                                                                                                                                                                                                                                                  |
|------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Electrostatic discharge (ESD) acc. to IEC 61000-4-2                                            | ± 8 kV contact discharge<br>± 15 kV air discharge                                                                                                                                                                                           | ± 8 kV contact discharge<br>± 15 kV air discharge                                                                                                                                                                                           | The relative humidity should be at least 30%.                                                                                                                                                                                                                                                                           |
| Electrical fast transients/bursts acc. to IEC 61000-4-4                                        | ± 2 kV for power lines<br>± 1 kV for input and output lines<br>100 kHz repetition                                                                                                                                                           | ± 2 kV/± 1 kV for power lines<br>± 1 kV for input and output lines<br>100 kHz repetition                                                                                                                                                    | The power supply quality should be that of a typical commercial or hospital environment.                                                                                                                                                                                                                                |
| Surges acc. to IEC 61000-4-5                                                                   | ± 1 kV line(s) to line(s)<br>± 2 kV line(s) to ground                                                                                                                                                                                       | ± 1 kV line(s) to line(s)<br>± 2 kV line(s) to ground                                                                                                                                                                                       | The power supply quality should be that of a typical commercial or hospital environment.                                                                                                                                                                                                                                |
| Voltage dips, short interruptions, and fluctuations of the power supply acc. to IEC 61000-4-11 | <u>Voltage dip:</u><br>Dip to 0% for 1 cycle @ 0° phase angle<br>Dip to 70% for 25/30 cycles @ 0° phase angle<br>Dropout to 0% for 0.5 cycles @ 0°, 45°, 90°, 135°, 180°, 225°, 270°, and 315° phase angles<br><u>Voltage interruption:</u> | <u>Voltage dip:</u><br>Dip to 0% for 1 cycle @ 0° phase angle<br>Dip to 70% for 25/30 cycles @ 0° phase angle<br>Dropout to 0% for 0.5 cycles @ 0°, 45°, 90°, 135°, 180°, 225°, 270°, and 315° phase angles<br><u>Voltage interruption:</u> | The power supply quality should be that of a typical commercial or hospital environment.<br><br>If the user of the equipment requires continued operation in the event of interruptions to the power supply network, it is recommended that the equipment be powered from an uninterruptible power supply or a battery. |

| Interference immunity tests                                                                 | EN/IEC 60601-1-2 test level                                                                                                                                                       | Compliance level                                                                                                                            | Electromagnetic environment – guidance                                                                                                                                                                                                                                                |
|---------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|                                                                                             | 100% for 250/300 cycles                                                                                                                                                           | 100% for 250/300 cycles                                                                                                                     |                                                                                                                                                                                                                                                                                       |
| Magnetic field at the supply frequency (50 Hz/60 Hz) acc. to IEC 61000-4-8                  | 30 A/m at 50 Hz/60 Hz                                                                                                                                                             | 30 A/m at 50 Hz/60 Hz                                                                                                                       | If image distortion occurs, it may be necessary to position the equipment further from sources of power frequency magnetic fields or to install magnetic shielding. Before installing the device, the electromagnetic field should be measured to ensure that it is sufficiently low. |
| Immunity test acc. to IEC 61000-4-3 for high-frequency electromagnetic fields               | 3 V/m 80 MHz to 2.7 GHz<br>see chapter <i>Table 2 – Test levels for proximity fields from RF wireless communications equipment</i> [p. 36] for wireless HF near field test levels | 3 V/m 80 MHz to 2.7 GHz                                                                                                                     | -                                                                                                                                                                                                                                                                                     |
| Immunity to conducted disturbances, induced by radio-frequency fields acc. to IEC 61000-4-6 | 3 V <sub>rms</sub> on 150 kHz to 80 MHz<br>1 kHz 80% AM modulation<br>6 V <sub>rms</sub> in ISM frequency bands between 0.15 MHz and 80 MHz                                       | 3 V <sub>rms</sub> on 150 kHz to 80 MHz<br>1 kHz 80% AM modulation<br>6 V <sub>rms</sub> in ISM frequency bands between 0.15 MHz and 80 MHz | -                                                                                                                                                                                                                                                                                     |
| Magnetic field in close proximity, IEC 61000-4-39                                           | 8 A/m @ 30 kHz (CW modulation)<br>65 A/m @ 134.2 kHz (pulse modulation)<br>7.5 A/m @ 13.56 kHz (pulse modulation)                                                                 | 8 A/m @ 30 kHz (CW modulation)<br>65 A/m @ 134.2 kHz (pulse modulation)<br>7.5 A/m @ 13.56 kHz (pulse modulation)                           | -                                                                                                                                                                                                                                                                                     |



**9.3.2 Table 2 – Test levels for proximity fields from RF wireless communications equipment**

| Test frequency MHz | Frequency band <sup>a)</sup> MHz | Radio service <sup>a)</sup>                                     | Modulation                                                | Immunity test level V/m | Compliance level V/m |
|--------------------|----------------------------------|-----------------------------------------------------------------|-----------------------------------------------------------|-------------------------|----------------------|
| 385                | 380–390                          | TETRA 400                                                       | Pulse modulation <sup>b)</sup> 18 Hz                      | 27                      | 27                   |
| 450                | 430–470                          | GMRS 460, FRS 460                                               | FM <sup>c)</sup> $\pm 5$ kHz deviation<br>1 kHz sine wave | 28                      | 28                   |
| 710                | 704–787                          | LTE band 13 & 17                                                | Pulse modulation <sup>b)</sup> 217 Hz                     | 9                       | 9                    |
| 745                |                                  |                                                                 |                                                           |                         |                      |
| 780                |                                  |                                                                 |                                                           |                         |                      |
| 810                | 800–960                          | GSM 800/900, TETRA 800, iDEN 820, CDMA 850, LTE band 5          | Pulse modulation <sup>b)</sup> 18 Hz                      | 28                      | 28                   |
| 870                |                                  |                                                                 |                                                           |                         |                      |
| 930                |                                  |                                                                 |                                                           |                         |                      |
| 1720               | 1700–1990                        | GSM 1800, CDMA 1900, GSM 1900, DECT, LTE band 1, 3, 4, 25, UMTS | Pulse modulation <sup>b)</sup> 217 Hz                     | 28                      | 28                   |
| 1845               |                                  |                                                                 |                                                           |                         |                      |
| 1970               |                                  |                                                                 |                                                           |                         |                      |
| 2450               | 2400–2570                        | Bluetooth, WLAN 802.11 b/g/n, RFID 2450, LTE band 7             | Pulse modulation <sup>b)</sup> 217 Hz                     | 28                      | 28                   |
| 5240               | 5100–5800                        | WLAN 802.11 a/n                                                 | Pulse modulation <sup>b)</sup> 217 Hz                     | 9                       | 9                    |
| 5500               |                                  |                                                                 |                                                           |                         |                      |
| 5785               |                                  |                                                                 |                                                           |                         |                      |

If necessary to achieve the immunity test level, the distance between the transmitting antenna and the product may be reduced to 1 meter. The 1 meter test distance is permitted by IEC 61000-4-3.

<sup>a)</sup> For some radio services, only the uplink frequencies are included.

<sup>b)</sup> The carrier shall be modulated using a 50% duty cycle square wave signal.

<sup>c)</sup> As an alternative to FM modulation, the carrier may be pulse modulated using a 50% duty cycle square wave signal at 18 Hz. While it does not represent actual modulation, it would be worst case.

**9.3.3 Table 3 – Emission class and group**
**Guidelines and manufacturer's declaration – electromagnetic emissions**

The product is intended for use in such an environment as specified below. The customer or user of the device should ensure that it is used in such an environment.



| Interference emission measurements                           | Conformity | Electromagnetic environment – Guidelines                                                                                                                                                                                |
|--------------------------------------------------------------|------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| RF emissions according to CISPR 11                           | Group 1    | The product uses RF energy for its internal function only. Therefore, its RF emissions are very low and are not likely to cause any interference affecting nearby electronic equipment.                                 |
| RF emissions according to CISPR 11                           | Class B    | The product is suitable for use in all establishments including domestic establishments and those directly connected to the public low voltage power supply network that supplies buildings used for domestic purposes. |
| Harmonic emissions acc. to IEC 61000-3-2                     | Class A    |                                                                                                                                                                                                                         |
| Voltage fluctuations/flicker emissions acc. to IEC 61000-3-3 | Compliant  |                                                                                                                                                                                                                         |

## 10 Errors and messages

### 10.1 Troubleshooting

| Fault                                                  | Possible causes                                      | Actions                                                                                                                 |
|--------------------------------------------------------|------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------|
| Product failed                                         | Power supply failure                                 | ▶ Have the power supply checked                                                                                         |
|                                                        | Defective fuse                                       | ▶ Replace fuses as described in the instruction manual. Make sure to use the correct fuse type                          |
|                                                        | Power plug and socket improperly connected           | ▶ Push the power plug firmly into the socket on the product                                                             |
| No light emission                                      | Electronics faulty                                   | ▶ Contact Service                                                                                                       |
|                                                        | Overheating due to covered air vents                 | ▶ Uncover air vents<br>▶ Switch off the product and let it cool down (10 – 15 min)<br>▶ Ensure adequate air circulation |
| No light emission, power switch lit (on)               | Significant vibration during operation               | ▶ Switch the product off and back on again                                                                              |
|                                                        | Power supply unit or LED defective                   | ▶ Send product to KARL STORZ for repair                                                                                 |
| Insufficient light                                     | Soiled end faces of the light cable or the endoscope | ▶ Clean the end faces of the light cable and the light outlet surfaces of the endoscope                                 |
|                                                        | Light cable or endoscope defective                   | ▶ Replace the light cable or endoscope                                                                                  |
|                                                        | LED service life has been exceeded                   | ▶ Send product to KARL STORZ for repair                                                                                 |
| Light intensity regulation displays “erratic behavior” | Internal error with input preparation                | ▶ Touch the <b>Standby</b> button once to reset to “Normal behavior”                                                    |

## 11 Overview of mitigating warnings

The original English warning text is as follows:

|                                                                                                |                                                                                                                                                                                                                                                                                                                                            |
|------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <br>WARNING   | <p>To avoid the risk of electric shock, this equipment must only be connected to a supply mains with protective earth.</p>                                                                                                                                                                                                                 |
| <br>WARNING   | <p>Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the [ME EQUIPMENT or ME SYSTEM], including cables specified by the manufacturer. Otherwise, degradation of the performance of this equipment could result.</p> |
| <br>WARNING   | <p>Use of this equipment adjacent to or stacked with other equipment should be avoided because it could result in improper operation. If such use is necessary, this equipment and the other equipment should be observed to verify that they are operating normally.</p>                                                                  |
| <br>WARNING  | <p>Use of accessories, transducers and cables other than those specified or provided by the manufacturer of this equipment could result in increased electromagnetic emissions or decreased electromagnetic immunity of this equipment and result in improper operation.</p>                                                               |
| <br>WARNING | <p>No modification of this equipment is allowed.</p>                                                                                                                                                                                                                                                                                       |

## 12 Subsidiaries

### KARL STORZ SE & Co. KG

Dr.-Karl-Storz-Straße 34, 78532 Tuttlingen/Germany  
Postfach 230, 78503 Tuttlingen/Germany  
Phone: +49 7461 708-0, Fax: +49 7461 708-105  
E-Mail: info@karlstorz.com

### KARL STORZ Endoskope Berlin GmbH

Scharnhorststr. 3, 10115 Berlin/Germany  
Phone: +49 30 30 69090, Fax: +49 30 30 19452

### KARL STORZ Endoscopy Canada Ltd.

7171 Millcreek Drive, Mississauga, Ontario L5N 3R3 Canada  
Phone: +1 905 816-4500, Fax: +1 905 816-4599  
Toll free phone: 1-800-268-4880 (Canada only)  
Toll free fax: 1-800-482-4198 (Canada only)  
E-Mail: info-canada@karlstorz.com

### KARL STORZ Endoscopy-America, Inc.

2151 East Grand Avenue, El Segundo, CA 90245-5017, USA  
Phone: +1 424 218-8100, Fax: +1 424 218-8525  
Toll free phone: 800 421-0837 (USA only)  
Toll free fax: 800 321-1304 (USA only)  
E-Mail: communications@ksea.com

### KARL STORZ Veterinary Endoscopy-America, Inc.

1 South Los Carneros Road, Goleta, CA 93117, USA  
Phone: +1 805 968-7776, Fax: +1 805 685-2588  
E-Mail: info@karlstorzvet.com

### KARL STORZ Endoscopia Latino-America, Inc.

815 N. W. 57th Avenue, Suite 480, Miami, FL 33126-2042, USA  
Phone: +1 305 262-8980, Fax: +1 305 262-8986  
E-Mail: info@ksela.com

### KARL STORZ Endoscopia México S.A. de C.V.

Av. Ejercito Nacional No. 453 Piso 2, Colonia Granada, Alcaldia Miguel Hidalgo, C.P. 11520 Ciudad de México  
Phone: +52 (55) 1101 1520  
E-Mail: mx-info@karlstorz.com

### KARL STORZ Marketing América Do Sul Ltda.

Rua Joaquim Floriano, nº. 413, 20º andar – Itaim Bibi, CEP-04534-011 São Paulo, Brasil  
Phone: +55 11 3526-4600, Fax: +55 11 3526-4680  
E-Mail: br-info@karlstorz.com

### KARL STORZ Endoscopia Argentina S.A.

Zufriategui 627 6º Piso, B1638 CAA - Vicente Lopez, Provincia de Buenos Aires, Argentina  
Phone: +54 11 4718 0919, Fax: +54 11 4718 2773  
E-Mail: info@karlstorz.com.ar

### KARL STORZ Endoskopi Norge AS

Stamveien1, 1483 Hagan, Norway  
Phone: +47 6380 5600, Fax: +47 6380 5601  
E-Mail: post@karlstorz.no

### KARL STORZ Endoskop Sverige AB

Storsätragränd 14, 127 39 Skärholmen, Sweden  
Phone: +46 8 505 648 00  
E-Mail: kundservice@karlstorz.se

### KARL STORZ Endoscopy Suomi OY

Taivaltie 5, 01610 Vantaa, Finland  
Phone: +358 (0)96824774, Fax: +358 (0)968247755  
E-Mail: asiakaspalvelu@karlstorz.fi

### KARL STORZ SE & Co. KG

Representative Office  
Žalgirio St. 94, LT9300 Vilnius, Lithuania  
Phone: +370 5 272 0448, Mobile: +370 685 67 000  
E-Mail: info-lt-lv@karlstorz.com

### KARL STORZ Endoskopi Danmark A/S

Skovlytoften 33, 2840 Holte, Denmark  
Phone: +45 45162600, Fax: +45 45162609  
E-Mail: marketing@karlstorz.dk

### KARL STORZ Endoscopy (UK) Ltd.

415 Perth Avenue, Slough, Berkshire, SL1 4TQ, United Kingdom  
Phone: +44 1753 503500, Fax: +44 1753 578124  
E-Mail: info-uk@karlstorz.com

### KARL STORZ Endoscopie Nederland B. V.

Displayweg 2, 3821 BT Amersfoort, Netherlands  
Phone: +31 (0)33 4545890  
E-Mail: info-nl@karlstorz.com

### KARL STORZ Endoscopy Belgium N. V.

Phone: +31 (0)33 4545890  
E-Mail: info-be@karlstorz.com

### KARL STORZ Endoscopie France S. A. S.

12, rue Georges Guynemer, Quartier de l'Europe, 78280 Guyancourt, France  
Téléphone: +33 1 30484200, Fax: +33 1 30484201  
E-Mail: marketing-fr@karlstorz.com

### KARL STORZ Endoskop Austria GmbH

Landstraßer Hauptstr. 148/1/G1, 1030 Wien, Austria  
Phone: +43 1 71 56 0470, Fax: +43 1 71 56 0479  
E-Mail: storz-austria@karlstorz.com

### KARL STORZ Endoscopia Ibérica S. A.

Parque Empresarial San Fernando, Edificio Munich – Planta Baja, 28830 Madrid, Spain  
Phone: +34 91 6771051, Fax: +34 91 6772981  
E-Mail: info-es@karlstorz.com

### KARL STORZ Endoscopia Italia S. r. l.

Via dell'Artigianato, 3, 37135 Verona, Italy  
Phone: +39 045 8222000, Fax: +39 045 8222001  
E-Mail: info-ita@karlstorz.com

### KARL STORZ Croatia d.o.o.

Capraška 6, 10000 Zagreb, Croatia  
Phone: +385 1 6406 070, Fax: +385 1 6406 077  
E-Mail: info-hrv@karlstorz.com

### KARL STORZ Endoskopija d.o.o.

Cesta v Gorice 34b, 1000 Ljubljana, Slovenia  
Phone: +386 1 620 5880, Fax: +386 1 620 5882  
E-Mail: pisarna@karlstorz.si

### KARL STORZ Polska Sp. z o.o.

ul. Hołubcowa 123, 02-854 Warszawa, Poland  
Phone: +48 22 2458 200, Fax: +48 22 2458 201  
E-Mail: info-pl@karlstorz.com

### KARL STORZ Endoszkóp Magyarország Kft.

Toberek utca 2. fsz. 17/b, HU-1112 Budapest, Hungary  
Phone: +36 195 096 31, Fax: +36 195 096 31  
E-Mail: info-hu@karlstorz.com

### KARL STORZ Endoscopia Romania srl

Str. Prof. Dr. Anton Colorian, nr. 74, Sector 4, 041393 Bukarest, Romania  
Phone: +40 (0)31 4250800, Fax: +40 (0)31 4250801  
E-Mail: info-ro@karlstorz.com

### KARL STORZ Endoskope Greece M.E.P.E.\*

Patriarhou Grigoriou E' 34, 54248 Thessaloniki, Greece  
Phone: +30 2310 304868, Fax: +30 2310 304862  
E-Mail: info-gr@karlstorz.com  
\*Repair & Service Subsidiary

### KARL STORZ Industrial\*\*

Gedik Is Merkezi B Blok, Kat 5, D 38-39, Bagdat Cad. No: 162, Maltepe Istanbul, Turkey  
Phone: +90 216 442 9500, Fax: +90 216 442 9030  
\*\*Sales for Industrial Endoscopy



**000 KARL STORZ Endoscopy – WOSTOK**

Derbenyevskaya nab. 7, building 4, 115114 Moscow, Russia  
Phone: +7 495 983 02 40, Fax: +7 495 983 02 41  
E-Mail: Info-ru@karlstorz.com

**TOV LLC KARL STORZ Ukraine**

Avenue Geroyiv Stalingrada Str. 2D, office 717 Kyiv, 04210/Ukraine  
Phone: +38 095 000-895-0, +38-097-000-895-0, +38 073 000-895-0  
E-Mail: marketing@karlstorz.com.ua

**KARL STORZ SE & Co. KG Representation Office**

Sabit Orudschow 1184, apt. 23, 1025 Baku, Azerbaijan  
Phone: +99 450 613 30 60  
E-Mail: info-az@karlstorz.com

**KARL STORZ ENDOSKOPE – East Mediterranean and Gulf (Offshore) S.A.L.**

Spark Tower 1st floor Charles Helou St., Horch Tabet – Sin El Fil, Beirut, Lebanon  
Phone: +961 1 501105, Fax: +961 1 501950  
E-Mail: info@karlstorz-emg.com

**KARL STORZ Endoscopy (South Africa) (Pty) Ltd.**

P.O. 6061, Roggebaai, 8012 Cape Town, South Africa  
Phone: +27 21 417 2600, Fax: +27 21 421 5103  
E-Mail: info@karlstorz.co.za

**T00 KARL STORZ Endoscopy Kasachstan**

Saryarka, 6, BC "Arman", off. 910, 010000 Astana, Republic of Kazakhstan  
Phone: +7 7172 552-549, 552-788, Fax: -444  
E-Mail: info@karlstorz.kz

**KARL STORZ ENDOSKOPE East Mediterranean & Gulf (branch)**

Building West Side 7A – Unit 7WA – 3008, Dubai Airport Free Zone, P.O. Box 54983, Dubai - United Arab Emirates  
Phone: +971 (0)4 2958887, Fax: +971 (0)4 3205282  
Service Hotline: +971 (0)4 3415882  
E-Mail: info-gne@karlstorz-emg.com

**KARL STORZ Endoscopy India Private Limited**

11th Floor, Dr. Gopal Das Bhawan, 28, Barakhamba Road, New Delhi 110001, India  
Phone: +91 11 4374 3000, Fax: +91 11 4374 3010  
E-Mail: corporate@karlstorz.in

**KARL STORZ SE & CO. KG**

Interchange 21 Tower, Level 32, Unit 3230, 399 Sukhumvit Road, North Klongtoey, Wattana, 10110 Bangkok, Thailand  
Phone: +66 2 660 3669  
E-Mail: info-th@karlstorz.com

**KARL STORZ SE & Co. KG**

Resident Representative Office  
14th Floor, MPlaza Saigon, 39 Le Duan, District 1, Ho Chi Minh City, Vietnam  
Phone: +84 28 3823 8000, Fax: +84 28 3823 8039  
E-Mail: infovietnam@karlstorz.com

**KARL STORZ Endoscopy China Ltd.**

Room 2503-05, 25F AXA Tower, Landmark East, No. 100 How Ming Street, Kwun Tong, Kowloon, Hong Kong, People's Republic of China  
Phone: +852 28 65 2411, Fax: +852 28 65 4114  
E-Mail: inquiry@karlstorz.com.hk

**KARL STORZ Endoscopy (Shanghai) Ltd., Beijing Branch**

Room 1805-1807, Building B, 18F Beijing IFC, No. 8, Jianguomenwai Street, Chaoyang District, 100022, Beijing, People's Republic of China  
Phone: +86 10 5638188, Fax: +86 10 5638199  
E-Mail: info@karlstorz.com.cn

**KARL STORZ Endoscopy (Shanghai) Ltd., Shanghai Branch**

Room 701A Building 5 & Room 501 Building 7, No. 3000 Longdong Avenue, Pilot Free Trade Zone, 201203, Shanghai, People's Republic of China  
Phone: +86 21 60339888, Fax: +86 21 60339808  
E-Mail: info@karlstorz.com.cn

**KARL STORZ Endoscopy (Shanghai) Ltd., Chengdu Branch**

Room 803-805, 8F Jin Jiang International Building, No. 1 West Linjiang Road, Wuhou District, 610041, Chengdu, People's Republic of China  
Phone: +86 28 86587977, Fax: +86 28 86587975  
E-Mail: info@karlstorz.com.cn

**KARL STORZ Endoscopy (Shanghai) Ltd., Shenyang Branch**

Room 2001-2005, 20F N-MEDIA International Center, No. 167 Youth Avenue, Shenhe District, 110014, Shenyang, People's Republic of China  
Phone: +86 24 23181118, Fax: +86 24 23181119  
E-Mail: info@karlstorz.com.cn

**KARL STORZ Endoscopy (Shanghai) Ltd., Guangzhou Branch**

Room 02B & 03 & 04A, 35F Teem Tower, No. 208 Tianhe Road, Tianhe District, 510620, Guangzhou, People's Republic of China  
Phone: +86 20 87321281, Fax: +86 20 87321286  
E-Mail: info@karlstorz.com.cn

**KARL STORZ Endoscopy Asia Marketing Pte Ltd.**

No. 8 Commonwealth Lane #03-02, Singapore 149555, Singapore  
Phone: +65 69229150, Fax: +65 69229155  
E-Mail: infoasia@karlstorz.com

**KARL STORZ Endoscopy Singapore Sales Pte Ltd**

No. 8 Commonwealth Lane #03-02, Singapore 149555, Singapore  
Phone: +65 69229150, Fax: +65 69229155  
E-Mail: infoasia@karlstorz.com

**KARL STORZ SE & Co. KG Representative Office Indonesia**

Sinarماس MSIG Tower Level 37, Jl. Jend. Surdirman No. Kav. 21, Jakarta Selatan DKI Jakarta 12920  
E-Mail: infoindonesia@karlstorz.com

**KARL STORZ Endoscopy Korea Co. Ltd.**

9F Hyowon-Building, 97, Jungdae-ro, Songpa-gu, 05719 Seoul, Korea  
Phone: +82-70-4350-7474, Fax: +82-70-8277-3299  
E-Mail: infokorea@karlstorz.com

**KARL STORZ Endoscopy Taiwan Ltd.**

12F, No. 192, Sec. 2, Chung Hsin Rd., Sindian District, New Taipei City, Taiwan  
Phone: +886 933 014 160, Fax: +886 2 8672 6399  
E-Mail: info-tw@karlstorz.com

**KARL STORZ Endoscopy Philippines Inc. Representative Office Philippines**

1901 Picadilly Star Bldg., 4th Avenue, BGC, Taguig City 1636, Philippines  
Phone: +63 2 317 45 00, Fax: +63 2 317 45 11  
E-Mail: philippines@karlstorz.com

**KARL STORZ Endoscopy Japan K. K.**

Stage Bldg. 8F, 2-7-2 Fujimi, Chiyoda-ku, Tokyo 102-0071, Japan  
Phone: +81 3 6380-8622, Fax: +81 3 6380-8633  
E-Mail: info-jp@karlstorz.com

**KARL STORZ Endoscopy New Zealand**

Ltd. 31 Morningside Drive Mt Albert Auckland, 1025, New Zealand PO Box 56 511, Dominion Rd Auckland, 1446, New Zealand  
Phone: +64 9 846 6044  
Toll free: +64 508 84 84 84 (New Zealand only)  
E-Mail: sales-nz@karlstorz.com

**KARL STORZ Endoscopy Australia Pty. Ltd.**

Suite 1, 68-72 Waterloo Road, Macquarie Park NSW 2113, Australia, PO Box 50 Lane Cove NSW 1595, Australia  
Phone: +61 (0)2 9490 6700, Fax: +61 (0)2 9420 0695  
Toll free: 1800 996 562 (Australia only)  
E-Mail: karlstorz@karlstorz.com.au

**KARL STORZ Endoscopy (UK) Limited**

415 Perth Avenue, Slough, Berkshire, SL1 4TQ, United Kingdom  
Phone: +44 (0)1753 503500  
E-Mail: info-uk@karlstorz.com

www.karlstorz.com









W XD587 • EN • V2.0 • 05-2024 • IFU • CE-MDR



## **KARL STORZ SE & Co. KG**

Dr.-Karl-Storz-Straße 34  
78532 Tuttlingen

Postfach 230  
78503 Tuttlingen  
Germany

Phone: +49 7461 708-0  
Fax: +49 7461 708-105  
E-mail: [info@karlstorz.com](mailto:info@karlstorz.com)  
[www.karlstorz.com](http://www.karlstorz.com)

