

CERTIFICATE OF ANALYSIS

Certificado de Análisis / Certificat de contrôle / Chargen-Prüfprotokoll

Pourquier® Brucellosis Negative Control [Brucellosis Negative Control]

Product code :	P00131	Expiration date :	08-SEP-2019
Batch :	144		
Tests per kit :	1		
Insert version :	0	Manufacturing date :	SEP-2014

Reconstitute* with 1ml distilled water

* After reconstitution, the control serum must be stored at a temperature $\leq -16^{\circ}\text{C}$ for up to 1 year

He can be frozen and unfrozen up to three times

Control Test	Spécifications	Results
EAT :	Negative : No agglutination	Pass
SAW :	Negative(< 30 UI/ml)	Pass
FC :	Negative(< 20 ICFTU/ml)	Pass
ELISA :	Negative	Pass

This product was performance tested and has met all quality control specifications required for release.

Title :

Signature :

Sélectionner

Sélectionner

Manufacturer: IDEXX Montpellier SAS - 326, rue de la Galéra - Parc Euromédecine - 34097 Montpellier Cedex 5 - France
 Tél. : 33(0) 4 99 23 24 25 - Fax : 33(0) 4 67 04 20 25 - info-montpellier@idexx.com
 EN 05/01/17 vB - Date 06/08/2012

Document printed on : 30/10/2014