

ACON BIOTECH (HANGZHOU) CO.,LTD

CERTIFICATE OF ANALYSIS

PRODUCT NAME: CMV IgG EIA Test Kit

PURCHASE NO.: 37740064A

BATCH NO.: 1911652

EXPIRATION DATE: 2021/06/18

QUANTITY: 20kits

SPECIFICATION	STANDARD	CONCLUSION	
Appearance	No obvious precipitates or floccules in liquid component and the color meets requirement.	Complies	
Blank Well	$OD < 0.050(450nm/630nm)$ or $OD < 0.100(450nm)$	Complies	
Calibrator 1#	$OD < 0.150$	Complies	
Calibrator 2#	$OD > 0.150$	Complies	
Calibrator 3#	Calibrator 2# $OD <$ Calibrator 3# $OD <$ Calibrator 4# OD	Complies	
Calibrator 4#	$OD > 1.200$	Complies	
R Value	$R \geq 0.98$	Complies	
Sensitivity Standard	L	$OD > 0.150$	Complies
	M	$OD \text{ of L} < OD < OD \text{ of H}$	Complies
	H	$OD > 1.000$	Complies
Compliance of Negative Standard	$OD < \text{Calibrator 2\#} OD$	Complies	
Precision Standard	$CV < 15\%$, $OD > 1.1 * \text{Calibrator 2\#} OD$	Complies	

Quality Department: Antibody. Hu

Date: Apr. 02, 2020