

**HiMedia Laboratories Private Limited**

23, Vadhani Industrial Estate, L.B.S. Marg, Mumbai - 400086

Website : www.himedialabs.com, Email : info@himedialabs.com

**Certificate of Analysis , Quality and Conformity**

Material Code : SD016	Material Name : Gentamicin GEN 10 mcg	Lot No : 0000417600
AR No.: 040000983368	Date of Report : 2020-01-03	Exp. Date : 2021-12

TEST	SPECIFICATIONS	RESULTS
<b><u>Appearance</u></b> 1 Appearance	Filter paper discs of 6 mm diameter with printed "GEN 10" on centre of each side of the disc.	Complies
<b><u>Cultural response</u></b> 1 Cultural response	Average diameter of zone of inhibition observed on Mueller Hinton Agar (M173) after 18 hours incubation at 35-37°C for standard cultures.	
<b><u>Escherichia coli ATCC 25922</u></b> 1 Zone of inhibition (mm)	19mm - 26mm	25mm
<b><u>Staphylococcus aureus ATCC 25923</u></b> 1 Zone of inhibition (mm)	19mm - 27mm	26mm
<b><u>Pseudomonas aeruginosa ATCC 27853</u></b> 1 Zone of inhibition (mm)	17mm - 23mm	21mm
<b><u>Staphylococcus aureus ATCC 29213</u></b> 1 Zone of inhibition (mm)*	19mm - 25mm	24mm

The performance of this batch has been tested and standardized as per the CLSI as well as EUCAST Standards. \*QC Limit as per EUCAST Standard

**STATUS OF THE MATERIAL : APPROVED**

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current HiMedia manual or product sheets.




**This document has been produced electronically and is valid**

**HiMedia Laboratories Private Limited**

23, Vadhani Industrial Estate, L.B.S. Marg, Mumbai - 400086

Website : [www.himedialabs.com](http://www.himedialabs.com), Email : [info@himedialabs.com](mailto:info@himedialabs.com)

<b>Material Code :</b> SD016	<b>Material Name :</b> Gentamicin GEN 10 mcg	<b>Lot No</b> : 0000417600
<b>AR No.:</b> 040000983368	<b>Date of Report :</b> 2020-01-03	<b>Exp. Date :</b> 2021-12

TEST	SPECIFICATIONS	RESULTS
 Sukanya Deshmukh Microbiologist/Sr.Executive Microbiologist	 Ujwala M. Kokate Asst./Dy/QC Manager	 Dr. Santosh Kaul Dy/QA Manager