

## Certificate of conformity to standards IEC 60118-0

This certificate confirms that OTICON hearing aid

(SIYA 2 BTE 13 2.4 G 85) presented

for LP no. MTender ID **ocds-b3wdp1-MD-1653050230290**

(Acquisition of medical devices (hearing aids), in order to achieve the National Program on hearing aids, for 2022 (repeatedly))

have been measured according to IEC 60118-0 (Measurement of the performance characteristics of hearing aids) and international standards.

SIYA 2 BTE 13 2.4 G 85

Maximum output: 130 dB;

Maximum gain: 66 dB;

Frequency range: 105 Hz - 7500 Hz;

Noise level (omni): 21 dB;

Harmonic distortion coefficient:

500 Hz - <2%

800 Hz - 2%

1600 Hz - <2%