

ACON BIOTECH (HANGZHOU) CO.,LTD

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Mission Liquid Urine Control for combo pack

PURCHASE NO.: 37740068A

BATCH NO.: UCL0090001

EXPIRATION DATE: 2022-06-22

QUANTITY: 70kits

ANALYSIS		SPECIFICATIONS	TEST RESULT
Specific gravity	Level 1	1.015~1.030	Complies
	Level 2	1.005~1.020	Complies
pH	Level 1	5.0~7.0	Complies
	Level 2	6.5~9.0	Complies
Glucose	Level 1	Negative	Complies
	Level 2	$\geq +/-(100\text{mg/dl})$	Complies
Protein	Level 1	Negative	Complies
	Level 2	$\geq 1+$ (30mg/dl)	Complies
Leukocyte	Level 1	Negative	Complies
	Level 2	$\geq 1+$ (70 Leu/ μl)	Complies
Nitrite	Level 1	Negative	Complies
	Level 2	Positive	Complies
Urobilinogen	Level 1	0.2(3.5) mg/dL($\mu\text{mol/L}$)	Complies
	Level 2	$\geq +/-(1.0\text{mg/dl})$	Complies
Blood	Level 1	Negative	Complies
	Level 2	$\geq 2+$	Complies
Ketone	Level 1	Negative	Complies
	Level 2	$\geq 1+$ (15mg/dl)	Complies
Bilirubin	Level 1	Negative	Complies
	Level 2	$\geq 1+$ (1mg/dl)	Complies
Ascorbic acid	Level 1	Negative	Complies
	Level 2	Negative	Complies
Albumin	Level 1	10~30mg/L	Complies
	Level 2	80~150mg/L	Complies
Creatinine	Level 1	10~100mg/dL	Complies
	Level 2	100~300mg/dL	Complies
Urine Calcium	Level 1	4~10mg/dL	Complies
	Level 2	20~40mg/dL	Complies

Quality Department: Hedy Zhang

Date: Sep. 24, 2020