



Monitor EIZO EV2460-BK 23.8" Color

Product Data

Monitor features

Display technology	IPS
Backlight	LED
Active screen size (diagonal)	60.5 cm (23.8")
Resolution	1920 × 1080 pixels
Aspect ratio (H:V)	16:9
Active screen size (H x V)	527 × 296.5 mm (H x V)
Pixel Pitch	0,275 mm
Pixel density	93 ppi
Display colors	16.77 million
Visibility angle	178°
Brightness (typical)	250 cd/m ²
Contrast ratio	1000:1
Response time (typical)	5 ms
Max refresh frequency	60 Hz
Compatibility	Compatible with VDX systems of Villa Sistemi Medicali

Connections

Video input	DisplayPort (HDCP 1.3) HDMI (HDCP 1.4) DVI-D (HDCP 1.4) D-Sub mini 15 pin
USB specifications	USB 3.1 Gen 1
USB port upstream	1x Type-B
USB port downstream	4x Type-A
Audio speakers	Integrated
Audio input terminals	Stereo mini jack, DisplayPort, HDMI
Audio output terminals	Headphones (Stereo mini jack)
Digital Scanning Frequency (H/V)	Dsub: 31-68 kHz/59-61 Hz; Display Port: 31-68 kHz/59-61 Hz; DVI-D: 31-68 kHz/59-61 Hz; HDMI: 49-51 Hz/59-61 Hz
Analog Scanning Frequency (H/V)	31 - 68 kHz / 59 - 61 Hz



Electrical features

Power requirements	AC 100-120 V / 200-240 V , 50 / 60 Hz AC
Power consumption	50 W
Typical power consumption	10 W
Energy-saving mode consumption	< 0,5 W

Mechanical features


Dimensions (Landscape, W x H x D)	537,8 x 333,6 - 473,6 x 233 mm
Dimensions (Portrait, W x H x D)	320,4 x 550,3 - 588,9 x 233 mm
Net weight	5,9 kg
Net weight without stand	3,8 kg
Tilt	344°
Pivot	90° (clockwise and anti-clockwise)
Downward/upward tilting	5° / 35°
Height adjustment range	140 mm
Hole spacing (VESA Standard)	100 x 100 mm

Standards and regulations

CE, CB, TCO Certified Generation 9, EPA ENERGY STAR ®, TÜV/GS, TÜV/Ergonomics (including ISO 9241-307), TÜV/Blue Low Light Content, TÜV/Flicker Free, EPEAT 2018 (US), cTÜVus, TÜV/S, EAC, FCC-B, CAN ICES-3 (B), VCCI-B, CCC, RoHS, China RoHS, WEEE



Dimensions (all quotes in mm)



Note: Products are continuously under review in the light of technical advancement. The actual specification may therefore be subject to improvement or modification without notice.

VILLA SISTEMI MEDICALI s.p.a.
 20090 BUCCINASCO (MI) – ITALY,
 Via delle Azalee, 3
 Tel. +39-02-488591, Fax +39-02-4881844

Company with Quality System certified by



ISO 9001:2015



ISO 13485:2016