

Certificate of Analysis



Anode Buffer Container 3500 Series

Product No. 4393927
Lot No. 2408964
Date of Manufacture 01OCT2024
Expiration Date 27NOV2025

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION Material is clear, colorless and free of particulates		
Appearance - Conformance		Pass
pH TEST Measured at 25.0°C ± 0.3°C		
pH Value	7.6 - 8.2	7.906
CONDUCTIVITY TEST Measured at 25.0°C ± 0.1°C		
Conductivity	1500 - 2000 µS/cm	1773 µS/cm

For Research Use Only.
Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system.
Site: Bedford, MA, USA

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

Luciana Torcasio
Quality Assurance
Issued 02OCT2024