

EU-Konformitätserklärung/*EU Declaration of Conformity*

Name und Anschrift des Herstellers: MAVIG GmbH
Name and Address of Manufacturer: Stahlgruberring 5
 81829 Munich
 Germany

Für die Ausstellung dieser EU-Konformitätserklärung trägt MAVIG GmbH die alleinige Verantwortung und bestätigt, dass die nachfolgend genannten Produkte die Anforderungen der Verordnung (EU) 2016/425 für Persönliche Schutzausrüstung sowie der Verordnung (EU) 2017/745 für Medizinprodukte erfüllen.

This EU declaration of conformity is issued under the sole responsibility of MAVIG GmbH who confirms that the products mentioned below meet the provision of the Regulation (EU) 2016/425 on personal protective equipment as well as of the Regulation (EU) 2017/745 on medical devices.

	Anforderungen an/ Requirements for:	
	Persönliche Schutzausrüstung/ <i>Personal Protective Equipment</i>	Medizinprodukte/ <i>Medical Devices</i>
Produktbezeichnung/ <i>Product name:</i>	Röntgenschutzkleidung Dental <i>Radiation Protective Clothing Dental</i>	
Modell/ <i>Model:</i>	RD614 + RD635 + RD642 + RD644	
Angewandte Normen/ <i>Applied standards:</i>	IEC 61331-1:2014 IEC 61331-3:2014	
Basis-UDI-DI/ <i>Basic UDI-DI:</i>	04251339900479WE	
Einmalige Registrierungsnummer/ <i>Single registration number (SRN):</i>	N/A	DE-MF-000006207
Risikoklasse/ <i>Risk class:</i>	Kategorie III <i>Category III</i>	Klasse I <i>Class I</i>
EU-Baumusterprüfung durch benannte Stelle/ <i>EU type-examination by notified body:</i>	NB 0302 ANCCP Certification Agency S.R.L. Via dello Struggino, 6 57121 Livorno (LI) – Italy PPE-1206-18319-Cert	
Konformitätsbewertungsverfahren/ <i>Conformity assessment procedure:</i>	Anhang I (Kategorie III, Punkt d); Anhang VIII (Modul D) durch <i>Annex I (Category III, point d); Annex VIII (Module D) by ANCCP Certification Agency S.R.L. Via dello Struggino, 6 57121 Livorno (LI) - Italy</i>	MDR, Art 52 (7)

MAVIG GmbH
 X-Ray Protection and
 Medical Suspension Systems

Company Address:
 Stahlgruberring 5
 81803 Munich
 Germany

Post Box:
 Postfach 820362
 81803 Munich
 Germany

Phone:
 +49 (0) 89 / 420 96 - 0

Fax:
 +49 (0) 89 / 420 96 - 200

e-Mail:
 info@mavig.com

Web:
 www.mavig.com

Bank Accounts:
 HypoVereinsbank Munich
 IBAN: DE80 7002 0270
 1620 2253 00
 SWIFT-BIC: HYVEDEMMXXX

Postbank Munich
 IBAN: DE12 7001 0080
 0071 3748 07
 SWIFT-BIC: PBNKDEFFXXX

Trade-Register Munich
 HRB-No. 76 403

Managing Directors:
 Christian Stoian
 Alexia Stoian-Lepère

VAT-ID No.:
 DE 811268845

Diese Erklärung wurde ausgestellt von/
This statement was issued by:


Irene Madl – Head of Regulatory Management

Ort und Datum der Ausstellung/
Place and date of issue:

Munich, 2021-07-01

Modellvarianten/ Model versions:

RD6145K61KL32	RD6145K62KL32	RD6145K25AK13	RD6145K29AK13	RD6145E29AK12
RD6145K61AK32	RD6145K62AK32	RD6145K25MA13	RD6145E25MA12	RD6145E29MA12
RD6145K61MA32	RD6145K62MA32	RD6145K29KL13	RD6145E29KL12	RD6145E30KL12
RD6145E30AK12	RD6145E22AK13	RD6145E28KL13	RD6145K30KL13	RD6145E32MA12
RD6145E30MA12	RD6145E22MA13	RD6145E28AK13	RD6145K30AK13	RD6145E61KL32
RD6145E31KL12	RD6145E27KL13	RD6145E28MA13	RD6145K30MA13	RD6145E61AK32
RD6145E31AK12	RD6145E27AK13	RD6145E25KL13	RD6145E32KL12	RD6145E61MA32
RD6145E22KL13	RD6145E27MA13	RD6145K29MA13	RD6145E32AK12	RD6145E62KL32
RD6145E62AK32	RD6145E29AK13	RD6145E31KL13	RD6145K32KL13	RD6145K27MA12
RD6145E62MA32	RD6145E29MA13	RD6145E31AK13	RD6145K22AK12	RD6145K28KL12
RD6145E25AK13	RD6145E30KL13	RD6145K31KL13	RD6145K22MA12	RD6145K28AK12
RD6145E25MA13	RD6145E30AK13	RD6145K31AK13	RD6145K27KL12	RD6145K28MA12
RD6145E29KL13	RD6145E30MA13	RD6145K31MA13	RD6145K27AK12	RD6145K25KL12
RD6145E31MA13	RD6145E61MA33	RD6145K61KL33	RD6145K30KL12	RD6145K22AK13
RD6145E32KL13	RD6145E62KL33	RD6145K61AK33	RD6145K30AK12	RD6145K22MA13
RD6145E32AK13	RD6145E62AK33	RD6145K25MA12	RD6145K30MA12	RD6145K27KL13
RD6145E32MA13	RD6145E62MA33	RD6145K29KL12	RD6145K31KL12	RD6145K27AK13
RD6145E61KL33	RD6145K32AK13	RD6145K29AK12	RD6145K31AK12	RD6145K27MA13
RD6145E61AK33	RD6145K32MA13	RD6145K29MA12	RD6145K22KL13	RD6145K28KL13
RD6145K28AK13	RD6145K62AK33	RD6145K25KL13	RD6145K61MA33	RD6145K62KL33
RD6145K28MA13	RD6145K62MA33	RD614-STD	RD614-SON	
RD6355E2212	RD6355E2712	RD6355E2812	RD6355E2512	RD6355E3112
RD6355E2213	RD6355E2713	RD6355E2813	RD6355E2513	RD6355E3113
RD6355E6132	RD6355E2912	RD6355E6232	RD6355E3212	RD6355E3012
RD6355E6133	RD6355E2913	RD6355E6233	RD6355E3213	RD6355E3013
RD6355K2212	RD6355K2712	RD6355K2812	RD6355K2512	RD6355K3112
RD6355K2213	RD6355K2713	RD6355K2813	RD6355K2513	RD6355K3113
RD6355K6132	RD6355K2912	RD6355K6233	RD6355K3212	RD6355K3012
RD6355K6133	RD6355K2913	RD6355K6232	RD6355K3213	RD6355K3013

MAVIG GmbH
X-Ray Protection and
Medical Suspension Systems

Trade-Register Munich
HRB-No. 76 403

Managing Directors:
Christian Stoian
Alexia Stoian-Lepère

VAT-ID No.:
DE 811268845

RD635-STD	RD635-SON	RD642-STD	RD642-SON	
RD6425E2212	RD6425E2712	RD6425E2812	RD6425E2512	RD6425E3112
RD6425E2213	RD6425E2713	RD6425E2813	RD6425E2513	RD6425E3113
RD6425E6132	RD6425E2912	RD6425E6232	RD6425E3212	RD6425E3012
RD6425E6133	RD6425E2913	RD6425E6233	RD6425E3213	RD6425E3013
RD6425K2212	RD6425K2712	RD6425K2812	RD6425K2512	RD6425K3112
RD6425K2213	RD6425K2713	RD6425K2813	RD6425K2513	RD6425K3113
RD6425K6132	RD6425K2912	RD6425K6232	RD6425K3212	RD6425K3012
RD6425K6133	RD6425K2913	RD6425K6233	RD6425K3213	RD6425K3013
RD6445E2212	RD6445E2512	RD6445E2712	RD6445E2812	RD6445E3112
RD6445E2213	RD6445E2513	RD6445E2713	RD6445E2813	RD6445E3113
RD6445E3012	RD6445E4812	RD6445E6132	RD6445E6232	RD6445E2912
RD6445E3013	RD6445E4813	RD6445E6133	RD6445E6233	RD6445E2913
RD6445E4912	RD6445E3212	RD6445K2712	RD6445K2512	RD6445K3112
RD6445E4913	RD6445E3213	RD6445K2713	RD6445K2513	RD6445K3113
RD6445K4812	RD6445K6232	RD6445K2912	RD6445K4912	RD6445K2812
RD6445K4813	RD6445K6233	RD6445K2913	RD6445K4913	RD6445K2813
RD6445K3212	RD6445K2212	RD6445K3012	RD6445K6132	RD644-STD
RD6445K3213	RD6445K2213	RD6445K3013	RD6445K6133	RD644-SON

EU-Konformitätserklärung EU Declaration of Conformity

Name und Anschrift des Herstellers
/Inverkehrbringer/ Name and Address of
Manufacturer/Distributor:

MAVIG GmbH
Stahlgruberring 5
81829 Munich
Germany

Für die Ausstellung dieser EU-Konformitätserklärung trägt MAVIG GmbH die alleinige Verantwortung und bestätigt, dass die nachfolgend genannten Produkte die Anforderungen der Verordnung (EU) 2016/425 für Persönliche Schutzausrüstung erfüllen.

This EU declaration of conformity is issued under the sole responsibility of MAVIG GmbH who confirms that the products mentioned below meet the provision of the Regulation (EU) 2016/425 on personal protective equipment.

Produktbezeichnung/
Product name/ *Röntgenschutzbrillen*
Radiation Protective Glasses

Modell/ *Model* *BR*

Angewandte Normen:
Applied standards: IEC 61331-1:2014
IEC 61331-3:2014

PSA entspricht den Bestimmungen der Verordnung (EU) 2016/425 und die Modelle erfüllen die Anforderungen der technischen Spezifikationen des Herstellers. Dies ist identisch mit der PSA, die Gegenstand der Baumusterprüfbescheinigung 523548/1 ist, ausgestellt von der Benannten Stelle BTTG 0338, Manchester M17 1EH, die die EU-Baumusterprüfung (Modul B) durchgeführt hat, die jetzt von Shirley gehalten wird, **Benannte Stelle 2895**, Port Tunnel Business Park, Office 13 Unit 21, Dublin 17, ROI und unterliegt dem in der Verordnung (EU) 2016/425, Modul-D festgelegten Verfahren unter der Aufsicht der Benannten Stelle Shirley Technologies (Europe) Limited (Sky Business Centre, Port Tunnel Business Park, Office 13 Unit 21, Dublin 17, ROI) durchgeführt bei Protech Leded Eyewear, Inc., 1360 N Killian Dr. Unit 2, Lake Park, Florida 33403.

*PPE is in conformity with the provisions of Regulation (EU) 2016/425 and the models satisfy the requirements of the manufacturer's technical specifications. This is identical to the PPE which is the subject of **Certificate 523548/1** issued by BTTG Notified Body 0338, Manchester M17 1EH whom performed the EU type-examination (Module B), which is now held by **Shirley, Notified Body 2895**, Port Tunnel Business Park, Office 13 Unit 21, Dublin 17, ROI and is subject to the procedure set out in Regulation (EU) 2016/425, Module D under the supervision of the notified body Shirley Technologies (Europe) Limited (Sky Business Centre, Port Tunnel Business Park, Office 13 Unit 21, Dublin 17, ROI), performed at Protech Leded Eyewear, Inc., 1360 N Killian Dr. Unit 2, Lake Park, Florida 33403.*

MAVIG GmbH
X-Ray Protection and
Medical Suspension Systems

Company Address:
Stahlgruberring 5
81829 Munich
Germany

Post Box:
Postfach 820362
81803 Munich
Germany

Phone:
+49 (0) 89 / 420 96 - 0

Fax:
+49 (0) 89 / 420 96 - 200

e-Mail:
info@mavig.com

Web:
www.mavig.com

Bank Accounts:
HypoVereinsbank Munich
IBAN: DE80 7002 0270
1620 2253 00
SWIFT-BIC: HYVEDEMMXXX


Postbank Munich
IBAN: DE12 7001 0080
0071 3748 07
SWIFT-BIC: PBNKDEFFXXX

Trade-Register Munich
HRB-No. 76 403

Managing Directors:
Christian Stoian
Alexia Stoian-Lepère

VAT-ID No.:
DE 811268845

Diese Erklärung wurde ausgestellt von:
This statement was issued by:


 Irene Madi – Head of Regulatory
 Management

Ort und Datum der Ausstellung:
Place and date of issue:

Munich, 2022-02-02

Modellvarianten/ *Model versions:*

BR115-BLACK			
BR118-BLACK	BR118-BLACK-PRESC-S	BR118-BLACK-PRESC-B	BR118-BLACK-PRESC-P
BR118-TAUPE	BR118-TAUPE-PRESC-S	BR118-TAUPE-PRESC-B	BR118-TAUPE-PRESC-P
BR118-BLUE	BR118-BLUE-PRESC-S	BR118-BLUE-PRESC-B	BR118-BLUE-PRESC-P
BR119-BLACK	BR119-BLACK-PRESC-S	BR119-BLACK-PRESC-B	BR119-BLACK-PRESC-P
BR119-SILVER	BR119-SILVER-PRESC-S	BR119-SILVER-PRESC-B	BR119-SILVER-PRESC-P
BR310	BR310-PRESC-S	BR310-PRESC-B	BR310-PRESC-P
BR321	BR321-PRESC-S	BR321-PRESC-B	BR321-PRESC-P
BR322	BR322-PRESC-S	BR322-PRESC-B	BR322-PRESC-P
BR331	BR331-PRESC-S	BR331-PRESC-B	BR331-PRESC-P

MAVIG GmbH
 X-Ray Protection and
 Medical Suspension Systems

Trade-Register Munich
 HRB-No. 76 403

Managing Directors:
 Christian Stoian
 Alexia Stoian-Lepère

VAT-ID No.:
 DE 811268845

EU-Konformitätserklärung *EU Declaration of Conformity*

Name und Anschrift des Herstellers/
Name and Address of Manufacturer:

MAVIG GmbH
Stahlgruberring 5
81829 Munich
Germany

Für die Ausstellung dieser EU-Konformitätserklärung trägt MAVIG GmbH die alleinige Verantwortung und bestätigt, dass die nachfolgend genannten Produkte die Anforderungen der Verordnung (EU) 2016/425 für Persönliche Schutzausrüstung erfüllen.

This EU declaration of conformity is issued under the sole responsibility of MAVIG GmbH who confirms that the products mentioned below meet the provision of the Regulation (EU) 2016/425 on personal protective equipment.

Produktbezeichnung/
Product name/

*Röntgenschutzkleidung Anwender
Radiation Protective Clothing Medical*

Modell/ *Model*

RA614

Angewandte Normen:
Applied standards:

IEC 61331-1:2014
IEC 61331-3:2014

Die benannte Stelle ANCCP Certification Agency srl., Via dello Struggino, 6 – 57121 Livorno (LI) – Italy hat die EU-Baumusterprüfung durchgeführt und die EU-Baumusterprüfbescheinigung PPE-1203-18319-Cert ausgestellt.

The notified body ANCCP Certification Agency srl., Via dello Struggino, 6 – 57121 Livorno (LI) – Italy performed the EU type-examination and issued the EU type-examination certificate PPE-1203-18319-Cert.

Die PSA unterliegt dem in Anhang VIII (Modul D) der Verordnung (EU) 2016/425 vorgesehenen Verfahren der regelmäßigen Überwachung unter der Aufsicht der folgenden Benannten Stelle:

The PPE is subject to the scheduled periodic surveillance procedure set out in Annex VIII (Module D) of Regulation (EU) 2016/425 under the supervision of the following Notified Body:

ANCCP Certification Agency srl., Via dello Struggino, 6 – 57121 Livorno (LI) – Italy – CE0302

Diese Erklärung wurde ausgestellt von:
This statement was issued by:



Irene Madi – Head of Regulatory
Management

Ort und Datum der Ausstellung:
Place and date of issue:

Munich, 2021-08-05

MAVIG GmbH
X-Ray Protection and
Medical Suspension Systems

Company Address:
Stahlgruberring 5
81829 Munich
Germany

Post Box:
Postfach 820362
81803 Munich
Germany

Phone:
+49 (0) 89 / 420 96 - 0

Fax:
+49 (0) 89 / 420 96 - 200

e-Mail:
info@mavig.com

Web:
www.mavig.com

Bank Accounts:
HypoVereinsbank Munich
IBAN: DE80 7002 0270
1620 2253 00
SWIFT-BIC: HYVEDEMMXXX

Postbank Munich
IBAN: DE12 7001 0080
0071 3748 07
SWIFT-BIC: PBNKDEFFXXX

Trade-Register Munich
HRB-No. 76 403

Managing Directors:
Christian Stoian
Alexia Stoian-Lepère

VAT-ID No.:
DE 811268845

Modellvarianten/ *Model versions:*

Beispielhafte Zusammensetzung des Artikelcodes/
Example for the formation of the article code:

RA614 5 22 KL 1 1
 ① ② ③ ④ ⑤

① Bleigleichwert/ Lead Equivalent	
	Code
0,50 mmPb	5

② Farbauswahl/ Colours			
	Code		Code
Curacao	22	Mandarin	30
Lagoon	25	Blue Admiral	31
Orchidee	27	Baccara	32
Regatta	28	Design Nautilus	61
Titan	29		

③ Verschlussart für Halsweiten/ Closure for neck sizes	
	Code
Klettverschluss/ <i>Hook & Burr</i>	KL
Austauschbarer Klettverschluss/ <i>Exchangeable Hook & Burr</i>	AK
Magnetverschluss/ <i>Magnetic</i>	MA

④ Außenmaterial – Hülle/ Outer Material - Cover	
	Code
ComforTex HPMF	1
ComforTex HPMF Hybrid	2
ComforTex HPMF Plus	3

⑤ Strahlenschutzmaterial – Innen/ X-Ray Protective Material - Inlay	
	Code
NovaLite 0,50 mmPb	1
Standard-Blei/ <i>Standard Lead</i>	2
Leadfree	3

MAVIG GmbH
 X-Ray Protection and
 Medical Suspension Systems

Trade-Register Munich
 HRB-No. 76 403

Managing Directors:
 Christian Stoian
 Alexia Stoian-Lepère

VAT-ID No.:
 DE 811268845

EU-Konformitätserklärung EU Declaration of Conformity

Name und Anschrift des Herstellers/ Name and Address of Manufacturer:	MAVIG GmbH Stahlgruberring 5 81829 Munich Germany
--	--

Für die Ausstellung dieser EU-Konformitätserklärung trägt MAVIG GmbH die alleinige Verantwortung und bestätigt, dass die nachfolgend genannten Produkte die Anforderungen der Verordnung (EU) 2016/425 für Persönliche Schutzausrüstung erfüllen.

This EU declaration of conformity is issued under the sole responsibility of MAVIG GmbH who confirms that the products mentioned below meet the provision of the Regulation (EU) 2016/425 on personal protective equipment.

Produktbezeichnung/ Product name/	Röntgenschutzkleidung Anwender <i>Radiation Protective Clothing Medical</i>
--------------------------------------	--

Modell/ Model	RA631
---------------	-------

Angewandte Normen: Applied standards:	IEC 61331-1:2014 IEC 61331-3:2014
--	--------------------------------------

Die benannte Stelle ANCCP Certification Agency srl., Via dello Struggino, 6 – 57121 Livorno (LI) – Italy hat die EU-Baumusterprüfung durchgeführt und die EU-Baumusterprüfbescheinigung PPE-1203-18319-Cert ausgestellt.


The notified body ANCCP Certification Agency srl., Via dello Struggino, 6 – 57121 Livorno (LI) – Italy performed the EU type-examination and issued the EU type-examination certificate PPE-1203-18319-Cert.

Die PSA unterliegt dem in Anhang VIII (Modul D) der Verordnung (EU) 2016/425 vorgesehenen Verfahren der regelmäßigen Überwachung unter der Aufsicht der folgenden Benannten Stelle:

The PPE is subject to the scheduled periodic surveillance procedure set out in Annex VIII (Module D) of Regulation (EU) 2016/425 under the supervision of the following Notified Body:

ANCCP Certification Agency srl., Via dello Struggino, 6 – 57121 Livorno (LI) – Italy – CE0302

Diese Erklärung wurde ausgestellt von:
This statement was issued by:


Irene Madl – Head of Regulatory
Management

Ort und Datum der Ausstellung:
Place and date of issue:

Munich, 2022-03-09

MAVIG GmbH
X-Ray Protection and
Medical Suspension Systems

Company Address:
Stahlgruberring 5
81829 Munich
Germany

Post Box:
Postfach 820362
81803 Munich
Germany

Phone:
+49 (0) 89 / 420 96 - 0

Fax:
+49 (0) 89 / 420 96 - 200

e-Mail:
info@mavig.com

Web:
www.mavig.com

Bank Accounts:
HypoVereinsbank Munich
IBAN: DE80 7002 0270
1620 2253 00
SWIFT-BIC: HYVEDEMMXXX

Postbank Munich
IBAN: DE12 7001 0080
0071 3748 07
SWIFT-BIC: PBNKDEFFXXX

Trade-Register Munich
HRB-No. 76 403

Managing Directors:
Christian Stoian
Alexia Stoian-Lepère

VAT-ID No.:
DE 811268845

Modellvarianten/ Model versions:

Beispielhafte Zusammensetzung des Artikelcodes/
Example for the formation of the article code:

RA631	K	3	2	Z	65	22	2	8
	①	②	③	④	⑤	⑥	⑦	⑧

① Ausführung/ Type	
	Code
Kostüm/ Costume	K
Weste/ Vest	W
Rock/ Skirt	R

② Bleigleichwert-Front/ Lead Equivalent-Front	
	Code
0,25 mmPb	2
0,35 mmPb	3
0,50 mmPb	5

③ Bleigleichwert-Rücken/ Lead Equivalent-Back	
	Code
0,25 mmPb	2

④ Größen/ Sizes			
	Code		Code
XS	Z	XL	X
S	S	XXL	D
M	M	3XL	T
L	L		

⑤ Rocklängen/ Skirt Length	
	Code
65 cm	65
70 cm	70
75 cm	75
80 cm	80

⑤ Farbauswahl/ Colours			
	Code		Code
Curacao	22	Mandarin	30
Lagoon	25	Blue Admiral	31
Orchidee	27	Baccara	32
Regatta	28	Design Nautilus	61
Titan	29		

⑥ Außenmaterial – Hülle/ Outer Material - Cover	
	Code
ComforTex HPMF	1
ComforTex HPMF Hybrid	2
ComforTex HPMF Plus	3

⑦ Strahlenschutzmaterial – Innen/ X-Ray Protective Material - Inlay	
	Code
NovaLite 0,25 mmPb and 0,50 mmPb	1
Novalite 0,35 mmPb	8
Standard-Blei/ Standard Lead	2
Leadfree	3

EU-Konformitätserklärung EU Declaration of Conformity

Name und Anschrift des Herstellers/ <i>Name and Address of Manufacturer:</i>	MAVIG GmbH Stahlgruberring 5 81829 Munich Germany
---	--

Für die Ausstellung dieser EU-Konformitätserklärung trägt MAVIG GmbH die alleinige Verantwortung und bestätigt, dass die nachfolgend genannten Produkte die Anforderungen der Verordnung (EU) 2016/425 für Persönliche Schutzausrüstung erfüllen.

This EU declaration of conformity is issued under the sole responsibility of MAVIG GmbH who confirms that the products mentioned below meet the provision of the Regulation (EU) 2016/425 on personal protective equipment.

Produktbezeichnung/ Product name/	Röntgenschutzkleidung Anwender <i>Radiation Protective Clothing Medical</i>
--------------------------------------	--

Modell/ Model	RA632
---------------	-------

Angewandte Normen: <i>Applied standards:</i>	IEC 61331-1:2014 <i>IEC 61331-3:2014</i>
---	---

Die benannte Stelle ANCCP Certification Agency srl., Via dello Struggino, 6 – 57121 Livorno (LI) – Italy hat die EU-Baumusterprüfung durchgeführt und die EU-Baumusterprüfbescheinigung PPE-1203-18319-Cert ausgestellt.


The notified body ANCCP Certification Agency srl., Via dello Struggino, 6 – 57121 Livorno (LI) – Italy performed the EU type-examination and issued the EU type-examination certificate PPE-1203-18319-Cert.

Die PSA unterliegt dem in Anhang VIII (Modul D) der Verordnung (EU) 2016/425 vorgesehenen Verfahren der regelmäßigen Überwachung unter der Aufsicht der folgenden Benannten Stelle:

The PPE is subject to the scheduled periodic surveillance procedure set out in Annex VIII (Module D) of Regulation (EU) 2016/425 under the supervision of the following Notified Body:

ANCCP Certification Agency srl., Via dello Struggino, 6 – 57121 Livorno (LI) – Italy – CE0302

Diese Erklärung wurde ausgestellt von:
This statement was issued by:



Irene Madl – Head of Regulatory Management

Ort und Datum der Ausstellung:
Place and date of issue:

Munich, 2022-03-09

MAVIG GmbH
X-Ray Protection and
Medical Suspension Systems

Company Address:
Stahlgruberring 5
81829 Munich
Germany

Post Box:
Postfach 820362
81803 Munich
Germany

Phone:
+49 (0) 89 / 420 96 - 0

Fax:
+49 (0) 89 / 420 96 - 200

e-Mail:
info@mavig.com

Web:
www.mavig.com

Bank Accounts:
HypoVereinsbank Munich
IBAN: DE80 7002 0270
1620 2253 00
SWIFT-BIC: HYVEDEMMXXX

Postbank Munich
IBAN: DE12 7001 0080
0071 3748 07
SWIFT-BIC: PBNKDEFFXXX

Trade-Register Munich
HRB-No. 76 403

Managing Directors:
Christian Stoian
Alexia Stoian-Lepère

VAT-ID No.:
DE 811268845

Modellvarianten/ *Model versions:*

Beispielhafte Zusammensetzung des Artikelcodes/
Example for the formation of the article code:

RA632	3	2	Z	12	22	2	8
	①	②	③	④	⑤	⑥	⑦

① Bleigleichwert-Front/ Lead Equivalent-Front	
	Code
0,25 mmPb	2
0,35 mmPb	3
0,50 mmPb	5

② Bleigleichwert-Rücken/ Lead Equivalent-Back	
	Code
0,25 mmPb	2

③ Größen/ Sizes			
	Code		Code
XS	Z	XL	X
S	S	XXL	D
M	M	3XL	T
L	L		

④ Längen/ Lengths	
	Code
100 cm	10
110 cm	11
120 cm	12
130 cm	13
140 cm	14
90 cm	90

⑤ Farbauswahl/ Colours			
	Code		Code
Curacao	22	Mandarin	30
Lagoon	25	Blue Admiral	31
Orchidee	27	Baccara	32
Regatta	28	Design Nautilus	61
Titan	29		

⑥ Außenmaterial – Hülle/ Outer Material - Cover	
	Code
ComforTex HPMF	1
ComforTex HPMF Hybrid	2
ComforTex HPMF Plus	3

⑦ Strahlenschutzmaterial – Innen/ X-Ray Protective Material - Inlay	
	Code
NovaLite 0,25 mmPb and 0,50 mmPb	1
Novalite 0,35 mmPb	8
Standard-Blei/ <i>Standard Lead</i>	2
Leadfree	3

EU-Konformitätserklärung EU Declaration of Conformity

Name und Anschrift des Herstellers/
Name and Address of Manufacturer:

MAVIG GmbH
Stahlgruberring 5
81829 Munich
Germany

Für die Ausstellung dieser EU-Konformitätserklärung trägt MAVIG GmbH die alleinige Verantwortung und bestätigt, dass die nachfolgend genannten Produkte die Anforderungen der Verordnung (EU) 2016/425 für Persönliche Schutzausrüstung erfüllen.

This EU declaration of conformity is issued under the sole responsibility of MAVIG GmbH who confirms that the products mentioned below meet the provision of the Regulation (EU) 2016/425 on personal protective equipment.

Produktbezeichnung/
Product name/

*Röntgenschutzkleidung Anwender
Radiation Protective Clothing Medical*

Modell/ Model

RA660

Angewandte Normen:
Applied standards:

IEC 61331-1:2014
IEC 61331-3:2014

Die benannte Stelle ANCCP Certification Agency srl., Via dello Struggino, 6 – 57121 Livorno (LI) – Italy hat die EU-Baumusterprüfung durchgeführt und die EU-Baumusterprüfbescheinigung PPE-1203-18319-Cert ausgestellt.


The notified body ANCCP Certification Agency srl., Via dello Struggino, 6 – 57121 Livorno (LI) – Italy performed the EU type-examination and issued the EU type-examination certificate PPE-1203-18319-Cert.

Die PSA unterliegt dem in Anhang VIII (Modul D) der Verordnung (EU) 2016/425 vorgesehenen Verfahren der regelmäßigen Überwachung unter der Aufsicht der folgenden Benannten Stelle:

The PPE is subject to the scheduled periodic surveillance procedure set out in Annex VIII (Module D) of Regulation (EU) 2016/425 under the supervision of the following Notified Body:

ANCCP Certification Agency srl., Via dello Struggino, 6 – 57121 Livorno (LI) – Italy – CE0302

Diese Erklärung wurde ausgestellt von:
This statement was issued by:



Irene Madl – Head of Regulatory
Management

Ort und Datum der Ausstellung:
Place and date of issue:

Munich, 2022-03-09

MAVIG GmbH
X-Ray Protection and
Medical Suspension Systems

Company Address:
Stahlgruberring 5
81829 Munich
Germany

Post Box:
Postfach 820362
81803 Munich
Germany

Phone:
+49 (0) 89 / 420 96 - 0

Fax:
+49 (0) 89 / 420 96 - 200

e-Mail:
info@mavig.com

Web:
www.mavig.com

Bank Accounts:
HypoVereinsbank Munich
IBAN: DE80 7002 0270
1620 2253 00
SWIFT-BIC: HYVEDEMMXXX

Postbank Munich
IBAN: DE12 7001 0080
0071 3748 07
SWIFT-BIC: PBNKDEFFXXX

Trade-Register Munich
HRB-No. 76 403

Managing Directors:
Christian Stoian
Alexia Stoian-Lepère

VAT-ID No.:
DE 811268845

Modellvarianten/ *Model versions:*

Beispielhafte Zusammensetzung des Artikelcodes/
Example for the formation of the article code:

RA660 3 2 L 13 22 2 8
 ① ② ③ ④ ⑤ ⑥ ⑦

① Bleigleichwert-Front/ <i>Lead Equivalent-Front</i>	
	Code
0,25 mmPb	2
0,35 mmPb	3
0,50 mmPb	5

② Bleigleichwert-Rücken/ <i>Lead Equivalent-Back</i>	
	Code
0,25 mmPb	2
Ohne Schutz/ <i>Without Protection</i>	0

③ Größen/ <i>Sizes</i>			
	Code		Code
XS	Z	XL	X
S	S	XXL	D
M	M	3XL	T
L	L		

④ Längen/ <i>Lengths</i>	
	Code
100 cm	10
110 cm	11
120 cm	12
130 cm	13
140 cm	14
90 cm	90

⑤ Farbauswahl/ <i>Colours</i>			
	Code		Code
Curacao	22	Mandarin	30
Lagoon	25	Blue Admiral	31
Orchidee	27	Baccara	32
Regatta	28	Design Nautilus	61
Titan	29		

⑥ Außenmaterial – Hülle/ <i>Outer Material - Cover</i>	
	Code
ComforTex HPMF	1
ComforTex HPMF Hybrid	2
ComforTex HPMF Plus	3

⑦ Strahlenschutzmaterial – Innen/ <i>X-Ray Protective Material - Inlay</i>	
	Code
NovaLite 0,25 mmPb and 0,50 mmPb	1
Novalite 0,35 mmPb	8
Standard-Blei/ <i>Standard Lead</i>	2
Leadfree	3



Product Service

Certificate

No. Q5 041090 0023 Rev. 01

Holder of Certificate: MAVIG GmbH
Stahlgruberring 5
81829 München
GERMANY

Facility(ies): MAVIG GmbH
Stahlgruberring 5, 81829 München, GERMANY

Certification Mark:



Scope of Certificate: Design and development, production and distribution of **Personal Protection in Medical X-Ray Application (Protective Clothing and Protection Systems), and Ceiling Suspension Systems (for Monitor Suspensions and other Medical Equipment). Distribution of Medical Devices.**

Applied Standard(s): EN ISO 13485:2016
Medical devices - Quality management systems - Requirements for regulatory purposes (ISO 13485:2016)
DIN EN ISO 13485:2016

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system, which meets the requirements of the listed standard(s). See also notes overleaf.

Report No.: 713176864

Valid from: 2020-04-01

Valid until: 2023-03-31

Date, 2020-03-05

Christoph Dicks
Head of Certification/Notified Body

TUV SUD TÜV SUD TÜV SUD TÜV SUD TÜV SUD TÜV SUD TÜV SUD TÜV SUD TÜV SUD TÜV SUD TÜV SUD TÜV SUD TÜV SUD TÜV SUD TÜV SUD
CERTIFICATE ◆ CERTIFICADO ◆ CERTИФИКАТ ◆ 認證證書 ◆ ZERTIFIKAT ◆ CERTIFICATE

EN

BR310

Design Tortoise

MAVIG



Model: BR310 Glasses with front and side X-ray protection of 0.75 mmPb, universal size, weight approx. 90 g

CE 2895

BR310 – X-Ray Protective Glasses

The X-ray protective glasses BR310 have slightly curved glasses and a protection of 0.75 mmPb. The frame is made of an extremely durable plastic that ensures easy cleaning and good resistance to chemicals (disinfectant agents). The BR310 comes with a storage case and a soft head band.

MAVIG uses only protective lenses of the highest quality. This guarantees the best light transmission values and a maximum lens break resistance. An ideal composition of the protective material results in a high level X-ray absorption, and optimized production processes achieve an absolute homogeneous allocation of the glass additives.

The quality features of our glasses make them perfect for the use in the medical environment:

- Great frame stability through high-grade synthetic materials
- Optimized balance of sturdy, light-weight frame and lenses
- Minimized risk of injury due to shape and used components
- Low allergy risk due to carefully chosen materials
- Easy cleaning and superior resistance against chemical influences (disinfection)
- Wide and flexible head band for a comfortable fit
- For this model prescription lenses are available: single, bifocal, and progressive
- Conformity: DIN EN 61331-1:2016 / DIN EN 61331-3:2016

MAVIG GmbH
X-Ray Protection and
Medical Suspension Systems

Stahlgruberring 5
81829 Munich
Germany

PO Box 82 03 62
81803 Munich
Germany

Phone +49 (0) 89 / 420 96 - 0
Fax +49 (0) 89 / 420 96 - 200
e-Mail info@mavig.com

RA631

BALANCE

ALL-ROUND PROTECTION: VEST & SKIRT

NEW

Available as ...

Premium Line - The Full Range
Comfort Line - The Economical Alternative

The two-part design optimally counteracts any stress on the body caused by heat or weight.

- Completely overlapping front parts for lead equivalent values of 0.25 mmPb, 0.35 mmPb or 0.50 mmPb on the front
- Ideal fit and amazing wearing comfort
- Suitable for long wearing times
- Clearly reduced weight on the spine
- Standard equipment includes shoulder pads and preparation for adaptable embroidery
- In the lower area, the skirt is also secured with a snap closure
- Type testing and certification in accordance with (EU) 2016/425, IEC 61331-1:2014 / DIN EN 61331-1:2016 and IEC 61331-3:2014 / DIN EN 61331-3:2016

Optionally, the vest can be fitted with a button-on stretch belt. This provides even more weight relief for the shoulders. Alternatively, the costume can be combined with our shoulder relief system BRYGGA.



Sizes*, recommended for:

XS	Chest girth ≤ 84 cm	Waist/hip girth ≤ 86 cm
S	Chest girth approx. 85 - 95 cm	Waist/hip girth approx. 87 - 97 cm
M	Chest girth approx. 96 - 106 cm	Waist/hip girth approx. 98 - 107 cm
L	Chest girth approx. 107 - 117 cm	Waist/hip girth approx. 108 - 118 cm
XL	Chest girth approx. 118 - 128 cm	Waist/hip girth approx. 119 - 129 cm
XXL	Chest girth approx. 129 - 139 cm	Waist/hip girth approx. 130 - 140 cm
3XL	Chest girth approx. 140 - 150 cm	Waist/hip girth approx. 141 - 151 cm

* As model RA631 consists of two pieces, the following applies:
 For the vest, the larger value of the chest and waist determines the size.
 For the skirt, the larger value of the waist and hip determines the size.



Back relief

The elastic insert in the back of the skirt makes it easier for users to maintain their posture in daily routines.

Free movement

For greater freedom of movement, the skirt has a cut-out at the back. Of course, the knees are still optimally protected at the front and sides.



Overlapping area

In order to ensure protection during movement (e.g. when bending over), the vest has sufficient length and it overlaps the skirt.



Elastic insert

The insert in the rear of the skirt makes it easy for the user to hold their posture and this can be easily replaced if necessary, thanks to a zip. High-quality hook-and-loop fasteners to close the vest and skirt are, of course, standard.



Patented safety

Ideal radiation protection is provided by colour-contrasting (Titan) safety zones on the sides and in the shoulder area. These are equipped with the full lead equivalent value of the front part and give you the certainty of being optimally protected in all situations.

If the front wing is within this clearly recognisable zone when the protective clothing is fastened, you are covering the full 60% of the girth with the selected protective value.

Protection/lead equivalent


NovaLite, Leadfree or standard lead



Front, optional:
0.25 mmPb, 0.35 mmPb or 0.50 mmPb
Back generally:
0.25 mmPb

Outer material

Colours/designs (see front page)



Optional:
ComforTex® HPMF, ComforTex® HPMF Hybrid, ComforTex® HPMF Plus (Nautilus design only)

Sizes

from XS to 3XL (see front page)



Skirt length / Body height:
65 cm / 150 - 170 cm
70 cm / 170 - 180 cm
75 cm / 180 - 190 cm
80 cm / 190 - 200 cm

Optional



On request:
BRYGGA Sleeves
Stretch belt (external)
Breast pocket
Embroidery
Made-to-measure protection

Accessories/Replacement

Bolero
Thyroid/sternum protection



Stretch insert:
(Adjustment via zip)
Stretch belt:
(Adjustment via press studs)

RA614

Thyroid Gland Protection

NEW

Available as ...
Premium Line
Comfort Line

Neck widths, depending on model:

Hook-and-loop fastener (RA614KL)
 Standard size fits neck sizes of 32 - 45 cm.

Magnetic closure (RA614MA)
 Standard size fits neck sizes of 32 - 45 cm.

Interchangeable hook-and-loop fastener (RA614AK)
 Standard size fits neck sizes of 37 - 50 cm.



Titan

Blue Admiral

Regatta

Curacao

Lagoon

Mandarin

Orchidee

Baccara

► "Nautilus"
 (Comfort® - ohne Plus only)

THYROID AND STERNUM PROTECTION

RA614

The thyroid is a radiation-sensitive organ of the human body and therefore in particular need of protection.

Wearing an appropriate shield should therefore be a matter of course.

Our thyroid and sternum protector variants are characterised by a first-class cut and the best fit, taking into account all aspects of radiation protection.

The press-stud connection allows secure adjustment of the thyroid and sternum protector on MAVIG X-ray protective clothing.

Model Variants RA614

The thyroid and sternum protector with press-stud connection to the vest/coat/apron is suitable for all model series.

As standard, it is manufactured with a firmly sewn-on hook-and-loop fastener. The other closure types, magnetic closure and interchangeable hook-and-loop fastener, are available as options.

RA614KL Thyroid/Sternum Shield with hook & burr closure

RA614AK Thyroid/Sternum Shield similar to RA614KL with an exchangeable hook & burr closure


RA614MA Thyroid/Sternum Shield similar to RA614KL with magnetic closure ^{*1}



Snap Fasteners
for a secure attachment of the thyroid and sternum protection to MAVIG protective clothing



Closure type – via magnet
The thyroid and sternum protection is also available with a practical magnetic closure (variant RA614MA).




Accessories: Hygienic cover
To meet the hygienic requirements in everyday life, we recommend using our washable hygienic cover in the colour Titan made of the special ComforTex® MF material.



Closure type – interchangeable
As a third alternative, an interchangeable closure can be chosen (RA614AK). Please note the different neck sizes here.

Protection/lead equivalent


Novalite, Leadfree or Standard lead



As standard:
0.50 mmPb

Outer material

Colours/designs (see front page)



Optional:
ComforTex® HPMF,
ComforTex® HPMF Hybrid,
ComforTex® HPMF Plus
(Nautilus design only)

Sizes/Neck widths

Uniform
Depending on the model



Optional

(pages 30/31)



On request:
Embroidery
Made-to-measure protection

Accessories/replacement

Hygienic cover (RA614HF)



Hook-and-loop fastener (R614AKV, adjustment via push buttons)

RA660

COMFORT

FRONT PROTECTION APRON

NEW

Available as ...

Premium Line - The Full Range
Comfort Line - The Economical Alternative

Our RA660 range offers up-to-date radiation protection clothing which accompanies your active working day and provides outstanding service. The extraordinarily good wearing comfort is supported by the wide stretch insert on the back wings.

- Suitable for a wide range of applications with only frontal radiation exposure
- Ideal fit, maximum comfort for the wearer
- Suitable for long wearing times
- Optional "without protection" in the back, for reduced weight of the apron
- Standard equipment includes shoulder pads and preparation for adaptable embroidery
- Type testing and certification in accordance with (EU) 2016/425, IEC 61331-1:2014 / DIN EN 61331-1:2016 and IEC 61331-3:2014 / DIN EN 61331-3:2016

Our customers report that this model is their favourite front apron. This is because it can be taken on and off easily, fits the body well, stays in place well and provides shoulder relief.



Sizes*, recommended for:

XS	Chest girth ≤ 84 cm	Waist/hip girth ≤ 91 cm
S	Chest girth approx. 85 - 95 cm	Waist/hip girth approx. 92 - 102 cm
M	Chest girth approx. 96 - 106 cm	Waist/hip girth approx. 103 - 112 cm
L	Chest girth approx. 107 - 117 cm	Waist/hip girth approx. 113 - 123 cm
XL	Chest girth approx. 118 - 128 cm	Waist/hip girth approx. 124 - 134 cm
XXL	Chest girth approx. 129 - 139 cm	Waist/hip girth approx. 135 - 145 cm
3XL	Chest girth approx. 140 - 150 cm	Waist/hip girth approx. 146 - 156 cm

* The size depends on the girth of the chest, waist, or hip, depending on which of the three measurements is the largest.





Special feature

The stretch insert, ideal relief especially for long periods of wear.



Stretch insert

The stretch insert running across the entire width of the back ensures a good fit and shoulder relief even when the X-ray protective clothing is worn for long periods.



Closure system

MAVIG uses only top quality, low-wear hook-and-loop fasteners.

Protection/lead equivalent

Novalite, Leadfree or Standard lead




Front, optional:
0.25 mmPb, 0.35 mmPb or 0.50 mmPb

Back, optional:
0.25 mmPb or without protection

Outer material

Colours/designs (see front page)



Optional:
ComforTex® HPMF, ComforTex® HPMF Hybrid, ComforTex® HPMF Plus (Nautilus design only)

Sizes

from XS to 3XL (see front page)



Skirt length / Body height:
90 cm / 150 - 160 cm
100 cm / 160 - 170 cm
110 cm / 170 - 180 cm
120 cm / 175 - 190 cm
130 cm / 185 - 195 cm
140 cm / 195 - 205 cm

Optional



On request:
Sleeves
Breast pocket
Embroidery
Made-to-measure protection

Accessories/replacement

Bolero
Thyroid/sternum protection



EN

BRV500 BRV501



Model BRV500



Model BRV501

Models: BRV500
BRV501

Continuous protection from forehead to the chin
Protection down to the cheek bones with a cut-out for the nose

BRV500 & BRV501 – X-Ray Protective Visors

Especially designed for the protection of medical professionals during radiological procedures. Protection of the eye lens against scattered X-rays. The visors provide a high reduction in radiation exposure to the eyes and most of the facial area (see table).

The quality features of our visors make them ideal for the use in the medical environment

- Pb Value (Weight): BRV500 0.10 mmPb (approx. 520 g)
BRV501 0.10 mmPb (approx. 480 g)
- Universal size / adjustable to fit different head sizes
- Multiple adjustments to conform to individual facial features
- Easy to set up for fitting the head of the wearer by simply pressing and turning the pivoting mechanism to prevent slipping
- Comfortable, exchangeable, and washable pads
- Curved lead acrylic glass with 0.10 mmPb
- Low allergy risk due to carefully chosen materials

X-Ray Tube Voltage	Attenuation Scattered Radiation
50 kV	96.9 %
60 kV	94.4 %
80 kV	86.3 %
100 kV	80.0 %
120 kV	75.0 %

MAVIG GmbH
X-Ray Protection and
Medical Suspension Systems

Stahlgruberring 5
81829 Munich
Germany

PO Box 82 03 62
81803 Munich
Germany

Phone +49 (0) 89 / 420 96 - 0
Fax +49 (0) 89 / 420 96 - 200
e-Mail info@mavig.com