


Test Report

– Eppendorf PCR clean –

Protein LoBind Tube 1.5 mL



EPPENDORF TUBES®
100 Safe-Lock Tubes
PCR clean
centrifugation up to 18 000 x g

LOT M209810G

2028-01-28

Order no.: 0030 108.118

Eppendorf SE
22331 Hamburg · Germany
Email: eppendorf@eppendorf.com

Cat. no.: 022431081

Distributed in North America by
Eppendorf North America, Inc. · USA
Email: info@eppendorf.com

• **Human DNA**

Method: real-time PCR acc. to SOP 99880/V2
Limit: < 2 pg; less than one human cell

• **DNase**

Method: DNA Digestion acc. to SOP 100570/V1
Limit: Not detectable (LOD: 1.0×10^{-6} Kunitz units)

• **PCR inhibition**

Method: real-time PCR acc. to SOP 100559/V1
Limit: Less than 10 targets amplifiable

• **RNase**

Method: RNA Digestion acc. to SOP 100571/V1
Limit: Not detectable (LOD: 1.0×10^{-9} Kunitz units)

Customer: Eppendorf SE, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 03.02.23

Testing of the above described lot showed conformity within the limits of detection.

10.02.23
Date of release

Epp 2302-2780
Registration number

i. V. B. Ullrich
Dipl.-Biol. Thomas Weigel (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.