



EUROPEAN COMMISSION
DIRECTORATE-GENERAL FOR HEALTH AND FOOD SAFETY

Public health, country knowledge, crisis management
Health Security

EU health preparedness:

**A common list of COVID-19 rapid antigen tests;
A common standardised set of data to be included in COVID-
19 test result certificates; and
A common list of COVID-19 laboratory based antigenic assays**

Agreed by the Health Security Committee

Common list of COVID-19 rapid antigen tests (Annex I)

Agreed by the Health Security Committee on 17 February 2021.

First update: 10 May 2021; Second update: 16 June 2021; Third update: 7 July 2021; Fourth update: 14 July 2021; Fifth update: 23 July 2021; Sixth update: 20 October 2021; Seventh update: 10 November 2021; Eighth update: 8 December 2021; Ninth update: 21 December 2021; Tenth update: 21 January 2022; Eleventh update: 10 February 2022.

Common standardised data set to be included in COVID-19 test result certificates (Annex II)

Agreed by the Health Security Committee on 17 February 2021.

An update to Annex II was agreed by the HSC on 19 March 2021

Common list of COVID-19 laboratory based antigenic assays (Annex III)

Agreed by the Health Security Committee on 20 October 2021

First update: 10 February 2022.

I. Introduction

Robust testing strategies are an essential aspect of preparedness and response to the COVID-19 pandemic, allowing for early detection of potentially infectious individuals and providing visibility on infection rates and transmission within communities. Moreover, they are a prerequisite to adequate contact tracing to limit the spread through prompt isolation. Also in the context of the circulation of SARS-CoV-2 variants of concern, surge testing in addition to existing testing deployment has proven to be key for controlling and suppressing further spread of the virus.

While the reverse transcription real-time polymerase chain reaction (RT-PCR) assay, which is a nucleic acid amplification test (NAAT), remains the ‘gold standard’ for COVID-19 diagnosis, rapid antigen tests, which detect the presence of viral proteins (antigens), are increasingly being used by Member States as a way of further strengthening countries’ overall testing capacity, particularly in case of limited NAAT capacities or where prolonged testing turnaround times results in no clinical utility.

The Health Security Committee (HSC) agreed on 17 September 2020 on Recommendations for a common EU testing approach for COVID-19¹, setting out various actions for consideration by countries when updating or adapting their testing strategies. The Recommendations included Member States’ first experiences with rapid antigen tests and their deliberations concerning the settings and situations in which these tests should be used. Since then, the HSC has been discussing the use and application of rapid antigen tests in great depth, and has brought together a wealth of (technical) information on the types of tests used in European countries and the conditions applied.

On 21 January 2021, Member States unanimously agreed on a Council Recommendation setting a common framework for the use of rapid antigen tests and the mutual recognition of COVID-19 test results across the EU². The Council Recommendation called on Member States to agree on three concrete deliverables:

1. **A common list of COVID-19 rapid antigen tests** that are considered appropriate for use in the context of the situations described in the Council Recommendation, that are in line with countries’ testing strategies.
2. A selection of rapid antigen tests of which Member States will **mutually recognise the test results for public health measures**.
3. **A common standardised set of data to be included in COVID-19 test result certificates**, further facilitating the mutual recognition of COVID-19 test results.

Based on the information collected by the HSC, and taking into consideration the current epidemiological situation and the testing strategies and approaches that have been put in place across the EU, this document sets out the deliverables as agreed by Member States.

¹ https://ec.europa.eu/health/sites/health/files/preparedness_response/docs/common_testingapproach_covid-19_en.pdf

² <https://data.consilium.europa.eu/doc/document/ST-5451-2021-INIT/en/pdf>

This document is based on the content of the Council Recommendation and further criteria agreed by Member States, and considers the relevant recommendations published by the Commission³ as well as the updated technical report issued the European Centre for Disease Prevention and Control (ECDC)⁴ and the World Health Organization (WHO)⁵.

II. Annex I: Common list of rapid antigen tests

Point 11 of the Council Recommendation of 21 January 2021, calls on Member States to, without prejudice to Directive 98/79/EC, agree on and maintain a common and updated list of COVID-19 rapid antigen tests that are considered appropriate for use in the context of the situations described under point 6 and are in line with countries' testing strategies.

This list should be shared with ECDC and the Commission to prevent duplication of work and to feed into ongoing initiatives, particularly the “COVID-19 In Vitro Diagnostic Devices and Test Methods Database⁶, hosted by the Joint Research Centre (JRC). **Annex I to this document sets out a common list of rapid antigen tests.** This list has been incorporated by the JRC in its COVID-19 In Vitro Diagnostic Devices and Test Methods Database.

The common list of rapid antigen tests is regularly being reviewed by Member States, and, if necessary, be updated in line with new results from independent validation studies becoming available and new tests entering the markets. These updates are also taking into account how mutations of the SARS-CoV-2 virus may affect the efficacy of rapid antigen tests, allowing for the removal of tests no longer deemed effective. The effect of SARS-CoV-2 mutations on the efficacy of NAAT, in particular RT-PCR assays, will also be kept under review.

A first update to Annex I was agreed by the Health Security Committee on 10 May 2021, a second update on 16 June 2021, a third update on 7 July 2021, a fourth update on 14 July 2021, a fifth update on 23 July 2021, a sixth update on 20 October 2021, a seventh update on 10 November 2021, an eight update on 8 December 2021, a ninth update on 21 December 2021, a tenth update on 21 January 2022, and a eleventh update on 10 February 2022.

As stipulated in point 15 of the Council Recommendation of 21 January 2021, Member States will agree on a selection of rapid antigen tests of which they will mutually recognise the test results for public health measures. The HSC agrees that, considering that *all* of the rapid antigen tests included in the EU common list are eligible for a test certificate issued as part of the EU Digital COVID Certificate⁷, the entire list is considered to consist of rapid antigen tests of which Member States mutually recognise the test results for public health measures.

³ <https://eur-lex.europa.eu/legal-content/EN/TXT/?uri=CELEX:32020H1595> and <https://eur-lex.europa.eu/legal-content/EN/TXT/PDF/?uri=CELEX:32020H1743&from=EN>

⁴ <https://www.ecdc.europa.eu/sites/default/files/documents/Options-for-the-use-of-rapid-antigen-tests-for-COVID-19-first-update.pdf>

⁵ <https://www.who.int/publications/i/item/9789240017740>

⁶ <https://covid-19-diagnostics.jrc.ec.europa.eu/devices>

⁷ <https://eur-lex.europa.eu/legal-content/EN/TXT/?uri=CELEX%3A32021R0953>

III. HSC Technical Working Group on COVID-19 Diagnostic Tests

Based on the increasing political and commercial interest in the HSC agreed common list of rapid antigen tests, particularly in the context of the EU Digital COVID Certificate⁸, there was a need to put in place a more structured, coherent and swift procedure for updating the common list of rapid antigen tests. As a first step, since 10 May 2021, it is possible for manufacturers to submit data and information concerning rapid antigen tests that they believe should be considered for inclusion in the HSC agreed common list. This information will thus be reviewed and considered alongside the proposals put forward by EU Member States.

Secondly, a HSC Technical Working Group on COVID-19 Diagnostic Tests was set up. This Working Group, consisting of technical experts from EU and EEA Member States, is responsible for reviewing the information submitted by countries and manufacturers, taking into account the latest result of independent validation studies and country practices and experiences. Based on this, the Technical Working Group presents proposals to the HSC for further updates to the common list of rapid antigen tests. The HSC thus remains the platform where agreement between Member States is reached for updates to the list.

Building on the interim definitions and criteria that were agreed by the experts on 29 June 2021, the Technical Working Group agreed on 21 September 2021 on further **definitions, scope, considerations and criteria** to be applied to independent validation studies assessing the clinical performance of rapid antigen tests for COVID-19 diagnosis. These further definitions, scope, considerations and criteria are used by the Technical Working Group in addition to the ones presented in Council Recommendation 2021/24/01 when assessing the proposals for new rapid antigen tests to be included in the EU common list. They have been applied to all proposals received after 12 July 2021. Concerning the rapid antigen tests that were included in the EU common list of rapid antigen tests before this date, the criteria will apply as of May 2022.

On 6 July 2021 the Technical Working Group agreed that that the common list should not include rapid antigen self-tests. Moreover, on 10 November 2021, it was agreed rapid antigen tests that are using a mix of different sampling materials (i.e. nasal, oropharyngeal and/or nasopharyngeal swabs as well as other specimen types such as saliva) can be included in the EU common list, however, only the validation results based on nasal, oropharyngeal and/or nasopharyngeal swabs of such devices will be reviewed by the Technical Working Group and assessed against the specified criteria. Only test results based on nasal, oropharyngeal and/or nasopharyngeal specimens should be valid for the issuance of test certificates for the EU Digital COVID Certificate. Rapid antigen tests that are solely based on sampling materials other than nasal, oropharyngeal and/or nasopharyngeal specimens (e.g. saliva, sputum, blood and/or faeces), are not included in the EU common list of antigen tests.

As of December 2021, the Technical Working Group has continuously been discussing the performance of rapid antigen tests in the context of the emerging **Omicron** variant of concern.

⁸ <https://eur-lex.europa.eu/legal-content/EN/TXT/PDF/?uri=CELEX:32021R0953&from=EN>.

In particular, concerns have been raised about rapid antigen devices that are solely targeting the spike protein (thus not combined with the nucleocapsid protein) as well as the viral load measured at different time points and at different sites (e.g. throat and nose) after Omicron infection. The Technical Working Group will continue monitoring the situation including emerging evidence on the potential impact of the Omicron variant of concern on the performance of COVID-19 rapid antigen tests and, if necessary, amend the agreed criteria accordingly.

For the time being, rapid antigen tests that target the S-protein or for which the target protein is unknown have been highlighted in the EU common list. Moreover, an additional column has been added where information is presented – either from published studies or data submitted by the manufacturer - on the performance of the device in relation to the detection of Omicron infection.

Agreed scope of the EU common list of rapid antigen tests:

- The EU common list includes rapid antigen tests that are used in practice in at least one EU Member State and that have been validated by at least one EU Member State.
- The EU common list includes rapid antigen tests for which their clinical performance was measured based on samples collected from nasal, oropharyngeal or nasopharyngeal specimens and that meet the criteria as further specified below.
- Rapid antigen tests that are solely based on other sampling materials, such as saliva, sputum, blood and/or faeces, are not included in the EU common list of antigen tests. This is in line with current evidence and the technical recommendations provided by the European Centre for Disease Prevention and Control (ECDC)⁹.
- Rapid antigen tests that are using a mix of different sampling materials (i.e. nasal, oropharyngeal and/or nasopharyngeal swabs as well as other specimen types such as saliva) can be included in the EU common list. However, the validation studies of such tests should present the results and data for each specimen type separately. Only the validation results based on nasal, oropharyngeal and/or nasopharyngeal swabs of such devices will be reviewed by the Technical Working Group and assessed against the specified criteria.
- Only test results based on nasal, oropharyngeal and/or nasopharyngeal specimens should be valid for the issuance of test certificates for the EU Digital COVID Certificate.
- The EU common list of antigen tests does not include rapid antigen self-tests. It only includes those rapid antigen tests that are conducted by trained healthcare personnel or trained operators where appropriate (in line with Commission Recommendation (EU) 2020/1743 of 18 November 2020).
- The EU common list of antigen tests does not include pooled rapid antigen tests,

⁹ <https://www.ecdc.europa.eu/sites/default/files/documents/covid-19-use-saliva-sample-material-testing.pdf>

which involve mixing of multiple samples together in a batch or pooled sample for testing.

- Only rapid antigen tests that carry CE marking are included in the EU common list of antigen tests.
- The Technical Working Group will continue to monitor the developments in the field of rapid antigen testing and will, if deemed necessary, reconsider the scope of the EU common list of rapid antigen tests once relevant evidence and data has become available. Particular attention will be paid to breakthrough infections among vaccinated individuals and the possible impact of such cases on the clinical performance of rapid antigen tests, as well as the performance of rapid antigen tests in the context of emerging SARS-CoV-2 variants.

Agreed definition and considerations of an independent validation study:

- A study that may involve collaborations with or that may involve funding by private entities, however, there is always a public body involved from an EU Member State.
- Such study should be carried out in an EU Member State, and be performed objectively and in the public interest.
- Such study should be performed by an independent laboratory, which is a laboratory not owned nor operated by the manufacturer or sponsor of the test, and which is not related to the operator by ownership, familial relationships, nor contractual or other relationships that result in the laboratory being controlled by or being under the common control of the operator.
- Such study should preferably be based on a **prospective clinical field study** design, testing *unselected* symptomatic and asymptomatic participants for SARS-CoV-2 infection. Until May 2022, validation studies carried out based on retrospective in vitro study designs, testing the clinical performance of rapid antigen tests using SARS-CoV-2 reference panels, will be accepted too¹⁰.
- “Unselected” means no prior knowledge of SARS-CoV-2 diagnosis (e.g. determined by PCR); inclusion is allowed based on general possible COVID-like symptoms (or close contact with COVID-19 cases); and exclusion is allowed of children (e.g. <16 years) or for medical ethical permission reasons.

Agreed clinical performance criteria for independent validation studies:

Prospective clinical field studies:

- A sensitivity over 80% when testing unselected symptomatic participants within the first seven days after symptom onset or asymptomatic participants, where the diagnosis is confirmed by RT-PCR in independent field studies, will be accepted.

OR

¹⁰ After May 2022, only rapid antigen tests of which their clinical performance has been evaluated through independent validation studies based on prospective clinical field study designs (in combination with retrospective in vitro studies) will be accepted for inclusion in the EU common list of rapid antigen tests.

In independent evaluations of unselected participants, assays should have a sensitivity of 90% or greater for subjects with a $Ct \leq 25$.

- The study population shall be clearly defined stating the inclusion criteria of participants (symptomatic individuals, close contacts or asymptomatic individuals without known exposure). Ideally, the sensitivity for each group should be discernible from the report. The RT-PCR protocol and the distribution of Ct values should be described. Samples should represent naturally occurring viral loads.
- Target population considered in the context of an independent validation study should be based on at least 100 RT-PCR positive samples and at least 300 RT-PCR negative samples. Each specimen type should be evaluated separately.
- In case of multiple smaller prospective clinical field studies that do not meet the minimum number of positive and/or negative samples separately but that do meet all the other criteria as agreed by the Technical Working Group, the number of samples may be combined, provided that the different studies applied the same or similar methodologies and that sufficient details are provided on their study design.
- Assays should have a specificity over 98%.
- In line with the *MDCG Guidance on performance evaluation of SARS-CoV-2 in vitro diagnostic medical devices*¹¹, preference is given to samples being compared against RT-PCR results on nasopharyngeal swabs. However, in independent validation studies, samples can also be compared against RT-PCR results on oropharyngeal or nasal swabs if reasoning is provided (e.g. when assessing the clinical performance of rapid antigen tests among children).

Retrospective in vitro studies:

- A sensitivity over 80% when testing all specimen in the reference panel will be accepted;

OR

Assays should have a sensitivity of 90% or greater for subjects with a $Ct < 25$.

- The composition of the reference panel should be as follows:
 - A panel of at least 50 pooled clinical specimens that cover naturally occurring viral loads with SARS-CoV-2 concentration ranging from approximately 1.1×10^9 to 4.2×10^2 genome copies per mL of specimen and Ct values between 17 and 36.
 - The whole evaluation panel should be subdivided into three subgroups: panel members, which are characterized by:
 - **Very high viral load** (Ct value 17-25; about 40% of the total number of pooled clinical specimens);
 - **High viral load** (Ct value 25-30; about 40% of the total number of pooled clinical specimens); and
 - **Moderate viral load** (Ct value 30-36; about 20% of the total number of pooled clinical specimens).

¹¹ https://ec.europa.eu/health/sites/default/files/md_sector/docs/mdcg_2021-21_en.pdf

- For each pool up to ten clinical respiratory specimens (nasopharyngeal/oropharyngeal) obtained for routine diagnostics with different virus loads may be used. The sample volume per panel member should be sufficient to allow comparative evaluation with different tests included in the evaluation.
 - RT-PCR needs to be applied to determine the RNA load per pool.
 - Ethical approval by an institutional review board is mandatory.
 - For each rapid antigen test and panel member, a pre-defined aliquot needs to be completely absorbed using the specimen collection device, e.g. swab, provided with the respective test.
 - Further steps needs to be strictly performed following the respective instructions for use (IFU).
 - The stability of the panel (antigen) must be considered throughout the preparation of the panel and the workflow up to the test.
- Assays should have a specificity over 98%, as measured through the retrospective in vitro evaluation study or as specified by the manufacturer in the IFU.
 - In line with the *MDCG Guidance on performance evaluation of SARS-CoV-2 in vitro diagnostic medical devices*¹¹, preference is given to samples being compared against RT-PCR results on nasopharyngeal swabs. However, in independent validation studies, samples can also be compared against RT-PCR results on oropharyngeal or nasal swabs if reasoning is provided (e.g. when assessing the clinical performance of rapid antigen tests among children).

As a wide range of different methodologies and protocols are being applied in countries, discussions on testing approaches will continue, with the overall goal for the Technical Working Group to develop and agree on an EU harmonised approach for validation studies assessing the clinical performance of COVID-19 rapid antigen tests. This work will take into account the ongoing work by the In Vitro Diagnostics Working Group of the Medical Device Coordination Group regarding guidance on the performance of COVID-19 tests in the context of CE-marking and common specifications under Article 9 of Regulation (EU) 2017/746¹².

Grace period

As of 1 January 2022, a grace period of 4 weeks applies whenever updates are made to Annex I. The grace period applies to both the inclusion of new devices as well as the removal of rapid antigen tests that are included in the EU common list or rapid antigen tests.

¹² The Medical Device Coordination Group is set up according to Art. 103 of Regulation (EU) 2017/745 and Art. 98 of Regulation (EU) 2017/746. This group is also responsible for overseeing the implementation of Directive 98/79/EC. See also Register of Commission Expert Groups and Other Similar Entities, code number X03565, and its subgroups.

IV. Annex II: Common standardised set of data for COVID-19 test certificates

In order to facilitate in practice the mutual recognition of results of rapid antigen tests as well as NAAT, including RT-PCR assays, point 18 of Council Recommendation 2020/1475 defines that Member States should agree on a common standardised set of data to be included in the form for test result certificates.

Based on information that was submitted by members of the Health Security Committee in response to a survey on mutual recognition on COVID-19 test results and further discussions that took place in the context of the Health Security Committee, Member States agree on **the common standardised set of data for COVID-19 test result certificates as presented in Annex II**. Member States agree that COVID-19 test results should be made available in the national language(s) of the country where the test was taken, as well as English.

An update to this Annex was agreed by the Health Security Committee on 19 March 2021, addressing input received from the eHealth Network and in particular the Semantic Subgroup and based on discussions that took place in the context of the EU Digital COVID Certificate. The Health Security Committee will discuss, whenever relevant, possible updates to the agreed common standardised set of data for COVID-19 test certificates, and publish, if necessary, an updated agreed document.

V. Annex III: Common list of laboratory-based antigenic assays

In addition to COVID-19 rapid antigen tests, as of 8 July 2021, it is possible for manufacturers and countries to put forward proposals for laboratory-based antigenic assays (e.g. enzyme immunoassays such as ELISA or automated tests) for review by the Technical Working Group. These proposals are reviewed by the experts against the same criteria used for the review of rapid antigen tests.

Annex III sets out those laboratory-based antigenic assays that meet these criteria. A first update to Annex I was agreed by the Health Security Committee on 10 February 2022. Further criteria for lab-based antigenic assays may be defined at a later stage.

Currently, a negative test results produced by a lab-based antigenic assays cannot be used for the issuance of the EU Digital COVID Certificate.

The Technical Working Group on COVID-19 diagnostic devices does not, at the moment, review proposals for inclusion of antibody tests in the EU common list.

ANNEX I: Common list of COVID-19 rapid antigen tests¹³

As agreed by EU Member States on 10 February 2022

Disclaimer: This list was agreed by the HSC based on a proposal by the Technical Working Group on COVID-19 Diagnostic Tests. Experts participating in the Technical Working Group strongly recommend that use of rapid antigen tests is primarily intended for preliminary testing for SARS-CoV-2 infection in symptomatic patients, and note that rapid antigen tests should in particular be used in the specific contexts and circumstances referred to by the Commission Recommendation (EU) 2020/1743 of 18 November 2020 and the updated technical report by ECDC on 26 October 2021. The content of the common list is based on the clinical performance data and information that is available at this moment in time. Updates to the common list are based on the criteria as described in Council Recommendation 2021/C 24/01 as well as the further criteria and definitions agreed by the Technical Working Group on 21 September 2021. The Medical Device Coordination Group Guidance on performance evaluation of SARS-CoV-2 in vitro diagnostic medical devices¹⁴, envisaged to form the basis for common specifications to be adopted according to Article 9 of Regulation (EU) 2017/746, has been taken into consideration in this regard.

Rapid antigen tests presented in boxes are so-called ‘twin tests’. These are rapid antigen tests that are identical in design and construction but, for example, branded or distributed under a different name. The results of independent validation studies may be transferred between twin tests.

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
AAZ-LMB	COVID-VIRO®	1833	Prospective clinical field study FR: Prospective study carried out in the “Centre Hospitalier d’Orléans” on NP swabs simultaneously tested by RT PCR: sensitivity <7 days after onset of symptoms: 94,7% (72/76), specificity: 100%.	96.6% sensitivity 100% specificity Nasal swab, NP swab	FR CH	Nucleo-protein	Nasal swab, Nasopharyngeal swab		10 May 2021

¹³ This is the list of rapid antigen tests as referred to in Article 3 of the Regulation (EU) 2021/953 of the European Parliament and of the Council of 14 June 2021 on a framework for the issuance, verification and acceptance of interoperable COVID-19 vaccination, test and recovery certificates (EU Digital COVID Certificate) to facilitate free movement during the COVID-19 pandemic, OJ L 211, 15.6.2021, p. 1–22.

¹⁴ https://ec.europa.eu/health/sites/default/files/md_sector/docs/mdcg_2021-21_en.pdf

¹⁵ As registered in and used by the JRC database, see: <https://covid-19-diagnostics.jrc.ec.europa.eu/>.

¹⁶ As reported in the JRC database, see: <https://covid-19-diagnostics.jrc.ec.europa.eu/>.

¹⁷ Only test results based on nasal, oropharyngeal and/or nasopharyngeal specimens should be valid for the issuance of test certificates for the EU Digital COVID Certificate. The information included in this column is based on the information provided by manufacturers to the JRC database.

¹⁸ The content in this column is based on information submitted by manufacturer via the JRC database as well as what has been published by scientific studies.

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Abbott Rapid Diagnostics	Panbio™ COVID-19 Ag Rapid Test	1232	<i>Prospective clinical field studies</i>						
			<p>BE: Small-scale head-to-head comparison of 5 RATs in Belgian hospital lab. Panbio overall sensitivity (Ct range 14,6 – 35,5): 45/57 samples (79%). Sensitivity for Ct≤25: 17/18 samples. Overall specificity 100%.</p> <p>NL: 1367 and 208 subjects were enrolled in Utrecht and Aruba, respectively. Specificity of the Panbio™ COVID-19 Ag Rapid Test was 100% (95%CI: 99.7–100%) in both settings. Test sensitivity was 72.6% (95%CI: 64.5–79.9%) in the Netherlands and 81.0% (95% CI: 69.0–89.8%) in Aruba. Restricting RT-qPCR test positivity to Ct-values <32 yielded test sensitivities of 95.2% (95%CI: 89.3–98.5%) in Utrecht and 98.0% (95%CI: 89.2–99.95%) in Aruba.</p> <p>PT: 83 samples from symptomatic individuals (27 PCR positive and 56 negative by PCR) were tested. Sensitivity 63% (95%IC 42-81); specificity 100% (95%IC 94-100). LoD TCID50/ml 1,38 x 10² and CT<24.</p> <p>SE: Karolinska hospital evaluation of Lot 41ADF061A. Patient samples: 95 PCR positive, 150 negative. No detailed sample description available. Sensitivity 59%, specificity 100%. Sensitivity Ct<25 = 90.2%.</p> <p>FIND evaluation studies DE (10 Dec 2020): 1108 samples, NP swab. Clinical sensitivities: Days ≤7: 90.8%; Ct ≤ 33: 88.3%; Ct ≤ 25: 95.8%. Clinical specificity: 99.9% CH (10 Dec 2020): 535 samples, NP swab. Clinical sensitivities: Days ≤7:</p>	<p>91.4% sensitivity 99.8% specificity NP swab (Ct ≤ 33)</p> <p>98.1% sensitivity 99.8% specificity Nasal swab (Ct ≤ 33)</p>	<p>BE, DE^[2], ES, FI, NL^[5], PT, SE</p> <p>CH, India, NO, UK</p>	Nucleo-protein	Nasal swab, Nasopharyngeal swab	<p>CH: Sensitivity was compared between Omicron and other VOCs (Alpha, Beta, Gamma and Delta) as well as an early-pandemic SARS-CoV-2 isolate.</p>	17 February 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
			85.6%; Ct ≤33: 89.7%; Ct ≤ 25: 96.8%. Clinical specificity: 100% India (25 June 2021): 526 samples, NP swab. Clinical sensitivities: Days ≤7: 61.3%-100%; Ct ≤33: 74.2%-86.7%; Ct ≤ 25: 91.9%-100%. Specificity: 100%						
			<i>Retrospective in vitro studies</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity of 99.8%						
ABIOTEQ	Cora Gentest-19	2374	<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.8%	Sensitivity 98,7%, Specificity 99,8%	DE ^[2]	Nucleo-capsid protein	Anterior nasal swab, Nasal swab, Nasopharyngeal swab, Oropharyngeal swab, Throat swab		20 October 2021
AccuBioTech Co.,Ltd	Accu-Tell SARS-CoV-2 Ag Cassette	2579	<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.2%	Sensitivity: 95.7% Specificity: 99.2%	DE ^[2]	Nucleo-capsid protein	Nasopharyngeal swab		20 October 2021
Acon Biotech (Hangzhou) Co., Ltd	Flowflex SARS-CoV-2 Antigen Rapid Test	1457	<i>Prospective clinical field study</i>						
			FIND evaluation CH (9 June 2021) 279 samples, nasal swab. Sensitivities: Days ≤7: 92.2%; Ct ≤ 33: 98.3%; Ct ≤ 25: 100%. Specificity: 99.5%	Nasal swab Clinical Sensitivity: 97.1 % Clinical Specificity: 99.5 %	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab	CH: Sensitivity was compared between Omicron and other VOCs (Alpha, Beta, Gamma and Delta) as well as an early-pandemic SARS-CoV-2 isolate.	14 July 2021
			<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99.54%	NP swab Clinical Sensitivity: 97.6 % Clinical Specificity: 99.4 %	CH, UK				
ACON Biotech(Hangzhou) Co., Ltd.	Flowflex SARS-CoV-2 Antigen Rapid Test (Nasal/Saliva)	1865	<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99.5%	Clinical Sensitivity 97.1 % (Nasal Swab) Clinical Specificity 99.5 % (Nasal Swab)	DE ^[2]	Nucleo-capsid protein	Nasal swab ! Saliva		10 February 2022

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
ACON Laboratories, Inc.	Flowflex SARS-CoV-2 Antigen Rapid Test	1468	<i>Retrospective in vitro study</i>	96.9% sensitivity 98.7% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 98,7%						
AESKU.DIAGNOSTICS GmbH & Co, KG	AESKU.RAPID SARS-CoV-2	2108	<i>Retrospective in vitro study</i>	96% sensitivity 98% specificity NP swab	DE ^[2]	Nucleo-protein	Nasal swab, Throat swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 84% at Ct ≤ 25; Manufacturer specificity: 98%						
Affimedix Inc.	TestNOW® - COVID-19 Antigen Test	2130	<i>Retrospective in vitro study</i>	NP swab: 95% sensitivity 99.2% specificity Nasal swab: 98.1% sensitivity 100% specificity	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99,2%						
AMEDA Labordiagnostik GmbH	AMP Rapid Test SARS-CoV-2 Ag	1304	<i>Retrospective in vitro study</i>	97.3% sensitivity NP swab 97.3% sensitivity Nasal swab 100% specificity	DE ^[2] CH, UK	Nucleo-protein	Nasal swab, Nasopharyngeal swab		17 February 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%						
Anbio (Xiamen) Biotechnology Co., Ltd	Rapid COVID-19 Antigen-Test (colloidal Gold)	1822	<i>Retrospective in vitro study</i>	99.27% sensitivity, 100% specificity Nasal swab	DE ^[2]	Unknown	Nasal swab, Throat swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%						
Anhui Deep Blue Medical Technology Co., Ltd	COVID-19 (SARS-CoV-2) Antigen Test Kit (Colloidal Gold)	1736	<i>Retrospective in vitro study</i>	Nasal/OP swab: 96,4% sensitivity, 99,8% specificity NP swab: 95,7% sensitivity, 99,3% specificity	DE ^[2] UK	Nucleo-protein	Nasal swab, ! Other		10 May 2021
	COVID-19 (SARS-CoV-2) Antigen Test Kit (Colloidal Gold) – Nasal swab	1815	<i>Retrospective in vitro study</i>						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Anhui Formaster Biosci Co., Ltd.	New Coronavirus (COVID-19) Antigen Rapid Test	2089	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.5%	sensitivity: 95.15%, specificity: 98.5%	DE ^[2]	Nucleo-protein	Nasopharyngeal swab, Oropharyngeal swab		20 October 2021
ArcDia International Ltd	mariPOC SARS-CoV-2	768	<i>Prospective clinical field study</i> FI: Clinical performance of the test was evaluated against qRT-PCR with nasopharyngeal swab specimens collected from patients suspected of acute SARS-CoV-2 infection. Sensitivity of the mariPOC test was 100.0% (13/13) directly from swab specimens and 84.4% (38/45) from swab specimens in undefined transport mediums. Specificity was 100.0% (201/201).	100% sensitivity 100% specificity Nasopharyngeal swab	FI	Nucleo-protein	Nasopharyngeal swab		10 May 2021
ArcDia International Oy Ltd	mariPOC Respi+	2078	<i>Prospective clinical field study</i> FI: Clinical performance of the test was evaluated against qRT-PCR with nasopharyngeal swab specimens collected from patients suspected of acute SARS-CoV-2 infection. Sensitivity of the mariPOC test was 100.0% (13/13) directly from swab specimens and 84.4% (38/45) from swab specimens in undefined transport mediums. Specificity was 100.0% (201/201).	100 % sensitivity 100 % specificity NP swab	FI	Nucleo-protein	Nasopharyngeal swab		14 July 2021
ArcDia International Oy Ltd	mariPOC Quick Flu+	2079	<i>Prospective clinical field study</i> FI: Clinical performance of the test was evaluated against qRT-PCR with nasopharyngeal swab specimens collected from patients suspected of acute SARS-CoV-2 infection. Sensitivity of the mariPOC test was 100.0% (13/13) directly from swab specimens and 84.4% (38/45) from swab specimens in undefined transport mediums. Specificity was 100.0%.	100 % sensitivity 100 % specificity NP swab	FI	Nucleo-protein	Nasopharyngeal swab		14 July 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Artron Laboratories Inc.	Artron COVID-19 Antigen Test	1618	<i>Retrospective in vitro study</i>	96.67% sensitivity, Nasal swab 91.67% sensitivity, NP swab 100 % specificity Nasal/NP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab		14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%						
Asan Pharmaceutical Co., Ltd	Asan Easy Test COVID-19 Ag	1654	<i>Retrospective in vitro study</i>	94.67% sensitivity, 97.71% specificity Nasal swab	DE ^[2]	Unknown	Nasal swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 97.71%						
Assure Tech. (Hangzhou) Co., Ltd.	ECOTEST COVID-19 Antigen Rapid Test Device	770	<i>Retrospective in vitro study</i>	92.5 % sensitivity 99.2 % specificity Nasal/NP/OP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		14 July 2021
	ECOTEST COVID-19 Antigen Rapid Test Device	2350	<i>Retrospective in vitro study</i>		UK				
Avalun	Ksmart® SARS-COV2 Antigen Rapid Test	1800	<i>Retrospective in vitro study</i>	Sensitivity: 93.18% Specificity: 99.32% NP swab	DE ^[2]	Unknown	Nasopharyngeal swab		7 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99,32%						
AXIOM Gesellschaft für Diagnostica und Biochemica mbH	COVID-19 Antigen Rapid Test	2101	<i>Retrospective in vitro study</i>	98% sensitivity 100% specificity NP/Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Throat swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Azure Biotech, Inc.	COVID-19 Antigen Rapid Test Device	1906	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (NP swab): Sensitivity of 86% at Ct < 25; Manufacturer specificity: 99.2%	95% sensitivity 99.2% specificity NP swab	DE ^[2]	Unknown	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		10 May 2021 ¹⁹
Becton Dickinson	BD Veritor™ System for Rapid Detection of SARS CoV 2	1065	<i>Prospective clinical field studies</i> ES: Prospective study in four Spanish hospitals (n = 476); 108 positive samples, 368 negative samples. Sensitivity: 92%, specificity: 98.6%. NL: Independent field study in symptomatic individuals (n=979, PCR positive n=161) - sampling was Nasal mid-turbinate + OP swab. Sensitivity overall: 79.5% - Sensitivity Ct<30: 93.2% - Specificity overall: 99.8% SE: Karolinska hospital evaluation of Lot 0255648. Patient samples: 95 PCR positive, 150 negative. No detailed sample description available. Sensitivity 45%, specificity 97%. Sensitivity Ct<25 = 87.8%. <i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 83% at Ct ≤ 25; Manufacturer specificity: 99.6%	Clinical Sensitivity: 91.1 % Clinical Specificity: 99.6 % Nasal swab	DE ^[2] , ES, NL, SE	Nucleo-protein	Nasal swab		7 July 2021

¹⁹ This rapid antigen test, device ID 1906, was removed from the EU common list on 21 December 2021. The grace period will end on 15 Feb 2022, 23:59 CET.

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Becton Dickinson	BD Kit for Rapid Detection of SARS-CoV-2	2282	<p><i>Prospective clinical field studies</i></p> <p>ES: Prospective study in four Spanish hospitals (n = 476); 108 positive samples, 368 negative samples. Sensitivity: 92%, specificity: 98.6%.</p> <p>NL: Independent field study in symptomatic individuals (n=979, PCR positive n=161) - sampling was Nasal mid-turbinate + OP swab. Sensitivity overall: 79.5% - Sensitivity Ct<30: 93.2% - Specificity overall: 99.8%</p>	Clinical Sensitivity: 91.1 % Clinical Specificity: 99.6 % Nasal swab	ES, NL	Nucleo-protein	Nasal swab		10 November 2021
Beijing Hotgen Biotech Co., Ltd	Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)	1870	<p><i>Prospective clinical field study</i></p> <p>FIND evaluation Brazil (15 September 2021) 453 samples, nasal swab. Clinical sensitivities: Days \leq 7: 90.1%; Ct \leq 33: 89.5%; Ct \leq 25: 95.5%. Clinical specificity: 100% UK (15 September 2021) 248 samples, NP swab. Clinical sensitivities: Days \leq 7: 84.4%; Ct \leq 33: 80.6%; Ct \leq 25: 82.8%. Clinical specificity: 99.4%</p> <p><i>Retrospective in vitro study</i></p> <p>DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct \leq 25; Manufacturer specificity: 99.76%</p>	97.1% sensitivity 99.76% specificity	DE ^[2]	Nucleo-protein	Nasal swabs, Throat swabs, ! Saliva		10 May 2021
Beijing Hotgen Biotech Co., Ltd	Coronavirus (2019-nCoV)-Antigentest	2807	<p><i>Retrospective in vitro study</i></p> <p>DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct \leq 25; Manufacturer specificity: 98.88%</p>	Clinical sensitivity: 96.95% Clinical specificity: 98.88%	DE ^[2]	Nucleo-capsid protein	Nasal swab		21 January 2022

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Beijing Jinwofu Bioengineering Technology Co.,Ltd.	Novel Coronavirus (SARS-CoV-2) Antigen Rapid Test Kit	2072	<i>Retrospective in vitro study</i>	96.88 % sensitivity 100 % specificity Nasal/ NP/ OP swab	DE ^[2]	Nucleo-protein	Anterior nasal swab, Nasal swab, Nasopharyngeal swab, Oropharyngeal swab ! Saliva		14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25 + Manufacturer specificity: 100%						
Beijing Kewei Clinical Diagnostic Reagent Inc	COVID19 Antigen Rapid Test Kit	1778	<i>Retrospective in vitro study</i>	Clinical Sensitivity: 96.18 % Specificity: 100%	DE ^[2]	Unknown	Nasal swab		21 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25 + Manufacturer specificity: 100%						
Beijing Lepu Medical Technology Co., Ltd	SARS-CoV-2 Antigen Rapid Test Kit	1331	<i>Retrospective in vitro study</i>	92.00% sensitivity, 99.26% specificity Nasal swab	DE ^[2]	Unknown	Nasal swab, Nasopharyngeal swab		17 February 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.26%						
Beijing O&D Biotech Co., Ltd.	COVID-19 Antigen Rapid Test	2494	<i>Retrospective in vitro study</i>	sensitivity: 92.17%, specificity: 98.67 %	DE ^[2]	Nucleo-capsid protein	Nasal swab, Oropharyngeal swab		20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.67%						
Beijing Wantai Biological Pharmacy Enterprise Co., Ltd	Wantai SARS-CoV-2 Ag Rapid Test (colloidal gold)	1485	<i>Prospective clinical field study</i>	93.2% sensitivity 98.2% specificity Nasal swab	CZ, DE ^[2]	Unknown	Anterior nasal swab, Nasal swab, Nasopharyngeal swab, Oropharyngeal swab ! Saliva		14 July 2021
			CZ: Independent prospective study by Public Health Institute Ostrava (CZ), including NP swabs from unselected symptomatic and asymptomatic participants. Sensitivity 80.6%, specificity 98.5% on 155 pos. and 325 neg. samples (as resulting by RT-PCR). Ct not reported. N total = 480						
			<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.2%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
BioGnost Ltd	CoviGnost AG Test Device 1x20	2247	<i>Retrospective in vitro study</i>	Sensitivity: 96%, Specificity: 99% NP swab	HR	Unknown	Nasopharyngeal swab		23 July 2021
			HR: 300 NP samples (retrospective), symptomatic (<7 dps): 200 PCR+ samples (range Ct 16-30), Ct<30: sensitivity 96.5%. 100 PCR- samples: specificity 100%						
Biohit Healthcare (Hefei) Co., Ltd.	SARS-CoV-2 Antigen Rapid Test Kit (Fluorescence Immunochromatography)	1286	<i>Retrospective in vitro study</i>	Sensitivity: 96.77% Specificity: 98.9% NP/OP swab	DE ^[2]	Nucleo-capsid protein	Anterior nasal swab		23 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.9%						
Biohit Healthcare (Hefei) Co., Ltd.	SARS-CoV-2 Antigen Rapid Test (Colloidal Gold Method)	2230	<i>Retrospective in vitro study</i>	Sensitivity: 96.12%, Specificity: 99.49 %	DE ^[2]	Nucleo-capsid protein	Nasal swab		8 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.49%						
BioMaxima SA	SARS-CoV-2 Ag Rapid Test	2035	<i>Prospective clinical field studies</i>	Sensitivity: 95% Specificity: 99% NP Swab	DE ^[2] , FR, PL	Nucleo-protein	Nasopharyngeal swab		23 July 2021
			FR: NP swabs, Diagnostic sensitivity: 96,4% (80/83) (95% CI: 89,8-99,2%); diagnostic specificity: 99,2%, (120/121) PL: Evaluation of the test was performed on 480 samples of NP swabs taken from patients with symptoms of COVID-19 and from people in contact with an infected person but without symptoms of infection. Positive results were obtained in 205 patients and in the molecular test 213 people. Negative results were obtained in 275 people and in the molecular test 267 people. The above results permitted calculation of diagnostic sensitivity, which was 93.43% (95% CI: 91.61%~97.19%) and diagnostic specificity, which was 97.75% (95% CI: 93.74%~98.92%)						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
			<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99%						
Biomerica Inc.	Biomerica COVID-19 Antigen Rapid Test (nasopharyngeal swab)	1599	<i>Retrospective in vitro study</i>	Clinical Sensitivity: 94.7%; Clinical specificity: 99.7% Nasal/NP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab		7 July 2021
			<i>Prospective clinical field study</i>						
			FIND evaluation Brazil (20 April 2021) 400 samples, NP swab. Clinical sensitivities: Days ≤ 7: 92.2%; Ct ≤ 33: 91.4%; Ct ≤ 25: 94.8%. Clinical specificity: 97.3% Brazil (30 March 2021) 218 samples, Nasal/NP swab. Clinical sensitivities: Days ≤ 7: 92.5% (N/NP); Ct ≤ 33: 97.2% (N/NP); Ct ≤ 25: 100% (N/NP); Clinical specificity: 98.6%	Clinical Sensitivity: 90.91 % Clinical Specificity: 99.43 % Nasal swab, NP swab	DE ^[2] Brazil	Unknown	Nasal swab, Nasopharyngeal swab		7 July 2021
			<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98,6%						
			<i>Prospective clinical field studies</i>						
BIO-RAD	CORONAVIRUS AG RAPID TEST CASSETTE	2031	ES⁷: <ul style="list-style-type: none"> Prospective study; 96 positive samples and 269 negative samples. Sensitivity 94%. Specificity 99.2%. No Ct distribution specified. NP swab: sensitivity 98,3%; specificity 99,6% (119 positive samples, 746 negative samples) Nasal swab: sensitivity 97,2%; specificity 100% (109 positive samples, 128 negative samples) 	Clinical Sensitivity: 98% (NP: 98,32% / Nasal: 97,25%) Clinical Specificity: 99% (NP: 99,6% / Nasal: 100%)	ES	Nucleo-protein	Nasal swab, Nasopharyngeal swab		7 July 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
BioSpeedia International	COVID19Speed-Antigen Test BSD_0503	2380	<i>Prospective clinical field studies</i>	Clinical sensitivity: 97.5% Clinical specificity: 99.3%	FR	Nucleo-capsid protein	Nasopharyngeal swab		21 January 2022
			FR: Independent prospective study by the University Hospital of Saint-Etienne: samples from unselected symptomatic and asymptomatic individuals (255 pos., 365 neg.), overall sensitivity: 95.29% (sensitivity Ct<25: 97.72%), specificity: 99.73%.						
BIOSYNEX SWISS S.A.	BIOSYNEX COVID-19 Ag BSS	1223	<i>Prospective clinical field studies</i>	96% sensitivity, 100% specificity, NP/Nasal swab	BE, DE ^[2] , FR, NL ^[5] , SE, CH	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab	DK: This RAT detects both Delta and Omicron on a comparable level with the wild-type (Wuhan).	17 February 2021
			BE^[6]: Small-scale head-to-head comparison of 5 RATs in Belgian hospital lab. Biosynex overall sensitivity (Ct range 14.6 – 35.5): 52/58 samples (89.7%). Sensitivity for Ct≤25: 18/18 samples. Overall specificity only 46.2%, probably linked to the use of transport medium instead of the swab included in the kit. FR: NP swabs, prospective study (71/71) : sensitivity 100% (45/45, specificity 100% NL: Independent field study, mainly symptomatic individuals (n=568, PCR positive n=39), NP swab; sensitivity Ct≤30: 96.0%, sensitivity ≤25: 100%; specificity overall: 100% NL: Independent field study, symptomatic individuals (n=270, PCR positive n=17), NP+OP swab; sensitivity Ct≤30: 94.1%, sensitivity Ct≤25: 100%; specificity overall: 100% SE: Karolinska hospital evaluation of Lot 20100103. Patient samples; 95 PCR positive, 150 negative. No detailed sample description available. Sensitivity 76%, specificity 96%. Sensitivity Ct<25 = 100%.						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
			<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%						
BIOSYNEX SA	BIOSYNEX COVID-19 Ag+ BSS	1494	<i>Prospective clinical field study</i>	Clinical Sensitivity: 97.5 % Specificity: 99% Nasal swab, NP swab	FR UK	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab		7 July 2021
			<i>Retrospective in vitro study</i>						
BIOTEKE CORPORATION (WUXI) CO., LTD	SARS-CoV-2 Antigen Test Kit (colloidal gold method)	2067	DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 95% at Ct ≤ 25; Manufacturer specificity: 99.28%	96.49 % sensitivity 99.28 % specificity OP/NP swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab, Oropharyngeal swab		14 July 2021
			<i>Retrospective in vitro study</i>						
Biotal Health S.L.U.BIOTAL HEALTH S.L.U	biotal SARS-CoV-2 Ag Card	2013	BE: Validation study 1: sensitivity 91.7% for Ct<25; Validation study 2: 94% for Ct<25. Manufacturer specificity: 99%	Sensitivity: 96%, Specificity: 99% NP swab	BE	Nucleo-protein	Nasopharyngeal swab		23 July 2021
			<i>Prospective clinical field study</i>						
Boditech Med Inc	AFIAS COVID-19 Ag	1989	NL: Independent field study in mild symptomatic (n= 427, PCR positive: 106); unknown swab, overall sensitivity: 81.1%, sensitivity Ct <30: 96.4%; specificity: 100%,	Sensitivity: 91.7%, Specificity: 98.7% NP swab	NL	Nucleo-protein	Nasopharyngeal swab		23 July 2021
			<i>Retrospective in vitro study</i>						
BTNX Inc	Rapid Response COVID-19 Antigen Rapid Test	1236	DE: Positive evaluation by Paul-Ehrlich-Institut (NP swab): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%	90.2% sensitivity 100% specificity NP swab, NP swab, OP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		10 May 2021
			<i>Prospective clinical field study</i>						
CerTest Biotec	CerTest SARS-CoV-2 Card test	1173	ES: Ct ≤ 25, sensitivity: 94,0%; sensitivity for samples within the first 5 days after symptom onset: 84,8%; 150 positive samples, 170 negative samples	92.9% sensitivity 99.6% specificity NP swab	DE ^[2] , ES	Nucleo-capsid protein	Nasopharyngeal swab		17 February 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Cesna Biyoteknoloji Araştırma Geliştirme Laboratuvar Sist. İnş. Müh. Dan. San. Tic. Ltd. Şti .	CHECK UP SARS-COV-2 NASAL ANTIGEN RAPID TEST	2696	<i>Retrospective in vitro study</i>	Clinical Sensitivity: 99.3 % Clinical Specificity: 98.8 %	DE ^[2]	Nucleo-capsid protein	Nasal swab		21 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.8%						
Cesna Biyoteknoloji Araştırma Geliştirme Laboratuvar Sist. İnş. Müh. Dan. San. Tic. Ltd. Şti .	CHECK UP SARS-COV-2 NASOPHARYNGEAL RAPID ANTIGEN TEST	2746	<i>Retrospective in vitro study</i>	Clinical Sensitivity: 99.3 % Clinical Specificity: 99.7 %	DE ^[2]	Nucleo-capsid protein	Nasopharyngeal swab		21 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.7%						
Chil Tıbbi Malzeme Sanayi ve Ticaret Limited Şirketi	CHIL COVID-19 Antigen Rapid Test (Nasopharyngeal / Oropharyngeal Swab-Casette)	1691	<i>Retrospective in vitro study</i>	Sensitivity 99.01% Specificity: 99.57%	DE ^[2]	Nucleo-protein	Nasopharyngeal swab, Oropharyngeal swab		20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.57%						
Chongqing M&D Biotechnology Co. Ltd	2019-nCoV Antigen Test Kit	2150	<i>Retrospective in vitro study</i>	sensitivity: 91.53%, specificity:100%	DE ^[2]	Nucleo-protein	Nasopharyngeal swab		20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 95% at Ct ≤ 25; Manufacturer specificity: 100%						
Core Technology Co., Ltd	Coretests COVID-19 Ag Test	1919	<i>Retrospective in vitro study</i>	98.1% sensitivity 99.6% specificity NP swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 88% at Ct ≤ 25; Manufacturer specificity: 99.6%						
CTK Biotech, Inc	OnSite COVID-19 Ag Rapid Test	1581	<i>Prospective clinical field study</i>	Clinical Sensitivity: 92.3 % Clinical Specificity: 100 % Nasal, NP swab	DK, ES	Nucleo-protein	Nasal swab, Nasopharyngeal swab	CH: Sensitivity was compared between Omicron and other VOCs (Alpha, Beta, Gamma and Delta) as well as an early-pandemic SARS-CoV-2 isolate.	7 July 2021
			DK: 107 samples; Nasal swab - clinical sensitivity 86%; (from asymptomatic and mild symptomatic individuals), Clinical specificity: 100%						
DDS DIAGNOSTIC	Test Rapid Covid-19 Antigen (tampon nazofaringian)	1225	<i>Prospective clinical field study</i>	98.77% sensitivity 99.03% specificity Nasal swab	RO China	Unknown	Nasal swab		10 May 2021
			RO: Clinical study based on 228 COVID-19 positive samples and 597 COVID-19 negative samples. All the samples were confirmed using PCR (Applied Biosystems™ 7500 and SLAN®- 96P) and clinical symptoms. The relative sensitivity of Rapid Test COVID-19						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
			Antigen (Nasopharyngeal Swab) was 99.56%, the relative specificity was 99.66%, and the accuracy was 99.64% compared to the qRT-PCR result.						
DNA Diagnostic	COVID-19 Antigen Detection Kit	2242	<i>Retrospective in vitro study</i>	Sensitivity: 93.8%, Specificity: 99.6% Nasal swab	DE ^[2] UK	Nucleo-protein	Nasal swab		23 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct < 25; Manufacturer specificity: 99.56%						
DNA Diagnostic	SARS-CoV-2 Antigen Rapid Test	2756	<i>Retrospective in vitro study</i>	Clinical sensitivity: 93.4% Clinical specificity: 99.3%	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab		21 January 2022
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct < 25; Manufacturer specificity: 99.3%						
Dräger Safety AG & Co. KGaA	Dräger Antigen Test SARS-CoV-2	2273	<i>Prospective clinical field studies</i>	Sensitivity: 96.1% (Ct values ≤25) Specificity: 99.6%	DE ^[2] CH	Nucleo-capsid protein	Nasal swab		20 October 2021
			DE: Independent prospective study, mainly symptomatic <7 dps (n=378, PCR positive = 70), self-collected nasal swab; sensitivity overall: 88.6%, sensitivity Ct<26: 96.8%; specificity overall: 99.7%						
			CH: Independent prospective study, mainly symptomatic ≤7 dps (n=464, PCR positive = 57), self-collected nasal swab; sensitivity Ct<30: 85.1%, sensitivity Ct<26: 90.0%; specificity overall: 100%						
			<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 95% at Ct < 25; Manufacturer specificity: 99.6%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Dynamiker Biotechnology(Tianjin) Co., Ltd.	Dynamiker SARS-CoV-2 Ag Rapid Test	2533	<i>Retrospective in vitro study</i>	sensitivity: 95.7%, specificity: 99.1%	DE ^[2]	Nucleo-capsid protein	Nasopharyngeal swab, Oropharyngeal swab		20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer Specificity: 99.1%						
Edinburgh Genetics Limited	Edinburgh Genetics ActivXpress+ COVID-19 Antigen Complete Testing Kit	1243	<i>Prospective clinical field study</i>	Clinical Sensitivity 97.27% NP swab Clinical Specificity 99.62% NP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		14 July 2021
			FIND evaluation Peru (26 April 2021) 120 samples, NP swab. Clinical sensitivities: Days ≤ 7: 62%; Ct ≤ 33: 75%; Ct ≤ 25: 100%. Clinical specificity: 100%						
			<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer Specificity: 99,24%	Clinical Sensitivity 95.63% OP swab Clinical Specificity 99.24% OP swab	Peru				
Eurobio Scientific	EBS SARS-CoV-2 Ag Rapid Test	1739	<i>Prospective clinical field study</i>	Clinical Sensitivity: 95.7 % Nasal swab	DE ^[2] , FR	Nucleo-protein	Nasal swab		7 July 2021
			FR: Validation study data: 119 positive and 125 negative samples; sensitivity 93%, specificity: 99%						
			<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99,1%						
Fujirebio	ESPLINE SARS-CoV-2	2147	<i>Prospective clinical field study</i>	Clinical Sensitivity: 87.8 % (n=98, Ct<33)) Clinical Specificity: 100 % NP swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab		7 July 2021
			FIND evaluation DE (29 March 2021) 723 samples, NP swab. Sensitivities: Days ≤ 7: 88.5%; Ct ≤ 33: 87.8%; Ct ≤ 25: 92.4%. Clinical specificity: 100% South Africa (6 Oct 2021) 494 samples, NP swab. Sensitivities: Days ≤ 7: 75%; Ct ≤ 33: 78.9%; Ct ≤ 25: 90.1%. Clinical specificity: 99.7%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
			<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99,13%						
GA Generic Assays GmbH	GA CoV-2 Antigen Rapid Test	1855	<i>Retrospective in vitro study</i>	Sensitivity: 97.059%, Specificity: 99.2% NP swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab		23 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.2%						
Genobio Pharmaceutical Co., Ltd.	Virusee® SARS-CoV-2 Antigen Rapid Test (Colloidal Gold)	2642	<i>Retrospective in vitro study</i>	OP: sensitivity: 97.14%, specificity: 99.28% NP: sensitivity: 97.22%, specificity: 99.23%	DE ^[2]	Nucleo-capsid protein	Oropharyngeal swab; Nasopharyngeal swab		8 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.2%						
Genrui Biotech Inc	SARS-CoV-2 Antigen Test Kit (Colloidal Gold)	2012	<i>Retrospective in vitro study</i>	Sensitivity: 91.15% Specificity: 99.02% Nasal/NP/OP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		7 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99,02%						
GenSure Biotech Inc	GenSure COVID-19 Antigen Rapid Test Kit	1253	<i>Retrospective in vitro study</i>	96.86% sensitivity 100% specificity Nasal swab	DE ^[2] UK	Unknown	Nasal swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 100%						
GenSure Biotech Inc.	GenSure COVID-19 Antigen Rapid Test Kit	2853	<i>Retrospective in vitro study</i>	Clinical Sensitivity 96.73 % Clinical Specificity 100 %	DE ^[2]	Nucleo-capsid protein	Nasal swab ! Saliva	Manufacturer: Lineages detected: B.1.1.7 (Alpha), B.1.351 (Beta), B.1.429 (Epsilon), B.1.525 (Eta), B.1.617.1 (Kappa), B.1.617.2 (Delta), B.1.617.3, B.1.621 (Mu), C.37 (Lambda), P.1 (Gamma)	10 February 2022
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 100%						
Getein Biotech, Inc	SARS-CoV-2 Antigen (Colloidal Gold)	1820	<i>Retrospective in vitro study</i>	97.06% sensitivity 98.71% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab ! Saliva		14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.71%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Getein Biotech, Inc.	One Step Test for SARS-CoV-2 Antigen (Colloidal Gold)	2183	<i>Retrospective in vitro study</i>	97.06% sensitivity 98.71% specificity Nasal swab	DE ^[2] UK	Nucleo-protein	Nasal swab ! Saliva		16 June 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 90% at Ct ≤ 30 and 100% at Ct ≤ 25; Manufacturer specificity: 98.71%						
Glallergen CO., LTD.	Novel Coronavirus (2019-nCoV) Antigen Test Kit (Colloidal gold immunochromatography)	2695	<i>Retrospective in vitro study</i>	Clinical Sensitivity: 94.44 % Clinical Specificity: 99.02 %	DE ^[2]	Nucleo-capsid protein	Nasal swab		21 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.02%						
Goldsite Diagnostic Inc.	SARS-CoV-2 Antigen Kit (Colloidal Gold)	1197	<i>Retrospective in vitro study</i>	93.04% sensitivity; 100% specificity Nasal swab	FR, DE ^[2] , ES UK	Unknown	Nasal swab ! Other		14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%						
Green Cross Medical Science Corp.	GENEDIA W COVID-19 Ag	1144	<i>Retrospective in vitro study</i>	100% sensitivity 90.1% sensitivity 100% specificity NP swab, Anterior nasal swab	DE ^[2]	Nucleo-protein	Anterior nasal swab, Nasopharyngeal swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 83% at Ct ≤ 25; Manufacturer specificity: 100%						
Guangdong Hecin Scientific, Inc.	2019-nCoV Antigen Test Kit (colloidal gold method)	1747	<i>Retrospective in vitro study</i>	96.23% sensitivity 99.07% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 82% at Ct ≤ 25; Manufacturer specificity: 99.07%						
Guangdong Longsee Biomedical Co., Ltd.	2019-nCoV Ag Rapid Detection Kit (Immuno-Chromatography)	1216	<i>Retrospective in vitro study</i>	Sensitivity: NP swab: 95.51%, OP swab: 95.22%, Nasal swab: 94.15% Specificity: NP swab: 99.72%, OP swab: 99.72%, Nasal swab: 99.68%	DE ^[2]	Nucleo-capsid protein	Nasopharyngeal swab, Oropharyngeal swab, Nasal swab	Manufacturer: Lineages detected: A.23.1;B.1.1.7 (Alpha); B.1.351 (Beta); B.1.427 (Epsilon); B.1.429 (Epsilon); B.1.525 (Eta); B.1.526 (Iota); B.1.617.1 (Kappa); B.1.617.2(Delta); B.1.617.3;B.1.621(Mu); P.1(Gamma); P.2(Zeta); P.3(Theta); C.37 (Lambda); B.1.616;B.1.1.529(Omicron); B.1.526.1; B.1.526.2	14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.5%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Guangdong Wesail Biotech Co. Ltd	COVID-19 Ag Test Kit	1360	<i>Retrospective in vitro study</i>	90% sensitivity 98% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab		17 February 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98%						
Guangzhou Decheng Biotechnology CO., Ltd	V-CHEK, 2019-nCoV Ag Rapid Test Kit (Immuno-chromatography)	1324	<i>Retrospective in vitro study</i>	<i>Clinical Sensitivity: 95.83% Specificity 99.57% Nasal swab</i>	DE ^[2]	Nucleo-protein	Nasal swab	Manufacturer: Lineages detected: B.1.1.7(Alpha); B.1.351 (Beta); B.1.429(Epsilon); B.1.617.1 (Kappa); B.1.617.2 (Delta); B.1.617.3; B.1.621 (Mu); P.1(Gamma); C.37(Lambda); B.1.1.529 (Omicron)	7 July 2021 ²⁰
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct < 25; Manufacturer specificity: 99,5%						
Guangzhou Wondfo Biotech Co., Ltd	Wondfo 2019-nCoV Antigen Test (Lateral Flow Method)	1437	<i>Prospective clinical field study</i>	Sensitivity: 98.11% Specificity: 99.72%	DE ^[2] CH, UK	Nucleo-capsid protein	Nasopharyngeal swab	CH: Sensitivity was compared between Omicron and other VOCs (Alpha, Beta, Gamma and Delta) as well as an early-pandemic SARS-CoV-2 isolate. Manufacturer: Lineages detected: B.1.1.7 (Alpha), B.1.351 (Beta), B.1.617.2 (Delta)	10 May 2021
			FIND evaluation CH (25 Feb 2020) 328 samples, NP swab. Clinical sensitivities: Days ≤ 7: 85.7%; Ct ≤ 33: 92.2%; Ct ≤ 25: 100%. Clinical specificity: 100%						
			Brazil (10 Oct 2021) 237 samples, NP swab. Clinical sensitivities: Days ≤ 7: 90.4%; Ct ≤ 33: 89.3%; Ct ≤ 25: 96.7%. Clinical specificity: 98.8%						
			<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 88% at Ct ≤ 25; Manufacturer specificity: 99.74%						
Hangzhou AllTest Biotech Co., Ltd	COVID-19 Antigen Rapid Test	1257	<i>Prospective clinical field study</i>	93,40% sensitivity, 99,90% specificity NP swab	FR	Nucleo-capsid protein	Nasopharyngeal swab	Manufacturer: Lineages detected: A.23.1; AT.1;B.1.1.7(Alpha); B.1.351 (Beta); B.1.427(Epsilon); B.1.429(Epsilon); B.1.525 (Eta);B.1.526(Iota); B.1.617.1 (Kappa); B.1.617.2 (Delta);	10 May 2021
			FR: Prospective study, sensitivity 96,4% (80/83), specificity 99,2% (120/121)						

²⁰ This rapid antigen test, device ID 1324, was removed from the EU common list on 21 December 2021. The grace period will end on 15 Feb 2022, 23:59 CET.

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
								B.1.617.3;B.1.621 (Mu);C.36; P.1(Gamma); P.2(Zeta); P.3 (Theta); C.37 (Lambda); B.1.616; B.1.1.529(Omicron); B.1.526.1; B.1.526.2;	
Hangzhou Biotest Biotech Co., Ltd	COVID-19 Antigen Rapid Test Cassette (Nasal Swab)	1876	<i>Retrospective in vitro study</i>	Sensitivity: 93.2%, Specificity: 99.2% Nasal swab	DE ^[2]	Nucleo-capsid protein	Nasal swab	Manufacturer: Lineages detected: B.1.1.7 (Alpha); B.1.351(Beta); B.1.617.1(Kappa); B.1.617.2 (Delta); P.1(Gamma); B.1.1.529(Omicron);	8 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.2%						
Hangzhou Clongene Biotech Co., Ltd.	COVID-19 Antigen Rapid Test Cassette	1610	<i>Retrospective in vitro study</i>	Clinical Sensitivity: 91.4 % Clinical Specificity: 100 % NP swab	DE ^[2] UK	Nucleo-protein	Nasopharyngeal swab		7 July 2021
		DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,4% at Ct ≤ 25; Manufacturer specificity: 100%							
	Covid-19 Antigen Rapid Test Kit	1363	<i>Retrospective in vitro study</i>	98.5% (Ct<33) sensitivity Nasal swab	DE ^[2] CH	Nucleo-protein	Nasal swab		17 February 2021
	COVID-19/Influenza A+B Antigen Combo Rapid Test	1365	<i>Retrospective in vitro study</i>	91% sensitivity 100% specificity NP swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab		10 May 2021
Hangzhou Immuno Biotech Co., Ltd	Immunobio SARS-CoV-2 Antigen ANTERIOR NASAL Rapid Test Kit (minimal invasive)	1844	<i>Retrospective in vitro study</i>	94% sensitivity 100% specificity Nasal swab, NP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab		10 May 2021
	SARS-CoV2 Antigen Rapid Test	2317	<i>Retrospective in vitro study</i>	Clinical Sensitivity: 98 % Clinical Specificity: 100 % Anterior nasal swab, NP swab, OP swab,	DE ^[2]	Nucleo-protein	Anterior nasal swab, Nasopharyngeal swab, Oropharyngeal swab ! Sputum		10 May 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Sigmed Sp. z o.o.	Redtest Professional Sars-CoV-2 Antigen Rapid Test (Covid-19 Ag)	2256	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 88% at Ct ≤ 25; Manufacturer specificity: 100%	sensitivity: 98,13%, specificity: 100%	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		8 December 2021
Hangzhou DIAN Biotechnology Co., Ltd.	COVID-19 Antigen Test Cassette	2629	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 90% at Ct ≤ 25; Manufacturer specificity: 98.4%	Clinical Sensitivity: 97.6 % Clinical Specificity: 98.4 %	DE ^[2]	Unknown	Nasal swab, Nasopharyngeal swab		21 December 2021
Hangzhou Laihe Biotech Co.	LYHER Novel Coronavirus (COVID-19) Antigen Test Kit (Colloidal Gold)	1215	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99,7%	Clinical Sensitivity: 95.07% % Clinical Specificity: 99.74% Nasal swab	DE ^[2] UK	Unknown	Nasal swab		10 May 2021
Hangzhou Lysun Biotechnology Co. Ltd	COVID-19 Antigen Rapid Test Device (Colloidal Gold)	2139	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%	96.46% sensitivity 100% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab		10 May 2021
Hangzhou Sejoy Electronics & Instruments Co.Ltd	SARS-CoV-2 Antigen Rapid Test Cassette	1945	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%	Sensitivity: 94.5%, Specificity:100% Nasal swab	DE ^[2]	Nucleo-capsid protein	Nasal swab		8 December 2021
Hangzhou Testsea Biotechnology Co., Ltd.	Covid-19 Antigen Test Cassette	1392	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (NP swab): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.4%	92.1% sensitivity 98.1% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab		10 May 2021
Healgen Scientific	Coronavirus Ag Rapid Test Cassette	1767	<i>Prospective clinical field studies</i> NL: 1): Clinical field study, symptomatic individuals (n=417, PCR positive n=70), NP swab; sensitivity overall: 75.7%, sensitivity Ct≤30: 85.2%, sensitivity Ct≤25: 90.7%; specificity: 100% 2): Clinical field study, symptomatic individuals (n=240, PCR positive n=21),	98.32 % sensitivity 99.6% specificity (NP swab) 97.25% sensitivity 100% specificity (Nasal swab)	DE ^[2] , NL ^[5]	Nucleo-proteins, S1, S1-RBD, S2	Nasal swab, Nasopharyngeal swab		17 February 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
			<p>NP+OP swab; sensitivity overall: 85.7%, sensitivity Ct_≤30: 89.5%, sensitivity Ct_≤25: 100%; specificity: 100%</p> <p>3): Clinical field study, symptomatic individuals (n=94, PCR positive n=18), NP+OP swab in VTM; sensitivity overall: 90.0%, sensitivity Ct_≤30: 100%, sensitivity Ct_≤25: 100%; specificity: 97.3%</p> <p style="text-align: center;"><i>Retrospective in vitro study</i></p> <p>DE: Positive evaluation by Paul-Ehrlich-Institut (NP swab): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%</p>						
Siemens Healthineers	CLINITEST Rapid COVID-19 Antigen Test	1218	<p style="text-align: center;"><i>Prospective clinical field studies</i></p> <p>NL: 1): Clinical field study, symptomatic individuals (n=417, PCR positive n=70), NP swab; sensitivity overall: 75.7%, sensitivity Ct_≤30: 85.2%, sensitivity Ct_≤25: 90.7%; specificity: 100% 2): Clinical field study, symptomatic individuals (n=240, PCR positive n=21), NP+OP swab; sensitivity overall: 85.7%, sensitivity Ct_≤30: 89.5%, sensitivity Ct_≤25: 100%; specificity: 100% 3): Clinical field study, symptomatic individuals (n=94, PCR positive n=18), NP+OP swab in VTM; sensitivity overall: 90.0%, sensitivity Ct_≤30: 100%, sensitivity Ct_≤25: 100%; specificity: 97.3%</p> <p>ES: Independent prospective study; 192 positive and 258 negative samples (NP swab). Sensitivity: 93,3%, Specificity. 99.2%, compared against NP PCR.</p>	<p>98.32% sensitivity (NP swab) 97.25% sensitivity (Nasal swab) 100% specificity</p>	<p>DE^[2], ES, IE, NL^[5]</p>	<p>Nucleo-proteins, S1, S1-RBD, S2</p>	<p>Nasal swab, Nasopharyngeal swab</p>		<p>17 February 2021</p>

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
			IE Independent prospective study: overall sensitivity 79,4% and among Ct ≤25 89,7% (34 samples). Overall specificity 98.8% (249 samples) and in only the symptomatic cohort 98% (153 samples). <i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (NP swab): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%						
Hoyotek Biomedical Co.,Ltd.	Corona Virus (COVID-19) Antigen Rapid Test (Colloidal Gold)	1929	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 90% at Ct ≤ 25; Manufacturer specificity: 99%	NP swab - Sensitivity: 96%, Specificity: 99% OP swab - Sensitivity: 93%, Specificity: 97.5%	DE ^[2]	Unknown	Nasopharyngeal swab, Oropharyngeal swab		20 October 2021
Hubei Jinjian Biology Co., Ltd	SARS-CoV-2 Antigen Test Kit	1759	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.3%	Sensitivity: 98.02% Nasal Swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab		23 July 2021
Humasis	Humasis COVID-19 Ag Test	1263	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (NP swab): Sensitivity of 88% at Ct ≤ 25; Manufacturer specificity: 100%	95.3% sensitivity 100% specificity Nasal swab	DE ^[2] UK	Unknown	Nasal swab		10 May 2021
Innova Medical Group.Inc	Innova SARS-CoV-2 Antigen Rapid Qualitative Test	1801	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99%	Sensitivity 94.0% : CI 95% (86.7%-98.0%) Specificity: 99.6% - CI:95%(99.4%-99.8%)	DE ^[2]	Nucleo-capsid protein	Anterior nasal swab, Nasal swab		20 October 2021
Innovation Biotech(Beijing) Co.Ltd	Coronavirus (SARS-Cov-2) Antigen Rapid Test Cassette (Nasal swab)	2278	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99%	Sensitivity: 95.6% Specificity: 100%	DE ^[2]	Nucleo-protein	Nasal swab		20 October 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
InTec PRODUCTS, INC.	Rapid SARS-CoV-2 Antigen Test (nasopharyngeal specimen)	2419	<i>Retrospective in vitro study</i>	Sensitivity 90.2% (95% CI: 83.1% to 95.0%); Specificity 100.0% (95% CI: 96.5% - 100.00%)	DE ^[2]	Nucleo-capsid protein	Nasopharyngeal swab		20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%						
Jiangsu Bioperfectus Technologies Co., Ltd.	Novel Corona Virus (SARS-CoV-2) Ag Rapid Test Kit	2107	<i>Retrospective in vitro study</i>	Sensitivity: NP: 96.67% (95% CI: 88.64%~99.08%), Nasal: 97.06% (95% CI: 93.30%~98.74%) Specificity: NP: 97.87% (95% CI: 95.12%~99.09%), Nasal: 99.15% (95% CI: 98.25%~99.59%)	DE ^[2] UK	Nucleo-protein	Nasal swab, Nasopharyngeal swab,		14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.15%						
Jiangsu Diagnostics Biotechnology Co., Ltd	COVID-19 Antigen Rapid Test Cassette (Colloidal Gold)	1920	<i>Retrospective in vitro study</i>	97.58 % sensitivity 100 % specificity Nasal/NP/ OP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab, Throat swab		14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%						
Jiangsu Konsung Bio-Medical Science and Technology Co.	H+W_COVID-19 Antigen Rapid Test Kit (Colloidal Gold)	1899	<i>Retrospective in vitro study</i>	Clinical Sensitivity 97.14 % Clinical Specificity 99.34 %	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		10 February 2022
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.34%						
Jiangsu Medomics medical technology Co.,Ltd.	SARS-CoV-2 antigen Test Kit (LFIA)	2006	<i>Retrospective in vitro study</i>	Sensitivity: 97.73% Specificity: 99.51% Anterior nasal swab, NP swab	DE ^[2]	Nucleo-protein	Anterior nasal swab, Nasopharyngeal swab, Throat swab		7 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99,51%						
Jiangsu Mole Bioscience CO., LTD.	SARS-CoV-2 Antigen Test Cassette	2586	<i>Retrospective in vitro study</i>	sensitivity: 98.31 %, specificity: 99.17 %	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab		8 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99,17%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Jiangsu Well Biotech Co., Ltd.	COVID-19 Ag Rapid Test Device	2144	<i>Retrospective in vitro study</i>	sensitivity: 94.74%, specificity: 99%	DE ^[2]	Nucleo-protein	Nasal swab		20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99%						
JINAN BABIO BIOTECHNOLOGY CO., LTD., China	SARS-CoV-2 Antigen Rapid Detection Kit (Colloidal Gold Method)	2151	<i>Prospective clinical field study</i>	Clinical Sensitivity 96.67 % Clinical Specificity 100 %	PL	Nucleo-capsid protein	Nasal swab		10 February 2022
			PL: Prospective study with nasal samples in a Polish hospital; 210 positive samples, overall sensitivity 96,7%; 450 negative samples, including 100 hospitalized patients and 50 potentially cross-reacting samples. Specificity 100%.						
Joinstar Biomedical Technology Co. Ltd	COVID-19 Rapid Antigen Test (Colloidal Gold)	1333	<i>Retrospective in vitro study</i>	96.1% sensitivity 98.1% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		17 February 2021
IEDAU INTERNATIONAL GMBH	Covid-19 Antigen Schnelltest (Colloidales Gold)	2555	<i>Retrospective in vitro study</i>	OP/Nasal: sensitivity: 96,1%, specificity: 99,2% NP: sensitivity: 97,1%, specificity: 99,2 %	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		8 December 2021
JOYSBIO (Tianjin) Biotechnology Co., Ltd.	SARS-CoV-2 Antigen Rapid Test Kit (Colloidal Gold immunochromatography)	1764	<i>Prospective clinical field studies</i>	98.13% sensitivity Nasal swab	CZ, DE ^[2] CH	Nucleo-capsid protein	Nasal swab		10 May 2021
			CZ: N=225 (90 RT-PCR positive), 60.3% symptomatic patients. Test parameters for a subgroup of symptomatic patients (estimates and 95% confidence intervals): sensitivity 92% (80.8–97.8), specificity 97.6% (91.5–99.7). Test parameters for a subgroup of asymptomatic patients (estimates and 95% confidence intervals): sensitivity 100% 100 (54.1–100), specificity 100% (95.5–100).						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
			FIND Evaluation CH (11 Feb 2021) 265 samples, Nasal swab. Clinical sensitivities: Days \leq 7: 74.2%; Ct \leq 33: 78.9%; Ct \leq 25: 91.3%; Clinical specificity: 99.1%						
Labnovation Technologies Inc.	SARS-CoV-2 Antigen Rapid Test Kit	1266	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94% at Ct \leq 25; Manufacturer specificity: 97.3%	96.3% sensitivity, 97.3% specificity NP/OP swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab, Oropharyngeal swab		10 May 2021
LINKCARE (NANTONG DIAGNOSIS BIO)	COVID-19 Antigen Test Kit (Colloidal Gold)	1353	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct \leq 25; Manufacturer specificity: 99.04%	Clinical Sensitivity: 92.59 % Specificity: 99.04%	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab		21 December 2021
Lumigenex (Suzhou) Co., Ltd	PocRoc [®] SARS-CoV-2 Antigen Rapid Test Kit (Colloidal Gold)	2128	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct \leq 25; Manufacturer specificity: 99,16%	93.33% sensitivity 99.16% specificity Nasal/NP/OP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		10 May 2021
LumiQuick Diagnostics Inc.	QuickProfile™ COVID-19 Antigen Test	1267	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct \leq 25; Manufacturer specificity: 98.8%	93.7% sensitivity, 98.8% specificity NP swab	DE ^[2]	Unknown	Nasopharyngeal swab		10 May 2021
LumiraDX	LumiraDx SARS-CoV-2 Ag Test	1268	<i>Prospective clinical field study</i> SKUP/2021/124: 448 samples: 83 positive samples and 365 negative samples. Nasal specimen: diagnostic sensitivity of 87% (79-92) and diagnostic specificity of 99,5% (98,3-99,9). NP specimen: diagnostic sensitivity of 90% (83-95) and diagnostic specificity of 97,8% (96,0-98,8) (Scandinavian evaluation of laboratory equipment for point of care testing)	97.6% sensitivity 96.6% specificity Nasal swab	DE ^[2] , ES SKUP CH	Nucleo-protein	Nasal swab		17 February 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
			FIND Evaluation DE (8 Oct 2021) 761 samples, NP swab. Clinical sensitivities: Days \leq 7: 86.4%; Ct \leq 33: 87.2%; Ct \leq 25: 92.6%; Clinical specificity: 99.3% Brazil (8 Oct 2021) 251 samples, NP swab. Clinical sensitivities: Days \leq 7: 85.7%; Ct \leq 33: 87.7%; Ct \leq 25: 94.1%; Clinical specificity: 99% <i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct \leq 25; Manufacturer specificity: 98.8%						
MEDsan GmbH	MEDsan SARS-CoV-2 Antigen Rapid Test	1180	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct \leq 25; Manufacturer specificity: 99.8%	92.5% sensitivity 99.8% specificity NP/OP swab	DE ^[2] CH	Unknown	Nasopharyngeal swab, Oropharyngeal swab		17 February 2021
Merlin Biomedical (Xiamen) Co., Ltd.	SARS-CoV-2 Antigen Rapid Test Cassette	2029	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 90% at Ct \leq 30 and 100% at Ct \leq 25; Manufacturer specificity: 98.99%	95.05% sensitivity 98.99% specificity Nasal/NP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab		16 June 2021
MEXACARE GmbH	MEXACARE COVID-19 Antigen Rapid Test	1775	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct \leq 25; Manufacturer specificity: 99,1%	Sensitivity: 96.17% Specificity: 99,1% Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab		7 July 2021
möLab	mö-screen Corona Antigen Test	1190	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct \leq 25; Manufacturer specificity: 99,99%	Sensitivity: 97.25% Specificity: 99.99% NP swab	DE ^[2] , IE	Unknown	Nasopharyngeal swab		10 May 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Mologic Ltd	COVIOS Ag COVID-19 Antigen Rapid Diagnostic Test	2640	<i>Prospective clinical field study</i>	Sensitivity: 90.6%, Specificity:100% Nasal swab	DE ^[2] UK	Nucleo-capsid protein	Nasal swab		8 December 2021
			FIND evaluation DE: Symptomatic and asymptomatic (n=649, PCR positive = 191), nasal and nasal-mouth-throat swab; sensitivity overall: 90.6%, sensitivity Ct ≤ 25: 96.4%; specificity: 100%						
MP Biomedicals	Rapid SARS-CoV-2 Antigen Test Card	1481	<i>Retrospective in vitro study</i>	96.17% sensitivity 99.16% specificity Nasal swab, Anterior nasal swab	DE ^[2] CH, UK	Nucleo-protein	Anterior nasal swab, Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		17 February 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.03%						
Multi-G bvba	Covid19Check-NAS	2260	<i>Retrospective in vitro study</i>	Clinical Sensitivity 97 % (99.35% for Ct values ≤25)) Clinical Specificity 99.5 %	DE ^[2]	Nucleo-capsid protein	Nasal swab	Manufacturer: Lineages detected: A.23.1, AT.1, B.1.1.7 (Alpha), B.1.351 (Beta), B.1.427 (Epsilon), B.1.429 (Epsilon), B.1.525 (Eta), B.1.526 (Iota), B.1.616, B.1.617.1 (Kappa), B.1.617.2 (Delta), B.1.617.3, B.1.621 (Mu), C.36, C.37 (Lambda), P.1 (Gamma), P.2 (Zeta), P.3 (Theta), B.1.526.1, B.1.526.2	10 February 2022
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.5%						
Nal von minden GmbH	NADAL COVID -19 Ag +Influenza A/B Test	2104	<i>Retrospective in vitro study</i>	97% sensitivity 98% specificity NP swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 83% at Ct ≤ 25; Manufacturer specificity: 99.9%						
Nal von minden GmbH	NADAL COVID -19 Ag Test	1162	<i>Prospective clinical field study</i>	97.6% sensitivity 99.9% specificity Nasal swab	DE ^[2] , FR China	Nucleo-protein	Nasal swab ! Serum, Whole blood		17 February 2021
			FIND evaluation CH (26 April 2021) 462 samples, NP swab. Clinical sensitivities: Days ≤ 7: 88.5%; Ct ≤ 33: 92.4%; Ct ≤ 25: 97.8%; Clinical specificity: 99.2%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
			<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 83% at Ct ≤ 25; Manufacturer specificity: 99.9%						
Nanjing Liming Bio-Products Co., Ltd.	StrongStep® SARS-CoV-2 Antigen Rapid Test	2301	<i>Retrospective in vitro study</i>	Sensitivity: 96.19 %, Specificity: 99.26 % Nasal swab	DE ^[2]	Nucleo-capsid protein	Nasal swab		8 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.26%						
Nanjing Norman Biological Technology Co., Ltd.	Novel Coronavirus (2019-nCoV) Antigen Testing Kit (Colloidal Gold)	2506	<i>Retrospective in vitro study</i>	Clinical sensitivity: 91.13% (Saliva); 93.02% (Anterior Nasal); 93.21% (NP) Clinical specificity: 93.02% (Anterior Nasal); 99.23% (Anterior Nasal); 99.29% (NP)	DE ^[2]	Nucleo-capsid protein	Anterior nasal swab, Nasopharyngeal swab, Oropharyngeal swab ! Saliva		10 November 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94% at Ct ≤ 25; Manufacturer specificity: 99.9%						
Nanjing Synthgene Medical Technology Co., Ltd.	SARS-COV-2 Nucleocapsid (N) Antigen Rapid Detection Kit (Colloidal gold method)	2164	<i>Retrospective in vitro study</i>	Clinical sensitivity: 99.33% Clinical specificity: 99.5%	DE ^[2]	Nucleo-capsid protein	Nasopharyngeal swab		21 January 2022
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.5%						
NanoEntek	FREND COVID-19 Ag	1420	<i>Retrospective in vitro study</i>	94.12% sensitivity 100% specificity NP swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 88% at Ct ≤ 25; Manufacturer specificity: 100%						
NanoRepro AG	NanoRepro SARS-CoV-2 Antigen Rapid Test	2200	<i>Retrospective in vitro study</i>	97.2 % sensitivity 98.4% specificity Nasal/NP/OP swab	DE ^[2]	Nucleo-protein	Anterior nasal swab, Nasopharyngeal swab, Oropharyngeal swab		14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 98.4%						
Nantong Egens Biotechnology Co.,Ltd	COVID-19 Antigen Rapid Test Kit	1573	<i>Retrospective in vitro study</i>	sensitivity: 95.8 %, specificity: 99.5 %	DE ^[2]	Nucleo-protein	Nasal swab		10 February 2022
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.5%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
NESAPOR EUROPA SL	MARESKIT COVID-19 ANTIGEN RAPID TEST KIT	2241	<i>Prospective clinical field study</i>	Sensitivity: 95.24% (95% CI: 83.84% to 99.42%), Specificity: 100% (95% CI: 97.22% to 100.00%) Nasal swab	ES	Nucleo-protein	Nasal swab		23 July 2021
			ES: Independent validation study; Nasal test compared to nasal PCR. Sensitivity 95.24% (Ct<30), Specificity 100%.						
Neo-nostics (Suzhou) Bioengineering Co., Ltd.	COVID 19 Antigen Test Kit (Colloidal Gold Method)	2608	<i>Retrospective in vitro study</i>	Clinical Sensitivity 95.93 % Clinical Specificity 99.19 %	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		10 February 2022
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.19%						
New Gene (Hangzhou) Bioengineering Co., Ltd.	COVID-19 Antigen Detection Kit	1501	<i>Retrospective in vitro study</i>	98% sensitivity 99.2% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab ! Saliva, Sputum		16 June 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 92,5% at Ct ≤ 30 and 100% at Ct ≤ 25; Manufacturer specificity: 99.2%						
NG Biotech	Ninonasal	1880	<i>Prospective clinical field study</i>	Clinical sensitivity: 98%, Clinical specificity: 99%	FR	Nucleo-protein	Nasal swab, Nasopharyngeal swab	Manufacturer: Lineages detected: B.1.1.7 (Alpha); B.1.351 (Beta); B.1.427 (Epsilon); B.1.429 (Epsilon); B.1.525 (Eta); B.1.526 (Iota); B.1.617.1 (Kappa); B.1.617.2 (Delta); B.1.621 (Mu); P.1 (Gamma); C.37 (Lambda); B.1.1.529 (Omicron);	10 November 2021
			FR: Prospective validation study for NP and nasal swabs: NP sensitivity 89% (75/84), specificity 99% (92/93). Nasal sensitivity 98% (125/128), specificity 99% (388/390)						
Novatech	SARS-CoV-2 Antigen Rapid Test	1762	<i>Retrospective in vitro study</i>	95 % sensitivity 100% specificity Nasal/ NP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab		14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 100%						
Oncosem Onkolojik Sistemler San. ve Tic. A.S.	CAT	1199	<i>Retrospective in vitro study</i>	93.75% sensitivity 98.04% specificity Nasal swab	DE ^[2]	Nucleo-capsid protein	Nasal swab	Manufacturer: Lineages detected: - B.1.1.7 (Alpha); B.1 617 (Delta); C.37 (Lambda); B.1.1.529 (Omicron)	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 98,04%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
OSANG Healthcare Co., Ltd.	GeneFinder COVID-19 Ag Plus Rapid Test	2741	<i>Prospective clinical field study</i>	Clinical Sensitivity: 94% (95% CI: 87.52% ~ 97.22%) Clinical Specificity: 100% (95% CI: 99.05% ~ 100.00%)	IT	Nucleo-capsid protein	Nasopharyngeal swab	Manufacturer: Lineages detected: B.1.1.7 (Alpha); B.1.351 (Beta); B.1.617.2 (Delta); P.1 (Gamma); P.2 (Zeta); B.1.1.529 (Omicron)	21 December 2021
			IT: Independent prospective evaluation study carried out in Hospital Pugliese Ciaccio, Italy. Sample type: NP swab; sample size: 100 pos., 400 neg.; Sensitivity: 94%; Specificity: 100%						
PCL Inc.	PCL COVID19 Ag Rapid FIA	308	<i>Prospective clinical field study</i>	94,92% sensitivity, 99,99% specificity	DE ^[2] , FR	Unknown	Nasopharyngeal Swab		10 May 2021
			FR: Validation study data: NP swabs, sensitivity 94.29% (33/35) and specificity 100% (70/70)						
PCL Inc.	PCL COVID19 Ag Gold	2243	<i>Prospective clinical field study</i>	Clinical Sensitivity: 90.83 % Clinical Specificity: 99,5 %	FR	Nucleo-protein	Nasal swab, Nasopharyngeal swab ! Saliva		7 July 2021
			FR: Validation study data: 120 positive and 200 negative samples; sensitivity 92%, specificity: 100%						
PerGrande Bio Tech Development Co., Ltd.	SARS-CoV-2 Antigen Detection Kit (Colloidal Gold Immunochromatographic Assay)	2116	<i>Retrospective in vitro study</i>	94.28% sensitivity 99.11% specificity NP/Nasal/OP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.11%						
Precision Biosensor Inc.	Exdia COVI-19 Ag	1271	<i>Retrospective in vitro study</i>	93.9% sensitivity 98% specificity NP swab	DE ^[2] CH	Unknown	Nasopharyngeal swab		17 February 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.3%						
Prognosis Biotech	Rapid Test Ag 2019-nCov	1495	<i>Retrospective in vitro study</i>	Clinical Sensitivity: 95.56 % Specificity: 99,58% Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab		7 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99,58%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Qingdao Hightop Biotech Co. Ltd	SARS-CoV-2 Antigen Rapid Test (Immunochromatography)	1341	<i>Retrospective in vitro study</i>	95% sensitivity 99.75% specificity Nasal swab	DE ^[2]	Nucleo-protein	Anterior nasal swab, Nasal swab		17 February 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 30 and 100% at Ct ≤ 25; Manufacturer specificity: 99.75%						
Qingdao Hightop Biotech Co., Ltd.	SARS-CoV-2/Flu A+B/RSV Antigen Rapid Test	2754	<i>Retrospective in vitro study</i>	Clinical Sensitivity: 100 % (SARS-CoV-2 at Ct lower or equal to 25) Clinical Specificity: 99.75 % (SARS-CoV-2)	DE ^[2]	Nucleo-capsid protein	Nasopharyngeal swab		21 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 30 and 100% at Ct ≤ 25; Manufacturer specificity: 99.75%						
Quidel Corporation	Sofia SARS Antigen FIA	1097	<i>Prospective clinical field studies</i>	96.7% sensitivity 100% specificity NP/Nasal swab	DE ^[2] , NL ^[5] , PT CH	Nucleo-protein	Nasal swab, Nasopharyngeal swab		17 February 2021
			FR: Validation study data: NP swabs sensitivity 84,44% (76/90), specificity 99,19 (491/495)						
			NL: Independent prospective clinical field study in symptomatic (n=733, PCR positive 144); NP swab; sensitivity overall: 84.0%, sensitivity Ct<30: 90.1%, sensitivity Ct<25: 92.5%; specificity overall: 99.8%. PT: 80 samples from symptomatic individuals (27 PCR positive and 53 negative by PCR) were tested. Sensitivity 70% (95%IC50-86); specificity 100% (95%IC 93-100). TCID50/ml 0,68x 102 and CT<25.						
			<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 89% at Ct ≤ 25; Manufacturer specificity: 100%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Rapid Pathogen Screening, Inc	LIAISON® Quick Detect Covid Ag Assay	2290	<i>Retrospective in vitro study</i>	Sensitivity: 96.1%, Specificity: 97% NP and Nasal swab	IT	Nucleo-protein	Nasal swab, Nasopharyngeal swab		23 July 2021
			IT: Independent validation study, 100 pos. and 100 neg. samples; sensitivity: 92.7% with Ct<25; specificity: 100%.						
Roche (SD BIOSENSOR)	SARS-CoV-2 Rapid Antigen Test	1604	<i>Prospective clinical field study</i>	96.52% sensitivity 99.2% specificity NP swab	DE ^[2] , FI, NL, PT, SE UK	Nucleo-protein	Nasopharyngeal swab	NL: Performance is similar and not affected by the Omicron variant.	10 May 2021
			NL: Independent prospective clinical field study in symptomatic (n=970, PCR positive 186); NP swab; sensitivity overall: 84.9%, sensitivity Ct≤30: 94.3%, sensitivity Ct≤25: 99.1%; specificity overall: 99.5%						
			SE: Karolinska hospital evaluation of Lot QCO3020109. Patient samples: 95 PCR positive, 150 negative. No detailed sample description available. Sensitivity 43%, specificity 100%. Sensitivity Ct<25 = 80.5%.						
			<i>Retrospective in vitro studies</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 89% at Ct ≤ 25; Manufacturer specificity: 99.68%						
Roche (SD BIOSENSOR)	SARS-CoV-2 Rapid Antigen Test Nasal	2228	<i>Prospective clinical field studies</i>	Clinical Sensitivity: 89.6 % (Ct ≤ 30) 93.1 % (Ct ≤ 27) Clinical Specificity: 99.1 % Nasal swab	DE ^[2] Brazil, UK	Nucleo-protein	Nasal swab	DK: This RAT detects both Delta and Omicron on a comparable level with the wild-type (Wuhan).	7 July 2021
			FIND evaluation DE (12 April 2021) 179 samples, nasal swab. Clinical sensitivities: Days ≤ 7: 81.2%; Ct ≤ 33: 87.5%; Ct ≤ 25: 100%; Clinical specificity: 99.3% Brazil (12 April 2021) 214 samples, nasal swab. Clinical sensitivities: Days ≤ 7: 81.2%; Ct ≤ 33: 91.7%; Ct ≤ 25: 100%; Clinical specificity: 99.3%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
			<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 89.6% at Ct ≤ 30; Manufacturer specificity: 99.1%						
Safecare Biotech (Hangzhou) Co. Ltd	COVID-19 Antigen Rapid Test Kit (Swab)	1489	<i>Retrospective in vitro study</i>	97.27% sensitivity 99.42% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab		17 February 2021
	Multi-Respiratory Virus Antigen Test Kit (Swab) (Influenza A+B/COVID-19)	1490	<i>Retrospective in vitro study</i>	97.04% sensitivity 99.44% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab		10 May 2021
Sansure Biotech Inc	SARS-CoV-2 Rapid Antigen Test (Colloidal Gold Method)	2097	<i>Retrospective in vitro study</i>	Clinical Sensitivity: 98.4 % Clinical Specificity: 98.1 %	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab		21 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.1%						
ScheBo Biotech	ScheBo SARS CoV-2 Quick Antigen	1201	<i>Retrospective in vitro study</i>	96.6% sensitivity (Ct ≤ 30) 99.00% specificity NP/ OP swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab, Oropharyngeal swab		16 June 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 95% at Ct ≤ 30 and 100% at Ct ≤ 25; Manufacturer specificity: 99%					! Serum	
ScheBo Biotech	ScheBo SARS CoV-2 Quick ANTIGEN (Colloidal Gold Method)	2763	<i>Retrospective in vitro study</i>	Clinical sensitivity: 96.12% Clinical specificity: 99.49%	DE ^[2]	Nucleo-capsid protein	Nasal swab		21 January 2022
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.49%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
SD Biosensor Inc	STANDARD Q COVID-19 Ag Test Nasal	2052	<i>Prospective clinical field studies</i>	Clinical Sensitivity: 97.12 % Clinical Specificity: 100 % Nasal swab	DE ^[2] , FI, FR Brazil, UK	Nucleo-protein	Nasal swab	<p>DK: This RAT detects both Delta and Omicron on a comparable level with the wild-type (Wuhan).</p> <p>CH: Sensitivity was compared between Omicron and other VOCs (Alpha, Beta, Gamma and Delta) as well as an early-pandemic SARS-CoV-2 isolate.</p>	7 July 2021
			<p>DE 146 symptomatic adults, 40 (27.4%) were RT-PCR-positive for SARS-CoV-2. Sensitivity with 85.0% (34/40; 95% CI 70.9-92.9) with professional testing. At high viral load (>7.0 log₁₀ SARS-CoV-2 RNA copies/ml), sensitivity was 96.6% (28/29; 95% CI 82.8-99.8) for professional testing.</p> <p>FIND evaluation DE (12 April 2021) 179 samples, nasal swab. Clinical sensitivities: Days ≤ 7: 81.2%; Ct ≤ 33: 87.5%; Ct ≤ 25: 100%; Clinical specificity: 99.3%</p> <p>Brazil (12 April 2021) 214 samples, nasal swab. Clinical sensitivities: Days ≤ 7: 81.2%; Ct ≤ 33: 91.7%; Ct ≤ 25: 100%; Clinical specificity: 99.3%</p>						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
SD BIOSENSOR Inc.	STANDARD F COVID-19 Ag FIA	344	<i>Prospective clinical field studies</i>	94,09% sensitivity 98.52% specificity NP swab	DE ^[2] , IT, NL ^[5] , DK Brazil, CH, India, UK	Nucleo-protein	Nasopharyngeal swab		17 February 2021
			<p>NL Independent prospective clinical field study in symptomatic (n=628, PCR positive 118); NP swab; sensitivity overall: 78.0%, sensitivity Ct<30: 84.4%, sensitivity Ct<25: 90.3%; specificity overall: 99.6%</p> <p>FIND evaluation DE (10 Dec 2020) 676 samples, NP swab. Clinical sensitivities: Days ≤ 7: 81.2%; Ct ≤ 33: 75%; Ct ≤ 25: 100%; Clinical specificity: 96.9%</p> <p>Brazil (10 Dec 2020) 453 samples, NP swab. Clinical sensitivities: Days ≤ 7: 80.2%; Ct ≤ 33: 80.9%; Ct ≤ 25: 87.9%; Clinical specificity: 97.9%</p> <p>India (25 June 2020) 417 samples, NP swab. Clinical sensitivities: Days ≤ 7: 61.8%; Ct ≤ 33: 53.6%; Ct ≤ 25: 68.5%; Clinical specificity: 99.5%</p>						
			<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.52%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
SD BIOSENSOR Inc.	STANDARD Q COVID-19 Ag Test	345	<i>Prospective clinical field studies</i>	96.52% sensitivity 99.68% specificity NP swab	DE ^[2] , ES, IT, NL ^[5] , DK, PT Brazil, CH, India, NO, UA, UK	<i>Unknown</i>	Nasopharyngeal swab		17 February 2021
			<p>PT 80 samples from symptomatic individuals (27 PCR positive and 53 negative by PCR) were tested. Sensitivity 70% (95%IC50-86); specificity 100% (95%IC 93-100). TCID50/ml 0,68x 10² and CT<25.</p> <p>FIND evaluation DE (10 Dec 2020) 1263 samples, NP swab. Clinical sensitivities: Days ≤ 7: 80%; Ct ≤ 33: 87.8%; Ct ≤ 25: 100%; Clinical specificity: 99.3% Brazil (10 Dec 2020) 400 samples, NP swab. Clinical sensitivities: Days ≤ 7: 90.7%; Ct ≤ 33: 91.9%; Ct ≤ 25: 95.9%; Clinical specificity: 97.6% CH (10 Dec 2020) 529 samples, NP swab. Clinical sensitivities: Days ≤ 7: 89.8%; Ct ≤ 33: 91.8%; Ct ≤ 25: 97.2%; Clinical specificity: 99.7% India (22 April 2021) 334 samples, NP swab. Clinical sensitivities: Days ≤ 7: 58.3%; Ct ≤ 33: 65.5%; Ct ≤ 25: 89.4%; Clinical specificity: 97.3% Peru (22 April 2021) 335 samples, NP swab. Clinical sensitivities: Days ≤ 7: 81.4%; Ct ≤ 33: 83.3%; Ct ≤ 25: 96.2%; Clinical specificity: 99.6%</p> <p style="background-color: #FF8C00; color: white; text-align: center;"><i>Retrospective in vitro studies</i></p> <p>DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 89% at Ct ≤ 25; Manufacturer specificity: 99.68%</p>						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
SGA Medikal	V-Chek SARS-CoV-2 Ag Rapid Test Kit (Colloidal Gold)	1319	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99,5%	96.6% sensitivity, 99.5% specificity, Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab		10 May 2021
	V-Chek SARS-CoV-2 Rapid Ag Test (colloidal gold)	1357	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99,5%	96.60% sensitivity: 99.5% specificity, Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab		7 July 2021
Shenzen Ultra-Diagnostics Biotech Co., Ltd	SARS-CoV-2 Antigen Test Kit	2017	<i>Prospective clinical field study</i> SI: Sensitivity in unselected symptomatic population: 86.4% (172 RAT pos. / 199 RT-PCR pos.), sensitivity of 97.8% at Ct≤25. Specificity: 99.1% (1972 RAT neg. / 1990 RT-PCR neg.), NP swab	Clinical Sensitivity: 95.33 % (Nasal), 95.48(NP) Clinical Specificity: 99.16 % (Nasal), 99.61 % (NP)	BE, SI	Nucleo-protein	Nasal swab, Nasopharyngeal swab ! Saliva		10 May 2021
Shenzhen CAS-Envision Medical Technology Co., Ltd.	SARS-CoV-2-Antigen Rapid Detection Kit	2152	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct < 25; Manufacturer specificity: 99.5%	OP: Sensitivity: 98.1% 94.7%-99.4%), Specificity: 99.5% 97.0%-99.9%) NP: Sensitivity: 98.1% 94.7%-99.4%), Specificity: 99.5% 97.0%-99.9%)	DE ^[2]	Nucleo-capsid protein	Oropharyngeal swab; Nasopharyngeal swab		8 December 2021
Shenzhen Dymind Biotechnology Co., Ltd	SARS-CoV-2 Antigen Test Kit (Colloidal Gold)	2415	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct < 25; Manufacturer specificity: 96.58%	Sensitivity: 96.58%, Specificity: 98.37%	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab		20 October 2021
Shenzhen Huian Biosci Technology Co., Ltd.	SARS-CoV-2 Antigen Test Kit (Colloidal Gold)	2414	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.1%	NP/OP swab: Sensitivity: 95.0%, Specificity: 99.1% Nasal swab: Sensitivity: 94.6%, Specificity: 99.1%	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab		20 October 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Shenzhen Kisshealth Biotechnology Co., Ltd	SARS-CoV-2 Antigen Test Kit (GICA)	1813	<i>Retrospective in vitro study</i>	NP swabs: Sensitivity: 96.43%, Specificity: 100%.	DE ^[2]	Nucleo-capsid protein	Anterior nasal swab, Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.2%	Nasal (Anterior) swabs: Sensitivity: 99.43%, Specificity: 99.23%.					
Shenzhen Lvshiyuan Biotechnology Co., Ltd.	Green Spring SARS-CoV-2 Antigen-Rapid test-Set	2109	<i>Retrospective in vitro study</i>	96.43% sensitivity 100% specificity NP/OP/Nasal swab	DE ^[2]	Nucleo-protein	Anterior nasal swab, Nasal swab, Nasopharyngeal swab, Oropharyngeal swab I Saliva		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%						
Shenzhen Microprofit Biotech Co., Ltd	SARS-CoV-2 Antigen Test Kit (Colloidal Gold Chromatographic Immunoassay)	1967	<i>Retrospective in vitro study</i>	Sensitivity: 92.93% Clinical Specificity: 100 % Nasal/NP/OP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		7 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct < 25; Manufacturer specificity: 100%						
Shenzhen Microprofit Biotech Co., Ltd.	SARS-CoV-2 Spike Protein Test Kit (Colloidal Gold Chromatographic Immunoassay)	1178	<i>Retrospective in vitro study</i>	Sensitivity: 86.3%, Specificity: 100% Nasal Swab	DE ^[2]	Spike protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		23 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct < 25; Manufacturer specificity: 100%						
Shenzhen Microprofit Biotech Co., Ltd	SARS-CoV-2 Spike Protein Test Kit (Fluorescence Immunoassay)	1228	<i>Retrospective in vitro study</i>	Sensitivity: 93.46%, Specificity: 100%	DE ^[2]	Nucleo-protein, S protein (S1)	Nasopharyngeal swab		8 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct < 25; Manufacturer specificity: 100%						
Shenzhen Reagent Technology Co.,Ltd.	SARS-CoV-2 antigen IVD kit SWAB	2026	<i>Retrospective in vitro study</i>	Sensitivity: 95.2 %, specificity: 98.1 %	DE ^[2]	Nucleo-protein	Nasopharyngeal swab, Oropharyngeal swab		20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct < 25; Manufacturer specificity: 98.1%						
Shenzhen Watmind Medical Co., Ltd	SARS-CoV-2 Ag Diagnostic Test Kit (Colloidal Gold)	1769	<i>Retrospective in vitro study</i>	NP/OP swab: Sensitivity 95.15%, specificity 99.12%. Nasal swab: Sensitivity: 91.51%	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct < 25; Manufacturer specificity: 99.12%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
				for onset of symptoms < 7 days, specificity: 99.02%.					
Shenzhen Watmind Medical Co., Ltd	SARS-CoV-2 Ag Diagnostic Test Kit (Immuno-fluorescence)	1768	<i>Retrospective in vitro study</i>	Clinical Sensitivity: 97.83 % (CT ≤ 33);	DE ^[2]	Nucleo-protein	Nasal swab		7 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99,13%	Clinical Sensitivity: 90.08 % (Ct ≤ 36); Specificity: 99,13% Nasal swab					
Shenzhen YHLO Biotech Co., Ltd.	GLINE-2019-nCoV Ag	1347	<i>Retrospective in vitro study</i>	Nasal: Sensitivity: 97.37% (95%CI: 92.50% - 99.45%); Specificity: 99.25% (95%CI: 97.82% - 99.85%)	DE ^[2]	Nucleo-capsid protein	Nasal swab; Nasopharyngeal swab		8 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 95% at Ct ≤ 25; Manufacturer specificity: 99.85%	NP: Sensitivity: 96.49% (95%CI: 91.26% - 99.04%); Specificity: 99.25% (95%CI: 97.82% - 99.85%)					
Shenzhen Zhenrui Biotech Co., Ltd	Zhenrui ®COVID-19 Antigen Test Cassette	1574	<i>Retrospective in vitro study</i>	96% sensitivity 97% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab ! Saliva		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 82% at Ct ≤ 25; Manufacturer specificity: 97%						
Sugentech, Inc.	SGTi-flex COVID-19 Ag	1114	<i>Retrospective in vitro study</i>	100% sensitivity 100% specificity OP/NP swab	DE ^[2]	Unknown	Nasopharyngeal swab, Oropharyngeal swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 30 and 100% at Ct ≤ 25; Manufacturer specificity: 99.0%						
SureScreen Diagnostics	SARS-CoV-2 Rapid Antigen Test Cassette	2297	<i>Retrospective in vitro study</i>	Sensitivity: 96.1%, Specificity: 99%	DE ^[2] UK	Nucleo-protein	Anterior nasal swab ! Other		20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99%						
Surge Medical Inc.	COVID-19 Antigen Test Kit	1942	<i>Retrospective in vitro study</i>	Clinical sensitivity: 93.33% Clinical specificity: 97.69%	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		21 January 2022
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 97.69%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
TODA PHARMA	TODA CORONADIAG Ag	1466	<i>Prospective clinical field study</i>	98.6% sensitivity Nasal swab	DE ^[2] , FR	Nucleo-protein	Nasal swab, Nasopharyngeal swab		10 May 2021
			FR: Validation data: NP swabs, sensitivity : 96,1-100%, specificity 99,2-100%						
			<i>Retrospective in vitro study</i>						
Triplex International Biosciences(China) Co., Ltd	SARS-CoV-2 Antigen Rapid Test Kit	2074	<i>Retrospective in vitro study</i>	98.51% sensitivity 99.91% specificity Nasal/OP/NP swab	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab ! Saliva		16 June 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 92,5% at Ct ≤ 30 and 100% at Ct ≤ 25; Manufacturer specificity: 99.91%						
Triplex International Biosciences Co., Ltd, China	SARS-CoV-2 Antigen Rapid Test Kit	1465	<i>Retrospective in vitro study</i>	98.51 % sensitivity 100% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab		14 July 2021
TÜRKLAB TIBBİ MALZEMELER SAN. ve TİC. A.Ş.	INFO Covid-19 Ag Test	2584	<i>Retrospective in vitro study</i>	Clinical Sensitivity: 92.71 % Clinical Specificity: 99.54 %	DE ^[2]	Nucleo-capsid protein	Nasal swab	Manufacturer: Linages detected: B.1.1.7 (United Kingdom), B.1.617.1 (India), B.1.617.2 (India), B.1.617.3 (India)	21 December 2021
		DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 90% at Ct ≤ 25; Manufacturer specificity: 99.54%							
TÜRKLAB TIBBİ MALZEMELER SAN. ve TİC. A.Ş.	Covid-19 Ag Test	1689	<i>Retrospective in vitro study</i>	Clinical sensitivity: 92.71% Clinical specificity: 99.54%	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab,	Manufacturer: Linages detected: B.1.1.7 (United Kingdom), B.1.617.1 (India), B.1.617.2 (India), B.1.617.3 (India)	21 January 2022
		DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 90% at Ct ≤ 25; Manufacturer specificity: 99.54%							
TÜRKLAB TIBBİ MALZEMELER SAN. ve TİC. A.Ş.	RAPIDAN TESTER Covid-19 Ag Test	1751	Nucleo-capsid protein	Clinical sensitivity: 92.71% Clinical specificity: 99.54%	DE ^[2]	Nucleo-capsid protein	Nasal swab	Manufacturer: Linages detected: B.1.1.7 (United Kingdom), B.1.617.1 (India), B.1.617.2 (India), B.1.617.3 (India)	21 January 2022
		DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 90% at Ct ≤ 25; Manufacturer specificity: 99.54%							
TÜRKLAB TIBBİ MALZEMELER SAN. ve TİC. A.Ş.	TOYO Covid-19 Ag Tes	1722	<i>Retrospective in vitro study</i>	Clinical sensitivity: 92.71% Clinical specificity: 99.54%	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab,	Manufacturer: Linages detected: B.1.1.7 (United Kingdom), B.1.617.1 (India), B.1.617.2 (India), B.1.617.3 (India)	21 January 2022
		DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 90% at Ct ≤ 25; Manufacturer specificity: 99.54%							

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Vitrosens Biotechnology Co., Ltd	RapidFor SARS-CoV-2 Rapid Ag Test	1443	<i>Retrospective in vitro study</i>	97.30% sensitivity 99.05% specificity NP swab	DE ^[2]	Nucleo-protein	Anterior nasal swab, Nasal swab, Nasopharyngeal swab, Oropharyngeal swab, Throat swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 30 and 100% at Ct ≤ 25; Manufacturer specificity: 99.05%						
VivaChek Biotech (Hangzhou) Co., Ltd, China	Verino Pro SARS CoV 2 Ag Rapid Test	2100	<i>Retrospective in vitro study</i>	Clinical Sensitivity: 97.42% Clinical Specificity: 99.9%	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	Manufacturer: Lineages detected: B.1.1.7 (Alpha); B.1.351 (Beta); B.1.427 (Epsilon); B.1.429 (Epsilon); B.1.617.2 (Delta); B.1.621 (Mu); P.1 (Gamma); C.37 (Lambda); B.1.1.529 (Omicron)	21 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.9%						
Wuhan EasyDiagnosis Biomedicine Co., Ltd.	COVID-19 (SARS-CoV-2) Antigen-Test Kit	2098	<i>Retrospective in vitro study</i>	96.1% sensitivity 100% specificity Nasal/OP/NP swab	DE ^[2] UK	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.26%						
Wuhan Life Origin Biotech Joint Stock Co., Ltd.	SARS-CoV-2 Antigen Assay Kit (Immuno-chromatography)	1773	<i>Retrospective in vitro study</i>	92.67% sensitivity Nasal swab	DE ^[2]	Unknown	Nasal swab		14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: %						
Wuhan UNscience Biotechnology Co., Ltd.	SARS-CoV-2 Antigen Rapid Test Kit	2090	<i>Retrospective in vitro study</i>	Sensitivity: 96.33% Specificity: 99.57% Nasal/NP/OP swab	DE ^[2] , FR	Nucleo-protein	Mid-turbinates swab, Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		7 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99,57%						
Wuxi Biohermes Bio & Medical Technology Co., Ltd.	SARS-CoV-2 Antigen Test Kit (Lateral Flow Assay)	2143	<i>Retrospective in vitro study</i>	Sensitivity: 95.31 %, Specificity: 98.02 %	DE ^[2]	Nucleo-protein	Nasopharyngeal swab, Oropharyngeal swab		20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.02%						

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Xiamen AmonMed Biotechnology Co., Ltd	COVID-19 Antigen Rapid Test Kit (Colloidal Gold)	1763	<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.55%	93.2% sensitivity 99.55% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab		10 May 2021
Xiamen Boson Biotech Co. Ltd	Rapid SARS-CoV-2 Antigen Test Card	1278	<i>Retrospective in vitro study</i>						
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.03%	96.08% sensitivity 99.14% specificity NP swab 96.19% sensitivity 99.2% specificity Nasal swab 96.23% sensitivity 99.2% specificity OP swab	DE ^[2] CH, UK	Nucleo-capsid protein	Nasopharyngeal swab, Oropharyngeal swab, Nasal swab	DK : This RAT detects both Delta and Omicron on a comparable level with the wild-type (Wuhan). Manufacturer: Linages detected: B.1.1.7 (Alpha), B.1.351 (Beta), P.1 (Gamma), B.1.525 (Eta), B.1.427 (Epsilon), B.1.429 (Epsilon), P.3 (Theta), B.1.616, B.1.617.1 (Kappa), B.1.617.2 (Delta), B.1.617.3, B.1.621 (Mu), A.23.1, C.37 (Lambda), C.36, AT.1, B.1.526 (Iota), B.1.1.529 (Omicron), P.2 (Zeta)	17 February 2021
Xiamen Wiz Biotech Co., Ltd	SARS-CoV-2 Antigen Rapid Test	1456	<i>Retrospective in vitro study</i>						
	DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 88% at Ct ≤ 25; Manufacturer specificity: 100%	96.3% sensitivity, 100% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab ! Other		10 May 2021		
Xiamen Wiz Biotech Co., Ltd	SARS-CoV-2 Antigen Rapid Test (Colloidal Gold)	1884	<i>Retrospective in vitro study</i>						
	DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 88% at Ct ≤ 25; Manufacturer specificity: 100%	95.91% sensitivity 100% specificity Nasal swab	DE ^[2]	Nucleo-protein (nucleo-capsid protein)	Anterior nasal swab		10 May 2021		
Zhejiang Anji Saianfu Biotech Co., Ltd	AndLucky COVID-19 Antigen Rapid Test	1296	<i>Retrospective in vitro study</i>						
DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99%	95.8% sensitivity, 99% specificity, Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab		10 May 2021			

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Detection of Omicron ¹⁸ (and other lineages)	Included in EU common list since:
Zhejiang Anji Saianfu Biotech Co., Ltd	reOpenTest COVID-19 Antigen Rapid Test	1295	<i>Retrospective in vitro study</i>	95.8% sensitivity, 99% specificity, Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99%						
Pantest SA	Pantest Coronavirus Ag (Nasopharyngeal Swab)	2271	<i>Retrospective in vitro study</i>	sensitivity: 95,70%, specificity: 99,10%	DE ^[2]	Nucleo-capsid protein	Nasopharyngeal swab		8 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 94% at Ct ≤ 25; Manufacturer specificity: 99.1%						
Zhejiang GENE SCIENCE Co., Ltd	Novel Coronavirus (COVID-19) Antigen Detection Kit (Swab)	2684	<i>Retrospective in vitro study</i>	OP: Sensitivity: 95.65%, Specificity: 99.17% NP: Sensitivity: 94.58%, Specificity: 98.73%	DE ^[2]	Nucleo-capsid protein	Oropharyngeal swab; Nasopharyngeal swab		8 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.73%						
Zhejiang Orient Gene Biotech Co., Ltd	Coronavirus Ag Rapid Test Cassette (Swab)	1343	<i>Retrospective in vitro study</i>	98.32 % sensitivity 99.6 % specificity Nasal/NP swab	DE ^[2] UK	Nucleo-protein	Nasal swab, Nasopharyngeal swab		17 February 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.22%						
Zhuhai Encode Medical Engineering Co.,Ltd	ENCODE SARS-COV-2 Antigen Rapid Test Device	1902	<i>Retrospective in vitro study</i>	Throat swab/Nasal Swab: Sensitivity 96.49%, Specificity 100% Anterior Swab: Sensitivity 94.74%, Specificity: 100%	DE ^[2] UK	Nucleo-capsid protein	Anterior nasal swab, Nasal swab, Throat swab		20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 95% at Ct ≤ 25; Manufacturer specificity: 100%						
Zhuhai Lituo Biotechnology Co., Ltd.	COVID-19 Antigen Detection Kit (Colloidal Gold)	1957	<i>Retrospective in vitro study</i>	96.12% sensitivity Nasal swab (CT≤33); 99.59% sensitivity NP swab; 100% specificity Nasal swab (CT≤33)	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab		14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%						

Notes:

[1] FR: Reference to validation study (not specifying which specific RAT is being recommended or was tested in practice): https://www.has-sante.fr/upload/docs/application/pdf/2020-10/synthese_tests_antigeniques_vd.pdf

[2] DE: Rapid antigen tests that have completed practical validation studies in Germany: See: https://www.pei.de/SharedDocs/Downloads/DE/newsroom/dossiers/evaluierung-sensitivitaet-sars-cov-2-antigentests-04-12-2020.pdf?__blob=publicationFile&v=43

[3] SE: Smaller evaluations ongoing in some of the regions.

[4] BE: In the clinical performance study performed in three different clinical laboratories during the ascendant phase of the epidemiological curve, we found an overall sensitivity and specificity of 57.6 and 99.5%, respectively with an accuracy of 82.6%.

[5] NL: Collected validation data from accredited laboratories in the Netherlands. The report includes evaluations of various RAT that labs performed at their own initiative. <https://lci.rivm.nl/antigeensneltesten>

[6] BE: Van Honacker E. et al., Comparison of five SARS-CoV-2 rapid antigen detection tests in a hospital setting and performance of one antigen assay in routine practice: a useful tool to guide isolation precautions? J Hosp Infect. In press.

ANNEX II: Common standardised set of data to be included in COVID-19 test result certificates, as agreed by Member States on 17 February 2021 and updated on 19 March 2021

Section	Data element	Description	Preferred Code System
Person identification	Person name	The legal name of the tested person. Surname(s) and forename(s), in that order.	
	Person identifier <i>(optional)</i>	An identifier of the tested person, according to the policies applicable in each country. Examples: citizen ID and/or document number (ID-card/passport).	
	Person date of birth <i>(optional)</i>	Tested person's date of birth. Mandatory if no Person identifier is provided.	Complete date, without time, following the ISO 8601.
Test information	Disease or agent targeted	Specification that it concerns the detection of SARS-CoV-2 infection.	ICD-10, SNOMED CT
	Type of test	Description of the type of test that was conducted, e.g. NAAT or rapid antigen test.	LOINC, NPU
	Test name <i>(optional for NAAT)</i>	Commercial or brand name of the test.	
	Test Manufacturer <i>(optional for NAAT)</i>	Legal manufacturer of the test.	
	Sample origin <i>(optional)</i>	The type of sample that was taken (e.g. nasopharyngeal swab, oropharyngeal swab, nasal swab).	SNOMED CT
	Date and time of the test sample collection	Date and time when the sample was collected.	Complete date, with time and time zone, following ISO 8601
	Date and time of the test result production <i>(optional)</i>	Date and time when the test result was produced.	Complete date, with time and time zone, following ISO 8601
	Result of the test	For example, negative, positive, inconclusive or void.	SNOMED CT
	Testing centre or facility <i>(mandatory for NAAT)</i>	Name/code of testing centre, facility or a health authority responsible for the testing event. <i>Optional:</i> address of the testing facility.	
	Health Professional identification <i>(optional)</i>	Name or health professional code responsible for conducting (and validating) the test. Surname(s) and forename(s), in that order.	
	Country where the test was taken	The country in which the individual was tested.	ISO 3166 Country Codes
	Test certificate metadata	Test result certificate issuer	Entity that issued the COVID-19 test result certificate (allowing to check the certificate).
Certificate identifier		Reference of the COVID-19 test result certificate (unique identifier).	

ANNEX III: Common list of COVID-19 laboratory based antigenic assays

As agreed by Member States on 10 February 2022

Manufacturer	RAT commercial name	Device ID # ²¹	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i>	Completed validation studies	SARS-CoV-2 Target protein	Specimen ²²	Detection of Omicron ²³ (and other lineages)	Included in EU common list since:
DIASORIN	LIAISON SARS-CoV-2 Ag assay	1960	<i>Prospective clinical field studies</i>						
			<p>BE: Independent prospective study (random selection), symptomatic and asymptomatic (n=414, PCR positive = 204, PCR negative = 210), NP swab; sensitivity Ct<35: 73.4%, sensitivity Ct<25: 96.4%; specificity: 100%</p> <p>FR: Independent prospective study, symptomatic and asymptomatic (n=378, PCR positive = 46), NP swab; overall sensitivity: 84.8%, sensitivity Ct=<25 100%; specificity: 99.4%</p> <p>IT: Independent prospective study, asymptomatic (n=1075, PCR positive = 23), NP swab; sensitivity Ct=<30 90.5%; specificity: 99.8%</p> <p>NL : Independent field study, mainly symptomatic individuals (n=980, PCR positive n=98), NP+OP swab; sensitivity overall 82.7%, sensitivity Ct<30: 91.9%; specificity overall: 99.1%.</p>	<p>Nasal Swab: Sensitivity: 99/101 (98.0%, 95% CI: 93.1 – 99.5%). Specificity: 210/211 (99.5%, 95% CI: 97.4 – 99.9%).</p> <p>NP Swabs: Sensitivity: 108/109 (99.1%, 95% CI: 95.0 – 99.8%). Specificity: 295/299 (98.7%, 95% CI: 96.6 – 99.5%).</p>	BE, FR, IT, NL	Nucleo-capside protein	Nasal swab, Nasopharyngeal swab	20 October 2021	

²¹ As registered in and used by the JRC database, see: <https://covid-19-diagnostics.jrc.ec.europa.eu/>.

²² The information included in this column is based on the information provided by manufacturers to the JRC database.

²³ The content in this column is based on information submitted by manufacturer via the JRC database as well as what has been published by scientific studies.

Manufacturer	RAT commercial name	Device ID # ²¹	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i>	Completed validation studies	SARS-CoV-2 Target protein	Specimen ²²	Detection of Omicron ²³ (and other lineages)	Included in EU common list since:
Ortho Clinical Diagnostics	VITROS Immunodiagnostic Products SARS-CoV-2 Antigen	1200	<i>Prospective clinical field studies</i>	Clinical Sensitivity 95 % (CI: 92.0% - 100%) Clinical Sensitivity 98 % (95% CI: 92.2–99.4%) Clinical Specificity 98.9 % (95% CI: 96.3-99.9%)	BE, FR	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab		10 February 2022
			BE: Independent prospective study: 80 positive NP samples (sensitivity 100%), 108 negative samples (specificity 100%).						
			FR: Independent prospective study: 107 positive NP samples with Ct _≤ 35 (sensitivity 93,5% for Ct _≤ 35), 1614 negative samples (specificity 100%).						
			<i>Retrospective in vitro study</i>						
			A retrospective study including 134 positive NP samples with Ct _{<} 35 (sensitivity 82,8% for Ct _{<} 35).						
Roche Diagnostics GmbH	Elecsys® SARS-CoV-2 Antigen	2156	<i>Prospective clinical field study</i>	Sensitivity: NP/OP: 94.5 % (95% CI: 90.4-97.2); Nasal swabs: 96.8% (95% CI: 88.8-99.6%) Specificity: 99.9% (95 % CI: 99.6-100%)	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	Manufacturer: Lineages detected: A.23.1, AT.1, B.1.1.7 (Alpha), B.1.351 (Beta), B.1.427 (Epsilon), B.1.429 (Epsilon), B.1.525 (Eta), B.1.526 (Iota), B.1.616, B.1.617.1 (Kappa), B.1.617.2 (Delta), B.1.617.3, B.1.621 (Mu), C.36, C.37 (Lambda), P.1 (Gamma), P.2 (Zeta), P.3 (Theta), B.1.526.1, B.1.526.2,	20 October 2021