


## Ureter Stents

- Standard Double-J
- Multiloop

Made of Polyurethane  
Tecoflex® USA  
Laying time up to 30 days



## Ureter Stents

Set consists of stent, pusher, suture and clamp. Made of Polyurethane Tecoflex® USA

### Standard Double-J

One side open	Both sides open	Size (Fr/CH)	Length (cm)	Guidewire
<b>SS-4716</b>	<b>SOT-4716</b>	4.7	16	0.032"
<b>SS-4724</b>	<b>SOT-4724</b>	4.7	24	0.032"
<b>SS-4726</b>	<b>SOT-4726</b>	4.7	26	0.032"
<b>SS-4728</b>	<b>SOT-4728</b>	4.7	28	0.032"
<b>SS-4730</b>	<b>SOT-4730</b>	4.7	30	0.032"

<b>SS-516</b>	<b>SOT-516</b>	5	16	0.035"
<b>SS-524</b>	<b>SOT-524</b>	5	24	0.035"
<b>SS-526</b>	<b>SOT-526</b>	5	26	0.035"
<b>SS-528</b>	<b>SOT-528</b>	5	28	0.035"
<b>SS-530</b>	<b>SOT-530</b>	5	30	0.035"

<b>SS-616</b>	<b>SOT-616</b>	6	16	0.035"
<b>SS-624</b>	<b>SOT-624</b>	6	24	0.035"
<b>SS-626</b>	<b>SOT-626</b>	6	26	0.035"
<b>SS-628</b>	<b>SOT-628</b>	6	28	0.035"
<b>SS-630</b>	<b>SOT-630</b>	6	30	0.035"

<b>SS-716</b>	<b>SOT-716</b>	7	16	0.038"
<b>SS-724</b>	<b>SOT-724</b>	7	24	0.038"
<b>SS-726</b>	<b>SOT-726</b>	7	26	0.038"
<b>SS-728</b>	<b>SOT-728</b>	7	28	0.038"
<b>SS-730</b>	<b>SOT-730</b>	7	30	0.038"

<b>SS-826</b>	<b>SOT-826</b>	8	26	0.038"
<b>SS-828</b>	<b>SOT-828</b>	8	28	0.038"
<b>SS-830</b>	<b>SOT-830</b>	8	30	0.038"

### Multiloop

One side open	Both sides open	Size (Fr/CH)	Length (cm)	Guidewire
<b>SS-47ML+ML</b>	<b>SOT-47ML+ML</b>	4.7	Multiloop 24-30	0.032"
<b>SS-5ML+ML</b>	<b>SOT-5ML+ML</b>	5	Multiloop 24-30	0.035"
<b>SS-6ML+ML</b>	<b>SOT-6ML+ML</b>	6	Multiloop 24-30	0.035"
<b>SS-7ML+ML</b>	<b>SOT-7ML+ML</b>	7	Multiloop 24-30	0.038"
<b>SS-8ML+ML</b>	<b>SOT-8ML+ML</b>	8	Multiloop 24-30	0.038"

other sizes available

Set with stainless steel guidewire add +G (for example SS-4716+G).  
Set with nitinol guidewire add +GTHS (for example SS-4716+GTHS).

## Multiband Ligator

**Barrel with 6 bands**

**Used for endoscopic ligation of esophageal varices at or above gastroesophageal junction**

Fits on wide range of endoscopes  
Second last band light colour  
Barrel with 7 bands available

### Standard Band Ligator

Requires 9.5 to 11.2 mm  
Endoscope diameter

### Paediatric Band Ligator

Requires 8.0 to 9.5 mm  
Endoscope diameter



	Art. No.	Description
Standard	<b>SGK-6</b>	Set incl. Applicator barrel, thread puller and flushing needle
	<b>SGB-6</b>	Barrel and thread puller only
Paediatric	<b>SGK-6-S</b>	Set incl. Applicator barrel, thread puller and flushing needle
	<b>SGB-6-S</b>	Barrel and thread puller only

Class IIa device – non sterile – for single use only

## Multiband Ligator Colonic

**Used for endoscopic ligation of colonic varices**

Fits on wide range of endoscopes  
Second last band light colour

### Band Ligator Colonic Set contains:

Applicator, Barrel loaded with 6 bands  
or Barrel loaded with 7 bands  
Atraumatic plastic hook with tube  
Flushing needle blunt tip  
Second last band with light colour  
Requires 12.8 to 13.5 mm  
Endoscope diameter

### Barrel Colonic contains:

Barrel loaded with 6 bands  
or Barrel loaded with 7 bands  
Atraumatic plastic hook with tube



	Art. No.	Description
Band Ligator	<b>SBLS-6-C</b>	Multiband Ligator Colonic Set incl. Applicator barrel, thread puller and flushing needle
	<b>SBL-6-C</b>	Multiband Ligator Colonic Barrel and thread puller only
Barrel Colonic	<b>SBLS-7-C</b>	Multiband Ligator Colonic Set incl. Applicator barrel, thread puller and flushing needle
	<b>SBL-7-C</b>	Multiband Ligator Colonic Barrel and thread puller only

Class IIa device – non sterile – for single use only

## Catheter Gel

For easy insertion of a catheter into the Urethra

Without Lidocaine anaesthetic  
6ml and 11ml syringes

Art. No.	Description	Volume
<b>CG 6</b>	Lubricant Gel	6ml
<b>CG 11</b>	Lubricant Gel	11ml



## Sterile Liquid with 10% Glycerine

to fill Catheter Balloons

Thanks to the Glycerine Oil, the balloon size of the catheter remains longer constant

Composition:  
90% Aquadest  
10% Glycerine Oil

Art. No.	Glycerine	Filling volume
<b>Block 10</b>	10%	10ml



## Hydrophilic Urinary Catheter for Paraplegic

"Do-it-yourself" with water sachet in the pouch

Male		Female		Paediatric	
Art. No.	Size (Fr/CH)	Art. No.	Size (Fr/CH)	Art. No.	Size (Fr/CH)
<b>ICM 4006</b>	6	<b>ICF 2012</b>	12	<b>ICP 2006</b>	6
<b>ICM 4008</b>	8	<b>ICF 2014</b>	14	<b>ICP 2008</b>	8
<b>ICM 4010</b>	10	<b>ICF 2016</b>	16	<b>ICP 2010</b>	10
<b>ICM 4012</b>	12	<b>ICF 2018</b>	18		
<b>ICM 4014</b>	14				
<b>ICM 4016</b>	16				
<b>ICM 4018</b>	18				



## Hydrophilic Urinary Catheter Kit for Paraplegic

"Do-it-yourself" KIT - with water sachet in the pouch and attached Urine-Bag

Male		Female		Paediatric	
Art. No.	Size (Fr/CH)	Art. No.	Size (Fr/CH)	Art. No.	Size (Fr/CH)
<b>IKM 4006</b>	6	<b>IKF 2012</b>	12	<b>IKP 2006</b>	6
<b>IKM 4008</b>	8	<b>IKF 2014</b>	14	<b>IKP 2008</b>	8
<b>IKM 4010</b>	10	<b>IKF 2016</b>	16	<b>IKP 2010</b>	10
<b>IKM 4012</b>	12	<b>IKF 2018</b>	18		
<b>IKM 4014</b>	14				
<b>IKM 4016</b>	16				
<b>IKM 4018</b>	18				

Class IIa device - for single use only



Benannt durch/Designated by  
Zentralstelle der Länder  
für Gesundheitsschutz  
bei Arzneimitteln und  
Medizinprodukten  
www.zlg.de  
ZLG-BS-244.10.08



Product Service

# EC Certificate

Full Quality Assurance System

Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4)  
(Devices in Class IIa, IIb or III)

**No. G1 106138 0002 Rev. 00**

**Manufacturer:**

**Marflow AG**

Soodstrasse 57  
8134 Adliswil, Zurich  
SWITZERLAND

**Product Category(ies):**

**Class IIb**  
Double J stent & set  
**Class IIa**  
PCN catheter & set  
Ureteral catheter  
Malecot catheter  
Re-entry malecot catheter  
Suprapubic catheter  
Braided shaft catheter  
Dual lumen catheter  
Facial dilator  
Amplatz dilator & set  
Ureteral dilator & set  
Ureteral balloon dilator  
Double J stent & set  
Mono J stent  
Endopyelotomy stent  
Guidewire  
IP Needle  
Chiba needle  
Stone basket  
Perk basket

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for design, manufacture and final inspection of the respective devices / device categories in accordance with MDD Annex II. This quality assurance system conforms to the requirements of this Directive and is subject to periodical surveillance. For marketing of class III devices an additional Annex II (4) certificate is mandatory. See also notes overleaf.

**Report No.:** IND20190101

**Valid from:** 2020-04-03

**Valid until:** 2024-05-26

*T. Wacker*

Page 1 of 3

TÜV SÜD Product Service GmbH is Notified Body with identification no. 0123

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany

TUV®

Legalization see reverse side

TÜV SÜD  
 ZERTIFIKAT ♦ CERTIFICATE ♦ 認證證書 ♦ CERTIFICADO ♦ CERTIFICAT

**Official Certification**

Seen for authentication of the foregoing signature, acknowledged in our presence by

Ms. **Tracey WALTHER**, born 28th December 1958, Swiss citizen of Oberentfelden AG, according to her information residing at Brunaustrasse 17, 8002 Zürich, identified by identity card.

Zürich, 8th April 2020  
BK no. 1027ff  
Fee CHF 20.00



**NOTARIAT ENGE-ZÜRICH**

Andreas Bachmann, Notary Public

**APOSTILLE**

(Convention de la Haye du 5 octobre 1961)

1. Land: Schweizerische Eidgenossenschaft, Kanton Zürich  
Country: Swiss Confederation, Canton of Zürich  
Diese öffentliche Urkunde / This public document

2. ist unterschrieben von  
has been signed by Andreas Bachmann

3. in seiner Eigenschaft als  
acting in the capacity of Notary Public

4. sie ist versehen mit dem Stempel/Siegel des (der) – bears the stamp/seal of  
Notariat Enge – Zürich Kanton Zürich

5. In / at 8090 Zürich / Zurich

6. am / the 08.04.2020

7. durch die Staatskanzlei des Kantons Zürich  
by the Chancellery of State of the Canton of Zurich

8. unter Nr. / under N° 1179274/2020

9. Stempel/Siegel, Stamp/seal

10. Unterschrift / Signature [Signature]



S. Overkott



Product Service

# Certificate

No. Q5 106138 0001 Rev. 00

**Holder of Certificate:** **Marflow AG**  
Soodstrasse 57  
8134 Adliswil, Zurich  
SWITZERLAND

**Facility(ies):** Marflow AG  
Soodstrasse 57, 8134 Adliswil, Zurich, SWITZERLAND

**Certification Mark:**



**Scope of Certificate:** Design and Development, Manufacture and Supply of Medical Disposables, Surgical Tools, Equipment & Accessories in the Field of Urology, Gastroenterology, Radiology, Gynaecology & Cardiology.

**Applied Standard(s):** EN ISO 13485:2016  
Medical devices - Quality management systems - Requirements for regulatory purposes (ISO 13485:2016)  
DIN EN ISO 13485:2016

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system, which meets the requirements of the listed standard(s). See also notes overleaf.

**Report No.:** IND20190101

**Valid from:** 2020-04-03

**Valid until:** 2023-04-02

**Date,** 2020-04-03

Christoph Dicks  
Head of Certification/Notified Body

TUV SUD TÜV SÜD  
ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ CERTIFICADO ◆ CERTIFIKAT ◆ CERTIFICATE ◆ CERTIFICATE

**Official Certification**

Seen for authentication of the foregoing signature, acknowledged in our presence by

Ms. **Tracey WALTHER**, born 28th December 1958, Swiss citizen of Oberentfelden AG, according to her information residing at Brunastrasse 17, 8002 Zürich, identified by identity card.

Zürich, 8th April 2020  
BK no. 1027ff  
Fee CHF 20.00



**NOTARIAT ENGE-ZÜRICH**

Andreas Bachmann, Notary Public

**APOSTILLE**

(Convention de la Haye du 5 octobre 1961)

1. Land: Schweizerische Eidgenossenschaft, Kanton Zürich  
Country: Swiss Confederation, Canton of Zürich  
Diese öffentliche Urkunde / This public document
2. ist unterschrieben von  
has been signed by Andreas Bachmann
3. in seiner Eigenschaft als  
acting in the capacity of Notary Public
4. sie ist versehen mit dem Stempel/Siegel des (der) – bears the stamp/seal of  
Notariat Enge – Zürich Kanton Zürich
5. In / at 8090 Zürich / Zurich
6. am / the 08.04.2020
7. durch die Staatskanzlei des Kantons Zürich  
by the Chancellery of State of the Canton of Zurich
8. unter Nr. / under N° 1179275/2020
9. Stempel/Siegel, Stamp/seal
10. Unterschrift / Signature



S. Overkott



Benannt durch/Designated by  
Zentralstelle der Länder  
für Gesundheitsschutz  
bei Arzneimitteln und  
Medizinprodukten  
www.zlg.de  
ZLG-BS-244.10.08



Product Service

# EC Certificate

Full Quality Assurance System

Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4)

(Devices in class I in sterile conditions, sterilized systems or procedure packs)

**No. G1S 106138 0003 Rev. 00**

**Manufacturer:**

**Marflow AG**

Soodstrasse 57  
8134 Adliswil, Zurich  
SWITZERLAND

**Product**

**Category(ies):**

**Class Is**

Urine bag connector

Penile clamp

Evacuator

IUI catheter without syringe

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for manufacture in accordance with MDD Annex II. This quality assurance system covers those aspects of manufacture concerned with securing and maintaining sterile conditions of the respective devices / device categories and conforms to the requirements of this Directive. It is subject to periodical surveillance. See also notes overleaf.

**Report No.:**

IND20190101

**Valid from:**

2020-04-03

**Valid until:**

2024-05-26

**Date,**

2020-04-03

Christoph Dicks  
Head of Certification/Notified Body

Page 1 of 2

TÜV SÜD Product Service GmbH is Notified Body with identification no. 0123

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany

TUV®

Legalization see reverse side

TÜV SÜD  
 ZERTIFIKAT ♦ CERTIFICATE ♦ 認 證 證 書 ♦ CERTIFICADO ♦ CERTIFICAT



**Official Certification**

Seen for authentication of the foregoing signature, acknowledged in our presence by

Ms. **Tracey WALTHER**, born 28th December 1958, Swiss citizen of Oberentfelden AG, according to her information residing at Brunastrasse 17, 8002 Zürich, identified by identity card.

Zürich, 8th April 2020  
BK no. 1027ff  
Fee CHF 20.00



**NOTARIAT ENGE-ZÜRICH**

Andreas Bachmann, Notary Public

**APOSTILLE**  
(Convention de la Haye du 5 octobre 1961)

1. Land: Schweizerische Eidgenossenschaft, Kanton Zürich  
Country: Swiss Confederation, Canton of Zürich  
Diese öffentliche Urkunde / This public document

2. ist unterschrieben von \_\_\_\_\_  
has been signed by Andreas Bachmann

3. in seiner Eigenschaft als \_\_\_\_\_  
acting in the capacity of Notary Public

4. sie ist versehen mit dem Stempel/Siegel des (der) – bears the stamp/seal of  
Notariat Enge – Zürich Kanton Zürich

Bestätigt / Certified

5. In / at 8090 Zürich / Zurich 6. am / the 08.04.2020

7. durch die Staatskanzlei des Kantons Zürich  
by the Chancellery of State of the Canton of Zurich

8. unter Nr. / under N° 1179273/2020

9. Stempel/Siegel, Stamp/seal 10. Unterschrift / Signature



S. Overkott