



Nanchang Micare Medical Equipment Co.,Ltd

Address: No 666 Yaohu West 5th Road, Hi-Tech Zone, Nanchang, Jiangxi, China
Email address: sales3@micare.cn Website: www.micare-med.com/www.surgicallight.com
MP: +86-18979105197 Contact: Eric sang

ATT: Tender No. ocds-b3wdpl-MD-1752741146369
Achiziționarea Dispozitivelor medicale (lămpi chirurgicale)
conform necesităților instituțiilor medico-sanitare publice (repetat)

DECLARATION

We, Nanchang Micare Medical Equipment Co.,Ltd, address: Building 16th No.666 Yaohu West 5th Road, Hi-Tech Zone, Nanchang, Jiangxi, China, do hereby confirm the below specification of the customized surgical light model E700/500 according to the requirement of the tender, version of the Catalog Customized Version 2025-07-25, offered by our distributor MedPunkt SRL.

Specification	E700/500
Voltage	95~245V · 50/60HZ
Illuminance at distance of 1 m (LUX)	60,000-180,000/40,000-160,000
Brightness Adjustable	0-100% (10steps)
Lamp Head Diameter	650MM/650MM
Quantity of LEDS	61PCS/39PCS
Color Temperature Adjustable	3500-5500K
Color rendering index Ra	>96
Color rendering index R9 (Red)	98
Light Field Size Adjustable	150-350 MM
Total radiant flux density	364W/m2
Endoscopy Mode	YES
Endoscopy Mode LEDS	6PCS/6PCS
Illumination For Endo-Mode	12%
LED service life	80,000hrs
Working Distance	70-140 CM
Rotating Angle of Arm	>540°
Temperature at Surgeon's Head	<1° C
Depth of illumination L1+L2	1300 MM
Light head protection rating	IP54+FireProof
Shadow dilution with I mask	73%
Shadow dilution with I tube	90%
Control panel on each lamp head	YES
Back Up Battery (4 – 6 hrs)	YES
Internal / External Sony Camera (20X) resolution 1920x1080	YES
Monitor 24 inches resolution 1920x1080 with possibility to install on a pendant	YES

南昌迈柯尔医疗器械有限公司
NANCHANG MICARE MEDICAL EQUIPMENT CO.,LTD

李永强

Authorized Signature(s)

Flower design lamp heads	YES
Laminar Air Flow Compliancy	YES
Removable and sterilizable handles	YES
Ceiling height (M)	2.9 - 4

The performances, features and parameters of our surgical lights can be customized according to the requirements of our clients and not all information is placed on the website of our company.

南昌迈柯尔医疗器械有限公司
NANCHANG MICARE MEDICAL EQUIPMENT CO.,LTD

Signed and Stamped:

Name: Eric Sang

Title: Sales Manager

Authorized Signature(s)

Date: July 28th, 2025