




Certificate of Analysis

29/11/2024

DONPREP®	
Product code	P50, P50B, P50/500
Product size	10, 50 and 500 columns
Batch number	ML 963
Date of manufacture	21 November 2024
Expiry	21 November 2026
Storage recommendations	2 - 8 °C. Do not freeze. Alternatively this product can be stored at room temperature (21 - 25 °C) for 12 months from the date of manufacture.

This document is electronically created and therefore valid without a signature

QC/CERTIFICATES/TEMPLATES/V9/11.2023





Certificate of Analysis

This document is electronically created and therefore valid without a signature


Tested to Specification: Quality Control Product Release Inspection and Test Specification			
Recovery testing method	500 ng of DON is applied in 2 ml of water.		
Recovery specifications	≥85 % to ≤110 % of DON, according to CEN specification (CEN Method No. EN-15891).		
Recovery results (n = 6)	DON: 89 %	% RSD (n = 6)	DON: 2.7 %
Capacity testing method	2500 ng of DON is applied in 2 ml of water.		
Capacity specification	Contains antibody against DON with a capacity ≥1000 ng for DON, according to CEN specification (CEN Method No. EN-15891).		
Capacity results	Recovery ≥70 % when challenged with 2500 ng of DON. For optimum results, samples with greater amounts should be diluted.		

This is to certify that DONPREP® batch ML 963 passes the required specifications and is released for sale.

R-Biopharm Rhône Ltd.
Quality Control Department



Kirsten Stirling
Quality Manager



QC/CERTIFICATES/TEMPLATES/V9/11.2023