




Centrifuges



USER MANUAL

DM0408 Clinical Centrifuge



CE FC

12301638

Ver.20211215

Before operating and using the centrifuge, please carefully read this user manual and fully understand safety-related precautions.





Contents

Safety NOTICE	1
Common safety precautions	1
1 INSTRUCTIONS	5
4.1 Basic operational conditions	8
4.2 Transport and storage conditions	8
5 INSTALLATION	8
5.1 Location	9
6 STRUCTURE	10
7 OPERATION PANEL	11
8 ROTOR PREPARATION	13
8.4 Inspect the rotor	14
9 OPERATION	15
9.1 Normal operation	15
10 MAINTENANCE	20
10.1 Cleaning	20
10.2 Rotor installation	22
11.1 List of common failures	23
11.2 How to open the upper cover	24
12.1 Introduction to rotor	26
12.2 Centrifuge tube	27
13 RCF CALCULATION	30
14 ACTORY REPAIR AND SCRAPPING	30
14.1 Factory repair	30
14.2 Scrapping	31
AFTER -SALES SERVICES	31
DLAB Scientific Co., Ltd.	33







Safety NOTICE

Common safety precautions

Carefully read the following safety precautions for a thorough understanding.

- Follow the instructions and procedures described in this manual to operate this centrifuge safely.
- Carefully read all safety messages in this manual and the safety instructions on the instrument.
- Safety messages are labeled as indicated below. They are in combination with signal words of “WARNING” and “CAUTION” with the safety alert symbol to call your attention to items or operations that could be dangerous to you or other persons using this instrument. The definitions of signal words are as follows:

 **WARNING: Personal Danger**

Warning notes indicate any condition or practice, which if not strictly observed, could result in personal injury or possible death.

 **CAUTION: Possible damage to instrument**



Caution notes indicate any condition or practice, which if not strictly observed or remedied, could result in damage or destruction of the instrument.

NOTE: Notes indicate an area or subject of special merit, emphasizing either the product's capability or common errors in operation or maintenance.

- Do not operate this centrifuge in any manner not described in this User manual. When in doubt or have any troubles with this centrifuge, ASK FOR HELP.
- The precautions described in this User manual are carefully developed in an attempt to cover all the possible risks. However, it is also important that you are alert for unexpected incidents. Be carefully operating this centrifuge.



WARNING:

- This centrifuge is not explosion-proof. Never use explosive or flammable samples.
- Do not install the centrifuge in or near places where inflammable gases are generated or chemicals are stored.
- Do not place dangerous material within 30cm around the centrifuge.
- Make sure to prepare necessary safety measures before using samples that



are toxic, radioactive or contaminated with pathogenic micro-organisms at your own responsibility.

- If the instrument, rotor and/or accessories that has been contaminated by solutions with toxic, radioactive or pathogenic materials, clean it according to the decontamination procedure that you are specified.
- If you require services at site, please sterilize and decontaminate it in advance, and then notice the service center involved in the details of the particular materials.
- Do not handle the power cord or turn on or off the POWER switch with wet hands to void electrical shocks.
- For safety purposes, do not enter within 30cm around this centrifuge while it is in operation.
- While the rotor is rotating, never forcedly release the door lock.
- Unauthorized repairs, disassembly, and other services to the centrifuge except by our service center are strictly prohibited.



CAUTION

- This centrifuge must be located on one firm and level table.
- Make sure the centrifuge is horizontal before running.
- Make sure the angle between the door and cover is greater than 70 degrees when open the door.
- Be careful not put your fingers or hands between the door and cover when the



door off.

- Do not move or relocate this centrifuge while it is running.
- If fluid spills in the rotor chamber, please promptly clean and dry with a dry cloth to avoid sample contamination.
- Ensure to remove any objects and fragments of the tubes dropped inside the rotor chamber before running this centrifuge.
- Cautions on rotors
 - (1) Always check for corrosion and damages on the rotor surface before using it. Do not use the rotor if an abnormality is found.
 - (2) Do not set the centrifuge speed beyond the allowable minimum speed of the rotor kits (rotor or adapters). Make sure to run it below the allowable minimum speed.
 - (3) Do not exceed the allowable imbalance.
 - (4) Use the rotor and tubes within their actual capacities.
 - (5) If the rotor is attached with a lid, ensure it is tightened before operation.
- If any abnormal condition occurs during operation, please stop it immediately and contact our service center. Notify the service center is a warning code if displayed.
- Vibrations are likely to damage the centrifuge, contact our service center if abnormality observed.



1 INSTRUCTIONS

This device is a medical device (laboratory centrifuge) and this centrifuge is mainly used to separate blood, urine or feces samples. Before using this instrument, please undergo specialized training and operate it strictly according to the instructions.

2 SPECIFICATIONS

Max. rotation speed	4000rpm(300-4000rpm), step: 100rpm
Max. relative centrifugal acceleration	1900×g
Capacity	15ml×8,10ml×12
Timing	1 min-99 min-HOLD (continuous operation)





Noise	≤60dB(A)	
Drive motor	Brushless DC motor	
Safety mechanism	Single door locks, over speed and status	
Power supply	Single phase, 100V-240V, 50Hz/60Hz , 50W	
Ambient conditions		
-place	Indoor	
-altitude	No more than 2,000 m	
-ambient temperature	2°C - 40°C	
-humidity	80%	
-over-voltage category	II	
-contamination level	2	
Protection class	I	
EMC		
-radiated interference	EN/IEC 61326-1	FCC Class A
-radiation immunity	Class A	
Dimensions (mm)	286 (L) × 367 (D) × 227 (H)	



Weight	9.2kg
Other functions	operation process display, sound reminder

3 DECLARATION OF CONFORMITY

In compliance with the following safety standards:

UL 61010-1:2012 R4.16

CAN/CSA-C22.2 NO.61010-1-12+GI1+GI2

IEC 61010-2-020:2016

In compliance with the following electromagnetic compatibility standards:

FCC Part 15 Subpart B: 2017

IECS-001: 2006



4 REQUIRED OPERATIONAL CONDITION

4.1 Basic operational conditions

- (1) Power supply: 100V-240V, 50Hz/60Hz, 0.95A/115VAC, 0.56A/230VAC.
- (2) Ambient temperature: 2°C~40°C.
- (3) Relative humidity: ≤80%.
- (4) No vibration or airflow around.
- (5) No airborne dust with charge, explosive and corrosive gases around.

4.2 Transport and storage conditions

- (1) Storage temperature: -40°C~55°C.
- (2) Relative humidity: ≤93%.

5 INSTALLATION

Users must strictly comply with the installation instructions contained in this chapter.



 Warning

- Improper power supply may damage centrifuge.
- Make sure the power source conforms to the requirement before connecting.

5.1 Location

(1) Place the centrifuge on a firm, flat and level bench top, ensure the four feet of this centrifuge stand on the table firmly. Avoid slippery surface or surface prone to vibration.

(2) The recommended ambient temperature is $20^{\circ}\text{C} \pm 5^{\circ}\text{C}$. Avoid temperature over 30°C or direct exposure to sunlight.

(3) Keep clear of the centrifuge at least 10cm on both sides and at least 30cm behind it to guarantee the cooling efficiency.

(4) Keep away from heat source or liquid leakage to avoid elevated sample temperature or centrifuge failures.



5.2 Connection of power cord and grounding

⚠ Warning

- To avoid electrical shocks, ensure your hands are dry when touching the power cord.
- This centrifuge must be properly grounded.

A power outlet rated for 10A or above, with proper earth protection and in compliance with municipal safety requirements is compulsory.

6 STRUCTURE




Figure 6.1 Front view of the centrifuge




Figure 6.2 Rear view of centrifuge

7 OPERATION PANEL












Figure 7.1 Layout of operation panel



Item	Markings	Name	Function
1		Blood program key	Press this key to load the parameter settings for blood separation.
2		Urine program key	Press this key to load the parameter settings for urine separation.
3		Fecal program key	Press this key to load the parameter settings for fecal separation.
4		SET key	Press this key to choose the parameters to be input.
5		Parameter input key	Parameter increase
6		Parameter input key	Parameter decrease
7		ENTER key	Confirm the parameter.
8		Start/Stop key	When the speed is zero, it will start running. While the centrifuge is running, press this key to stop.



Speed display area

Time display area




Figure 7.2 LCD Screen Display

The screen display is shown in Figure 7-2, in which the speed is set at 3000rpm and the set running time is 88 min.

8 ROTOR PREPARATION

8.1 Prepare the samples

8.2 Transfer the samples into tubes

 CAUTION

- Do not over fill the tube with sample and avoid liquid leakage.
- Do not over load the rotor.



8.3 Keep the tubes in balance

- Although this centrifuge may be used with visually confirmed balance, we recommend that samples be weighed using a balance to ensure balanced centrifuge tube in order to prolong the service life of centrifuge, with the deviation of no more than $\pm 1.5g$.
- Never intentionally run the centrifuge with rotor imbalance, no matter within the tolerance or not.

8.4 Inspect the rotor

Check the rotor for corrosion or structural damage before using.

 CAUTION

- If any corrosion or scratch is found on the rotor, please stop using it.
- Only use rotors from the manufacturer.

8.5 Symmetrically load centrifuge tubes onto the rotor

 CAUTION

- Be sure to tighten the rotor and the spindle firmly and ensure that the cover is securely fixed on the rotor. Otherwise, the rotor might fall off while the centrifuge is in operation, resulting in damage to the centrifuge or rotor.
- Tighten the rotor cover and rotor firmly.



- Never press the unlocking button while the rotor is in operation.

9 OPERATION

CAUTION

- While the centrifuge is in operation, don't move the centrifuge or lean against it.
- Before running the centrifuge, keep the centrifugal chamber clean and be sure to remove any foreign object from the centrifugal chamber, such as fragments of centrifuge tube.
- If any abnormal phenomenon such as strange noise occurs during the operation, please immediately stop the centrifuge, contact the service center and inform them of the error code on the LCD display.

9.1 Normal operation

Turn on the power switch and the self-test will launch automatically. Then the most recent program setting is loaded and displayed as shown in Figure 8-1.







Figure 9-1 The most recent operation interface

- Speed setting: 3000rpm. Time setting: 88 min.





1) Set the operating parameters

Press the SET key  gently to switch between different parameters. The parameter will blink on the LCD display when it is chosen. Press the parameter input keys to increase or decrease the selected parameter. Press and hold the parameter input keys, the value of the selected parameter will change faster. Press the parameter input keys briefly to change the parameter value slowly. The increment is 100rpm for the rotation speed, and 1 minute for the timer. Press the ENTER key  to confirm the parameters.




(a) Set the rotation speed

- Press the SET key  to choose the rotation speed in rpm.
- When the speed parameter is selected, it will blink on the LCD screen.
- Increase the speed from 500rpm by 100rpm each time.






- Press the parameter input key  to increase the parameter; press the parameter input key  to decrease the parameter.
- Press and hold the parameter input key  or  to accelerate the parameter change.

(b) Set the running time







- Press the SET key  to make the time value blink and enter into the time setting mode.
- Press the parameter input key  or  to input the set time within a range of 1min-99 min.
- HD refers to unlimited time for continuous running.

(c) Preset program setting


- Press the blood program key  to load parameter settings for the blood sample.
- Press the urine program key  to load the parameter settings for the urine sample.
- Press the fecal program key  to load the parameter settings of the fecal sample.

(d) Modify the preset mode parameter



- After setting the speed and running time, press and hold the SET key  + blood program key  for over 5s to save the current parameter settings as the blood presetting program.
- After setting the speed and running time, press and hold the SET key  + urine program key  for over 5s to save the current parameter settings as the urine presetting program.
- After setting the speed and running time, press the SET key  + fecal program key  for over 5s to save the current parameter settings as the fecal presetting program.


2) Start running

- (a) Press the running key  to start operation.
- The upper cover is locked, the rotor starts rotating and the red LED indicator is lit.
 - Begin timing after startup, with the time display indicating the remaining running time.
- (b) Error display
- It will automatically stop if any failure occurs in the running process, with the error code indicated on the time display window. By looking up it from table 11-



1, the cause of error can be found and appropriate action may be taken.

3) Stop running

(a) When the running time is out or by pressing the key , the centrifuge begins deceleration to stop.

- When the rotor fully stops, the buzzer will beep to remind the user that the operation is over.
- After the rotor fully stops, the red LED and green LED indicators will blink alternately and you will see a blinking text “OPEN” in the time display area to remind you that the operation is over. If there is no operation within 1 minute, the red LED indicator goes out and the green LED indicator stays on. If there is any operation within 1 minute, both the red and the green LED indicators will go out.

(b) The upper cover lock is opened.

- After the end of operation, the upper cover may be opened by pressing the door-opening button.
- After the end of operation, the program will automatically save the set parameters of this operation. When the device is started up again, the program will automatically retrieve the set parameters of the last operation.



(c) Open the upper cover and take out the samples and rotor.

10 MAINTENANCE

10.1 Cleaning


 CAUTION

- Cleaning or sterilizing the centrifuge without complying with the instructions contained herein might cause damage to the centrifuge.

(1) Centrifuge

- The color of housing might change and the label thereon might fall off if the centrifuge is exposed to ultraviolet for a prolonged period of time. After use, please cover the centrifuge with cloth to avoid exposure to light.
- Please clean the centrifuge using cloth or sponge soaked with neutral cleaning agent.
- The centrifuge may be sterilized using cloth soaked with 70% alcohol.

(2) Centrifugal chamber

 CAUTION


- Never pour water or other solvents directly into the centrifugal chamber, otherwise these solvents might enter into the drive unit and cause corrosion or damage to the bearing.
- Clean the chamber using cloth or sponge soaked with neutral cleaning agent and sterilize the centrifuge using cloth soaked with 70% alcohol.
- (3) Drive shaft
- We recommend to clean the drive shaft with a piece of soft cloth and apply a thin layer of silicone grease on it periodically.
- (4) Upper cover
- Clean or sterilize the upper cover in the same manner as (1) centrifuge.
- (5) Rotor
- If the rotor is put out of service for a prolonged period, please remove the rotor from the centrifugal chamber, remove the rotor's lid and place the rotor upside down to empty the rotor hole in order to prevent corrosion.
 - If any sample leaks into the rotor hole, flush the rotor hole with clean water and apply a thin layer of silicone grease on the rotor surface after it dries up.
 - The rotor needs periodical maintenance. We recommend to clean the rotor every three months by cleaning the tube hole and spindle hole and applying a




thin layer of silicone grease thereafter.

10.2 Rotor installation

(1) Installation



(2) Adjustment



11 COMMON FAILURES AND SOLUTIONS

11.1 List of common failures

This centrifuge is capable of self diagnosis. When the centrifuge fails, the time display area will show the error code to identify the cause of failure.

Failure	Possible cause	Solution
No display after powering on	No power supply to power socket.	Check the power and reconnect the power supply.





Abnormal vibration of centrifuge		The rotor does not match with the drive shaft. The sample installed asymmetrically.	Reinstall the rotor. Weigh using a balance and install it symmetrically.
Error code shown in the time display area	E-02 Upper cover indoor failed	Door opened during operation Press the key  when the door is open.	Immediately close the cover door. Close the cover door and start operation.
	E-10-86	See the service manual.	Contact the service representative.

Table 11-1 Common failure and solutions

- Error code E-02 is related to erroneous operation. The centrifuge may continue in operation after elimination of the failure.

11.2 How to open the upper cover

1) When the centrifuge is powered on

 **CAUTION**

- Open the upper cover only when the centrifuge is powered on and the rotor is not running.



(1) The centrifuge is powered on and the upper cover locking electromagnet automatically opens. When the rotor fully stops, press the uncovering button to open the upper cover.

(2) When the centrifuge is powered off

When the upper cover cannot be opened due to sudden power outage, the upper cover may be opened as follows.

(a) Check whether the rotor is running.

- Listen carefully to confirm that the rotor fully stops.

(b) Press the upper cover opening button to open the upper cover.

- The upper cover opening button is located on the front position on the right side of the centrifuge.
- Confirm the rotor has stopped running and press the opening button to open the upper cover.

12. INTRODUCTION TO ROTOR AND CENTRIFUGE TUBE

CAUTION

- Carefully read the user manual and correctly install and use the rotor.
- Don't exceed the max. speed allowed by the rotor, test tube and adaptor. The max. speed allowed by certain adaptors is lower than the max. speed of the



rotor.

12.1 Introduction to rotor

1) Rotor structure




Figure 12-1 Rotor structure

2) Rotor and adaptor

Rotor model	Centrifuge tube	Holding capacity	Max. rotation speed (rpm)	Max. relative centrifugal acceleration (×g)	Adaptive
A8-15P	15ml	8	4000	1900	Standard



A12-10P	10ml	12	4000	1900	Optional
---------	------	----	------	------	----------

Table 12 -1 List of rotors and adaptors

3) Precautions

The density of sample that centrifuge rotor can separate is less than 2.0g/ml. If the density of sample to be separated exceeds 2.0g/ml, please calculate the permissible rotation speed using the following formula:

$$\text{Permissible speed (rpm)} = \text{Max. speed} \times \sqrt{\frac{2.0(\text{g/ml})}{\text{Sample density (g/ml)}}}$$

4) Sterilization

A8-15P rotor body is made of high -strength plastics and cannot be autoclaved or radiated by ultraviolet beam.

12.2 Centrifuge tube

1) Cleaning and sterilization



Conditions		Materials	PA	PC	PP
Cleaning	Fluid cleaning	Acidcleaner(pH5orlower)	X	X	X
		Acidcleaner(higherthanpH5)	O	O	O
		Alkalinecleaner(higherthanpH9)	O	X	O
		Alkalinecleaner(pH9orlower)	O	O	O
		Neutralcleaner(pH7)	O	O	O
		70°C hotwater	O	O	O
	Ultrasoniccleaning	Neutralcleaner(pH7)	O	O	O
Sterilization	Autoclaving	115°C(0.7kg/cm ²)30min	O	O	O
		121°C(1.0kg/cm ²)20min	X	O	O
		126°C(1.4kg/cm ²)15min	X	X	X
	Boilingsterilization	15minto30min	O	O	O
	Ultravioletsterilization	200-300nm	X	X	X
	Gassterilization	Ethyleneoxide	O	X	O
		Formaldehyde	O	O	O

PA: Polyallomer

PC: Polycarbonate

PP: Polypropylene

2) PC centrifuge tube cleaning

PC material has relatively low chemical stability to alkaline solvent, therefore any use of cleaning agent with pH value of over 9 should be avoided. Some neutral cleaning agents still have pH value of over 9 after being diluted as recommended by the vendor, therefore use of cleaning agent with pH value of 7-9 is recommended.

3) Autoclaving of PA, PC and PP centrifuge tube

PA begins softening at the temperature of 120°C, while PC and PP begin softening



at 130°C. Generally, PA may be sterilized for 30 min at 115°C (0.7 kg/cm²), while PC and PP may be sterilized for 20 min at 121 °C (1.0 kg/cm²). Too high temperature would result in deformation of centrifuge tube. When autoclave is used, take the following steps.

- (1) Place the centrifuge tube upright with opening facing upward. If the centrifuge tube is placed in an inclined or horizontal manner, it will deform due to gravity.
- (2) Remove the threaded cover and inner cover to prevent deformation or crack of the centrifuge tube.
- (3) Take the centrifuge tube only when the autoclave cools down to the room temperature.

4) Service life of centrifuge tube

The service life of plastic centrifuge tube depends upon the nature of sample, rotor speed and centrifugation temperature. When the plastic centrifuge tube is used for centrifugation of conventional neutral samples (pH5-pH9), its estimated service life at the max. rotation speed is as follows:

When used at the max. rotation speed:

High-quality centrifuge tube (PA, PC, PP): 30-50 times

Conventional centrifuge tube: about 10 times (number of uses may be increased in



case of low-speed use).

The service life of centrifuge tube is also related to the cleaning and sterilization conditions and might decrease somewhat.

Note: Never use any centrifuge tube with cracks on it.

13 RCF CALCULATION

Relative centrifugal force (RCF) can be calculated using the following formula:

$$RCF = 1.118 \times r \times n^2 \times 10^{-5}$$

r-rotation radius, unit: cm; n -rotation speed, unit: rpm

14 ACTORY REPAIR AND SCRAPPING

14.1 Factory repair

 CAUTION

- Before returning the instrument to factory for repair, please confirm it has been securely fixed for transport.



To protect human health, environment and material safety, please sterilize and clean the equipment and accessories before return to the factory for repair.

14.2 Scrapping

To protect human health and environmental safety, please decontaminate the equipment before scrapping it and comply with applicable laws when scrapping it. According to 2002/96/EC (WEEE), equipment purchased after August 13, 2005 may not be scrapped as domestic article. These pieces of equipment fall within type 8 (medical device) and company-to-company domain. The trash can icon with cross mark indicates that this equipment may not be scrapped as domestic article. Different EC countries have different standards for disposal of this type of wastes. If necessary, please contact the supplier.

AFTER -SALES SERVICES

To ensure safe and efficient operation of centrifuge, periodical maintenance is required. If the centrifuge fails, don't try to repair it.

Please contact the service representative center.

WARRANTY



Unit warranty

The entire unit will have one-year warranty period commencing from delivery date under the conditions of normal maintenance.

Rotor warranty

The rotor will have 5-year warranty period from the date of delivery. Don't use any rotor damaged due to corrosion or fatigue. The damage to the entire unit or rotor due to any of the following reasons is outside the scope of warranty:

- (1) Damage due to improper installation;
- (2) Damage due to brutal or improper operation;
- (3) Damage due to relocation or transport after completion of installation;
- (4) Damage due to dismantling or modification by any unauthorized entity or individual;
- (5) Damage due to use any parts not supplied by our company, such as rotor and adaptor;
- (6) Damage due to natural disasters, including fire and earthquake;
- (7) Wearing parts and parts with warranty period.

CONTACT US



Centrifuges



DLAB Scientific Co., Ltd

Yu An Road 31, Beijing Airport Economic core zone,

Shun yi District, Beijing, China 101318

Tel: 86-10- 85653451/52/53

Fax: 86-10-85653383

E- mail: info@dlabsci.com

web: www.dlabsci.com