

Model	Distal Size	Working distance	Endoscope Model	Picture
XT-DL-092-40	9.0~9.2mm	4.0mm	EG-2790K, EG-2770; GIF-Q260, GIF-H190	
XT-DL-098-40	9.6~9.8mm	4.0mm	GIF-H260,GIF-H180,GIF-HQ190,PCF-PH190L/I	
XT-DL-099-40	9.7~10.1mm	4.0mm	EC-2990Li, EC-2990Fi, EC-2990Mi, EG-2990K	
XT-DL-105-40	10.3~10.7mm	4.0mm	PCF-Q260J, EC-3490TLi, EG-2990Zi, EG-2990K	
XT-DL-108-40	10.6~11.0mm	4.0mm	GIF-H260Z, EG-590ZW, EG-450CT5, EG-2990i	
XT-DL-115-40	11.3~11.7mm	4.0mm	PCF-H180AI/L, PCF-H190DI/L, EG-3470K, EC-3470MK/FK/LK, EG-3490K, EC-3490FK/Fi/LK, EG-	
XT-DL-118-40	11.6~12.0mm	4.0mm	GIF-2T240	
XT-DL-122-40	12.0~12.4mm	4.0mm	CF-Q260A, CF-Q260D	
XT-DL-128-40	12.6~13.0mm	4.0mm	EC-3890LZi/FZi/MZi, EC-380Kp, EC-250WM5, EC-590WM/ZW, EC-530WM/WI, EC-450WM5/WI5	
XT-DL-132-40	13.0~13.4mm	4.0mm	CF-Q180AL/I, CF-H190L/I, CF-HQ190L/I, EC-3890TFK/TLK, EC-3890Mi/Fi/Li, EG-3890TK	
XT-DL-148-40	14.6-15.0mm	4.0mm	CF-H260AZL/I	

EC Certificate
Directive 93/42/EEC Annex II, excluding Section 4
Full Quality Assurance System
Medical Devices

Registration No.: HD 60127938 0001

Report No.: 16801992 009

Manufacturer: Shangxian Minimal Invasive
Inc.
1st Floor,Block B2-2
China medicine innovation park
Mulan road,Hi-tech development zone
Benxi
117004 Liaoning
China

Products: Medical Devices

(see attachment for products and additional site included)

Replaces Approval, Registration No.: HD 60114910 0001


Expiry Date: 2023-07-27

The Notified Body hereby declares that the requirements of Annex II, excluding section 4 of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex II, section 5 of the aforementioned directive. For placing on the market of class III devices covered by this certificate an EC design-examination certificate according to Annex II, section 4 is required.

Effective Date: 2018-07-28

Date: 2018-06-26

Notified Body



X. Ren

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg
TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.

TÜV Rheinland
LGA Products GmbH
Tillystraße 2, 90431 Nürnberg

**Attachment to
Certificate**

Registration No.: HD 60127938 0001
Report No.: 16801992 009

Manufacturer: Shangxian Minimal Invasive
Inc.
1st Floor,Block B2-2
China medicine innovation park
Mulan road,Hi-tech development zone
Benxi
117004 Liaoning
China

Products:

- Disposable Biopsy Forceps
- Disposable Endoscopic Distal Attachments
- Disposable Hemostasis Clips and Delivery Systems
- Capsule Endoscopy Systems
- Disposable Hemostasis Clips
- Disposable Grasping Forceps
- Disposable Injection Needles
- Disposable Polypectomy Snares

Aspects of manufacture concerned securing and
maintaining sterile conditions:

- Disposable Guide Wires

Site included:

501, IT Int'l Building, No. 3 West Yuanhang Road,
Hunnan New District, Shenyang, 110179 Liaoning, China

Date: 2018-06-26

Notified Body

