

TopPette Mechanical Pipettes

Features for TopPette

- Light weight, ergonomic, low force design
- Digital display clearly reads volume setting
- The pipettes cover volume range of 0.1 μ L to 10mL
- Easy to calibrate and maintain with tool supplied
- Design helps avoid repetitive strain injuries
- Calibrated in accordance with ISO8655. Each pipette supplied with individual test certificate
- The lower section can be autoclaved





Features for multi-channel TopPette

- 8 and 12 channel pipettes are available for standard 96-well plate
- Dispensing head rotates for optimum pipetting convenience
- Individual piston and tip cone assemblies allow easy repair and maintenance
- Compound material tip cone design allows visual seal verification
- Can be used with universal style pipette tips



12 channels

8 channels



Pipette with switch

DRAGONLAB Pipettes can be stopped with an on/off switch and cable. This can be used with coagulation analyzers or any other instruments that require precise timing.



MicroPette Plus Autoclavable Pipettes

Specifications

TopPette/ MicroPette / MicroPette Plus Mechanical Pipettes (Adjustable and Fixed Volume)

Single-channel Adjustable Volume Pipettes						
Volume Range	Increment	Test Volume	Maximum permissible systematic error (Inaccuracy)		Maximum permissible random error (Imprecision)	
			%	µL	%	µL
0.1-2.5µL	0.05µL	2.5µL	2.50%	0.0625	2.00%	0.05
		1.25µL	3.00%	0.0375	3.00%	0.0375
		0.25µL	12.00%	0.03	6.00%	0.015
0.5-10µL	0.1µL	10µL	1.00%	0.1	0.80%	0.08
		5µL	1.50%	0.075	1.50%	0.075
		1µL	2.50%	0.025	1.50%	0.015
2-20µL	0.5µL	20µL	0.90%	0.18	0.40%	0.08
		10µL	1.20%	0.12	1.00%	0.1
		2µL	3.00%	0.06	2.00%	0.04
5-50µL	0.5µL	50µL	0.60%	0.3	0.30%	0.15
		25µL	0.90%	0.225	0.60%	0.15
		5µL	2.00%	0.1	2.00%	0.1
10-100µL	1µL	100µL	0.80%	0.8	0.15%	0.15
		50µL	1.00%	0.5	0.40%	0.2
		10µL	3.00%	0.3	1.50%	0.15
20-200µL	1µL	200µL	0.60%	1.2	0.15%	0.3
		100µL	0.80%	0.8	0.30%	0.3
		20µL	3.00%	0.6	1.00%	0.2
50-200µL	1µL	200µL	0.60%	1.2	0.15%	0.3
		100µL	0.80%	0.8	0.30%	0.3
		50µL	1.00%	0.5	0.40%	0.2
100-1000µL	5µL	1000µL	0.60%	6	0.20%	2
		500µL	0.70%	3.5	0.25%	1.25
		100µL	2.00%	2	0.70%	0.7
200-1000µL	5µL	1000µL	0.60%	6	0.20%	2
		500µL	0.70%	3.5	0.25%	1.25
		200µL	0.90%	1.8	0.30%	0.6
1000-5000µL	50µL	5000µL	0.50%	25	0.15%	7.5
		2500µL	0.60%	15	0.30%	7.5
		1000µL	0.70%	7	0.30%	3
2-10mL	0.1mL	10mL	0.60%	60	0.20%	20
		5mL	1.20%	60	0.30%	15
		2mL	3.00%	60	0.60%	12



8-channel Adjustable Volume Pipettes						
Volume Range	Increment	Test Volume	Maximum permissible systematic error (Inaccuracy)		Maximum permissible random error (Imprecision)	
			%	μL	%	μL
0.5-10μL	0.1μL	10μL	1.50%	0.15	1.50%	0.15
		5μL	2.50%	0.125	2.50%	0.125
		1μL	4.00%	0.04	4.00%	0.04
5-50μL	0.5μL	50μL	1.00%	0.5	0.50%	0.25
		25μL	1.50%	0.375	1.00%	0.25
		5μL	3.00%	0.15	2.00%	0.1
50-300μL	5μL	300μL	0.70%	2.1	0.25%	0.75
		150μL	1.00%	1.5	0.50%	0.75
		50μL	1.50%	0.75	0.80%	0.4

12-channel Adjustable Volume Pipettes						
Volume Range	Increment	Test Volume	Maximum permissible systematic error (Inaccuracy)		Maximum permissible random error (Imprecision)	
			%	μL	%	μL
0.5-10μL	0.1μL	10μL	1.50%	0.15	1.50%	0.15
		5μL	2.50%	0.125	2.50%	0.125
		1μL	4.00%	0.04	4.00%	0.04
5-50μL	0.5μL	50μL	1.00%	0.5	0.50%	0.25
		25μL	1.50%	0.375	1.00%	0.25
		5μL	3.00%	0.15	2.00%	0.1
50-300μL	5μL	300μL	0.70%	2.1	0.25%	0.75
		150μL	1.00%	1.5	0.50%	0.75
		50μL	1.50%	0.75	0.80%	0.4

Fixed Volume Pipettes						
Volume Range	Increment	Test Volume	Maximum permissible systematic error (Inaccuracy)		Maximum permissible random error (Imprecision)	
			%	μL	%	μL
5μL	-	5μL	1.3%	0.065	1.2%	0.06
10μL	-	10μL	0.8%	0.08	0.8%	0.08
20μL	-	20μL	0.6%	0.12	0.5%	0.1
25μL	-	25μL	0.5%	0.125	0.3%	0.075
50μL	-	50μL	0.5%	0.25	0.3%	0.15
100μL	-	100μL	0.5%	0.5	0.3%	0.3
200μL	-	200μL	0.4%	0.8	0.2%	0.4
250μL	-	250μL	0.4%	1.0	0.2%	0.5
500μL	-	500μL	0.3%	1.5	0.2%	1.0
1000μL	-	1000μL	0.3%	3.0	0.2%	2.0
2000μL	-	2000μL	0.3%	6.0	0.15%	3.0
5000μL	-	5000μL	0.3%	15	0.15%	7.5