



Ferritin (Fer) Assay Reagent Kit (CMIA) Package Insert

INTENDED USE

The Ferritin (Fer) assay Reagent Kit (CMIA) is a chemiluminescent microparticle immunoassay (CMIA) for the quantitative determination of Ferritin (Fer) in human serum and plasma.

PACKING SIZE

24 Device/Kit,30Device/Kit,48 Device/Kit,60 Device/Kit.

SUMMARY

Ferritin is a macromolecule with a molecular weight of at least 440 kDa (depending on the iron content) and consists of a protein shell (apo ferritin) of 24 subunits and an iron core containing an average of approx.

Ferritin is known as iron-storage protein and is synthesized by many body cells. It is found mainly in the liver, spleen, muscle and bone marrow, with only a small fraction found in blood. The amount of ferritin in serum serves as an indicator of the iron reserves, showing if too little (e.g., iron deficiency anemia) or too much iron is available (e.g., hemochromatosis). The protein is involved in the cellular uptake, storage and release of iron. Ferritin has a double function: storage of iron in its bioavailable form and at the same time protection of the cells from the toxic effects of iron, caused by iron's capacity to generate reactive species which can directly damage DNA and proteins. ferritin is a good indicator of storage iron in the body; however it does not provide information about the amount of iron actually available for erythropoiesis.

Clinically, a threshold value of 20 µg/L (ng/mL) has proved useful in the detection of prelatent iron deficiency. This value provides a reliable indication of exhaustion of the iron reserves that can be mobilized for hemoglobin synthesis. Latent iron deficiency is defined as a fall below the 12 µg/L (ng/mL) ferritin threshold. These two values necessitate no further laboratory elucidation, even when the blood picture is still morphologically normal. If the depressed ferritin level is accompanied by hypochromic, microcytic anemia, then manifest iron deficiency is present.

When the ferritin level is elevated and the possibility of a distribution disorder can be ruled out, this is a manifestation of iron overloading in the body. 400 µg/L (ng/mL) ferritin is used as the threshold value. Elevated ferritin values are also encountered with the following tumors: acute leukemia, Hodgkin's disease and carcinoma of the lung, colon, liver and prostate. The determination of ferritin has proved to be of value in liver metastasis. Studies indicate that 76 % of all patients with liver metastasis have ferritin values above 400 µg/L (ng/mL). Reasons for the elevated values could be cell necrosis, blocked erythropoiesis or increased synthesis in tumor tissue.

PRINCIPLE OF TEST

The Ferritin assay is a two-step immunoassay for the quantitative measurement of Ferritin (Fer) in human serum, plasma using CMIA technology, with flexible assay protocols.

In the first step, sample and anti-Fer coated paramagnetic microparticles are combined. The Ferritin present in the sample binds to the anti-Fer coated microparticles. After the wash cycle, ALP-labeled anti-Fer conjugate is added to create a reaction mixture in the second step. Following another wash cycle, substrate is added to the reaction mixture. The resulting chemiluminescent reaction is measured as relative light units (RLUs). A direct relationship exists between the amount of Fer in the sample and the RLUs detected by the system optics.

REAGENTS

The device is pre-dispensed with buffer needed for single use.

The device is constituted with Buffers described below is the main reagent

Object	Content
Micro-particles Buffer	Anti-Fer (mouse, monoclonal) coated Micro-particles in TRIS buffer with protein (bovine) stabilizer. Minimum concentration: 0.1% solid. Preservative: ProClin-300.
Conjugate Buffer	Anti-Fer (mouse, monoclonal) alkaline phosphatase (ALP) labeled conjugate in TRIS buffer with protein (bovine) stabilizer. Preservative: ProClin-300.
Wash Buffer	TRIS buffer with surfactant. Preservative: ProClin-300.
Substrate Buffer	AMPPD, Enhancer, Surfactant, ProClin-300.

Reagent Handling

The reagents in the kit have been assembled into a ready-for-use unit that cannot be separated.

All information required for correct operation is read in from the respective reagent barcodes.

MATERIALS PROVIDED

- Ferritin Test Device
- Product Insert
- Calibration Solution (optional)
- Control Solution (optional)

MATERIALS REQUIRED BUT NOT PROVIDED

- Analyzer

STORAGE AND STABILITY

· Store at 2-8°C and avoid light.

· Do not freeze.

· Store the reagent kit upright before to use.

· Expiration date: up to the stated expiration date.

Note: The Ferritin Assay Reagent Kit must be stored at 2-8°C in an upright position, and must be used immediately after removal from 2-8°C storage or the device was opened. Unused reagents should be put back into the kit in time.

SPECIMEN COLLECTION AND STORAGE

Specimen Types

Validated specimen types to be used with this assay:

Specimen Types	Collection Tubes
Human serum	Serum Serum separator tubes
Human plasma	Sodium heparin Lithium heparin Potassium EDTA Sodium EDTA

Other anticoagulants have not been validated for use with this assay.

The instrument does not provide the capability to verify specimen type. It is the responsibility of the operator to verify that the correct specimen types are used in the assay.

Specimen Conditions

➤ Do not use specimens with the following conditions:

- heat-inactivated
- pooled
- grossly hemolyzed
- obvious microbial contamination

➤ For optimal results, serum and plasma specimens should be free of fibrin, red blood cells or other particulate matter.

➤ Ensure that complete clot formation in serum specimens has taken place prior to centrifugation. Some specimens especial those from patients receiving anticoagulant or thrombolytic therapy may exhibit increased clotting time. If the specimen is centrifuged before a complete clot forms, the presence of fibrin may cause erroneous results.

➤ To prevent cross contamination, use of disposable pipettes or pipette tips is recommended.

Preparation for Analysis

➤ Follow the tube manufacturer's processing instructions for specimen collection tubes.

➤ Specimens must be mixed THOROUGHLY after thawing, by LOW speed vortex, and centrifuged prior to use to remove red blood cells or particulate matter to ensure consistency in the results.

➤ Inspect all specimens for bubbles. Remove bubbles with an applicator stick before analysis. Use a new applicator stick for each specimen to prevent cross contamination.

Specimen Storage

Specimen Type	Storage Temperature	Maximum Storage Time
Serum/Plasma	2-8°C	6 days

➤ If testing will be delayed more than 24 hours, remove serum or plasma from the clot, serum separator or red blood cells.

➤ If testing will be delayed more than 6 days, specimens should be frozen at -10°C or colder.

➤ Specimens stored frozen at -10°C or colder for 3 months showed no performance difference.

➤ Avoid more than 3 freeze/thaw cycles.

Specimen Shipping

➤ Before shipping specimens, it is recommended that specimens be removed from the clot, red blood cells, or separator gel.

➤ When shipping specimens, package and label specimens in compliance with applicable state, federal and international regulations covering the transport of clinical specimens and infectious substances.

➤ Specimens may be shipped ambient, at 2-8°C (wet ice), or frozen (dry ice). Do not exceed the storage time limitations listed above.

INSTRUMENT

The Ferritin Test Device is designed for use on the REALY Analyzer System.

TEST PROCEDURE

Assay Procedure

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer specific assay instructions. Resuspension of the microparticles takes place automatically prior to use. Read in the test-specific parameters via the reagent barcode. If in exceptional cases the barcode cannot be read, enter the digit sequence of numbers. Bring the cooled reagents to approx. 20°C and place on the reagent disk of the analyzer. Avoid foam formation. The system automatically regulates the temperature of the reagents.

For this test device, the transfer volume of specimens, calibrators or controls into the sample hole is 50 µL. (No less than 50 µL.)

Reagent strips should be left at room temperature between 20 and 25 °C for more than 30 minutes before use and kept away from light.

In order to avoid the magnetic beads adsorbed on the side wall and top due to the upside down and side placement of the reagent strip during transportation, the reagent strip should be mixed by shaking and mixing before use. The reagent strip should be mixed upside down for about 30 seconds, and then the reagent strip should be mixed upward for about 30 seconds. The reagent strip was then gently shaken so that the magnetic beads fell completely to the bottom of the strip.

Calibration

Every Test Device has a barcode label containing specific information for calibration of the particular reagent lot. The pre-defined master curve is adapted to the analyzer using

the relevant CalSet.

Calibration frequency: Calibration must be performed before new lot of device is used.

Renewed calibration is recommended as follows:

- After 90 days (when using the same reagent lot on the analyzer);
- As required: e.g. quality control findings outside the defined limits.

Note: Refer to Instruction of Analyzer for the procedure of calibration.

Quality Control

For quality control, please use Control of REALY or Control Universal.

In addition, other suitable control material can be used. Controls for the various concentration ranges should be run individually at least once every 24 hours when the test is in use, once per reagent kit, and following each calibration.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

Specimen Dilution Procedures

Specimens with a Fer concentration greater than 2000 ng/mL will be flagged as " >2000 ng/mL " and may be diluted using Manual Dilution Procedure. Use the 1:10 dilutions is recommended. The operator must enter the dilution factor in the Patient or Control order screen. The system will use this dilution factor to automatically calculate the concentration of the sample before dilution.

EXPECTED VALUES

Reference value:

Male:30-400ng/ml;

Female:13-150ng/ml;

Results may differ between laboratories due to variations in population and test method.

If necessary, advice each laboratory set up your own reference range.

INTERPRETATION OF RESULTS

As interpret the results, the patient's overall clinical situation, including symptoms, medical history and other related data, should be referred to.

LIMITATIONS

- If the Ferritin results are inconsistent with clinical evidence, additional testing is suggested to confirm the result.
- For diagnostic purposes, results should be used in conjunction with other data; e.g., symptoms, results of other tests, clinical impressions, etc.
- Specimens from patients who have received preparations of mouse monoclonal antibodies for diagnosis or therapy may contain human anti-mouse antibodies (HAMA). Specimens containing HAMA may produce anomalous values when tested with assay kits such as REALY Ferritin that employ mouse monoclonal antibodies.
- Heterophilic antibodies in human serum can react with reagent immunoglobulins, interfering with in vitro immunoassays. Patients routinely exposed to animals or to animal serum products can be prone to this interference and anomalous results may be observed. Additional information may be required for diagnosis.
- Although the REALY Ferritin assay is specifically designed to minimize the effects of HAMA and heterophilic antibodies, assay results that are not consistent with other clinical observations may require additional information for diagnosis.

PERFORMANCE CHARACTERISTICS

Linearity

The linearity of Ferritin Reagent Kit was determined by use Ferritin calibrator to prepare 6 different specimens, measuring all these specimens follow the test instruction and then do linear fitting, the results show that the linear correlation coefficient (r) was not less than 0.9900.

Precision / Reproducibility

Intra-assay coefficient of variation was evaluated on 3 different levels of control serum.

Repeatedly measured 20 times, calculating the coefficient of variation.

Intra-assay Precision

Control	Mean (ng/mL)	SD (ng/mL)	CV
Level 1	6.93	0.36	5.23%
Level 2	115.82	4.48	3.87%

Level 3	535.56	25.28	4.72%
---------	--------	-------	-------

Inter-assay coefficient of variation was evaluated on three batches of kits. Repeatedly measured 3 different levels of control serum 30 times, calculating the coefficient of variation.

Inter-assay Precision			
Control	Mean (ng/mL)	SD (ng/mL)	CV
Level 1	6.87	0.50	7.28%
Level 2	115.37	5.41	4.69%
Level 3	536.78	29.95	5.58%

Analytical Sensitivity

The analytical sensitivity is defined as the concentration of Ferritin equivalent to the mean RLU of 20 replicates of the zero standard plus two standard deviations corresponding to the concentration from the standard curve. The analytical sensitivity is typically less than 1 ng/mL.

Analytical Specificity

The specificity of the Ferritin assay system was assessed by measuring the apparent response of the assay to various potentially cross-reactive analytes.

Compound	Concentration	Cross-reactivity
Human heart ferritin	1000 ng/mL	< 1%
Hemoglobin	500 mg/dL	< 0.5%
AFP	1000 ng/mL	< 0.5%
CEA	1000 ng/mL	< 0.5%
CA125	1000 U/mL	< 0.5%
CA19-9	1000 U/mL	< 0.5%
CA15-3	100 U/mL	< 1%

Interference

The following compounds in both low-level specimen and high-level specimen show no cross-reactivity when tested with the Ferritin Assay Reagent Kit at a concentration below:

Compound	Concentration
Bilirubin	20 mg/dL
Biotin	50 ng/mL
Triglycerides	1000 mg/dL
Human albumin	200 g/L

Method Comparison

The comparison between the Ferritin Assay Reagent Kit (y) and a commercially available Ferritin test kit (x), using clinical samples gave the following correlations (ng/mL):

Linear regression

$y=0.9732x-0.2375$

$r=0.9803$

Number of samples measured: 107

The sample concentrations were between about 2.35 – 1936.75 ng/mL.

WARNINGS AND PRECAUTIONS

For *In Vitro* Diagnostic Use.

Do not use expired or clearly damaged kits.

Operating according to the steps described, can make the risk of daily handling patients' samples and blood products into a minimum, however, no matter what the source of the products, handling mode or the previous proof, these potentially infectious substances were used shall be in accordance with the unified considerations and Good Laboratory Practice (GLP).

Proper disinfectant should be used to eliminate pollution.

Follow local rules and regulations to keep and dispose of these items and containers for these items.

The ProClin-300 is a potential skin sensitizer. Avoid dumping or splashing this reagent on your skin and clothing. In case of contact with this reagent, wash thoroughly with soap and water.

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

The reagents should be kept away from light, and unused reagents should be put back into the kit in time and be careful to avoid light.

BIBLIOGRAPHY

- Wick M, Pinggera W, Lehmann P. Ferritin in Iron Metabolism -Diagnosis of Anemias (second edition). Springer-Verlag 1995;ISBN3-211-82525-8 and ISBN 0-387-82525-8.
- Arosio P, Levi S, Gabri E, et al. Heterogeneity of ferritin II: Immunological aspects. In: Albertini A, Arosio P, Chiancone E, Drysdale J (eds). Ferritins and isoferritins as biochemical markers. Elsevier, Amsterdam 1984;33-47.
- Kaltwasser JP, Werner E. Serumferritin: Methodische und Klinische Aspekte. Springer Verlag (1980).
- Morikawa K, Oseko F, Morikawa S. A role for ferritin in hematopoiesis and the immune system. Leuk-Lymphoma 1995;18(5-6):429-433.
- Borch-Johnson B. Determination of Iron status: brief review of physiological effects on iron measures. Analyst 1995;120(3):891-903.
- Cook JD, Skikne BS, Baynes RD. Iron deficiency: the global perspective. Adv-Exp-Med-Biol 1994;356:219-228.
- Albertini A, Arosio P, Chiancone E, et al. (eds). Functional aspects of isoferritins. In: Jacobs A, Hodges J, Hoy TG. Ferritins and isoferritins as biochemical markers. Elsevier, Amsterdam 1984;113-127.
- Wick M, Pinggera W, Lehmann P. Clinical aspects and laboratory. Iron metabolism, anemias. Springer Verlag, Wien, New York, 7th edition, 2011:pp.8-11.
- Knoch MA, Storey JA, Coffman LG, et al. Ferritin for the clinician. Blood Rev 2009;23(3):95-104.
- Sargent PJ, Farnaud S, Evans RW. Structure/function overview of proteins involved in iron storage and transport. Curr Med Chem 2005;12(23):2683-2693.
- Ferraro S, MozzI R, Panteghini M. Reevaluating serum ferritin as a marker of body iron stores in the traceability era. Clin Chem Lab Med 2012;50(11):1911-1916.
- Malyszko J, Levin-laina N, Myśliwiec M, et al. Iron metabolism in solid-organ transplantation: how far are we from solving the mystery? Pol Arch Med Wewn 2012;122(10):504-511.
- Polin V, Coriat R, Perkins G, et al. Iron deficiency: from diagnosis to treatment. Dig Liver Dis 2013;45(10):803-809.

SYMBOLS

Symbol	Meaning	Symbol	Meaning
IVD	In vitro diagnostic medical device		Storage temperature limit
	Manufacturer		Authorized representative in the European Community /European Union
	Date of Manufacture		Use-by date
	Do not re-use		Consult instructions for use or consult electronic instructions for use
	Batch code		Do not use if package is damaged and consult instructions for use
	Catalogue number		Contains sufficient for n tests

Hangzhou Realy Tech Co., Ltd.

#2 Building, No. 763, Yuansha Village, Xinjie Street, Xiaoshan District, 311200 Hangzhou City, Zhejiang Province, PEOPLE'S REPUBLIC OF CHINA
Website: www.realytech.com

Luxus Lebenswelt GmbH

Kochstr.1,47877, Willich, Germany

Number:1100050204
Version:1.6
Effective Date:2023-08-10