



ASAP INTERNATIONAL SDN. BHD. (1157980-X)

No. 1, Jalan Sitar 33/6, Seksyen 33, 40400 Shah Alam, Selangor, Malaysia.

Tel: +6 03 5191 0166

Fax: +6 03 5191 0702

Website: www.whyasap.com

Email: info@whyasap.com

EU DECLARATION of CONFORMITY



Manufacturer:

ASAP International Sdn. Bhd.

No. 1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor, Malaysia.

SRN: MY-MF-000004500

Medical Devices &
Basic UDI- DI:

- Powdered Latex Examination Gloves
(955589090ASAPPL5E)
- Powder Free Latex Examination Gloves
(955589090ASAPPFL4G)
- Powder Free Nitrile Examination Gloves
(955589090ASAPPFN4L)

Classification:

Class I Medical Devices according to Medical Device
Regulation (EU) 2017/745, Annex VIII

Harmonised Standards:

EN455-1:2000, EN 455-2:2015, EN455-3:2015,
EN455-4:2009



European Authorized
Representative:

ASAP INNOVATIONS LIMITED

7 Saggart Lakes, Saggart,
Dublin, D24 PY01, Ireland.

SRN: IE-AR-000002548

This EU declaration of conformity is issued under the sole responsibility of manufacturer, ASAP International Sdn. Bhd. We hereby declare that device covered by the present declaration is in conformity with the Medical Device Regulation (EU) 2017/745 and with the above-mentioned standards. All supporting documentation is retained at the premises of the manufacturer.

Place, Date of Declaration:

Malaysia., May 26, 2021

Signature:

Mr. Chin Tze Weng

Senior Global Business Development Manager





Art & Science of Amazing Protection

ASAP INNOVATIONS LIMITED (625986)

Registered Address: 7 Saggart Lakes, Saggart, Dublin D24 PY01, Ireland

Office: Unit 17, Ace Enterprise Park, Bawnogue Road, Dublin D22 V272, Ireland

Tel: +353 1 466 1660

Website: www.whyasap.ie

Email: info@whyasap.ie



CERTIFICATE OF CE (MDD) NOTIFICATION

We: ASAP Innovations Ltd.

7 Saggart Lakes, Saggart, Dublin D24 PY01, Ireland

HPRA Registered number: IE/CA01/R/GM/1268

Performed all notification duties and responsibilities according to the European Council Directive 93/42/EEC as the Authorised representative (EC REP) of:

Manufacturer: ASAP International SDN BHD (1157980-X)

No. 1, Jalan Sitar 33/6, Seksyen 33, 40400 Shah Alam, Selangor, Malaysia.

AS STIPULATED AND DEMANDED BY THE AFOREMENTIONED DIRECTIVE

The Manufacturer declares that the Class I devices comply with the directive including all essential requirements.

The Manufacturer has provided ASAP Innovations Ltd. with all the appropriate declarations as per the European Council Directive 93/42/EEC article 14 requirements, including the EC Declaration of Conformity (according to Annex VII) confirming that their Class I medical devices, as stipulated here below, are fulfilling the applicable requirements of the European Council Directive 93/42/EEC.

The notification of the following medical devices has been completed by ASAP Innovations Ltd. with the Health Products Regularity Authority in Ireland (www.hpra.ie) on the 24/4/2020 in compliance with the European Council Directive 93/42/EEC – article 14 requirements.

CLASS I MEDICAL DEVICES: Please see Annex A – List of Devices (1 page, 4 devices)

As of the 24/04/2020 and as long as the manufacturer will continue complying with the hereabove mentioned requirements, they therefore:

- Are required to affix the CE marking on these devices.
- May place these devices in the European Community territory.

Mr S. Keller, Managing Director
(authorised signature)

Date of issue: 24/04/2020





Art & Science of Amazing Protection

ASAP INNOVATIONS LIMITED (625986)

Registered Address: 7 Saggart Lakes, Saggart, Dublin D24 PY01, Ireland

Office: Unit 17, Ace Enterprise Park, Bawnogue Road, Dublin D22 V272, Ireland

Tel: +353 1 466 1660

Website: www.whyasap.ie

Email: info@whyasap.ie

Annex A – List of devices

No.	Make	Generic Name	Description and intended use	HPRA Registration number	GMDN Code	Class
1.	ASAP	ASAP Powder Free Nitrile Examination Gloves	<p>Non-sterile, Powder Free, Ambidextrous, Single Use, Beaded Cuff, Finger Textured Surface or Textured Surface, Nitrile Synthetic Rubber Examination Glove.</p> <p>A powder free nitrile examination glove is a disposable glove made of nitrile synthetic rubber intended to wear on the hand for medical purposes to provide a barrier against potentially infections materials and other contaminants. In addition, this product demonstrated reduced potential for sensitizing users to chemical activities.</p>	IE/CA01/R/ GM/1268/15016	56286	I
2.	ASAP	ASAP Powdered Latex Examination Gloves	<p>Non-sterile, Ambidextrous, Single Use, Beaded Cuff, Textured or Smooth Surface, Natural Rubber Latex Examination Gloves.</p> <p>A powdered latex examination glove is a disposable glove made of natural rubber latex intended to wear on the hand for medical purposes to provide a barrier against potentially infections materials and other contaminants.</p>	IE/CA01/R/ GM/1268/15018	47173	I
3.	ASAP	ASAP Latex Powder Free Examination Gloves	<p>Non-sterile, Powder Free Ambidextrous, Single Use, Beaded Cuff, Textured Surface, Natural Rubber Latex Examination Gloves</p> <p>A powder free latex examination glove is a disposable glove made of natural rubber latex intended to wear on the hand for medical purposes to provide a barrier against potentially infections materials and other contaminants</p>	IE/CA01/R/ GM/1268/15017	47172	I
4.	ASAP	ASAP Face Mask	<p>Non-sterile, Surgical Face Mask with Filter, 3 ply, Ear loops. Intended to be used as isolation face mask, procedure mask and dental face mask.</p>	IE/CA01/R/ GM/1268/17491	35177	I



Certificate of Registration

QUALITY MANAGEMENT SYSTEM - ISO 13485:2016

This is to certify that:

ASAP International Sdn. Bhd.
No.1 , Jalan Sitar 33/6
Seksyen 33
Shah Alam
Selangor
40400
Malaysia


Holds Certificate No:

MD 676150

and operates a Quality Management System which complies with the requirements of ISO 13485:2016 for the following scope:

Manufacture, Marketing and Distribution of latex and nitrile examination gloves, face masks and sterile surgical gloves.

For and on behalf of BSI:



Graeme Tunbridge, Senior Vice President Medical Devices

Original Registration Date: 2017-12-13

Latest Revision Date: 2022-04-08

Effective Date: 2022-04-23

Expiry Date: 2025-04-22



Page: 1 of 1

...making excellence a habit.™

This certificate was issued electronically and remains the property of BSI and is bound by the conditions of contract.

An electronic certificate can be authenticated [online](#).

Printed copies can be validated at www.bsi-global.com/ClientDirectory or telephone +603 9212 9638.

Further clarifications regarding the scope of this certificate and the applicability of ISO 13485:2016 requirements may be obtained by consulting the organization.

This certificate is valid only if provided original copies are in complete set.

Certificate of Registration

QUALITY MANAGEMENT SYSTEM - ISO 9001:2015

This is to certify that:

ASAP International Sdn. Bhd.
No.1 , Jalan Sitar 33/6
Seksyen 33
Shah Alam
Selangor
40400
Malaysia

Holds Certificate No:

FM 679348

and operates a Quality Management System which complies with the requirements of ISO 9001:2015 for the following scope:

Manufacture, Marketing and Distribution of latex and nitrile examination gloves, face masks and sterile surgical gloves.

For and on behalf of BSI:

Poon Cheong Yuen, Managing Director

Original Registration Date: 2017-11-13

Latest Revision Date: 2022-03-28

Effective Date: 2022-04-25

Expiry Date: 2025-04-24

Page: 1 of 1



...making excellence a habit.™

This certificate was issued electronically and remains the property of BSI and is bound by the conditions of contract.

An electronic certificate can be authenticated [online](#).

Printed copies can be validated at www.bsi-global.com/ClientDirectory or telephone +603 9212 9638.

Further clarifications regarding the scope of this certificate and the applicability of ISO 9001:2015 requirements may be obtained by consulting the organization.

This certificate is valid only if provided original copies are in complete set.

EU Type-Examination Certificate

Certificate number: 2777/11077-01/E03-01

This EU Type-Examination Certificate covers the following product group(s) supported by testing to the relevant standards/technical specifications and examination of the technical file documentation:

Following the EU Type-Examination this product group has been shown to satisfy the applicable essential health and safety requirements of Annex II of the PPE Regulation (EU) 2016/425 as a Category III product.

Product reference:

AS LPF

Description:

Latex Examination powder free gloves.

Sizes:

6	XS
7	S
8	M
9	L
10	XL

Classification:

EN ISO 374-1:2016/Type B	Level	EN374-4:2013
40% Sodium Hydroxide	1	-46.1%
65% Nitric Acid	2	30.9%
37% Formaldehyde	2	-93.8%
40% Hydrofluoric Acid	2	X
EN ISO 374-5:2016		
Resistance to Bacteria and Fungi	Pass	
Resistance to Virus	Pass	

Standards/Technical specifications applied:

EN 420: 2003+A1: 2009; EN ISO 374-1:2016; EN ISO 374-5:2016

Technical reports/Approval documents:SATRA: CHM0265590/1750/CL/A, CHM0265590/1750/CL/B, CHM0265590/1750/SPT
TUV: 7191154274-CHM16-01-RC

Signed on behalf of SATRA:



Hannah Coe



Geoff Graham

Date of issue: 17/04/2019**Expiry date:** 14/08/2023

TERMS AND CONDITIONS

The following conditions apply in addition to SATRA's standard terms and conditions of business and those given in the current certification agreement.

The certificate holder is licensed to mark the products detailed within this certificate in accordance with Annex V (Module B) of the Regulation (EU) 2016/425 of the European Parliament and of the council of 9th March 2016 on personal protective equipment once you have drawn up an EU declaration of product conformity. Please note:

1. Where the product is classified as category III then CE Marking of production is reliant on current compliance with Regulation 2016/425 module C2 or Module D. (Except that specifically produced to fit an individual user).
2. Full details of the certification and product are contained within the manufacturer's technical documentation.
3. Where a translation of this certificate exists, the English language version shall be considered as the authoritative text.
4. Certification is limited to production undertaken at the sites listed in the manufacturers technical documentation.
5. Ongoing manufactured product shall be consistent with the product(s) certified and listed on this certificate.
6. The Manufacturer shall inform SATRA of any changes to the certified product or technical documentation.
7. This certificate shall be kept together with the relevant technical documentation in a safe place by the client named on this certificate. Production of this certificate and other documentation may be required by a representative of the EC member state government.
8. This certificate relates only to the condition of the testable items at the time of the certification procedure and is subject to the expiry date shown.
9. SATRA Technology reserves the right to withdraw this certificate if it is found that a condition of manufacture, design, materials or packaging have been changed and therefore no longer comply with the requirements of Regulation 2016/425.

EU Type-Examination Certificate

Certificate number: 2777/10905-01/E08-01

This EU Type-Examination Certificate covers the following product group(s) supported by testing to the relevant standards/technical specifications and examination of the technical file documentation:

Following the EU Type-Examination this product group has been shown to satisfy the applicable essential health and safety requirements of Annex II of the PPE Regulation (EU) 2016/425 as a Category III product.

Product reference:

AS LPP

Description:

Latex examination powdered gloves

Sizes:

6 XS
7 S
8 M
9 L
10 XL

Classification:

EN ISO 374-1:2016/Type B	Level	EN374-4:2013
40% Sodium Hydroxide	2	13.8%
65% Nitric Acid	1	46.8%
37% Formaldehyde	2	-5.1%
30% Hydrogen Peroxide	1	16.1%
40% Hydrofluoric Acid	2	Not Tested

EN ISO 374-5:2016

Resistance to Bacteria and Fungi	Pass
Resistance to Virus	Pass

Standards/Technical specifications applied:

EN 420: 2003+A1: 2009; EN ISO 374-1:2016; EN ISO 374-5:2016

Technical reports/Approval documents:

SATRA: 26530/0201/SPC-0, CHM0145939/0649, SPC0228864/1441/Issue 2, CHM0230439/1448/DRWM-Issue 2, SPC0234310/1518/2 Issue 2, SPC0235570/1524, PRC0251482/1645/SMcD, CHM0265592/1750/EN/A, CHM0265592/1750/EN/B, CHM0265592/1750/SPT
TUV: 7191152447-CHM16-01-RC

Signed on behalf of SATRA:



Hannah Coe



Geoff Graham

Date of issue: 17/04/2019**Expiry date:** 19/07/2023

TERMS AND CONDITIONS

The following conditions apply in addition to SATRA's standard terms and conditions of business and those given in the current certification agreement.

The certificate holder is licensed to mark the products detailed within this certificate in accordance with Annex V (Module B) of the Regulation (EU) 2016/425 of the European Parliament and of the council of 9th March 2016 on personal protective equipment once you have drawn up an EU declaration of product conformity. Please note:

1. Where the product is classified as category III then CE Marking of production is reliant on current compliance with Regulation 2016/425 module C2 or Module D. (Except that specifically produced to fit an individual user).
2. Full details of the certification and product are contained within the manufacturer's technical documentation.
3. Where a translation of this certificate exists, the English language version shall be considered as the authoritative text.
4. Certification is limited to production undertaken at the sites listed in the manufacturers technical documentation.
5. Ongoing manufactured product shall be consistent with the product(s) certified and listed on this certificate.
6. The Manufacturer shall inform SATRA of any changes to the certified product or technical documentation.
7. This certificate shall be kept together with the relevant technical documentation in a safe place by the client named on this certificate. Production of this certificate and other documentation may be required by a representative of the EC member state government.
8. This certificate relates only to the condition of the testable items at the time of the certification procedure and is subject to the expiry date shown.
9. SATRA Technology reserves the right to withdraw this certificate if it is found that a condition of manufacture, design, materials or packaging have been changed and therefore no longer comply with the requirements of Regulation 2016/425.

EU Type-Examination Certificate

Certificate number: 2777/10648-04/E04-01

This EU Type-Examination Certificate covers the following product group(s) supported by testing to the relevant standards/technical specifications and examination of the technical file documentation:
 Following the EU Type-Examination this product group has been shown to satisfy the applicable essential health and safety requirements of Annex II of the PPE Regulation (EU) 2016/425 as a Category III product.

Product reference:

AS NPF

Description:

Nitrile examination powder free gloves

Sizes:

6 (XS) – 10 (XL)

Classification:

EN ISO 374-1:2016/Type B	Level	EN374-4:2013
37% Formaldehyde	6	3.1%
40% Sodium Hydroxide	6	-25.6%
30% Hydrogen Peroxide	2	17.0%
EN ISO 374-5:2016		
Resistance to Bacteria and Fungi	Pass	
Resistance to Virus	Pass	

Standards/Technical specifications applied:

EN 420: 2003+A1: 2009; EN ISO 374-1:2016; EN ISO 374-5:2016

Technical reports/Approval documents:

SATRA: CHM0265112/1749/EN/A, CHM0265112/1749/EN/B, CHM0265112/1749/SPT, CHM0272621/1826/JS, CHM0275215/1836/LH, CHM0275215/1836/LH/E, CHM0275215/1836/LH/D, CHM0275215/1836/LH/A/Final
 TUV: 7191143339-CHM16-01-RC

Signed on behalf of SATRA:



Hannah Coe



Geoff Graham

Date of issue: 17/04/2019

Expiry date: 25/06/2023

TERMS AND CONDITIONS

The following conditions apply in addition to SATRA's standard terms and conditions of business and those given in the current certification agreement.

The certificate holder is licensed to mark the products detailed within this certificate in accordance with Annex V (Module B) of the Regulation (EU) 2016/425 of the European Parliament and of the council of 9th March 2016 on personal protective equipment once you have drawn up an EU declaration of product conformity. Please note:

1. Where the product is classified as category III then CE Marking of production is reliant on current compliance with Regulation 2016/425 module C2 or Module D. (Except that specifically produced to fit an individual user).
2. Full details of the certification and product are contained within the manufacturer's technical documentation.
3. Where a translation of this certificate exists, the English language version shall be considered as the authoritative text.
4. Certification is limited to production undertaken at the sites listed in the manufacturers technical documentation.
5. Ongoing manufactured product shall be consistent with the product(s) certified and listed on this certificate.
6. The Manufacturer shall inform SATRA of any changes to the certified product or technical documentation.
7. This certificate shall be kept together with the relevant technical documentation in a safe place by the client named on this certificate. Production of this certificate and other documentation may be required by a representative of the EC member state government.
8. This certificate relates only to the condition of the testable items at the time of the certification procedure and is subject to the expiry date shown.
9. SATRA Technology reserves the right to withdraw this certificate if it is found that a condition of manufacture, design, materials or packaging have been changed and therefore no longer comply with the requirements of Regulation 2016/425.

Report No. : CRSSA/02640/18
Company : ASAP International Sdn Bhd
No. 1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor Darul Ehsan, Malaysia

TEST REPORT

Product Description : Powdered Latex Examination Gloves
Brand Name : ASAP
Country of Origin : Malaysia
Size : Medium
Quantity Tested : 200 pieces
Test Conducted : Freedom from holes
Test Method : EN455 Part 1:2000
Testing Period : 02 Mar 2018 – 08 Mar 2018

Based on submitted samples, the following results obtained :-

Acceptable Quality Limit (AQL) : 1.5 Accept : 7 Found : 1

Result : Within AQL

Note: Upon Customer's request, this report has been issued in more than one original. Only the first original is a legally binding document and may be used for any legal purpose, including payment. (Original 1-3)

SGS (MALAYSIA) SDN. BHD.



CHEE TUCK CHOON
B.Sc. MMIC
SECTION HEAD

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for seven days (perishable food sample) or three months only.

Report No. : CRSSA/02641/18
Company : ASAP International Sdn Bhd
No. 1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor Darul Ehsan, Malaysia

TEST REPORT

Product Description : Powdered Latex Examination Gloves
 Brand Name : ASAP
 Country of Origin : Malaysia
 Size : Medium
 Quantity Tested : 13 pieces
 Test Conducted : Dimensions
 Test Method : EN 455 Part 2:2015
 Testing Period : 02 Mar 2018 – 08 Mar 2018

Based on submitted samples, the following results obtained :-

Size	M	M	M	M	M	M	M	M	M	M	M	M	M	Median
Width	93	94	93	93	94	94	94	94	93	94	94	94	95	94
Length	245	245	247	246	252	255	250	247	246	250	252	245	245	247

Note: Upon Customer's request, this report has been issued in more than one original. Only the first original is a legally binding document and may be used for any legal purpose, including payment. (Original 1-3)

SGS (MALAYSIA) SDN. BHD.



CHEE TUCK CHOON
 B.Sc. MMIC
 SECTION HEAD

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for seven days (perishable food sample) or three months only.

Report No. : CRSSA/02642/18
 Company : ASAP International Sdn Bhd
 No. 1, Jalan Sitar 33/6, Seksyen 33,
 40400 Shah Alam, Selangor Darul Ehsan, Malaysia

TEST REPORT

Product Description : Powdered Latex Examination Gloves
 Brand Name : ASAP
 Country of Origin : Malaysia
 Size : Medium
 Quantity Tested : 13 pieces
 Test Conducted : Force at Break During Shelf Life and After Challenge
 Test Method : EN 455 Part 2:2015
 Ageing : 70 ± 2 Deg C for 168 hrs
 Testing Period : 02 Mar 2018 – 14 Mar 2018

SIZE	SAMPLE NO.	Force at Break, N	
		<u>BEFORE AGING</u>	<u>AFTER AGING</u>
M	1	8.9	7.7
	2	8.8	8.1
	3	8.8	8.1
	4	9.4	7.7
	5	9.1	7.8
	6	9.8	7.8
	7	10.2	8.0
	8	9.3	8.6
	9	8.8	7.8
	10	8.8	8.4
	11	8.7	7.4
	12	10.3	7.8
	13	8.6	8.1
Median		8.9	7.8
Requirement		≥ 6.0	≥ 6.0

Note: Upon Customer's request, this report has been issued in more than one original. Only the first original is a legally binding document and may be used for any legal purpose, including payment. (Original 1-3)

SGS (MALAYSIA) SDN. BHD.



CHEE TUCK CHOON
 B.Sc. MMIC
 SECTION HEAD

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for seven days (perishable food sample) or three months only.

Report No. : CRSSA/02643/18
Company : ASAP International Sdn Bhd
No. 1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor Darul Ehsan, Malaysia

TEST REPORT

Product Description : Powdered Latex Examination Gloves
Brand Name : ASAP
Country of Origin : Malaysia
Size : Medium
Quantity Tested : 8 pieces
Test Conducted : Aqueous Extractable Protein Content
Test Method : EN455 Part 3:2015
Testing Period : 02 Mar 2018 – 12 Mar 2018

EXTRACTABLE PROTEIN CONTENT

Analysis had been carried out as per your request. We report the following results:

SIZE	AQUEOUS EXTRACTABLE PROTEIN CONTENT, µg/g
M	64

- Note: 1) Detection limit = 10µg/g
2) Upon Customer's request, this report has been issued in more than one original. Only the first original is a legally binding document and may be used for any legal purpose, including payment. (Original 1-3)

SGS (MALAYSIA) SDN. BHD.



CHEE TUCK CHOON
B.Sc. MMIC
SECTION HEAD

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for seven days (perishable food sample) or three months only.

Report No. : CRSSA/02644/18
Company : ASAP International Sdn Bhd
No. 1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor Darul Ehsan, Malaysia

TEST REPORT

Product Description : Powdered Latex Examination Gloves
Brand Name : ASAP
Country of Origin : Malaysia
Size : Medium
Quantity Tested : 2 pieces
Test Conducted : Powder Content
Test Method : EN455 Part 3:2015
Testing Period : 02 Mar 2018 – 08 Mar 2018

On testing the samples, the following results were obtained:-

<u>SIZE</u>	<u>Average Powder Mass per Glove</u>
M	11.6 mg

Note: Upon Customer's request, this report has been issued in more than one original. Only the first original is a legally binding document and may be used for any legal purpose, including payment. (Original 1-3)

SGS (MALAYSIA) SDN. BHD.



CHEE TUCK CHOON
B.Sc. MMIC
SECTION HEAD

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for seven days (perishable food sample) or three months only.

Test Report No : CPSA/210663477-CA64942
Company : ASAP International Sdn. Bhd.
No.1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor,
Malaysia.

TEST REPORT

Sample Description : Powder Free Latex Examination Gloves,
Device Basic UDI-DI: 955589090ASAPPFL4G
Brand Name : ASAP
Country of Origin : Malaysia
Country of Destination : All countries
Size : M
Quantity Tested : 200 pieces
Test Conducted : Freedom from holes
Test Method : EN455 Part 1:2020
Testing Period : 21 May 2021 – 04 June 2021

Based on submitted samples, the following results obtained :-

Acceptable Quality Limit (AQL) : 1.5 Accept : 7 Found : 4

Result : Within AQL

SIGNED FOR AND ON BEHALF OF
SGS (MALAYSIA) SDN BHD

****End of Report****



CHEE TUCK CHOON
SECTION HEAD
IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

Test Report No : CPSA/210663478-CA64943
Company : ASAP International Sdn. Bhd.
No.1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor,
Malaysia.

TEST REPORT

Sample Description : Powder Free Latex Examination Gloves,
 Device Basic UDI-DI: 955589090ASAPPFL4G
 Brand Name : ASAP
 Country of Origin : Malaysia
 Country of Destination : All countries
 Size : M
 Quantity Tested : 13 pieces per test
 Test Conducted : Force at Break During Shelf Life and After Challenge
 Test Method : EN 455 Part 2:2015
 Ageing : 70 ± 2 Deg C for 168 hrs
 Testing Period : 21 May 2021 – 04 June 2021

SIZE	SAMPLE NO.	Force at Break, N	
		<u>BEFORE AGING</u>	<u>AFTER AGING</u>
M	1	10.7	8.8
	2	10.7	10.0
	3	10.3	9.1
	4	11.1	9.5
	5	9.8	9.1
	6	10.6	9.8
	7	10.1	9.1
	8	10.7	9.1
	9	10.0	9.4
	10	10.7	10.0
	11	10.7	9.1
	12	9.4	8.4
	13	8.9	9.0
Median		10.6	9.1
Requirement		≥ 6.0	≥ 6.0

****End of Report****

SIGNED FOR AND ON BEHALF OF
 SGS (MALAYSIA) SDN BHD



CHEE TUCK CHOON
 SECTION HEAD
 IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

Test Report No : CPSA/210663479-CA64944
Company : ASAP International Sdn. Bhd.
No.1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor,
Malaysia.

TEST REPORT

Sample Description : Powder Free Latex Examination Gloves,
Device Basic UDI-DI: 955589090ASAPPFL4G
Brand Name : ASAP
Country of Origin : Malaysia
Country of Destination : All countries
Size : M
Quantity Tested : 13 pieces
Test Conducted : Dimensions
Test Method : EN 455 Part 2:2015
Testing Period : 21 May 2021 – 04 June 2021

Based on submitted samples, the following results obtained :-

Size	M	M	M	M	M	M	M	M	M	M	M	M	M	Median
Width Median: 95±10mm	95	94	95	95	94	95	94	95	95	95	95	94	95	95
Length Median: ≥ 240mm	251	248	251	252	252	254	250	252	252	251	251	250	251	251

SIGNED FOR AND ON BEHALF OF
SGS (MALAYSIA) SDN BHD

****End of Report****



CHEE TUCK CHOON
SECTION HEAD
IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

Test Report No : CPSA/210663480-CA64945
Company : ASAP International Sdn. Bhd.
No.1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor,
Malaysia

TEST REPORT

Sample Description : Powder Free Latex Examination Gloves,
Device Basic UDI-DI: 955589090ASAPPFL4G
Brand Name : ASAP
Country of Origin : Malaysia
Country of Destination : All countries
Size : M
Quantity Tested : 8 pieces
Test Conducted : Aqueous Extractable Protein Content
Test Method : EN455 Part 3:2015
Testing Period : 21 May 2021 – 04 June 2021

EXTRACTABLE PROTEIN CONTENT

Analysis had been carried out as per your request. We report the following results:

SIZE	AQUEOUS EXTRACTABLE PROTEIN CONTENT, µg/g
M	23

Note: Detection limit = 10 µg/g

SIGNED FOR AND ON BEHALF OF
SGS (MALAYSIA) SDN BHD

End of Report

CHEE TUCK CHOON
SECTION HEAD
IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

Test Report No : CPSA/210663481-CA64946
Company : ASAP International Sdn. Bhd.
No.1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor,
Malaysia.

TEST REPORT

Sample Description : Powder Free Latex Examination Gloves,
Device Basic UDI-DI: 955589090ASAPPFL4G
Brand Name : ASAP
Country of Origin : Malaysia
Country of Destination : All countries
Size : M
Quantity Tested : 5 pieces
Test Conducted : Powder Content
Test Method : EN455 Part 3:2015
Testing Period : 21 May 2021 – 04 June 2021

On testing the samples, the following results were obtained:-

SIZE

Average Powder Mass per Glove

M

0.12 mg

SIGNED FOR AND ON BEHALF OF
SGS (MALAYSIA) SDN BHD

****End of Report****



CHEE TUCK CHOON
SECTION HEAD
IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

Test Report No : CPSA/210663482-CA64947
Company : ASAP International Sdn. Bhd.
No.1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor,
Malaysia

TEST REPORT

Sample Description : Powder Free Nitrile Examination Gloves,
Device Basic UDI-DI: 955589090ASAPPFN4L
Brand Name : ASAP
Country of Origin : Malaysia
Country of Destination : All countries
Size : M
Quantity Tested : 200 pieces
Test Conducted : Freedom from holes
Test Method : EN455 Part 1:2020
Testing Period : 21 May 2021 – 04 June 2021

Based on submitted samples, the following results obtained:-

Acceptable Quality Limit (AQL) : 1.5 Accept : 7 Found : 3

Result : Within AQL

SIGNED FOR AND ON BEHALF OF
SGS (MALAYSIA) SDN BHD

**** End of Report ****



CHEE TUCK CHOON
SECTION HEAD
IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

Test Report No : CPSA/210663483-CA64948
Company : ASAP International Sdn. Bhd.
No.1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor,
Malaysia

TEST REPORT

Sample Description : Powder Free Nitrile Examination Gloves,
Device Basic UDI-DI: 955589090ASAPPFN4L
Brand Name : ASAP
Country of Origin : Malaysia
Country of Destination : All countries
Size : M
Quantity Tested : 13 pieces per test
Test Conducted : Force at Break During Shelf Life and After Challenge
Test Method : EN 455 Part 2:2015
Ageing : 70 ± 2 Deg C for 168 hrs
Testing Period : 21 May 2021 – 04 June 2021

SIZE	SAMPLE NO.	Force at Break, N	
		<u>BEFORE AGING</u>	<u>AFTER AGING</u>
M	1	8.4	8.1
	2	8.6	8.3
	3	9.1	8.9
	4	7.5	9.8
	5	8.2	8.8
	6	8.6	9.9
	7	8.9	8.9
	8	8.7	8.2
	9	8.1	8.2
	10	7.9	8.6
	11	9.4	9.1
	12	8.9	9.3
	13	8.5	9.3
Median		8.6	8.9
Requirement		≥ 6.0	≥ 6.0

SIGNED FOR AND ON BEHALF OF
SGS (MALAYSIA) SDN BHD

**** End of Report ****



CHEE TUCK CHOON
SECTION HEAD
IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

Test Report No : CPSA/210663484-CA64949
Company : ASAP International Sdn. Bhd.
 No.1, Jalan Sitar 33/6, Seksyen 33,
 40400 Shah Alam, Selangor,
 Malaysia

TEST REPORT

Sample Description : Powder Free Nitrile Examination Gloves,
 Device Basic UDI-DI: 955589090ASAPPFN4L
 Brand Name : ASAP
 Country of Origin : Malaysia
 Country of Destination : All countries
 Size : M
 Quantity Tested : 13 pieces
 Test Conducted : Dimensions
 Test Method : EN 455 Part 2:2015
 Testing Period : 21 May 2021 – 04 June 2021

Based on submitted samples, the following results obtained :-

Size	M	M	M	M	M	M	M	M	M	M	M	M	M	Median
Width Median: 95±10mm	96	96	96	96	96	96	96	96	96	95	95	96	96	96
Length Median: ≥ 240mm	245	240	246	244	241	245	241	246	242	240	248	249	245	245

SIGNED FOR AND ON BEHALF OF
 SGS (MALAYSIA) SDN BHD

**** End of Report ****



CHEE TUCK CHOON
 SECTION HEAD
 IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

Test Report No : CPSA/210663485-CA64950
Company : ASAP International Sdn. Bhd.
No.1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor,
Malaysia

TEST REPORT

Sample Description : Powder Free Nitrile Examination Gloves,
Device Basic UDI-DI: 955589090ASAPPFN4L
Brand Name : ASAP
Country of Origin : Malaysia
Country of Destination : All countries
Size : M
Quantity Tested : 5 pieces
Test Conducted : Powder Content
Test Method : EN455 Part 3:2015
Testing Period : 21 May 2021 – 04 June 2021

On testing the samples, the following results were obtained:-

SIZE

Average Powder Mass per Glove

M

0.08 mg

SIGNED FOR AND ON BEHALF OF
SGS (MALAYSIA) SDN BHD

**** End of Report ****



CHEE TUCK CHOON
SECTION HEAD
IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.



Art & Science of Amazing Protection

DATA SHEET

LATEX
POWDERED
EXAMINATION GLOVES



LATEX POWDERED EXAMINATION GLOVES

Features

Smooth surface

With smooth exterior, it enhances tactile sensitivity for improved feedback and precision.

Superior Strength

Capable to withstand against physical damage and harmful chemicals for increased hand security.

Enhanced dexterity

The enhanced dexterity and flexibility improve hand comfort for comfortable long period wear.

Lightly powdered

Interior powdered with absorbable corn-starch (USP) to ensure easy donning even with wet hands.

Ambidextrous

Fits well on both hands making it convenient and economical.

100% Latex

Latex powder-free with low protein rate to minimize exposure to allergens and skin irritants.

Application

ASAP Latex Powdered gloves are recommended to be used in industries that require:

- Handling chemicals and hazardous objects in medical and construction industries
- Protection from blood and other fluid contaminants
- Overall protection for workers such as nurses, caregivers and other healthcare providers

Industry icons:



Examination & Surgery (E&S)



Research & Development (R&D)



Building & Construction (B&C)



Hygiene Matters

At ASAP, we are committed to hygiene control and quality assurance. Proper hygiene standard is practiced throughout the development of all ASAP products – from raw materials handling, processing, production, inspection, to our finished product to deliver high quality products while limiting the risk of cross contamination.

Look for the Hygiene Matters logo, quality and hygiene you can trust.

Technical Specification Sheet

Product Information

Series Number	10410
Description	Latex, Powdered, Examination Gloves
Material	100% Latex Powder
Design	Ambidextrous, Textured & Beaded Cuff
Color	Natural
Quality Assurance	ISO 9001 ISO 13485 ISO 14001
Quality Control	EN 455-1:2000 EN 455-2:2015 EN 455-3:2015 EN 455-4:2009
Food Safe	Tested in accordance with EU standard EN 1186 under Food Commission Regulation 10/2011
Acceptable Quality Level (AQL)	1.5
Packaging	100 pcs per box, 10 boxes per carton
Handling & Storage	Keep out of direct sunlight; store in a cool and dry place
Country of Origin	Malaysia

Physical Properties Specification

Dimension Specification

Test	Particular	Specification	
Dimension (mm)	Palm Width	Extra Small	75 ± 5
		Small	85 ± 5
		Medium	95 ± 5
		Large	105 ± 5
		Extra Large	115 ± 5
	Length		Min. 240
	Thickness (Single wall)	Cuff (25 ± 5 from bead)	Min. 0.06
Palm (centre of palm)		Min. 0.08	
Finger (13 ± 3 from tip)		Min. 0.10	
Physical Properties	Force at break (N)	Before Aging	Min. 6.0
		After Aging	Min. 6.0
Physical Appearance	Rolled beaded cuff Natural Latex Colour Smooth Surface Ambidextrous		

Packaging Dimensions

Inner	220 x 110 x 65 mm
Carton	340 x 230 x 230 mm



Product Size Codes

Small	Medium	Large	Extra Large
10412	10413	10414	10415

Instructions for Use

Use

This guideline is to be used in combination with the specific information that is mentioned on each packaging. These gloves are designed as single and transient use gloves and should be disposed of after use. The gloves are liquid proof, and can therefore be used for splash protection against certain chemicals. Kindly ensure the gloves are used for the designated purposes only.

Precautions

This product contains latex rubber which may cause an allergic reaction in some individuals. Do not expose these gloves to any person with a known or suspected sensitivity to nitrile rubber. If allergic reaction occurs, discontinue use immediately and consult a physician. These gloves should not be used for thermal or harsh chemical protection. They are not intended for application involving prolonged direct exposure to harsh chemical where heavy-duty industrial gloves are required.

Variability in material thickness, gloves integrity, chemical concentration, temperature, and length of exposure to chemicals may affect performance. User is responsible for determining glove suitability for individual application. Before usage, inspect the gloves for any faults or deficiencies. If the gloves are ripped or punctured during use, dispose them immediately.

Storage

Store in a dry place at room temperature. Open product should be shielded from exposure to direct sunlight, intense artificial light, x-ray machines, and other sources of zone.

Disposal

Dispose according to Local Authority Regulations. Landfill or incinerate under controlled conditions.

ASAP INTERNATIONAL SDN BHD

No. 1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor, Malaysia.

T : +603 5191 0166

E : info@whyasap.com

W : www.whyasap.com

ASAP INNOVATIONS LTD

UNIT 7, Fonthill Business Park, THE
COURTYARD, Fonthill Rd, Ballyowen,
Dublin, D22 XA07, Ireland

T : +353 1 4661660

E : info@whyasap.ie

W : www.whyasap.ie

ASAP INNOVATIONS UK LTD

Atterbury Lakes, Fairbourne Drive,
Milton Keynes, MK10 9RG, United
Kingdom

T : +44 (0)1908 909146

E : info@whyasap.co.uk

W : www.whyasap.co.uk



Art & Science of Amazing Protection

DATA SHEET

LATEX

POWDER FREE

EXAMINATION GLOVES



LATEX POWDER FREE EXAMINATION GLOVES

Features

Micro textured surface

Provides excellent grip in all conditions with micro textures on the surface of the gloves.

Superior Strength

Capable to withstand against physical damage and harmful chemicals for increased hand security.

Enhanced dexterity

The enhanced dexterity and flexibility improve hand comfort for comfortable long period wear.

Powder Free

Retains donning efficiency through the use of smooth polymer interior that improves moist or dry Donning.

Ambidextrous

Fits well on both hands making it convenient and economical.

100% Latex

Latex powder-free with low protein rate to minimize exposure to allergens and skin irritants.

Application

ASAP Latex Powder Free gloves are recommended to be used in industries that require:

- Handling chemicals and hazardous objects in medical and construction industries
- Protection from blood and other fluid contaminants
- Overall protection for workers such as nurses, caregivers and other healthcare providers

Industry icons:



Examination & Surgery (E&S)



Research & Development (R&D)



Building & Construction (B&C)



Hygiene Matters

At ASAP, we are committed to hygiene control and quality assurance. Proper hygiene standard is practiced throughout the development of all ASAP products – from raw materials handling, processing, production, inspection, to our finished product to deliver high quality products while limiting the risk of cross contamination.

Look for the Hygiene Matters logo, quality and hygiene you can trust.



Technical Specification Sheet

Product Information

Series Number	10390
Description	Latex, Powder Free, Examination Gloves
Material	100% Latex Powder
Design	Ambidextrous, Textured & Beaded Cuff
Color	Natural
Quality Assurance	ISO 9001 ISO 13485 ISO 14001
Quality Control	EN 455-1:2000 EN 455-2:2015 EN 455-3:2015 EN 455-4:2009
Food Safe	Tested in accordance with EU standard EN 1186 under Food Commission Regulation 10/2011
Acceptable Quality Level (AQL)	1.5
Packaging	100 pcs per box, 10 boxes per carton
Handling & Storage	Keep out of direct sunlight; store in a cool and dry place
Country of Origin	Malaysia

Physical Properties Specification

Dimension Specification

Test	Particular	Specification		
Dimension (mm)	Palm Width	Extra Small	75 ± 5	
		Small	85 ± 5	
		Medium	95 ± 5	
		Large	105 ± 5	
		Extra Large	115 ± 5	
	Length		Min. 240	
		Thickness (Single wall)	Cuff (25 ± 5 from bead)	Min. 0.06
			Palm (centre of palm)	Min. 0.08
Finger (13 ± 3 from tip)	Min. 0.10			
Physical Properties	Force at break (N)	Before Aging	Min. 6.0	
		After Aging	Min. 6.0	
Physical Appearance	Rolled beaded cuff Natural Latex Colour Textured Surface Ambidextrous			



LATEX POWDER FREE EXAMINATION GLOVES

Packaging Dimensions

Inner	220 x 110 x 65mm
Carton	340 x 230 x 230mm



Product Size Codes

Small	Medium	Large	Extra Large
10392	10393	10394	10395

Instructions for Use

Use

This guideline is to be used in combination with the specific information that is mentioned on each packaging. These gloves are designed as single and transient use gloves and should be disposed of after use. The gloves are liquid proof, and can therefore be used for splash protection against certain chemicals. Kindly ensure the gloves are used for the designated purposes only.

Precautions

This product contains latex rubber which may cause an allergic reaction in some individuals. Do not expose these gloves to any person with a known or suspected sensitivity to nitrile rubber. If allergic reaction occurs, discontinue use immediately and consult a physician. These gloves should not be used for thermal or harsh chemical protection. They are not intended for application involving prolonged direct exposure to harsh chemical where heavy-duty industrial gloves are required.

Variability in material thickness, gloves integrity, chemical concentration, temperature, and length of exposure to chemicals may affect performance. User is responsible for determining glove suitability for individual application. Before usage, inspect the gloves for any faults or deficiencies. If the gloves are ripped or punctured during use, dispose them immediately.

Storage

Store in a dry place at room temperature. Open product should be shielded from exposure to direct sunlight, intense artificial light, x-ray machines, and other sources of zone.

Disposal

Dispose according to Local Authority Regulations. Landfill or incinerate under controlled conditions.

ASAP INTERNATIONAL SDN BHD

No. 1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor, Malaysia.

T : +603 5191 0166

E : info@whyasap.com

W : www.whyasap.com

ASAP INNOVATIONS LTD

UNIT 7, Fonthill Business Park, THE
COURTYARD, Fonthill Rd, Ballyowen,
Dublin, D22 XA07, Ireland

T : +353 1 4661660

E : info@whyasap.ie

W : www.whyasap.ie

ASAP INNOVATIONS UK LTD

Atterbury Lakes, Fairbourne Drive,
Milton Keynes, MK10 9RG, United
Kingdom

T : +44 (0)1908 909146

E : info@whyasap.co.uk

W : www.whyasap.co.uk



Art & Science of Amazing Protection

DATA SHEET

LITE NITRILE
POWDER FREE
EXAMINATION GLOVES



LITE NITRILE POWDER FREE EXAMINATION GLOVES

Features

Low extractable particulate level

Manufactured powder-free to avoid exposure to allergens and skin irritants.

Finger textured

Strategically placed finger textured surfaces increase grip to reduce gripping effort needed, ultimately reducing hand strain.

Superior comfort and natural fit

Soft yet three times the puncture resistance of latex with high resistance to chemicals with increased safety in various jobs scopes.

Ambidextrous

Fits either hand conveniently without needing to determine left or right for simpler donning.

100% Nitrile

Nitrile powder-free with no plasticizers, ideal solution when latex allergies are concerned.

Colourful options

Coloured glove for easy identification as a synthetic glove.

Application

ASAP Nitrile gloves are recommended to be used in industries for:

- Barrier protection against harmful chemicals and infectious fluid
- Protection from bodily fluids such as blood or mucus.
- Procedures that require tactile sensitivity
- Increased protection with overall great comfort for industries such as Healthcare, Dental, Cleaning, Beauty etc.

Industry icons:



Examination & Surgery (E&S)



Research & Development (R&D)



Building & Construction (B&C)



Hygiene Matters

At ASAP, we are committed to hygiene control and quality assurance. Proper hygiene standard is practiced throughout the development of all ASAP products – from raw materials handling, processing, production, inspection, to our finished product to deliver high quality products while limiting the risk of cross contamination.

Look for the Hygiene Matters logo, quality and hygiene you can trust.

Technical Specification Sheet

Product Information

Series Number	20540
Description	Nitrile, Powder Free, Examination Gloves
Material	100% Nitrile
Design	Ambidextrous, Finger Textured & Beaded Cuff
Color	Blue
Quality Assurance	ISO 9001 ISO 13485 ISO 14001
Quality Control	EN 455-1:2000 EN ISO 21420:2020 EN 455-2:2015 EN 374-1:2016+A1:2018 EN 455-3:2015 EN 374-4:2019 EN 455-4:2009 EN 374-5:2016
Food Safe	Tested in accordance with EU standard EN 1186 under Food Commission Regulation 10/2011
Acceptable Quality Level (AQL)	1.5
Packaging	100 pcs per box, 10 boxes per carton
Handling & Storage	Keep out of direct sunlight; store in a cool and dry place
Country of Origin	Malaysia

Physical Properties Specification

Dimension Specification

Test	Particular	Specification	
Dimension (mm)	Palm Width	Extra Small	75 ± 5
		Small	85 ± 5
		Medium	95 ± 5
		Large	105 ± 5
		Extra Large	115 ± 5
	Length		Min. 240
	Thickness (Single wall)	Cuff (25 ± 5 from bead)	Min. 0.04
Palm (centre of palm)		Min. 0.05	
Finger (13 ± 3 from tip)		Min. 0.06	
Physical Properties	Force at break (N)	Before Aging	Min. 6.0
		After Aging	Min. 6.0
Physical Appearance	Rolled beaded cuff Blue Colour Finger Textured Surface Ambidextrous		



LITE NITRILE POWDER FREE EXAMINATION GLOVES

Packaging Dimensions

Inner	220 x 110 x 65mm
Carton	340 x 230 x 230mm



Product Size Codes

Extra Small	Small	Medium	Large	Extra Large
20541	20542	20543	20544	20545

Instructions for Use

Use

This guideline is to be used in combination with the specific information that is mentioned on each packaging. These gloves are designed as single and transient use gloves and should be disposed of after use. The gloves are liquid proof, and can therefore be used for splash protection against certain chemicals. Kindly ensure the gloves are used for the designated purposes only.

Precautions

This product contains nitrile rubber which may cause an allergic reaction in some individuals. Do not expose these gloves to any person with a known or suspected sensitivity to nitrile rubber. If allergic reaction occurs, discontinue use immediately and consult a physician. These gloves should not be used for thermal or harsh chemical protection. They are not intended for application involving prolonged direct exposure to harsh chemical where heavy-duty industrial gloves are required.

Variability in material thickness, gloves integrity, chemical concentration, temperature, and length of exposure to chemicals may affect performance. User is responsible for determining glove suitability for individual application. Before usage, inspect the gloves for any faults or deficiencies. If the gloves are ripped or punctured during use, dispose them immediately.

Storage

Store in a dry place at room temperature. Open product should be shielded from exposure to direct sunlight, intense artificial light, x-ray machines, and other sources of zone.

Disposal

Dispose according to Local Authority Regulations. Landfill or incinerate under controlled conditions.

ASAP INTERNATIONAL SDN BHD

No. 1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor, Malaysia.

T : +603 5191 0166

E : info@whyasap.com

W : www.whyasap.com

ASAP INNOVATIONS LTD

UNIT 7, Fonthill Business Park, THE
COURTYARD, Fonthill Rd, Ballyowen,
Dublin, D22 XA07, Ireland

T : +353 1 4661660

E : info@whyasap.ie

W : www.whyasap.ie

ASAP INNOVATIONS UK LTD

Atterbury Lakes, Fairbourne Drive,
Milton Keynes, MK10 9RG, United
Kingdom

T : +44 (0)1908 909146

E : info@whyasap.co.uk

W : www.whyasap.co.uk