



## Certificate of Analysis

29/05/2025

<b>EASIMIP™ PATULIN</b>	
Product code	P250, P250B
Product size	10 and 50 columns
Batch number	NJ 182
Date of manufacture	02 September 2024
Expiry	02 September 2027
Storage recommendations	22°C. Room temperature

This document is electronically created and therefore valid without a signature

QC/CERTIFICATES/TEMPLATES/V9/11.2023





## Certificate of Analysis

<b>Tested to Specification:</b>			
<b>Quality Control Product Release Inspection and Test Specification</b>			
<b>Visual inspection</b>	The columns were inspected for powder colour and column integrity.		
<b>Visual inspection result</b>	Pass		
<b>Recovery testing method</b>	80 ng of patulin is applied in 4 ml of 1 % acetic acid/water solution.		
<b>Recovery specifications</b>	Greater than or equal to 70 % and less than or equal to 105 % of patulin should be recovered with standard solutions. This meets the specification required by (EC) 466/2001.		
<b>Recovery results (n = 6)</b>	81 %	<b>% RSD (n = 6)</b>	3.9 %

This document is electronically created and therefore valid without a signature

This is to certify that EASIMIP™ PATULIN batch NJ 182 passes the required specifications and is released for sale.

R-Biopharm Rhône Ltd.  
Quality Control Department



Kirsten Stirling  
Quality Manager



QC/CERTIFICATES/TEMPLATES/V9/11.2023