



# EC-CERTIFICATE

(Full quality assurance system)



This is to certify that the company

## Feuerstein GmbH

Gustav-Krone-Straße 3  
14167 Berlin  
Germany

has implemented and maintains a full quality assurance system which applies to the products at every stage from design to final controls.

Through an audit, documented in a report, performed by DQS Medizinprodukte GmbH, it was verified that the management system fulfills the requirements of

### Annex II – excluding Section 4 of Council Directive 93/42/EEC concerning medical devices

with respect to the following medical devices:

Device	Class
Surgical Sutures (Silk and Polyester) with and without needle	Ila
Surgical Needles steril	Ila

The manufacturer is subject to surveillance according to Annex II, Section 5. The CE marking with the Notified Body Identification Number (0297) may be affixed on the devices listed in the certificate. An EC Design Examination Certificate according to Annex II, Section 4 is required for class III devices covered by this certificate. The certificate is in the case of class I(s) devices (I(s) = class I products placed on the market in sterile conditions) limited to the aspects of manufacture concerned with securing and maintaining sterile conditions. The certificate is in the case of class I(m) devices (I(m) = class I devices with a measuring function) limited to the aspects of manufacture concerned with the conformity of the products with the metrological requirements.

Certificate registration no. 092558 MR2  
Certificate unique ID 170653076  
Effective date 2017-01-20  
Expiry date 2022-01-19  
Frankfurt am Main 2017-01-20

### DQS Medizinprodukte GmbH

Sigrid Uhlemann  
Managing Director

Dr. Thomas Feldmann  
Head of Certification Body



August-Schanz-Straße 21, 60433 Frankfurt am Main,  
Tel. +49 (0) 69 95427-300, [medical.devices@dqs-med.de](mailto:medical.devices@dqs-med.de)

DQS Medizinprodukte GmbH is a Notified Body according to Council Directive 93/42/EEC concerning medical devices with the Identification Number 0297.

