

## **VOMIT BAG**

TD-58-I.2.c-1

Revision 2

date: 2021-07-01

amendment of: 2024-04-23

MANUFACTURER: ZARYS International Group sp. z o.o. sp.k.

ADDRESS: ul. Pod Borem 18, 41-808 Zabrze, Poland; T: +48 32 376 07 00-02,06 F: +48 32 370 38 08; www.zarys.com

**CLASSIFICATION** 

Class I medical device compliant with Medical Devices Regulation (EU) 2017/745 (MDR)

## **INTENDED PURPOSE**

 Disposable device intended for the collection of gastric contents during vomiting in the patient in order to observe the nature of the secretion for further diagnostic action

## **FEATURES**

- made of durable, transparent LDPE foil in blue or red colour, allowing the secretion to be observed
- wide bag's inlet protected by polypropylene collar
- collar's circular rim provides a firm grip and makes the bag easier to manoeuvre, reducing the risk of content contamination
- collar is equipped with a special notch enabling closing of the bag and hygienic disposal of vomitus content - "twist and hitch"
- total capacity of the bag: 2000 ml
- precise measuring graduation (from 10 to 100 ml every 10 ml (numerically every 20 ml) and from 100 to 2000 ml every 100 ml) placed on the bag allows accurate estimation of liquid volume
- reading direction of the graduation for small volumes is indicated by a arrowhead
- device name, manufacturer's name and pictorial instruction for use printed directly on the bag
- latex-free
- phthalate-free
- single-use
- non-sterile
- shelf life: 5 years







CATALOGUE NUMBER	CAPACITY	COLOUR	INTERMEDIATE PACKAGING	TRANSPORT PACKAGING
WW-01	2000 ml	blue	20 pcs (foil)	30 x 20 pcs (carton)
WW-02 *	2000 ml	red	20 pcs (foil)	30 x 20 pcs (carton)
WW-03	2000 ml	white	20 pcs (foil)	30 x 20 pcs (carton)

* _	avai	lable	on	request	t

Marketing material intended for the health care professionals.  Prepared by: Ł.W. – Medical Device Specialist ZARYS International Group sp. z o.o. sp.k.	
No part of this document may be changed or copied without the permission of ZARYS International Group sp. z o.o.	. <i>sp.k.</i> rev.2022.01