



Certificate of Conformity

VWR TWIST SEAL BAG, CLOSURE WITH 1 ROUND AND 1 FLAT WIRE, WITHOUT WRITE ON PANEL

VWR International certifies that the products comply in every respect with the relevant particular specifications and/or drawings and/or relevant standards and regulations in force, mentioned in this document.

VWR International certifie que les produits sont conformes à tous égards aux spécifications particulières appropriées et / ou dessins, normes et règlements en vigueur, mentionnées dans le présent document

VWR International bestätigt, dass das Produkte allen zutreffenden Spezifikations- und/oder Zeichnungs- und/oder Standardanforderungen sowie den im Dokument aufgeführten gültigen Vorschriften entspricht.

Cat. No.	Capacity (ml)	Thickness (µm)	Dimensions (mm)	Pk
129-0010	2250	101	254×305	250
129-0011	4000	101	254×382	250
129-0417	1080	76	140×305	500
129-0418	1650	76	178×305	250
129-9019	130	63	76×178	500
129-9801	800	63	114×305	500
129-9802	1200	63	114×382	500
129-9803	450	63	114×229	500
129-9804	1500	76	140×382	500
129-9805	650	76	140×229	500

Bag material : Virgin polyethylene tubing (FDA approved)

Closure material : Steel wires covered with FDA approved pressure sensitive tape with a backing of polypropylène or PVC.

Sterility :

Sterility is assured due to extrusion temperatures exceeding 220°C which eliminates any present bacteria. During manufacturing all bags are sealed at both top and bottom in a continuous format before being cut into individual bags. This method assures that the bag's interior will never come in contact with its surroundings. Periodic validation tests, based on USP 34 – Sterility testing, are performed in our facility to guaranty the sterility of our processes.

Shelf life : 5 years sterility guaranteed if package undamaged

RNase, DNase and pyrogen : products are free of RNase, DNase and pyrogen



Certificate of Conformity

Compliant to: FDA

This is to confirm that the polyethylene used in the production of our sterile sampling bags conforms to the applicable regulations to the Food and Drug Administration [21 CFR 177.1520 (c)3.1a, (c)3.2a and 178.2010] and EU No 10/2011 for use in contact with food.

Other characteristics:

- Provide a secure, contaminant-free pliant container that ensures dependable analysis results
- Offer an economical and efficient way to collect, contain and carry samples
- Wire tab keeps bag open for filling
- Unique fold over system providing a temporary leak-proof seal
- Sterile material.

Disclaimer: VWR wants to state that this declaration will not discharge the user from his obligation to ensure the product is suitable for the intended use. - If there is any doubt ask us.

Date of issue: **June 2017**