## **Specification: V3**



## SHENZHEN COMEN MEDICAL INSTRUMENTS CO.,LTD

No.2 of FIYATA Timepiece Building, Nanhuan Avenue, Gongming sub-district, Guangming New District, Shenzhen, P.R. China

Tel: +86-755-26408879 Fax: +86-755-26431232 Email: info@szcomen.com Web: www.comen.com



## Ventilator

**Technical Specification** 

## V3



<b>Physical Characteristics</b>		
Dimensions	1389mm × 528 mm × 697 mm	
	(Height $\times$ Width $\times$ Depth) (including	
	trolley);	
	343.5mm × 312.5 mm × 258 mm	
	(Height $\times$ Width $\times$ Depth)	
	(excluding trolley)	
Weight	30±0.5 kg (including trolley)	
	10±0.5 kg (excluding trolley)	
Air Supply	Integrated ultra-silent turbine	
Screen		
Screen Size:	12.1" TFT touch screen	
Resolution	1280 × 800	
Brightness:	Adjustable	
<b>Ventilator Specification</b>		
Ventilation mode	V-A/C (Volume assist/control)	

	343.5mm × 312.5 mm × 258 mm
	(Height $\times$ Width $\times$ Depth)
	(excluding trolley)
Weight	30±0.5 kg (including trolley)
	10±0.5 kg (excluding trolley)
Air Supply	Integrated ultra-silent turbine
Screen	
Screen Size:	12.1" TFT touch screen
Resolution	1280 × 800
Brightness:	Adjustable
<b>Ventilator Specification</b>	
Ventilation mode	V-A/C (Volume assist/control)
	P-A/C (Pressure assist/control)
	V-SIMV (Volume - Synchronized
	Intermittent Mandatory
	Ventilation)
	P-SIMV (Pressure - Synchronized
	Intermittent Mandatory
	Ventilation)
	CPAP/PSV,
	DuoVent,
	APRV,
	PRVC
	PRVC-SIMV
	VS
	PSV-S/T
Invasive Mode	V-A/C, P-A/C, V-SIMV, P-SIMV,
	CPAP/PSV, DuoVent, PRVC, APRV,
	PRVC-SIMV, VS

Non-invasive Mode	P-A/C, P-SIMV, CPAP/PSV,
	DuoVent, APRV, PSV-S/T

Controlled parameter r	anges
O <sub>2</sub> %:	21 - 100% (increments of 1 %)
TV (Tidal Volume):	Adult: 100 - 2200 mL (increments
	of 10 mL)
	Pediatric: 20 - 300 mL (increments
	of 1 mL)
Respiratory Rate (RR):	0 - 100 bpm (increments of 1 bpm)
fSIMV (Ventilation	1 - 60 bpm (increments of 1 bpm).
frequency in SIMV	
mode):	
I:E range:	1:10~4:1.
Tinsp (Inspiratory time):	0 - 10 s (increments of 0.05 s).
Tslope (Time of Pressure	0 - 2.00 s (increments of 0.05 s).
Rising):	
High Pressure Time	0.2 - 30 s (increments of 0.1 s)
(Thigh)]:	
Tlow (Low Pressure	0.2 - 30 s (increments of 0.1 s)
Time):	
Max inspiratory Time	0 ~ 15.00 s(increments of 0.1 s)
(Timax):	
Tpause:	5 % - 60 % (increments of 5 %), Off
ΔPinsp(Inspiratory	0 - 80 cmH <sub>2</sub> O (increments of 1
pressure):	cmH₂O)
ΔPsupp:	0 - 85 cmH <sub>2</sub> O (increments of 1
	cmH₂O)
Phigh (High Pressure	0 - 80 cmH <sub>2</sub> O (increments of 1
Level):	cmH₂O)
Plow (Low Pressure	0 - 50 cmH <sub>2</sub> O (increments of 1
Level):	cmH₂O)
PEEP:	$0-50 \text{ cmH}_2\text{O}$ (increments of 1
	cmH₂O), Off



			22. (2.22
Flow trigger	0.5 -15 L/min (increments of 0.1	Compliance	Cstat, Cdyn (0 - 300 mL/cmH <sub>2</sub> O)
	L/min), Off	Inspired Oxygen (FiO₂)	15 - 100 %
Pressure trigger	-10 to - 0.5 cmH <sub>2</sub> O (increments of	WOB (Work of	0 – 100.0 J/min
Form O/ /Formitantiana	0.5 cmH <sub>2</sub> O), Off	Breathing)	0.40
Exp % (Expiration	10 - 85% (increments of 5%), Auto	RCexp (Expiratory Time	0 - 10 s
termination/trigger		Constant)	Airman Florida
level)		Waveforms	Airway pressure - time, Flow - time, Volume - time
Apnea Ventilation TVapnea	Adult: 100 - 2200 mL (increments	Mooning indicator	volume - time
Γναμπεα	of 10 mL) Pediatric: 20 - 300 mL	Weaning indicator P0.1	-20 - 0 cmH <sub>2</sub> O
	(increments of 1 mL)	NIF (Maximum negative	-20 - 0 CITIH <sub>2</sub> O
ΔPapnea	5 - 80 cmH₂O (increments of 1	inspiratory pressure)	-45 - 0 cmH₂O
ді арпеа	cmH <sub>2</sub> O)	RSBI (Rapid Shallow	
RRapnea (Apnea	1 - 80 bpm (increments of 1 bpm)	Breathing Index)	0 - 999 /(L•min)
Respiratory Rate)	1 oo op (e. ee. o. 1 op,	Special Function	
Apnea Tinsp	0.20 - 10 s (increments of 0.05 s)	Manual Breath	
Sigh	, , , , , , , , , , , , , , , , , , , ,	Expiration Hold	
Sigh Switch	ON, Off	Inspiration Hold	
Interval	20 s - 180 min (increments of 1 s	Nebulizer	
	from 20 to 59 s, increments of 1	O2个(O2 enrichment)	
	min from 1 to 180 min)	Sputum Suction	
Cycles Sigh	1 - 20 (increments of 1)	Smart Pulmonary View	
Δint.PEEP	1 - 45 cmH₂O (increments of 1	Lung Recruitment Tool	Sustained Insufflation
	cmH₂O), Off	PEEPi monitoring	
Automatic Tube Resistance Compensation		<del>-</del>	Paw - Volume, Flow - Volume, Paw
Tube Type	endotracheal intubation and	P-V Tool	- Flow
	tracheotomy tube	Tube Resistance	
Tube I.D.	Adult: 5.0 - 12.0 mm (increments of	Compensation	TRC
	0.5 mm) Pediatric: 2.5 - 8.0 mm	Smart Sync	IntelliSynTec
	(increments of 0.5 mm)	O2 Therapy	2-60 L/min
Compensate	1 -100 % (increments of 1 %) off		EtCO2, Vdaw, VDaw/Tve, Vtalv,
Expiration Compensation	ON, Off	CO2	V'alv, SlopeCO2, V'CO2, VeCO2,
Switch			ViCO2
Monitoring		Control accuracy	
Airway pressure range	Ppeak, Pplat, Pmean (Range -20 -	O <sub>2</sub> %	± (3 vol.% +1 % of setting)
	120 cmH₂O)	TV	$\pm$ (10 mL + 10% of the set value)
PEEP	$0{\sim}120\mathrm{cmH2O}$	Tinsp	$\pm$ 0.1 s or $\pm$ 10 % of setting,
Tidal volume range:	0∼4000 mL		whichever is greater
Respiratory Rate	ftotal, fmand, fspn (Range 0 - 200	I: E	$1:4^{2}:1: \pm 10\%$ of the set value;
	bpm)		Other range(1:10~4:1): ± 15% of
Minute volume range	MV, MVspn, MVleak (Range 0 –		the set value.
	100.0 L/min)	RR	±1 bpm
Resistance	Rinsp, Rexp (0 - 600 cmH <sub>2</sub> O/L/s)	fSIMV	±1 bpm



T. (D. ( m. )	. (0.0		
Tslope (Rising Time)	± (0.2s + 20% of the set value)		Within the range of 5
Phigh	± (2 cmH2O + 5% of the set value)		cmH2O/(L/s) $\sim$ 20 cmH2O/(L/s),
Plow	± (2 cmH2O + 5% of the set value)		±10 cmH2O/(L/s);
Thigh	±0.2s or ± 10% of the set value,		Within the range of 20
	whichever is larger		cmH2O/(L/s) $\sim$ 500 cmH2O/(L/s)
Tlow	±0.2s or ± 10% of the set value,		(exclude 20), ±50% of the actual
	whichever is larger		reading).
Pressure Trigger	$\pm$ (1 cmH2O + $\pm$ 10% of the set	Compliance	Within the range of 0 mL/
	value)		cmH2O $\sim$ 300 mL/ cmH2O, $\pm$ (2 mL/
Flow Trigger	± (1 L/min + 10% of the set value)		cmH2O + 20% of the actual
Δint.PEEP	2-45cmH2O± (2 cmH2O + 5% of the		reading).
	set value) (exclude 2)1-2cmH2O±	RSBI	Within the range of 0 $$ /(min·L) $\sim$
	(1% of the set value)		999 /(min·L) , ± (3 /(min·L)+15%
Exp %	±10% (absolute error)		of the actual reading).
Fapnea (Apnea	±1bpm	WOB	Within the range of 0.0 J/min $\sim$
Frequency)			100.0 J/min, ± (1 J/min + 15% of
ΔPapnea	$\pm$ (2 cmH2O + 5% of the set value)		the actual reading).
TVapnea	± (10 mL + 10% of the set value)	NIF	Within the range of -45.0 cmH2O $\sim$
Apnea Tinsp	±0.1s or ± 10% of the set value,		0.0 cmH2O, ± (2 cmH2O + 4% of
	whichever is larger		the actual reading)
Monitoring Accuracy		P0.1	Within the range of -20.0 cmH2O $\sim$
Airway pressure (Ppeak,	Within the range of -20cmH2O $\sim$		0.0 cmH2O, $\pm$ (2 cmH2O + 4% of
Pplat, Pmean, PEEP)	120 cmH2O, ± (2 cmH2O + 4% of		the actual reading).
	the actual reading)	RCexp	Within the range of 0.0s $\sim$ 10.0s,
Tidal Volume (TVi, TVe,	idal Volume (TVi, TVe, Within the range of 0 mL∼100 mL,		$\pm$ (0.2s + 20 % of the actual
TVe/IBW, TVe spn)			reading).
	reading);	Alarm Settings	
	Within the range of 100 mL $\sim$ 4000	Tidal Volume	Upper alarm limit Adult:
	mL, ± (3 mL + 10% of the actual		110~4000 mL, OFF
	reading)		Pediatric: 25~600 mL, OFF
Minute Volume (MV,	Within the range of 0.0 L/min $\sim$		Lower alarm limit Adult: OFF,
MVspn, MVleak)	100.0 L/min, ± (0.2 L/min + 10% of		50~3995 mL
, ,	the actual reading)		Pediatric: OFF, 10~595 mL
Frequency (ftotal,	Within the range of 0 bpm $\sim$ 200	Minute Volume	Upper alarm limit: Adult: 0.2~100.0
fmand, fspn)	bpm, ±1 bpm or ±5% of the actual		L/min
···-/ ·/	reading, whichever is larger		Pediatric: 0.2~60.0 L/min
Inspired Oxygen (FiO₂)	Within the range of 15 vol.% $\sim$ 100		Lower alarm limit: Adult: 0.1~50.0
- p	vol.%, ± (2.5 vol. % + 2.5% of the		L/min
	actual reading).		Pediatric: 0.1~30.0 L/min
Resistance	Within the range of 0	Airway pressure	
	cmH2O/(L/s) $\sim$ 5 cmH2O/(L/s), the	Frequency (Respiratory	
	accuracy is not defined;	Rate)	
	· · · · · · · · · · · · · · · · · · ·	Inspired oxygen (FiO₂)	Upper alarm limit: 10~90 cmH2O.



	Lower alarm limit: OFF,5~ (upper alarm limit -5) cmH2O		CO2 accuracy (under all conditions):
Apnea alarm time	Upper alarm limit: 2~160 bpm, OFF Lower alarm limit: OFF,1~(upper alarm limit -1) bpm		a) Within the range of 0 mmHg $\sim$ 114 mmHg, $\pm$ (2.25 mmHg + 4% of the reading).
SideStream CO2 modu	ile		b) Within the range of 115
Displayed numerics	EtCO <sub>2</sub>		mmHg $\sim$ 190 mmHg, the accuracy
Measurement Range	Comen SideStream: 0 mmHg $\sim$ 150		is not defined
	mmHg, 0% $\sim$ 19.7%, 0 kPa $\sim$ 20 kPa	Waveforms	EtCO <sub>2</sub> - time
	(at 760 mmHg)	Resolution	
	Respironics Capno SideStream: 0	Sampling rate and	Comen SideStream: sampling rate:
	mmHg $\sim$ 99 mmHg, 0.0 % $\sim$ 13.0 %,	accuracy	50 mL/min; sampling rate control
	0 kPa $\sim$ 13.2 kPa (at 760 mmHg)		accuracy: ± 10mL/min;
	Masimo ISA Capno SideStream: 0		Respironics Capno SideStream:
	mmHg $\sim$ 190 mmHg, 0 vol $\%~\sim~$ 25		sampling rate: 50 mL/min;
	vol% (at 760 mmHg)		sampling rate control accuracy: ±
Measurement accuracy	Comen SideStream:		10 mL/min.
	a) Within the range of 0 mmHg $\sim$		Masimo ISA Capno SideStream:
	40 mmHg, ± 2 mmHg;		sampling rate: 50mL/min; sampling
	b) Within the range of 41 mmHg $\sim$		rate control accuracy: ± 10 mL/min.
	70 mmHg, ± 5% of the reading;	System response time	Masimo mainstream: < 1 s; Masimo
	c) Within the range of 71 mmHg $\sim$		ISA Capno sidestream: < 3s (use a 2
	100 mmHg, ± 8% of the reading;		m sampling line)
	d) Within the range of 101	Rise time (Response	Masimo mainstream: < 1 s; Masimo
	mmHg $\sim$ 150 mmHg, $\pm$ 10% of the reading.	time)	ISA Capno sidestream: < 3s (use a 2 m sampling line)
	Respironics Capno SideStream:	EtCO₂ Alarm Upper	Comen sidestream: (lower alarm
	(Note: the gas temperature is $25^{\circ}$ C, if respiratory rate is greater than 80 rpm, the accuracy is 12% of the reading):	Limits	limit + 2 mmHg) $\sim$ 150 mmHg Respironics Capno sidestream: (lower alarm limit + 2 mmHg) $\sim$ 99 mmHg
	0 mmHg $\sim$ 38 mmHg: $\pm$ 2 mmHg,		Masimo ISA Capno sidestream:
	39 mmHg $\sim$ 99 mmHg: $\pm$ 10% of the		(lower alarm limit + 2 mmHg) $\sim$
	actual reading.		190 mmHg
	Masimo ISA Capno SideStream:	EtCO₂ Alarm Lower	Comen sidestream:0 mmHg $\sim$
	CO2 accuracy (under the condition:	Limits	(upper alarm limit - 2 mmHg)
	$22^{\circ}$ C $\pm 5^{\circ}$ C $1013 \pm 40$ hPa; gas		Respironics Capno sidestream: 0
	mixture of CO2 and N2.)  a) Within the range of 0		mmHg $\sim$ (upper alarm limit $$ - $$ 2 mmHg)
	mmHg $\sim$ 114 mmHg, $\pm$ (1.52		Masimo ISA Capno sidestream: 0
	mmHg + 2% of the reading). b) Within the range of 114		mmHg $\sim$ (upper alarm limit $$ - $$ 2 mmHg)
	mmHg $\sim$ 190 mmHg, the accuracy	MainStream CO₂ Mod	
	is not defined.		



Displayed numerics	EtCO <sub>2</sub>		CO2 accuracy (under all
EtCO <sub>2</sub> Measurement	Comen mainstream: 0 mmHg $\sim$ 150		conditions):
range	mmHg,0% $\sim$ 19.7%,0 kPa $\sim$ 20 kPa		a) Within the range of 0 mmHg $\sim$
	(at 760 mmHg);		114 mmHg, $\pm$ (2.25 mmHg + 4%
	Respironics CAPNOSTAT 5: 0		of the reading);
	mmHg $\sim$ 150 mmHg, 0% $\sim$ 19.7%, 0		b) Within the range of 114
			mmHg $\sim$ 190 mmHg, the accuracy
	kPa~20 kPa (at 760 mmHg);		is not defined;
	Masimo IRMATM mainstream: 0	Decelution	is not defined;
	mmHg $\sim$ 190 mmHg, 0 vol% $\sim$ 25 vol% (at 760 mmHg);	Resolution Waveforms	EtCO <sub>2</sub> - time, V - CO <sub>2</sub>
EtCO₂ Measurement	Comen mainstream:	EtCO <sub>2</sub> Alarm Upper	Comen mainstream: (lower alarm
	a) Within the range of 0mmHg $\sim$	Limits	·
Accuracy	40mmHg, $\pm$ 2 mmHg;	LIIIIILS	limit + 2 mmHg) ∼150 mmHg
	b) Within the range of 41mmHg $\sim$		Respironics CAPNOSTAT 5
			mainstream: (lower alarm limit
	70mmHg, $\pm$ 5% of the reading;		+2mmHg) ~150 mmHg
	c) Within the range of 71mmHg~		Masimo IRMATM mainstream:
	100mmHg, $\pm$ 8% of the reading; d) Within the range of 101mmHg $\sim$		(lower alarm limit + 2 mmHg) $\sim$ 190 mmHg
	150mmHg, $~\pm~$ 10% of the reading.	EtCO₂ Alarm Lower	Comen mainstream: 0 mmHg $\sim$
	Respironics CAPNOSTAT 5	Limits	(upper alarm limit - 2 mmHg)
	mainstream: CO2 accuracy (Note:		Respironics CAPNOSTAT 5
	Temperature :35 $^{\circ}\mathrm{C}$ ):		mainstream: 0 mmHg $\sim$ (upper
	a) Within the range of 0		alarm limit - 2 mmHg)
	mmHg $\sim$ 40 mmHg, $\pm$ 2 mmHg;		Masimo IRMATM mainstream: 0
	b) Within the range of 41		mmHg $\sim$ (upper alarm limit - 2
	mmHg $\sim$ 70 mmHg, $~\pm~$ 5% of the		mmHg)
reading;	reading;	SpO₂ module:	
	c) Within the range of 71	Display	Pulse rate (PR)
	mmHg $\sim$ 100 mmHg, $\pm$ 8% of the		waveform/parameter, SpO2
	reading;	SpO2 measurement	Nellcor SpO2: 0%~100%
	d) Within the range of 101	range	Masimo SpO2: $1\%$ $\sim$ 100%
	mmHg $\sim$ 150 mmHg, $~\pm~$ 10% of		Comen SpO2: 0%~100%
	the reading.	SpO2 accuracy	Nellcor SpO2: Within the range of
	Masimo IRMATM mainstream: CO2		70% $\sim$ 100%, Adult/Pediatric
	accuracy (under the condition:		measurement accuracy is $\pm 2\%$
	$22^{\circ}$ C $\pm$ 5 $^{\circ}$ C 1013 $\pm$ 40 hPa;		(during non-motion state); Within
	gas mixture of CO2 and N2.):		the range of 0% $\sim$ 69%,
	a) Within the range of 0 mmHg $\sim$		measurement accuracy is not
	114 mmHg, $\pm$ (1.52 mmHg + 2%		defined.
	of the reading);		Masimo SpO2: Within the range of
	b) Within the range of 114		70% $\sim$ 100%, Adult/Pediatric
	mmHg $\sim$ 190 mmHg, the accuracy		measurement accuracy is $\pm 2\%$
	is not defined;		(during non-motion state), $\pm 3\%$
	,		(during non-motion state), ±3%



	(during motion state); Within the range of $1\%\sim69\%$ , the measurement accuracy is not defined.  Comen SpO2: Within the range of $70\%\sim100\%$ , Adult/ Pediatric measurement accuracy is $\pm2\%$ (during non-motion state); Within	Lower SpO2 alarm limit	+1%)~100%  Comen SpO2: (Lower alarm limit +1%)~100%  Nellcor SpO2: 20%~(Upper alarm limit -1%)  Masimo SpO2: 1%~(Upper alarm limit -1%)  Comen SpO2: 0%~(Upper alarm
DP measurement range	the range of 0%~69%, the measurement accuracy is not defined.	Upper PR alarm limit	limit -1%)  Nellcor SpO2: (Lower alarm limit +1  bpm)~300 bpm  Masima SpO3: (Lower alarm limit +1
PR measurement range PR measurement	Nellcor SpO2: 20 bpm~300 bpm  Masimo SpO2: 25 bpm~240 bpm  Comen SpO2: 20 bpm~300 bpm  Nellcor SpO2: resolution: 1 bpm		Masimo SpO2: (Lower alarm limit $+1$ bpm) $\sim$ 240 bpm Comen SpO2: (Lower alarm limit $+1$ bpm) $\sim$ 254 bpm
resolution	Masimo SpO2: resolution: 1 bpm Comen SpO2: resolution: 1 bpm	Lower PR alarm limit	Nellcor SpO2: 25bpm $\sim$ (Upper alarm limit -1bpm)
PR measurement Nellcor SpO2: 20 bpm $\sim$ 250 bpm: the measurement error should be $\pm$ 3 bpm; 251 $\sim$ 300 bpm: measurement accuracy is not		Masimo SpO2: 25bpm $\sim$ (Upper alarm limit -1bpm) Comen SpO2: 20bpm $\sim$ (Upper alarm limit -1bpm)	
	defined.	Trend	
	Masimo SpO2: the measurement	Туре	Tabular, Graphic
error should be ±3 bpm (during non-motion state) and ±5 bpm (during motion state)  Comen SpO2: the measurement error should be ±2 bpm	non-motion state) and ±5 bpm (during motion state) Comen SpO2: the measurement	Length Content	72 hours  Monitor Parameters, Setting  Parameters (Setting Ventilation mode and Parameters) includes parameter alarm events and
modul	Nellcor SpO2: / (Note: Nellcor SpO2 module has no perfusion index.)		parameter waveforms related to the alarm time
	Masimo SpO2: $0.02\%\sim20\%$ , the accuracy is not defined. Comen SpO2: $0.05\%\sim20\%$ , the accuracy is not defined.	Data Review Event logs	Up to 8000 event logs can be saved, including alarm logs and operation logs. The alarm log
Data update period	≤ 2 s		includes parameter alarm events
Signal Quality Index (SIQ) indication function	Masimo SpO2 and Comen SpO2 should come with SIQ indication function should conform to the requirements of YY0784-2010		and parameter waveforms related to the alarm time.
Regulatory compliance		Freeze the waveform review	Freeze the waveform of the interface at the current time and use the knob to review the data.
Upper SpO2 alarm limit	Nellcor SpO2: (Lower alarm limit +1%)~100% Masimo SpO2: (Lower alarm limit		When freezing, 30 most recent historical waveforms can be



	reviewed by sliding the screen or
	rotating the knob.
Freeze the loop review	Up to 5 reference loops can be saved.
O <sub>2</sub> Therapy	
O₂ %	15 - 100 % (increments of 1 %) ± (3 vol.% +1 % of setting)
Flow	2 - 60 L/min ± (1.5 L/min +10 % of setting) (BTPS)
Gas Circuit Specification	
Gas type	Air, O2
Gas source requirement	Medical compressed oxygen
High-pressure O2 source	
Gas source pressure ange	280∼600 kPa
Rated flow rate requirement	120 L/min
nput connector	NIST (ISO 5356-1) or DISS (CGA 1240)
tandards compliant	YY/T 0799-2010 EN ISO5359:2008
Low-pressure O2 source	
nput pressure range	< 100 kPa
laximum flow rate	15 L/min
put connector	CPC quick coupling
nspiratory module	
eak flow rate	≥ 210 L/min
ebulizer connector	Flow rate: 5 L/min∼8 L/min
afety pressure of espiration	≤ 12.5 kPa
nspiratory-side external	Coaxial 22 mm/15 mm conical
onnector	connector
emovable, sterilizable	can be entirety removed quickly;
	and can be entirety cleaned and
	disinfected.
egulatory compliance	YY1040.1-2003
	EN ISO5356-1:2004
Expiratory module	
xpiratory-side external	Coaxial 22 mm/15 mm conical
onnector	connector
Removable, sterilizable	can be entirety removed quickly;

	and can be entirety cleaned and
	disinfected.
Regulatory compliance	YY1040.1-2003
	EN ISO5356-1:2004
System compliance and i	resistance
Compliance	Adult disposable circuit (including
	inspiratory safety valve, adult
	disposable breathing tube, water
	collection cup, expiratory valve): ≤
	4 mL/cmH2O;
	Adult reusable circuit (including
	inspiratory safety valve, adult
	reusable breathing tube, water
	collection cup, expiratory valve, Y-
	joint): ≤ 2 mL/cmH2O;
	Pediatric disposable circuit
	(including inspiratory safety valve,
	pediatric disposable breathing
	tube, water collection cup,
	expiratory valve): ≤ 2 mL/cmH2O;
	Pediatric reusable circuit (including
	inspiratory safety valve, pediatric
	reusable breathing tube, water
	collection cup, expiratory valve, Y-
	joint): ≤ 2 mL/cmH2O;
	Neonate reusable circuit (including
	inspiratory safety valve, neonate
	reusable breathing tube, water
	collection cup, expiratory valve, Y-
	joint): ≤ 1 mL/cmH2O.
nspiratory resistance	≤ 6 cmH2O at the flow rate of 60
	L/min (Adult);
	≤ 6 cmH2O at the flow rate of 30
	L/min (Pediatric);
	≤ 6 cmH2O at the flow rate of 5



	L/min (Neonate).	External AC power su	pply
Expiratory resistance	≤ 6 cmH2O at the flow rate of 60	Input voltage	100 - 240 V
	L/min (Adult);	Input frequency	50/60 Hz
	≤ 6 cmH2O at the flow rate of 30	Input current	1.2 – 0.5 A
	L/min (Pediatric);	Fuse	T3AL/250 V
	≤ 6 cmH2O at the flow rate of 5	External DC power su	pply
	L/min (Neonate).	Input voltage	12V
Basic performance		Input current	10A
Pressure monitoring	-20~120 cmH2O	Internal battery	
range		Number of batteries	One or Two
Safety pressure of	In ventilation state: ≤ 125cmH2O	Battery type	Build-in Lithium-ion battery, 14.4
system	In non-ventilation state or		VDC, 6700mAh
	power failure or	Battery life	140 min (when a new fully charged
	gas source failure (<0.12MPa): ≤ 14		battery is used in standard operating mode)
	cmH2O		280 min (when two new fully
<b>Environmental specifi</b>	ications		
Temperature	5 - 40 °C (operating); -20 to 60 °C		charged batteries are used in
	(storage and transport, O₂ sensor: -		standard operating mode)
	20 to 50°C)	1/0	
Relative Humidity	5 - 95 % (operating); 5 - 95 %	Communication	Rs232, Ethernet, VGA, USB port,
Barometric Pressure	(storage and transport) 62 - 106 kPa (operating); 50 -106	interface	Nurse call
barometric Pressure	kPa (storage and transport)	Trolley MC100	
Power Specification		Dimensions	
		Weight	20 kg

\*Notice: Specifications subject to changes without prior notice. All rights reserved by Comen