

# SRI TRANG GLOVES (THAILAND) PUBLIC COMPANY LIMITED

10 Soi 10 Phetkasem Rd., Hatyai, Songkhla, Thailand, 90110  
Tel. (66) 074-344-663 Fax. (66) 074-344-677

SCT.PD.SP.05.483-220421 R.04  
PRODUCT SPECIFICATION SHEET : NOOOGF-E-EU-M-NS

DAR-SCT-2021-0558  
PAGE 1 of 1

## GLOVE SPECIFICATION

**Product Specification Group Code :** NOOOGF-E-EU-M-NS  
**Product Code :** DNOFEOG (With Color)  
**Product name :** Nitrile Examination Gloves, Powder Free, Chlorinated, Non Sterile  
**Grade :** EU Spec, Medical Grade

### GENERAL INFORMATION

**Type :** Single-use disposable examination or protective glove  
**Marking :** Gloves are not marked to designated size.  
**Shape :** Ambidextrous  
**Material :** Nitrile Butadiene Rubber  
**Cuff :** Rolled cuff  
**Surface :** Textured at Finger Tip

<b>Thickness :</b> (single wall according to ASTM D6319, ASTM D3767)	<b>Finger :</b> Min. 0.05mm	<b>Palm :</b> Min. 0.05mm
---	-----------------------------	---------------------------

<b>Residual powder :</b> (according to EN 455-3, EN ISO 21171)	Max. 2.0mg/glove
--	------------------

<b>Force at break :</b> (according to EN 455-2)	Before Aging : Med. $\geq$ 6.0N After Aging : Med. $\geq$ 6.0N
--	---

<b>Length :</b> (according to EN 455-2)	XS-XL : Med. $\geq$ 240mm
--	---------------------------

<b>Tensile Strength :</b> (according to ASTM D6319, ASTM D412, ASTM D573)	Before Aging : Min. 14Mpa After Aging : Min. 14Mpa
---	---


<b>% Elongation :</b> (according to ASTM D6319, ASTM D412, ASTM D573)	Before Aging : Min. 500% After Aging : Min. 400%
---	---

<b>Available Sizes &amp; Palm width (mm) :</b> XS (5-6) : $\leq$ 80, S (6-7) : 80+/-10, M (7-8) : 95 +/-10, L (8-9) : 110+/-10, XL (9-10) : $\geq$ 110 (according to EN 455-2)
---

**Storage:** 3 years storage life (date of expiration see on the box). Keep area cool, dry and dust free, avoid ventilation and storage close to photocopy equipment. Copper-ions discolor the glove. Protect gloves against ultraviolet light sources, as sunlight and oxidizing agents. Storage above 30°C will lead to accelerated aging and should be avoided.

### STANDARDS, GUIDELINES AND QUALITY CERTIFICATION

(Every mentioned standard is used in the latest edition.)

<b>Certification :</b>	ISO 9001, ISO 13485		
<b>Conformity to guidelines :</b>	Medical Device Regulation (EU)2017/745 : Class I under Rule1 and Rule 5 according to Annex VIII (Subject to EU Declaration of Conformity to brand owner) Personal Protective Equipment Regulation (EU) 2016/425 : Category I or Category III (Subject to EU Type-Examination Certificate)		
<b>Conformity to standards :</b>	<u>Medical</u> EN 455 ASTM D6319	<u>PPE</u> CAT I EN 420	<u>PPE</u> CAT III EN 420 EN 374 EN 16523-1
<b>Conformity to food regulation :</b>	1935/2004		

### WARNING AND REPORTING SYSTEM

**Potentially allergenic ingredients :** Dithiocarbamate type.  
(For further information, the actual version of the ingredients list is given on request.)

Please report any incidents with the products to [STGT\\_TA@sritranggroup.com](mailto:STGT_TA@sritranggroup.com) and [regulatory@sritranggroup.com](mailto:regulatory@sritranggroup.com) without undue delay. Our Product Safety Officers will review all information provided and might contact the original informant for further clarification.

# SRI TRANG GLOVES (THAILAND) PUBLIC COMPANY LIMITED

10 Soi 10 Phetkasem Rd., Hatyai, Songkhla, Thailand, 90110  
Tel. (66) 074-344-663 Fax. (66) 074-344-677

SCT.PD.SP.05.483-220421 R.04

PRODUCT SPECIFICATION SHEET : NOOOGF-E-EU-M-NS

DAR-SCT-2021-0558

Appendix of specification

## APPENDIX : 4

**Product Specification Code :** NOOOGF-E-DOB-EU-MC-NS  
**Product Code :** DNOFEOG DARK OCEAN BLUE  
**Material :** Nitrile Butadiene Rubber  
**Treatment and Feature :** Powder Free, Online Chlorinated

## SAMPLING, INSPECTION AND FINAL RELEASE INFORMATION

Product Specification Code	Glove color	Characteristics			Test-Method
		Barrier Defects*	Major Visual Defects*	Minor Defects*	
NOOOGF-E-DOB-EU-MC-NS	DARK OCEAN BLUE	AQL ≤ 1.5	AQL ≤ 1.5	AQL ≤ 4.0	EN 455-1

### **Barrier Defects :**

Inspection level G I (ISO 2859-1)

Holes that results in the malfunction of glove as a protection barrier either physically or not physically affect the appearance.

### **Major Visual Defects :**

Inspection level G I (ISO 2859-1)

Defined defect that disrupt the glove's function or defect that cause an uncomfortable feeling to the user.

### **Minor Defects :**

Inspection level G I (ISO 2859-1)

Visual defects aggregated are non-conformities of lower degree of concern, which do not prevent correct use of gloves. Minor cosmetic defect which does not affect on the glove's performance

Remark :

\* Defect Characteristics following In-house practice.