

MBD-BT-SPEC-0130

Page 1 of 4

OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION COLUMBIA BLOOD AGAR BASE (CM0331)

COLUMBIA BLOOD AGAR BASE	CM0331	
Typical Formula*		
Special peptone	grams per litre	23.0
Soluble starch		1.0
Sodium chloride		5.0
Agar		10.0

^{*} adjusted as required to meet performance standards

Directions

Suspend 39g in 1 litre of distilled water. Bring to the boil to dissolve completely. Sterilize by autoclaving at 121°C for 15 minutes. Cool to 50°C. Mix well and pour into sterile Petri dishes. For blood agar, enrich with 5% v/v sterile defibrinated blood.

Physical Characteristics

Straw, free-flowing powder
Colour on reconstitution - straw 2-3
Moisture level - less than or equal to 7%
pH 7.3 ± 0.2 at 25°C
Clarity - clear
Gel strength - firm, comparable to 10.0g/litre of agar

The medium is tested for compatibility using 7% v/v oxalated horse blood, defibrinated horse blood and defibrinated sheep blood. There shall be no evidence of lysis or darkening, after incubation at 37°C , 25°C and 4°C for 72 hours.

Microbiological Tests Using Optimum Inoculum Dilution

Control Medium: Columbia Blood Agar Base

Plain plates

Reactions after incubation at 37°C for 24 hours

Medium is challenged with 10-100 colony-forming units

Staphylococcus aureus

ATCC®25923

1-2mm cream colonies



MBD-BT-SPEC-0130

Page 2 of 4

OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION COLUMBIA BLOOD AGAR BASE (CM0331)

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium.

Reactions after incubation at 37°C for 48 hours under anaerobic conditions

Medium is challenged with 10-100 colony-forming units

Clostridium sporogenes ATCC®19404 1-2mm pale straw colonies Clostridium sporogenes ATCC®11437 1-2mm pale straw colonies

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium.

Enriched with 5% v/v horse blood

Reactions after incubation at 37°C for 24 hours

Medium is challenged with 10-100 colony-forming units

Streptococcus pyogenes ATCC $^{\circ}$ 19615 0.25-1mm pale straw colonies, β haemolysis Streptococcus pneumoniae ATCC $^{\circ}$ 6305 0.5-1mm grey/green colonies, α haemolysis

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium.

Reactions after incubation at 37°C for 48 hours under microaerophilic conditions

Neisseria gonorrhoeae NCTC11148 1-2mm grey/brown colonies

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium.

Reactions after incubation at 37°C for 18 hours

Zones of growth/no growth surrounding X, V and X+V factor discs (DD0003, DD0004 and DD0005) when plain plates are inoculated with the following organisms and incubated at 37°C for 18 hours:

		X	V	X+V
Haemophilus influenzae	ATCC®49247	0	0	≥ 15mm
Haemophilus parainfluenzae	ATCC®33392	0	≥ 20mm	≥ 20mm



MBD-BT-SPEC-0130

Page 3 of 4

OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION COLUMBIA BLOOD AGAR BASE (CM0331)

Reactions after incubation at 37°C for 18 hours

Zones of inhibition with Bacitracin discs (DD0002) shall be 10-20mm when 7% v/v horse blood plates are inoculated with *Streptococcus pyogenes* ATCC®19615 and incubated at 37°C for 18 hours.

Testing performed in accordance with ISO11133:2014

Enriched with 5% v/v sheep blood

Reactions after incubation at 41.5 \pm 2°C for 44 \pm 4 hours under microaerophilic conditions

Medium is challenged with 50-120 colony-forming units

Campylobacter jejuni	ATCC®29428	WDCM00156	1-3mm grey, mucoid colonies
Campylobacter jejuni	ATCC®33291	WDCM00005	1-3mm grey, mucoid colonies
Campylobacter coli	ATCC®43478	WDCM00004	1-3mm grey, mucoid colonies

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium.

Testing performed in accordance with current CLSI M22 A

Enriched with 5% v/v sheep blood

Reactions after incubation at 35°C for 18-24 hours

Medium is challenged with 1E+02 to 1E+04 colony-forming units

Streptococcus pyogenes	ATCC®19615	0.25-1mm pale straw colonies, β haemolysis
Streptococcus pneumoniae	ATCC®6305	0.5-1mm grey/green colonies, α haemolysis
Staphylococcus aureus	ATCC®25923	1-2mm cream colonies
Escherichia coli	ATCC®25922	1-3mm cream colonies



MBD-BT-SPEC-0130

Page 4 of 4

OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION COLUMBIA BLOOD AGAR BASE (CM0331)

Revision History

Section / Step	Description of Change	Reason for Change	Reference
Entire document	Update to new document format and correction of minor/typographical errors	N/A	N/A
Microbiological characteristics	Removal of obsolete BSAC testing	Change control	MOC-2022-0741