

Specificații tehnice

[Acest tabel va fi completat de către ofertant în coloanele 2, 3, 4, 6, 7, iar de către autoritatea contractantă – în coloanele 1, 5,]

Numărul procedurii de achiziție: ocds-b3wdp1-MD-1658687141472 / 21060477 din: 24.07.2022 conform SIARSAP Mtender
Obiectul achiziției: Echipamente medicale

Nr.	Denumirea bunurilor/ serviciilor	Denumirea modelului/ bunului/ serviciului	Țara de origine	Producătorul	Specificarea tehnică deplină solicitată de către autoritatea contractantă	Specificarea tehnică deplină propusă de către ofertant	Standarde de referință
	Bunuri/servicii						
1	Disposable FFP2/N95 NR Filtering half Mask with Exhalation Valve, (Particulate respirators), to protect against particles.	E-830V	China	Weini Technology	<ul style="list-style-type: none"> •To be used for medical purposes for the protection of personnel working in Tuberculosis laboratories; •Can help protect against airborne micro-organisms such as mycobacterium tuberculosis, and SARS; •Stretchable headbands easy to put on and provide a secure fit; <ul style="list-style-type: none"> •Adjustable nose clip; •Protection Level – FFP2 (N95); minimum 94-95% according to “Chapter 7.2, WHO Tuberculosis Laboratory Biosafety Manual. (the Year 2012)”. •Assigned Protection Factor - min. 10; •Includes an exhalation valve fitted to the front of the respirator which minimizes heat and moisture build-up. •At least 3 samples of the offered product to conduct Fit Test (sealing property test). Will be considered only the products that pass the Fit Test. 	<ul style="list-style-type: none"> •To be used for medical purposes for the protection of personnel working in Tuberculosis laboratories; •Can help protect against airborne micro-organisms such as mycobacterium tuberculosis, and SARS; •Stretchable headbands easy to put on and provide a secure fit; <ul style="list-style-type: none"> •Adjustable nose clip; •Protection Level – FFP3; minimum 94-95% according to “Chapter 7.2, WHO Tuberculosis Laboratory Biosafety Manual. (the Year 2012)”. •Assigned Protection Factor - min. 10; •Includes an exhalation valve fitted to the front of the respirator which minimizes heat and moisture build-up. •At least 3 samples of the offered product to conduct Fit Test (sealing property test). Will be considered only the products that pass the Fit Test. 	CE

Semnat: _____

Numele, Prenumele Jighili Tatiana

În calitate de: administrator

Ofertantul: SRL Triumf-Motiv

Adresa: str. Grenoble 193 of 1301