


Certificate of Analysis

r-biopharm®



RIDASCREEN® Aflatoxin Total

REF Art. No.	R4701
LOT Lot	24155
Expiry	2026-11-06



	Lot. No.	Expiry		Lot. No.	Expiry
Microtiter plate M	21454	2027-02-03	Stop solution	21114	2029-03-10
Standards	25455	2026-11-06			
Wash buffer salt Tween	21422	2027-09-30			
Conjugate	21145	2027-03-30			
Antibody	21145	2027-03-30			
Substrate/Chromogen (Red Chromogen Pro)	25334	2027-07-31			

IFU (and other accompanying documents where applicable)

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Please note: The absorbance for the standards may decrease during the shelf life of the kit. The general shape of the curve will remain similar, while the slope might change slightly. Furthermore refer to product leaflet section 8. Indication of instability or deterioration of reagents.

Quality Control Department

Date: 2025-11-10



Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com