

POP-4™, POP-6™, and POP-7™ Polymer for 3500/3500xL Genetic Analyzers

Catalog Number A26070, 4393715, 4393710, A26071, 4393717, 4393712, A26073, 4393708, 4393714

Pub. No. 4408234 Rev. F

Item	Cat. no. (96 samples)	Cat. no. (384 samples)	Cat. no. (960 samples)	Storage (for all polymers)
POP-4™ Polymer	A26070	4393715 ^[1]	4393710 ^[1]	2°C to 8°C
POP-6™ Polymer	A26071	4393717	4393712	
POP-7™ Polymer	A26073	4393708	4393714	

^[1] The polymer has been validated for HID applications.

WARNING! Read the Safety Data Sheets (SDSs) and follow the handling instructions. Wear appropriate protective eyewear, clothing, and gloves. Safety Data Sheets (SDSs) are available from thermofisher.com/support.

Product description

The polymer is the separation matrix for capillary electrophoresis. It is supplied as a ready-to-use pouch with a radio frequency identification (RFID) tag incorporated into the label. The instrument uses the RFID tag to track polymer usage and expiration.

Expiration date/on-instrument supported limits

The on-instrument life is determined by the limit that is reached first—number of days after first installation, samples run, injections performed, or expiry date. Usage is tracked by the system software.

IMPORTANT! The usage limits are determined by the system software. The limits shown below are for Data Collection Software v3.1. If you are running v3.0 or earlier, refer to the user guide provided with the software or instrument.

IMPORTANT! For the POP-4™ and POP-7™ Polymers (Cat. nos. A26070, 4393715, 4393710, A26073, 4393708, and 4393714), the on-instrument supported limit is 14 days only when the instrument operating temperature is 15 to ≤ 25°C. When the instrument operating temperature is > 25°C, the supported limit is 7 days.

For the POP-6™ Polymers (Cat. nos. A26071, 4393717, and 4393712), the on-instrument supported limit is 14 days when the instrument operating temperature is 15 to 30°C.

Pouch size	Instrument	On-instrument supported limits ^[1] Lower of:	Guidelines
96 samples	8-capillary	14 days, 96 samples, 12 injections, or expiry date	The polymer has been verified for use for up to 14 days on the instrument. The software displays a warning message when a usage limit is met and allows you to continue running. Before doing so, see "Important notice regarding use of consumables that exceed supported limits" on page 2.
	24-capillary	14 days, 96 samples, 5 injections, or expiry date	
384 samples	8-capillary	14 days, 384 samples, 60 injections, or expiry date	
	24-capillary	14 days, 384 samples, 20 injections, or expiry date	
960 samples	8-capillary	14 days, 960 samples, 120 injections, or expiry date	
	24-capillary	14 days, 960 samples, 50 injections, or expiry date	

^[1] The pouch has adequate polymer to support the stated number of samples or injections, plus additional volume to accommodate installation and wizard operations. Multiple pouch installations and/or excessive use of wizards reduce the number of remaining samples and injections. For example, if you run the **total bubble remove** option in the Remove Bubbles wizard more than four times, the number of remaining samples and injections is reduced.

Precautions for use

- Do not reuse a polymer pouch that has been installed on another type of instrument. For example, if you remove a partially used polymer pouch from an 8-capillary instrument, do not reuse that polymer on a 24-capillary instrument.
- If you remove a polymer pouch for storage (2–8°C), place a pouch cap (Cat. no. 4412619) onto the pouch, then place an empty pouch (or conditioning reagent) on the connector to prevent desiccation of any residual polymer on the connector. Follow the instructions in the wizard to ensure proper operation of the pouch and the instrument.

Replenish polymer or change polymer type

1. Check the expiration date on the label to ensure that the polymer is not expired and will not expire during intended use.

IMPORTANT! Do not use if the product is expired, if the pouch or label is damaged, or if the top seal is missing or damaged.

2. Allow the refrigerated polymer to equilibrate to ambient temperature (15–30°C) before use.
3. In the Dashboard, click **Wizards**, then click **Replenish Polymer** (requires 10 to 20 minutes) or **Change Polymer Type** (requires 60 to 70 minutes).
4. Follow the prompts in the Wizard window.
5. When instructed to install the polymer, peel off the seal at the top of the pouch fitment.
Note: You may notice a tiny droplet of polymer inside the fitment (residual from the pouch filling process). This is *not* expected to cause any performance issues.
6. With the RFID label *facing* the instrument, slide the pouch fitment onto the slot of the lever assembly. Push the lever up to snap the pouch into the connector end of the instrument pump.
Note: The RFID label must face the instrument (away from you) to ensure that the RFID information is read accurately by the instrument.
7. In the Dashboard, click **Refresh**, then check the Quick View section for the updated polymer status.

Refer to the instrument user guide for instructions on initiating the runs.

Important notice regarding use of consumables that exceed supported limits

BEFORE DISMISSING THE WARNING THAT THE CONSUMABLES HAVE REACHED SUPPORTED LIMITS AND CONTINUING WITH OPERATION OF THE INSTRUMENT, PLEASE READ AND UNDERSTAND THE FOLLOWING IMPORTANT NOTICE AND INFORMATION:

Life Technologies does not recommend the use of consumables that exceed supported limits. The recommended limits are designed to promote the production of high quality data and minimize instrument downtime. Reagent and consumable lifetime minimum performance are based on testing and studies that use reagents and consumables that have not exceeded supported limits.

The use of consumables beyond the supported limits may impact data quality or cause damage to the instrument or capillary array. The cost of repairing such damage is **NOT** covered by any Life Technologies product warranty or service plan. Customer use of expired consumables is at customer's own risk and without recourse to Life Technologies. For example, product warranties do not apply to defects resulting from or repairs required due to misuse, neglect, or accident including, without limitation, operation outside of the environmental or use specifications or not in conformance with Life Technologies instructions for the instrument system, software, or accessories.

Please see your specific service contract or limited product warranty for exact language regarding coverage and ask your Life Technologies representative if you have further questions.

Limited product warranty

Life Technologies Corporation and/or its affiliate(s) warrant their products as set forth in the Life Technologies' General Terms and Conditions of Sale found on Life Technologies' website at www.thermofisher.com/us/en/home/global/terms-and-conditions.html. If you have any questions, please contact Life Technologies at www.thermofisher.com/support.

The information in this guide is subject to change without notice.

DISCLAIMER

TO THE EXTENT ALLOWED BY LAW, LIFE TECHNOLOGIES AND/OR ITS AFFILIATE(S) WILL NOT BE LIABLE FOR SPECIAL, INCIDENTAL, INDIRECT, PUNITIVE, MULTIPLE, OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH OR ARISING FROM THIS DOCUMENT, INCLUDING YOUR USE OF IT.

Limited Use Label License No. 358: Research Use Only: Notice to Purchaser: The purchase of this product conveys to the purchaser the limited, non-transferable right to use the purchased amount of the product only to perform internal research for the sole benefit of the purchaser. No right to resell this product or any of its components is conveyed expressly, by implication, or by estoppel. This product is for internal research purposes only and is not for use in commercial applications of any kind, including, without limitation, quality control and commercial services such as reporting the results of purchaser's activities for a fee or other form of consideration. For information on obtaining additional rights, please contact outlicensing@lifetech.com or Out Licensing, Life Technologies Corporation, 5791 Van Allen Way, Carlsbad, California 92008.

Limited Use Label License No. 481: Sequencing or Fragment Analysis Intellectual Property: Notice to Purchaser: This product is optimized for use in the DNA sequencing or fragment analysis methods covered by patents owned and/or controlled by Life Technologies Corporation ("LTC"). LTC does not convey any right or license under these patents, whether expressly, by implication, by estoppels, or otherwise, to the purchaser by the purchase of this product to use the DNA sequencing or fragment analysis methods. Notwithstanding the foregoing, a limited license to use the DNA sequencing or fragment analysis methods covered by such patents can be obtained for certain research and development activities (a) through the purchase of certain LTC reagents when such reagents are used in conjunction with an authorized LTC instrument, or (b) directly from LTC. For information on obtaining additional rights to practice the DNA sequencing or fragment analysis methods, please contact outlicensing@lifetech.com or Out Licensing, Life Technologies Corporation, 5791 Van Allen Way, Carlsbad, California 92008.

Corporate entity: Life Technologies | Carlsbad, CA 92008 USA | Toll Free in USA 1.800.955.6288

©2015 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific and its subsidiaries unless otherwise specified.

For support visit thermofisher.com/support or email techsupport@lifetech.com

thermofisher.com

ThermoFisher
SCIENTIFIC