

Manufacturer: Zhenjiang Huayang Latex Products Co., Ltd. Adress: Lianhe Village, Xinba Town, Yanzhong City, Jiangsu

Province, 212212, P.R. China

Product Technical Specifications

Sterile Latex Surgical Gloves
Reference Number: SCMDG001



Manufacturer: Zhenjiang Huayang Latex Products Co., Ltd. Adress: Lianhe Village, Xinba Town, Yanzhong City, Jiangsu Province, 212212, P.R. China

1. Types/specifications of sterile latex surgical gloves and instructions of its classification

1.1 Sterile Latex Surgical Gloves

1.2 Sizes:

6# 6.5# 7# 7.5# 8# 8.5# 9#

1.3 Types and specifications of Sterile Latex Surgical Gloves should meet the requirements of sheet 1

Sheet 1 Surgical Gloves specifications and basic sizes

Order	6	6.5	7	7.5	8	8.5	9	Franchise
A	76	83	89	95	102	108	114	±6mm
В	0.13	0.13	0.13	0.13	0.13	0.13	0.13	min
С	265	265	265	265	265	265	265	min


1.4 Powder:

Powdered or Powder-free



Manufacturer: Zhenjiang Huayang Latex Products Co., Ltd. Adress: Lianhe Village, Xinba Town, Yanzhong City, Jiangsu

Province, 212212, P.R. China





Manufacturer: Zhenjiang Huayang Latex Products Co., Ltd. Adress: Lianhe Village, Xinba Town, Yanzhong City, Jiangsu Province, 212212, P.R. China

2. Material Descriptions

Sterile latex surgical gloves are mainly made of natural rubber.

All material information--Natural rubber.

3. Requirement of the Product appearance

3.1 **Product appearance**

Sterile latex surgical gloves have no obvious stains, holes and edges.

3.2Main Product Performance

	Requirement	Results
BEFORE AGING TENSILE	≥ 24	≥ 29.2
STRENGTH Mpa		
BEFORE AGING ULTIMA	≥ 750	≥ 835
ELONGATION %		
AFTER ACCELERATED AGING	≥ 18	≥ 22.3
TENSILE STRENGTH		
AFTER ACCELERATED ULTIMATE	≥ 560	≥ 645
ELONGATION		

3.3 Standards:

EN 455-1 EN 455-2, EN 455-3, EN 455-4, ASTM D3577



Manufacturer: Zhenjiang Huayang Latex Products Co., Ltd. Adress: Lianhe Village, Xinba Town, Yanzhong City, Jiangsu Province, 212212, P.R. China

3.4 Sterile

Sterile Latex Surgical Gloves should be supplied aseptically and should be made sterile by an sterilization process .

Sterilization - gamma rays

4. Package

pairs of surgical+1 inne paper+1 pouch /paper 50pairs pouch/1paper box ,10 paper boxes/paper carton.

5. Certificates:

ISO 13485, CE

6.Valid day

Under the specified storage conditions, the period of validity is 3 years from the date of production.