

Technical Specification

Product Name	Transfer pipette			
Code	Length	Drop	Material	Package
MF7110135055S	350mm, 0.5ml	55ul	PE	Individual pack, 5000pcs/ctn
Specification	0.5ML			
Items of inspection	Technical requirements			
Outward appearance	Translucent and smooth. Enough intensity.			
Size	Should conform to the requirements of drawings			
Leakage test	No leaking hole and crack. No leakage after watering.			
Diaphaneity	Water line in the container should be clear by eyes under natural light.			
Sterilization	Yes, Ethylene Oxide			

Ningbo 2022-12-22

Place, date
and function

宁波蓝柏医疗器械有限公司
NINGBO MFLAB MEDICAL INSTRUMENTS CO.,LTD

Yazeng Li General Manager

Name

李雅增