


Test report No. id4019

EVALUATION OF BACTERICIDAL ACTIVITIES OF A DISINFECTANTS AND ANTISEPTICS USED IN THE MEDICAL AREA (EN 13727)

Name of the product: STERISEPT INSTRU
Batch number: 211310518
Order number: 18014
Manufacturer: Chemi-Pharm Ltd
Client, representative: Chemi-Pharm Ltd., Põllu 132, Tallinn, 10917, ESTONIA
Maris Millner, +372-51-77-090
Date of delivery: 04.10.2018
Test material conditions: No specific features, sample in the manufacturers tare
Storage conditions: In room temperature, dark
Active substance – conc.: N-(3-aminopropyl)-N-dodecylpropane-1,3-diamine 15%; Didecyl-Dimethyl-Ammonium Chloride (DDAC) 15%
Appearance of the product: Transparent liquid
Test concentration: 0.25%; 0.50%
Contact time: 15 min for 0.25%; 5 min, 10 min and 15 min for 0.50%
Interfering substance: 3.0 g/l bovine albumin + 3ml/l sheep blood erythrocytes = dirty conditions
Neutralizer: -
Rinsing liquid: Tryptone 1 g/l + NaCl, 9 g/l
Test organisms: *Pseudomonas aeruginosa* ATCC 15442
Staphylococcus aureus ATCC 6538
Enterococcus hirae ATCC 10541
Staphylococcus aureus MRSA ATCC 33592
Enterococcus faecium VRE ATCC 700221
Testing method: EVS-EN 13727:2012+A2:2015
Chemical disinfectants and antiseptics - Quantitative suspension test for the evaluation of bactericidal activity in the medical area - Test method and requirements (phase 2, step 1)
Testing date: 26.11.2018 – 27.11.2019
Results: look appendix 1-6



Allar Laaneleht
Chief specialist

Date of test report: 15.01.2019

Appendix 1

TEST RESULTS (bactericidal suspension test)

EVS-EN 13727:2012+A2:2015; Phase 2, step 1;

Membrane filtration method;

Rinsing liquid: Tryptone 1 g/l + NaCl 9 g/l;

Test organism: *Staphylococcus aureus* ATCC 6538;

Test temperature: +20° C; Incubation temperature: +37 °C

Interfering substance: 3.0 g/l bovine albumin + 3ml/l sheepblood erythrocytes = dirty conditions;

Nordic Tersus Laboratory LLC.;

Date of test: 26.11.2018

Responsible person: Allar Laaneleht

Validation and controls

Dirty conditions

Validation suspension N_{vo}			Experimental conditions control (A)			Filtration control (B)			Method validation (C)		
V_{C1}	49	$\bar{x} = 45.5$	V_{C1}	36	$\bar{x} = 34.5$	V_{C1}	38	$\bar{x} = 36.5$	V_{C1}	32	$\bar{x} = 31$
V_{C2}	42		V_{C2}	33		V_{C2}	35		V_{C2}	30	
$30 \leq \bar{x} N_{vo} \leq 160$? yes X; no ☐			$\bar{x} A \geq 0.5 \bar{x} N_{vo}$? yes X; no ☐			$\bar{x} B \geq 0.5 \bar{x} N_{vo}$? yes X; no ☐			$\bar{x} C \geq 0.5 \bar{x} N_{vo}$? yes X; no ☐		

Test suspension and test

Testsuspension: N and N_0	N	V_{C1}	V_{C2}	$\bar{x}_{wm} = 1.60 \times 10^8$; $\log N = 8.20$ $N_0 = N/10$; $\log N_0 = 7.20$ $7.17 \leq \log N_0 \leq 7.70$; yes X; no ☐
	10^{-6}	155	163	
	10^{-7}	16	18	

Experimental results

Concentration of the product %	Dilution step	V_{C1}	V_{C2}	Na ($=\bar{x} \cdot 10$)	$\log Na$	$\log R$	Contact time	Conditions
0.05%	-	>165	>165	>1650	>3.22	<3.98	15 min	Dirty
0.25%	-	<14	<14	<140	<2.15	>5.05	15 min	Dirty
0.50%	-	<14	<14	<140	<2.15	>5.05	5 min	Dirty
0.50%	-	<14	<14	<140	<2.15	>5.05	10 min	Dirty
0.50%	-	<14	<14	<140	<2.15	>5.05	15 min	Dirty
2.50%	-	<14	<14	<140	<2.15	>5.05	5 min	Dirty

Explanations:

V_C = count per ml (one plate or more)

$\bar{x}$ = average of V_{C1} and V_{C2} (1. + 2. duplicate)

N = cfu/ml microbes in testsuspension

N_0 = cfu/ml at the start of the contact time (t=0)

N_{vo} = cfu/ml in the validation suspension (t=0)

Na = surviving microbes after the test

R = reduction factor ($R = N_0 / Na$; $\log R = \log N_0 - \log Na$)

Appendix 2

TEST RESULTS (bactericidal suspension test)

EVS-EN 13727:2012+A2:2015; Phase 2, step 1;
Membrane filtration method;
Rinsing liquid: Tryptone 1 g/l + NaCl 9 g/l;
Test organism: *Enterococcus hirae* ATCC 10541;
Test temperature: +20° C; Incubation temperature: +37 °C
Interfering substance: 3.0 g/l bovine albumin + 3ml/l sheepblood erythrocytes = dirty conditions;
Nordic Tersus Laboratory LLC.;
Date of test: 26.11.2018
Responsible person: Allar Laaneleht

Validation and controls

Dirty conditions

Validation suspension N_{vo}			Experimental conditions control (A)			Filtration control (B)			Method validation (C)		
V_{C1}	72	$\bar{x} = 67.5$	V_{C1}	62	$\bar{x} = 58.5$	V_{C1}	58	$\bar{x} = 54.5$	V_{C1}	53	$\bar{x} = 49.5$
V_{C2}	63		V_{C2}	55		V_{C2}	51		V_{C2}	46	
$30 \leq \bar{x} N_{vo} \leq 160$? yes X; no ☐			$\bar{x} A \geq 0.5 \bar{x} N_{vo}$? yes X; no ☐			$\bar{x} B \geq 0.5 \bar{x} N_{vo}$? yes X; no ☐			$\bar{x} C \geq 0.5 \bar{x} N_{vo}$? yes X; no ☐		

Test suspension and test

Testsuspension:	N	V_{C1}	V_{C2}	$\bar{x}_{wm} = 1.87 \times 10^8$; $\log N = 8.27$ $N_0 = N/10$; $\log N_0 = 7.27$ $7.17 \leq \log N_0 \leq 7.70$; yes X; no ☐
N and N_0	10^{-6}	179	193	
	10^{-7}	21	18	

Experimental results

Concentration of the product %	Dilution step	V_{C1}	V_{C2}	N_a ($=\bar{x} \times 10$)	$\log N_a$	$\log R$	Contact time	Conditions
0.05%	-	>165	>165	>1650	>3.22	<4.05	15 min	Dirty
0.25%	-	<14	<14	<140	<2.15	>5.12	15 min	Dirty
0.50%	-	<14	<14	<140	<2.15	>5.12	5 min	Dirty
0.50%	-	<14	<14	<140	<2.15	>5.12	10 min	Dirty
0.50%	-	<14	<14	<140	<2.15	>5.12	15 min	Dirty
2.50%	-	<14	<14	<140	<2.15	>5.12	5 min	Dirty

Explanations:

V_C = count per ml (one plate or more)
 $\bar{x}$ = average of V_{C1} and V_{C2} (1. + 2. duplicate)
 N = cfu/ml microbes in testsuspension
 N_0 = cfu/ml at the start of the contact time ($t=0$)
 N_{vo} = cfu/ml in the validation suspension ($t=0$)
 N_a = surviving microbes after the test
 R = reduction factor ($R = N_0 / N_a$; $\log R = \log N_0 - \log N_a$)

Appendix 3

TEST RESULTS (bactericidal suspension test)

EVS-EN 13727:2012+A2:2015; Phase 2, step 1;

Membrane filtration method;

Rinsing liquid: tryptone 1 g/l + NaCl 9 g/l;

Test organism: *Pseudomonas aeruginosa* ATCC 15442

Test temperature: +20° C; Incubation temperature: +37 °C

Interfering substance: 3.0 g/l bovine albumin + 3ml/l sheepblood erythrocytes = dirty conditions;

Nordic Tersus Laboratory LLC.;

Date of test: 26.11.2018

Responsible person: Allar Laaneleht

Validation and controls

Dirty conditions

Validation suspension N_{vo}			Experimental conditions control (A)			Filtration control (B)			Method validation (C)		
V_{C1}	84	$\bar{x} = 79.5$	V_{C1}	58	$\bar{x} = 61$	V_{C1}	67	$\bar{x} = 69$	V_{C1}	72	$\bar{x} = 70$
V_{C2}	75		V_{C2}	64		V_{C2}	71		V_{C2}	68	
$30 \leq \bar{x} N_{vo} \leq 160$? yes X; no ☐			$\bar{x} A$ is $\geq 0.5 \bar{x} N_{vo}$? yes X; no ☐			$\bar{x} B$ is $\geq 0.5 \bar{x} N_{vo}$? yes X; no ☐			$\bar{x} C$ is $\geq 0.5 \bar{x} N_{vo}$? yes X; no ☐		

Test suspension and test

Testsuspension: N and N_0	N	V_{C1}	V_{C2}	$\bar{x}_{wm} = 2.09 \times 10^8$; $\log N = 8.32$ $N_0 = N/10$; $\log N_0 = 7.32$ $7.17 \leq \log N_0 \leq 7.70$; yes X; no ☐
	10^{-6}	204	215	
	10^{-7}	22	19	

Experimental results

Concentration of the product %	Dilution step	V_{C1}	V_{C2}	Na ($=\bar{x} \cdot 10$)	$\log Na$	$\log R$	Contact time	Conditions
0.05%	-	>165	>165	>1650	>3.22	<4.10	15 min	Dirty
0.25%	-	<14	<14	<140	<2.15	>5.17	15 min	Dirty
0.50%	-	<14	<14	<140	<2.15	>5.17	5 min	Dirty
0.50%	-	<14	<14	<140	<2.15	>5.17	10 min	Dirty
0.50%	-	<14	<14	<140	<2.15	>5.17	15 min	Dirty
2.50%	-	<14	<14	<140	<2.15	>5.17	5 min	Dirty

Explanations:

V_C = count per ml (one plate or more)

$\bar{x}$ = average of V_{C1} and V_{C2} (1. + 2. duplicate)

N = cfu/ml microbes in testsuspension

N_0 = cfu/ml at the start of the contact time ($t=0$)

N_{vo} = cfu/ml in the validation suspension ($t=0$)

Na = surviving microbes after the test

R = reduction factor ($R = N_0 / Na$; $\log R = \log N_0 - \log Na$)

Appendix 4

TEST RESULTS (bactericidal suspension test)

EVS-EN 13727:2012+A2:2015; Phase 2, step 1;

Membrane filtration method;

Rinsing liquid: tryptone 1 g/l + NaCl 9 g/l;

Test organism: *Staphylococcus aureus* MRSA ATCC 33592

Test temperature: +20° C; Incubation temperature: +37 °C

Interfering substance: 3.0 g/l bovine albumin + 3ml/l sheepblood erythrocytes = dirty conditions;

Nordic Tersus Laboratory LLC.;

Date of test: 26.11.2018

Responsible person: Allar Laaneleht

Validation and controls

Dirty conditions

Validation suspension N_{vo}			Experimental conditions control (A)			Filtration control (B)			Method validation (C)		
V_{C1}	60	$\bar{x} = 57$	V_{C1}	42	$\bar{x} = 45$	V_{C1}	51	$\bar{x} = 46.5$	V_{C1}	51	$\bar{x} = 55$
V_{C2}	54		V_{C2}	48		V_{C2}	42		V_{C2}	59	
$30 \leq \bar{x} N_{vo} \leq 160$? yes X; no ☐			$\bar{x} A$ is $\geq 0.5 \bar{x} N_{vo}$? yes X; no ☐			$\bar{x} B$ is $\geq 0.5 \bar{x} N_{vo}$? yes X; no ☐			$\bar{x} C$ is $\geq 0.5 \bar{x} N_{vo}$? yes X; no ☐		

Test suspension and test

Testsuspension:	N	V_{C1}	V_{C2}	$\bar{x}_{wm} = 1.79 \times 10^8$; $\log N = 8.25$ $N_0 = N/10$; $\log N_0 = 7.25$ $7.17 \leq \log N_0 \leq 7.70$; yes X; no ☐
N and N_0	10^{-6}	184	175	
	10^{-7}	16	18	

Experimental results

Concentration of the product %	Dilution step	V_{C1}	V_{C2}	Na ($=\bar{x} \cdot 10$)	$\log Na$	$\log R$	Contact time	Conditions
0.05%	-	>165	>165	>1650	>3.22	<4.03	15 min	Dirty
0.25%	-	<14	<14	<140	<2.15	>5.10	15 min	Dirty
0.50%	-	<14	<14	<140	<2.15	>5.10	5 min	Dirty
0.50%	-	<14	<14	<140	<2.15	>5.10	10 min	Dirty
0.50%	-	<14	<14	<140	<2.15	>5.10	15 min	Dirty
2.50%	-	<14	<14	<140	<2.15	>5.10	5 min	Dirty

Explanations:

V_C = count per ml (one plate or more)

$\bar{x}$ = average of V_{C1} and V_{C2} (1. + 2. duplicate)

N = cfu/ml microbes in testsuspension

N_0 = cfu/ml at the start of the contact time ($t=0$)

N_{vo} = cfu/ml in the validation suspension ($t=0$)

Na = surviving microbes after the test

R = reduction factor ($R = N_0 / Na$; $\log R = \log N_0 - \log Na$)

Appendix 5

TEST RESULTS (bactericidal suspension test)

EVS-EN 13727:2012+A2:2015; Phase 2, step 1;

Membrane filtration method;

Rinsing liquid: tryptone 1 g/l + NaCl 9 g/l;

Test organism: *Enterococcus faecium* VRE ATCC 700221

Test temperature: +20° C; Incubation temperature: +37 °C

Interfering substance: 3.0 g/l bovine albumin + 3ml/l sheepblood erythrocytes = dirty conditions;

Nordic Tersus Laboratory LLC.;

Date of test: 26.11.2018

Responsible person: Allar Laaneleht

Validation and controls

Dirty conditions

Validation suspension N_{vo}			Experimental conditions control (A)			Filtration control (B)			Method validation (C)		
V_{C1}	58	$\bar{x} = 54.5$	V_{C1}	40	$\bar{x} = 40$	V_{C1}	36	$\bar{x} = 37.5$	V_{C1}	41	$\bar{x} = 38.5$
V_{C2}	51		V_{C2}	40		V_{C2}	39		V_{C2}	36	
$30 \leq \bar{x} N_{vo} \leq 160$? yes X; no ☐			$\bar{x} A$ is $\geq 0.5 \bar{x} N_{vo}$? yes X; no ☐			$\bar{x} B$ is $\geq 0.5 \bar{x} N_{vo}$? yes X; no ☐			$\bar{x} C$ is $\geq 0.5 \bar{x} N_{vo}$? yes X; no ☐		

Test suspension and test

Testsuspension:	N	V_{C1}	V_{C2}	$\bar{x}_{wm} = 1.75 \times 10^8$; $\log N = 8.24$ $N_0 = N/10$; $\log N_0 = 7.24$ $7.17 \leq \log N_0 \leq 7.70$; yes X; no ☐
N and N_0	10^{-6}	167	181	
	10^{-7}	17	20	

Experimental results

Concentration of the product %	Dilution step	V_{C1}	V_{C2}	Na ($=\bar{x} \cdot 10$)	$\log Na$	$\log R$	Contact time	Conditions
0.05%	-	>165	>165	>1650	>3.22	<4.02	15 min	Dirty
0.25%	-	<14	<14	<140	<2.15	>5.09	15 min	Dirty
0.50%	-	<14	<14	<140	<2.15	>5.09	5 min	Dirty
0.50%	-	<14	<14	<140	<2.15	>5.09	10 min	Dirty
0.50%	-	<14	<14	<140	<2.15	>5.09	15 min	Dirty
2.50%	-	<14	<14	<140	<2.15	>5.09	5 min	Dirty

Explanations:

V_C = count per ml (one plate or more)

$\bar{x}$ = average of V_{C1} and V_{C2} (1. + 2. duplicate)

N = cfu/ml microbes in testsuspension

N_0 = cfu/ml at the start of the contact time ($t=0$)

N_{vo} = cfu/ml in the validation suspension ($t=0$)

Na = surviving microbes after the test

R = reduction factor ($R = N_0 / Na$; $\log R = \log N_0 - \log Na$)

Appendix 6

Interpretation:

The product STERISEPT INSTRU (batch no. 211310518) was tested according to the test method EVS-EN 13727:2012+A2:2015. The test was performed at 20 °C ± 1 °C, under dirty conditions with the contact times of 15 min for 0.25% and 0.50% ; 10 min for 0.50%; 5 min for 0.50%. The membrane filtration method was used for testing the product's effectiveness against the reference strains: *Pseudomonas aeruginosa* ATCC 15442, *Enterococcus hirae* ATCC 10541, *Staphylococcus aureus* ATCC 6538, *Enterococcus faecium* VRE ATCC 700221 and *Staphylococcus aureus* MRSA ATCC 33592. Under dirty conditions the 0.25% solution of the tested product was effective against all the reference strains within 15 min and the 0.50% solution was active within 5 min

Conclusion:

The surviving count of bacterial reference strains showed at least 5lg reduction meaning that under dirty conditions the 0.25% solution of product STERISEPT INSTRU has a bactericidal effect within 15 min the 0.50% solution of the product has a bactericidal effect within 5 min.



.....

Allar Laaneleht

Chief Specialist

15.01.2019