

## EU DECLARATION OF CONFORMITY

Manufacturer Grena Limited  
1000 Great West Road, Brentford, Middlesex, TW8 9HH  
United Kingdom

SRN GB-MF-000007429

Authorised representative MDML INTL LTD.  
Unit 7, Argus House Greenmount Office Park, Harold's Cross  
Road Dublin 6W, DUBLIN, Ireland D6W PP38

SRN IE-AR-000001156

Product LigaV® Ligating Clip Applier

REF number(s)

0301-02L20	0301-02M28	0301-02MLEB
0301-02L28	0301-02ME	0301-02S15
0301-02LE	0301-02MEB	0301-02S185
0301-02LEB	0301-02ML20	0301-02S19
0301-02M15	0301-02ML275A45	0301-02S20
0301-02M185	0301-02ML28	0301-02S28
0301-02M19	0301-02MLEA25	0301-05LLC
0301-02M20	0301-02MLE	

Basic UDI-DI 5060171280204W

Classification Class Ir, Rule 6.2 according to Annex VIII of the (EU) MDR 2017/745

Product intended purpose Grena LigaV® Ligating Clips Appliers are indicated for use as delivery devices for Grena LigaV® titanium ligating clips during laparoscopic and thoracoscopic procedures as well as in open surgery. Conformity of the size of the occluded tissue and the clips is required.

We herewith declare under our sole responsibility that the above mentioned products meet the provisions of the (EU) MDR 2017/745 concerning medical devices which apply to them. The conformity assessment procedure was performed according to the Annex IX, Chapter I, Section 2 and 3 of the Regulation. All supporting documentation is retained under the premises of the manufacturer, including Product Release Reports.

**Standards Applied**

All applicable standards and common specifications required by the (EU) MDR 2017/745. The detailed list in the Medical Device Files.

### CE 0197

Notified body

TÜV Rheinland LGA Products GmbH  
Tillystrasse 2. D-90431 Nürnberg, Germany

EC Approval number

HZ 1554225-1

Brentford, 27.08.2025  
Rev. 06Iwona Michalska  
Quality Director