



Bio-Rad  
Certificate of Analysis

---

Material Description: Buffered Peptone Water Standard, Dehydrated 500 g  
Material Number: 12013259  
Batch Number: 402171  
Manufacture Date: 2023-02-17  
Expiration Date: 2027-02-17

---

<u>Characteristic</u>	<u>Specification</u>		<u>Results</u>
	Lower Limit	Upper Limit	
Labelling	Specifications see above		Conforms
Appearance	Fine powder, white cream-slightly toasted		Conforms
pH	6.8	7.2	7.1
.....			
Productivity Test after 18 ±2h at 37±1 °C			
.....			
Salmonella Typhimurium WDCM_00031	Growth		Conforms
Salmonella Enteritidis WDCM_00030	Growth		Conforms
.....			
Survival Test			
Count at T0 and T45 to 60 min at 20-25 °C			
.....			
Escherichia coli WDCM_00012	Colonies count at T45 to T60 = ± 30% colonies count at T0		24.2%, Conforms



Bio-Rad  
Certificate of Analysis

---

Material Description: Buffered Peptone Water Standard, Dehydrated 500 g  
Material Number: 12013259  
Batch Number: 402171  
Manufacture Date: 2023-02-17  
Expiration Date: 2027-02-17

---

<u>Characteristic</u>	<u>Specification</u>		<u>Results</u>
	Lower Limit	Upper Limit	
Staphylococcus aureus WDCM_00034  ..... Count at T0 and T55 to 65 min at 18-22°C  .....	Colonies count at T45 to T60 = ± 30% colonies count at T0		-7.7%, Conforms
Listeria monocytogenes 4b_WDCM_00021	Colonies count at T55 to T65 = ± 30% colonies count at T0		2%, Conforms

Controlled according to EN ISO 11133

Written by: Magali Brossier

Date/Time: 2023-03-06

Verified and approved by an  
authorized person:

Date/Time: 2023-03-06



Nicolas Sauvage