

# Declaration of Conformity



**Manufacturer:** Shenzhen Mindray Bio-Medical Electronics Co., Ltd.  
Mindray Building, Keji 12th Road South, Hi-tech Industrial  
Park, Nanshan, Shenzhen, 518057, P. R. China

**EC-Representative:** Shanghai International Holding Corp. GmbH (Europe)  
Eiffestraße 80  
20537 Hamburg, Germany

**Product:** **Auto Hematology Analyzer**

**Model:** **BC-20s**  
Including reagents as following:  
**M-30D DILUENT**  
**M-30CFL LYSE**  
**PROBE CLEANSER**

**Classification:** The device not in IVDD annex II and not for self  
testing/performance evaluation

**Conformity Assessment Route:** IVDD Annex III(excluding Section 6)

**We herewith declare that the above mentioned products meet the provisions of the Directive 98/79/EC on In Vitro Diagnostic Medical Devices. All supporting documentations are retained under the premises of the manufacturer.**

**Standards Applied:**

List of (harmonized) standards for which documented evidence for compliance can be provided as attachment.

**Start of CE-Marking:** 2015-3-31

**Place, Date of Issue:** Shenzhen, 2015-3-31

**Signature:** 

**Name of Authorized Signatory:** Mr.tan ChuanBin

**Position Held in Company:** Manager ,Technical Regulation

# Declaration of Conformity



**Manufacturer:** Shenzhen Mindray Bio-Medical Electronics Co., Ltd.  
Mindray Building, Keji 12th Road South, Hi-tech Industrial  
Park, Nanshan, Shenzhen, 518057, P. R. China

**EC-Representative:** Shanghai International Holding Corp. GmbH (Europe)  
Eiffestraße 80  
20537 Hamburg, Germany

**Product:** **Auto Hematology Analyzer**

**Model:** **BC-30s**  
Including reagents as following:  
**M-30D DILUENT**  
**M-30CFL LYSE**  
**PROBE CLEANSER**

**Classification:** The device not in IVDD annex II and not for self  
testing/performance evaluation

**Conformity Assessment Route:** IVDD Annex III(excluding Section 6)

**We herewith declare that the above mentioned products meet the provisions of the Directive 98/79/EC on In Vitro Diagnostic Medical Devices. All supporting documentations are retained under the premises of the manufacturer.**

**Standards Applied:**

List of (harmonized) standards for which documented evidence for compliance can be provided as attachment.

**Start of CE-Marking:** 2015-3-31

**Place, Date of Issue:** Shenzhen, 2015-3-31

**Signature:** \_\_\_\_\_ 

**Name of Authorized Signatory:** Mr.tan ChuanBin  
**Position Held in Company:** Manager ,Technical Regulation

Declaration of Conformity V 1.0

## Applied Standards List

**Product:** Auto Hematology Analyzer

**BC-20s, BC-30s**

Including reagents as following:

**M-30D DILUENT**

**M-30CFL LYSE**

**PROBE CLEANSER**

### Applied Standards:

EN ISO 18113-1:2011	In vitro diagnostic medical devices — Information supplied by the manufacturer (labelling) Part 1: Terms, definitions and general requirements
EN ISO 18113-2:2011	In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 2: In vitro diagnostic reagents for professional use
EN ISO 18113-3:2011	In vitro diagnostic medical devices — Information supplied by the manufacturer (labelling) Part 3: In vitro diagnostic instruments for professional use
EN ISO 15223-1:2012	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements
EN 13612: 2002	Performance evaluation of in vitro diagnostic medical devices
ISO 14971:2012	Medical devices – Application of risk management to medical devices
EN 61010-1:2001	Safety requirements for electrical equipment for measurement, control, and laboratory use Part 1: General requirement
EN 61010-2-081:2002+A1: 2003+A1: 2003	Safety requirements for electrical equipment for measurement, control and laboratory use - Part 2-081: Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes
EN 61010-2-101: 2002	Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment
IEC 61010-2-010: 2005	Safety requirements for electrical equipment for measurement, control and

## Declaration of Conformity V 1.0

	laboratory use - Part 2-010: Particular requirements for laboratory equipment for the heating of materials
EN 61326-1:2006	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements
EN 61326-2-6:2006	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-6: Particular requirements - In vitro diagnostic (IVD) medical equipment
EN 62304:2006	Medical device software- Software life cycle processes
EN 62366:2008	Medical devices — Application of usability engineering to medical devices
EN 13640: 2002	Stability testing of in vitro diagnostic medical devices
EN ISO13485:2012	Medical devices - Quality management systems - Requirements for regulatory purposes