

## WUXI MEDICAL INSTRUMENT FACTORY CO., LTD

Add: No.43 Xixin Road,ZhangJing,Xibei Town,Wuxi City,Jiangsu 214194,ChinaTel: 0086-510-83791431Fax:0086-510-83796580Web:www.wuximedical.comE-mail:wuximedical@vip.163.com

## EC Declaration of Conformity

Manufacturer:

Wuxi Medical Instrument Factory Co.,ltd Address: No.43 Xixin Road,Zhangjing, Xibei Town,Wuxi City,Jiangsu,214194 China. whose single Authorized Representative: Lotus NL B.V. Address: Koningin Julianaplein 10,1e Verd,2595AA, The Hague, Netherlands

We, the manufacturer, herewith declare that the products Mercury-free Thermometer Model: CR.W00 UMDNS-Code: 14028

meet the provisions of Directive 93/42/EEC which apply to them. The medical device has been assigned to class Im according to Annex IX of the

Directive.93/42/EEC. It bears the mark

## CE 1023

The product concerned has been manufactured under a quality management system according to Annex V of Directive 93/42/EEC.

Compliance of the designated product with the Directive 93/42/EEC has been assessed and certified by the Notified Body

ITC INSTITUTE FOR TESTING AND CERTIFICATION, Inc.

Zlin,Czech Republic-www.ltczlin cz Certificate No.: 19 0248 QS/NB Issue date: 22<sup>th</sup>.May.2019 Expiry date: 21<sup>st</sup> .May.2024

following the procedure relating to the EC Declaration of Conformity set out in Annex V of Directive 93/42/EEC.

This Declaration of conformity is valid in connection with the release document for the respective batch of produced devices.

The above mentioned declaration of conformity is exclusively under the responsibility of Company: Wuxi Medical Instrument Factory Co.,Itd

Address: No.43 Xixin Road,Zhangjing,, Xibei Town, Wuxi City, Jiang u 214194 China. WUXI,CHINA

Place, date Legally Function