

Section 2.0 HHS Compliance Compatibility List

Table 14-1 HHS Compliance Compatibility List

Product Category	Product Description	Part Number
Master X-ray Control - Console	Magic PC	5263305
Generator	JEDI Generator 50 KW	2212259-2
X-ray Tube	Toshiba E7843X HHS tube	5331186/ 5370023
Collimator	Collimator	5429399- ROHS/ 5189248
Table	Table	5653685/ 5669723
Wallstand	Wallstand	S39232NM/ S39232PC

Section 3.0 Customer Information

Fill in the Customer Information data sheet.

GE Service Personnel: Be sure that you return FDA2579. Failure to do so, except in case of a tube change, is a violation of FDA regulations and GE Healthcare policy.

Section 4.0 Test Equipment

Fill in the Test Equipment data sheet.

Table 14-2

Item	Manufacture	Model No.	Serial No.	Cal. Due Date mm/dd/ yyyy
Radiation Meter	MDH	2025		
Non-invasive KV Meter	Keithley	35080A		
Light Meter				
DVM				
HHS Test Kit	GEMS	46-303879G1 or 46-315694G1	N/A	N/A
mA Meter				