

Product Specification Sheet

VWR PLAIN BLENDER BAGS STERILE

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

Cat. No.	Description	Capacity (ml)	Height (mm)	Width (mm)	Thickness (mm)	Pk
129-0728	Plain Blender bags sterile	400	300±3	180±3	0.07±0.005	20bags of 25pcs
129-0729	Plain Blender bags sterile	400	300±3	180±3	0.07±0.005	10 bags of 50pcs
129-0731	Plain Blender bags sterile	3500	510±3	380±3	0.07±0.005	20 bags of 25pcs
129-0732	Plain Blender bags sterile	80	150±3	105±3	0.07±0.005	60 bags of 25pcs

Material: Natural LDPE (Low Density Polyethylene (PE)) + HDPE (High Density Polyethylene (PE))

Product Requirements

Absence of heavy material

Raw material is compliant with the EC/94/62 directive

Food contact

Raw material is compliant with regulation EC N°1935/2004

Based on ISO 11137

Sterilization of health care products -- Radiation, SAL10⁻³

Blender test

Bags have to be resistant for 3 minutes blending (frequency: 4 times / second)

Use temperature main body

-80°C / +75°C

Expiry: 3 years



VWR International bvba/sprl, Researchpark Haasrode 2020, Geldenaaksebaan 464, 3001 Leuven, Belgium
<http://www.vwr.com>

Product Specification Sheet

Physical Requirements

Flexible, good mechanical resistance.

Packing Requirements

Bags are zipper bag wrap and then heat-seal.
Color change label when sterile.

Disclaimer: VWR wants to state that these specifications can only be regarded as general information and will not discharge the user from his obligation to ensure the product is suitable for the intended use - If there is any doubt ask us.

Date of issue: **June 2013**