





AndraSnare®/LoopMaster Sochman Snare®

Gold plated atraumatic 3D nitinol snare /
Atraumatic 3D nitinol snare for foreign body retrieval

AndraSnare® & AndraSnare Micro®



- | | | | |
|---|--|---|---|
| 1 | Atraumatic nitinol loop with gold coating for better visualization in the X-rays | 6 | Luer connection |
| 2 | Pre-formed tip for better navigation | 7 | Insertion part for simple insertion of the loop into the catheter |
| 3 | X-ray marker at the shaft output | 8 | Large torquer for good operation |
| 4 | Shaft visible in X-rays | 9 | Continuous nitinol wire for 1:1 control of the loop |
| 5 | Bend protection for shaft reinforcement at the Luer | | |

Order Codes AndraSnare®


Order Code	Loop Ø in mm	Vessel Ø in mm	Loop/Guidewire Length in cm	Introducer Sheath Length in cm	Introducer Sheath Ø in Fr
AS-5	5	3 – 5	125	110	4
AS-10	10	6 – 10	125	110	4
AS-15	15	11 – 15	125	110	4
AS-20	20	16 – 20	125	110	4
AS-25	25	21 – 25	125	110	5
AS-25/65	25	21 – 25	65	50	5
AS-25/150	25	21 – 25	150	135	5
AS-30	30	26 – 30	125	110	5
AS-35	35	31 – 35	125	110	5

Order Codes AndraSnare Micro®

Order Code	Loop Ø in mm	Vessel Ø in mm	Loop/Guidewire Length in cm	Introducer Sheath Length in cm	Introducer Sheath Ø in Fr
ASM-2	2	1 – 2	175	150	2,3 – 3
ASM-4	4	3 – 4	175	150	2,3 – 3
ASM-7	7	5 – 7	175	150	2,3 – 3

Packaging Unit: 1


LoopMaster® Sochman Snare




- 1 Atraumatic distal end at the catching wire
- 2 Nitinol loop
- 3 Catching wire
- 4 X-ray marker
- 5 Flexible shaft

- 6 Bend protection at the Luer
- 7 Luer connector
- 8 Large torquers for simple control of the catching snare and the catching wire


Retrieval technique for foreign bodies (see graphics above)



- 1 Open the snare and place it around the foreign body to be removed.



- 2 By closing the snare, the foreign body is enclosed between the catching wire and the catching snare.



- 3 This makes it possible to remove the foreign body securely from the body.

Order Codes LoopMaster® Sochman Snare

Order Code	Loop Ø in mm	Loop/Guidewire Length in cm	Introducer Sheath Length in cm	Introducer Sheath Ø in Fr
ASL-25	25	130/150	110	8

Packaging Unit: 1

Andramed GmbH

Schiesswiesenstr. 18
72766 Reutlingen, Germany
Tel.: +49 (0) 7127 811 57 15
Fax: +49 (0) 7127 811 57 11
info@andramed.com
www.andramed.com

Made in Germany
Rev. 02/2019

