

HiMedia Laboratories Private Limited

23, Vadhani Industrial Estate, L.B.S. Marg, Mumbai - 400086

Website : www.himedialabs.com, Email : info@himedialabs.com

Certificate of Analysis , Quality and Conformity

Material Code : SD006	Material Name : Chloramphenicol C 30 mcg	Lot No : 0000417739
AR No.: 040000983630	Date of Report : 2020-01-04	Exp. Date : 2021-12

TEST	SPECIFICATIONS	RESULTS
Appearance 1 Appearance	Filter paper discs of 6 mm diameter with printed "C 30" on centre of each side of the disc.	Complies
Cultural response 1 Cultural response	Average diameter of zone of inhibition observed on Mueller Hinton Agar (M173) after 18 hours incubation at 35-37°C for standard cultures.	
Escherichia coli ATCC 25922 1 Zone of inhibition (mm)	21mm - 27mm	25mm
Staphylococcus aureus ATCC 25923 1 Zone of inhibition (mm)	19mm - 26mm	24mm
Staphylococcus aureus ATCC 29213 1 Zone of inhibition (mm)*	20mm - 28mm	27mm

The performance of this batch has been tested and standardized as per the CLSI as well as EUCAST Standards. *QC Limit as per EUCAST Standard


STATUS OF THE MATERIAL : APPROVED

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current HiMedia manual or product sheets.

This document has been produced electronically and is valid


Sujata Jagtap

Microbiologist/Sr.Executive Microbiologist


Ujwala M. Kokate

Asst./Dy/QC Manager


Dr. Santosh Kaul

Dy/QA Manager