Zibo Eastmed Healthcare Products Co., Ltd

Add: No.118 Huaguang Road, Zhangdian District, Zibo City 255000, Shandong China

TECHNICAL SPECIFICATION OF EXTENSION TUBE

REF NO.: DMES150A

-	Materials: A	BS + PVC
-	PVC tubing	150cm/120

PVC tubing 150cm/120cm/90cm/50cm/25cm

- Pressure 3BAR




- Latex-free

For single use

- ETO sterilized

0197

 $C \in$



		0171
No.	Designation	Materials
1	Luer cap	PE
2	Luer lock (male)	ABS
3	Tubing	Transparent PVC
4	Luer cap	PE
5	Luer lock (female)	ABS



PACKING:

- 1. Individual blister bag
- 2. Packing is adapted to the ethylene oxide sterilization. The cover mentions the necessary indications for content identification: designation, item code, manufacture date and expire date and lot number.
- 3. Packing details: single packed in blister, 25pcs/bag, 400pcs/ctn

QUALITY:

- 1. The product quality is guaranteed by inspections at all levels of manufacture, all our products are manufactured in 100,000 grade clean room, the final control is made at the level of packaging.
- 2. The packaged vaginal speculum is sterilized by ethylene oxide according to a validated method in the physical and microbiological field. The shelf life is 5 years from the manufacture date.
- 3. The product conforms to MDD 93/42/EEQ amended by 2007/47/EC and maintains a quality assurance system according to ISO 13485: 2016. 22ed Signature(s)