

# Declaration of Conformity

- for EU and UK -

**Manufacturer's Name:** SHIMADZU CORPORATION  
Analytical & Measuring Instruments Division  
**Address :** 1, Nishinokyo-Kuwabara-cho, Nakagyo-ku,  
Kyoto 604-8511, Japan

declares in sole responsibility that the following product

**Product Name** UV Visible Spectrophotometer  
**Model Name** UV-1900 / UV-1900i  
**P/N** Depend on configuration. See Appendix 1.

referred to in this declaration conforms with following legislations and standards

<b>EMC Directive 2014/30/EU</b> EN 61326-1:2013	<b>EMC Regulation S.I. 2016/1091</b> EN 61326-1:2013
<b>Low Voltage Directive 2014/35/EU</b> EN 61010-1:2010	<b>Electrical Equipment Regulation S.I. 2016/1101</b> EN 61010-1:2010
<b>RoHS Directive 2011/65/EU</b>	<b>RoHS Regulation S.I. 2012/3032</b>

Note 1) This declaration becomes invalid if technical or operational modifications are introduced without manufacturer's consent.

Note 2) This declaration is valid if this product is used alone or in combination with the accessories of this product which are mentioned in attached Appendix 1 or other instruments which fulfill with the requirement of mentioned legislations.

Note3) Authorised Representative in EU is as follows:

SHIMADZU EUROPA GmbH

Address :Albert-Hahn-Strasse 6-10,47269 Duisburg, F.R.Germany

Note4) Authorised Representative in UK is as follows:

SHIMADZU UK Limited

Address : Unit 1A, Mill Court, Featherstone Road Wolverton Mill South,  
Milton Keynes MK12 5RE, U.K

Kyoto, JAPAN 31, march, 2021

Place and date of issue



Signature

Toshiyuki Kawano

Name

General Manager of Quality Assurance Department

Analytical & Measuring Instruments Division

Title



## Appendix 1

**Product Name :** UV Visible Spectrophotometer

Model No.	P/N
UV-1900	207-23100-58
UV-1900i	207-25700-58

### Product Options

Model No.	P/N
SIPPER 160L	206-23790-51
SIPPER 160T	206-23790-52
SIPPER 160C	206-23790-53
SIPPER 160U	206-23790-54
SYRINGE SIPPER N	206-23890-51
SYRINGE SIPPER CN	206-23890-52
MULTI-CELL SAMPLE COMPARTMENT	206-69160-41
SOLENOID VALVE UNIT	204-06599-01
MMC-1600	206-23680-58
MMC-1600C	206-23690-58
DIS-8	206-24920-57/58
DIS-1200	206-24915-41
ANALOG OUTPUT I/F	206-25233-91
USB-CPS I/F	206-25234-91
USB-ASC I/F	206-25235-91
CPS-100	206-29500-58
TCC-100	206-29510-58
S-1700	206-23900-58
TMSPC-8	206-24350-58
Expanded memory	207-23119-41



**SHIMADZU**

新規・改訂内容一覧表

担当部署:分析計測品質保証部

改訂記号	改訂理由	改訂内容	作成者 日付	検討者 日付	承認者 日付
	制定		分析品証 18.4.12 村上	分析品証 18.4.12 金内	分析品証 18.4.12 逢坂
A	UV-1900i追加	UV-1900i と Expanded memory を追加 EMC Directive により IVD向け項目 (EN 61326-2-6:2006/2013) は削除	分析品証 村上 木村	分析品証 19.9.19 金内	分析品証 19.9.25 逢坂
B	誤記修正	CPS-100, TCC-100, TMSPC-8 の P/W E-58 に修正	分析品証 19.7.03 木村	分析品証 19.7.03 村上	分析品証 20.2.03 逢坂
C	UKCA対応	UK 指令, SHIMADZU UK 住所追記	分析品証 21.3.26 木村	分析品証 21.3.29 村上	分析品証 河野

配布先

名称 TITLE

Declaration of Conformity UV-1900

番号 No.

ZEAE-1099

SHEET

改定 REV