



HPi Verification Services

# EU-type Examination Certificate

This is to certify that the product listed below conforms to the requirements of the

## Pressure Equipment Directive 2014/68/EU

Annex III Module B(prod)

**Certificate Number**

HPiVS/P1057-048-I-01 rev.1

**Date of Issue**

15-Nov-2017

**Date of Expiry**

01-Oct-2027

**Designer**

**NÜVE SANAYİ MALZEMLERİ İMALAT VE TİCARET A.Ş.**

Saracalar Mah. Saracalar Kümeevleri No:4/2

Akyurt /Ankara /Turkey

**Description of Pressure Equipment**

Steam sterilizer Range : NC 23B, NC23S, NC32S, NC 100, NC 150, NC 150D, NC 40M, NC 90M, NC 300, NC 430, NC 430D, NC 570, NC 570D, NC 710, NC 710D;  
Steam generator Range: NC 100, NC 150, NC 300, NC 430, NC 570, NC 710.

**Drawing No**

N 8467, N 8466, N 8468, N 8475, N 8458, N 8460, N 8463, N 8462, N 8464, N 8461, N 8465, N 7118, N 8040, N 8036, N 6871, N 7987, N 8457, N 7539, N 8359, N 8456 Rev.0 dated 28 Sep 2017; N 8515 Rev.0 dated 19 Oct 2017; N 8497 Rev.0 dated 05 Oct 2017; N 8491 Rev.0 dated 26 Sep 2017

**Serial No**

n/a

**Design Pressure**

PS = 3 bar (for NC 23S, NC 23B, NC 32S PS = 2,6 bar)

**Design Temperature**

Max 144 °C (for NC 23S, NC 23B, NC 32S Max 140 °C)

**Standards Used**

EN 13445

**Report Reference**

HPiVS/P1057-048-DR01 to DR04

This Certificate is valid in any European Economic Area Member State.

This Certificate has been issued by HPi Verification Services Ltd which is a body notified to the European Commission according to the provisions of the Pressure Equipment Directive (Notified Body number 1521).

This Certificate is issued following the assessment of a representative sample of the Pressure Equipment detailed above in accordance with the provisions of the above regulations. The equipment must be subject to an appropriate conformity assessment module during manufacture prior to the CE Mark being affixed.

Check this certificate is genuine



Managing Director

Technical Manager



EU Notified Body No. 1521  
Company registered in England #7217086

© HPi Verification Services Ltd. 2017

Tel +44 1491 822818  
Fax +44 700 600 6831  
Email enquiries@eucertification.com  
www.eucertification.com

HPi Verification Services Ltd.  
The Manor House  
Howbery Park, Wallingford  
OX10 8BA, United Kingdom