

CODE:

ERP: 0202023

Aluminum Med. Oxygen Cylinder 2Lt

Oxygen therapy set

Portable Medical Oxygen Cylinder, made of Aluminum 2Lt, with 25E Valve. Flow Regulator is not included.



CHARACTERISTICS

Capacity	2Lt
Max Allowable Pressure	200 bar
Test Pressure	300 bar
Volume of Equipment	2lt
Cylinder Height	360±3.6 mm
Cylinder Diameter	111±0.8325 mm
Cylinder Neck Thread	25E
Valve Material	Brass
Paint Finish	White

Issue Date: 26/04/2021

Information may change

FACTORY: AKROTIRI, CHANIA, P.C. 73100, TEL. +30 2821063222, FAX: +30 28210 66260

CENTRAL EXHIBITION: 96-98 M. MPOTSARI str. P.C. 73136, CHANIA, GREECE, TEL. +30 28210 91561, FAX +30 28210 91644

ATHENS DISTRIBUTION CENTER: SXINEZA- ASPROPIRGOS, P.C. 19300, TEL. +30 210 4614006, FAX +30 210 5595986

N.GREECE DISTRIB. CENTER: 8 ARKADIOU str. KALOCHORI - THESSALONIKI, P.C. 57009, TEL. +30 2310 785668, FAX +30 2310 574793

CENT. & EAST CRETE DISTRIB. CENTER: P' str. INDUSTRIAL ZONE OF HERAKLION - CRETE, P.C. 71601, TEL./FAX +30 2810 333151

www.mobiak.com

CODE:

ERP: 0202018

Aluminum Med. Oxygen Cylinder 5Lt

Oxygen therapy set

Portable Medical Oxygen Cylinder, made of Aluminum 5Lt, with 25E Valve. Flow Regulator is not included.



CHARACTERISTICS

Capacity	5Lt
Max Allowable Pressure	200 bar
Test Pressure	300 bar
Volume of Equipment	5lt
Cylinder Height	525±5.25 mm
Cylinder Diameter	140±1.05 mm
Cylinder Neck Thread	25E
Valve Material	Brass
Paint Finish	White

Issue Date: 26/04/2021

Information may change

FACTORY: AKROTIRI, CHANIA, P.C. 73100, TEL. +30 2821063222, FAX: +30 28210 66260

CENTRAL EXHIBITION: 96-98 M. MPOTSARI str. P.C. 73136, CHANIA, GREECE, TEL. +30 28210 91561, FAX +30 28210 91644

ATHENS DISTRIBUTION CENTER: SXINEZA- ASPROPIRGOS, P.C. 19300, TEL. +30 210 4614006, FAX +30 210 5595986

N.GREECE DISTRIB. CENTER: 8 ARKADIOU str. KALOCHORI - THESSALONIKI, P.C. 57009, TEL. +30 2310 785668, FAX +30 2310 574793

CENT. & EAST CRETE DISTRIB. CENTER: P' str. INDUSTRIAL ZONE OF HERAKLION - CRETE, P.C. 71601, TEL./FAX +30 2810 333151

www.mobiak.com

Oxygen Therapy Set - Aphrodite




Oxygen Therapy Set with Regulator, Aphrodite



	Volume (lt)	m³	Pressure (bar)	Height (cm)	Diameter (cm)	Weight (kg)	Material
code 0808907	2	0.4	200	46	11	4.79	Aluminium
code 0808908	3	0.6	200	60	11	6.49	Aluminium
code 0808909	5	1.0	200	63	14	9.66	Aluminium
code 0808920	10	2.0	200	99	13.5	16.95	Steel
code 0808921	16	3.2	200	83	19	27.17	Steel
code 0808922	50	10.0	200	163	24	70.2	Steel

- Ready-to use Oxygen Therapy Sets with Valve 25E, Aphrodite Regulator and Aluminium (2-5Lt) or Steel (10-50Lt) Cylinder
- Flow Control Settings: 0 - 0.5 - 1 - 2 - 3 - 4 - 5 - 6 - 7 - 9 - 12 - 15Lt
- Flat Base
- Built-in Carrying Handle for 2Lt, 3Lt & 5Lt Cylinder
- Metal Cap for 10Lt, 16Lt & 50Lt Cylinder



Aluminium Oxygen Therapy Set Without Regulator

Ready-to use Oxygen Therapy Sets with Valve 25E and Aluminium Cylinder (2-5Lt)

Aluminium Cylinder Features

Code	Capacity	Operating Pressure	Dimensions (cm) (Height / Diameter)	Dimensions (Neck)	Weight (O ₂ -filled)
0202023	2Lt	200 bar	46 / 11	25E	4.3Kg
0202024	3Lt	200 bar	60 / 11	25E	6.0Kg
0202018	5Lt	200 bar	63 / 14	25E	9.18Kg



Steel Oxygen Therapy Set Without Regulator

Ready-to use Oxygen Therapy Sets with 25E Regulator and Steel Cylinder (10-50Lt)

Steel Cylinder Features

Code	Capacity	Operating Pressure	Dimensions (cm) (Height / Diameter)	Dimensions (Neck)	Weight (O ₂ -filled)
0202013	10Lt	200 bar	99 / 13,5	25E	16.6Kg
0202030	16Lt	200 bar	83 / 19	25E	26.6Kg
0202009	50Lt	200 bar	163 / 12	25E	69.6Kg



Regulators - Regulator Accessories

Oxygen Flow Regulator 1 - 25Lt/Min MediVital



	Depth	Height (cm)	Weight (gr)	Filling Inlet Dimensions	Pressure Range
cod. 0810350	11.8cm	16cm	1150gr	SI22.91*1.814 NF E 29-650-F	600 / 1200 mbar
Operating Temperature Range			Transport Temperature Range		
-20°C / +65°C			-40°C / +70°C		

- Compact and Lightweight Design, less than 1,150 gr
- Flow Control Settings: 0 - 1 - 2 - 3 - 4 - 5 - 6 - 7 - 9 - 12 - 15 - 25Lt
- 2-year Warranty

⚠ We Recommend Cylinder Assembly & Filling in Specially Certified Equipped Facilities for Utmost Reliability which Possess the Appropriate Adaptor Ref.: 0810542

► For Use with a Humidifier it is Necessary to Purchase Codes: 0810351, 0810352 (see p. 8)

Aphrodite Oxygen Flow Regulator

code 0807916



- Lightweight and Easy to Use
- 25E Valve Required
- Flow Control Settings:
0 - 0.5 - 1 - 2 - 3 - 4 - 5 - 6 - 7 - 9 - 12 - 15Lt/min
- Gauge with 0-300 bar Range
- 1-year Warranty



Hercules Oxygen Flow Regulator

code 0807915



- 25E Valve Required
- Flow Control Settings: 0-15Lt/min
- Gauge with 0-300 bar Range
- 1-year Warranty

Outlet Nozzle

code 0810162



- Aphrodite (Ref.: 0807916)
& Hercules (Ref.: 0807915) Outlet Nozzle

Rollators

"Platinum" Rollator

Ref: 0803160



2 Years
Warranty

New Product

Adjustable Height



- Frame: Aluminum
- Silver
- Safety Brakes on Handles that can Lock
- Height-Adjustable Handles
- With Seat and Basket
- With Walking Stick Holder
- Folding for Easy Storage
- Suitable for Indoor and Outdoor Use



82cm


With Walking
Stick Holder



67cm

25cm

47cm	60cm	64cm	52cm	81-94cm	8Kg	110Kg





Industrie Service

ZERTIFIKAT Certificate

Baumusterzulassung nach ADR/RID für nicht-UN-Flaschen
Type approval according to ADR/RID for non-UN cylinders

Zertifikat-Nr.: 0036-062-19 **Rev. 0**
Certificate No.:

Gültigkeit / Validity: 2029-12-18

Name und Anschrift des Antragstellers: Zhejiang Winner Fire Fighting Equipment Co., Ltd.
Name and Postal Address of Applicant: No. 260, Zhendong Road, Xinhuang, Fengqiao Town,
Nanhu District, Jiaxing City, Zhejiang Province, P.R. China

Hiermit wird bescheinigt, dass der unten spezifizierte Typ die technischen Anforderungen des ADR/RID 2019 und der Richtlinie 2010/35/EU (TPED) erfüllt. / We herewith certify that the type specified below meets the technical requirements of ADR/RID 2019 and the directive 2010/35/EU (TPED).

Prüfbericht Nr.: P-PRC-19-10-601539-1007, 2019-11-03
Test Report No.: TÜV SÜD Industrie Service GmbH

Typ, Norm, Zeichnung, Revision, Datum **Flasche / Cylinder** **EN ISO 9809-1:2010**
Type, Standard, Drawing, Revision, Date JXM315M/01-69, Rev. 0, 2019.8.23
(Anlage / Enclosure)

Fertigungsstätte **Zhejiang Winner Fire Fighting Equipment Co., Ltd.**
Manufacturing Plant No. 260, Zhendong Road, Xinhuang, Fengqiao Town,
Nanhu District, Jiaxing City, Zhejiang Province, P.R. China

Anmerkung: Cylinder Code "QTS" nach EN ISO 11114-1. Geeignet für UN 1066 Stickstoff, verdichtet und UN 1013 Kohlendioxid. Die weitere Gasverträglichkeit wurde nicht überprüft.

Remark: Cylinder code "QTS" according EN ISO 11114-1. Compatible with UN 1066 Nitrogen, compressed and UN 1013 Carbon Dioxide. Any further gas compatibility was not checked.

(Ort, Datum)
(Location, Date)

München, 2019-12-19



TÜV SÜD Industrie Service GmbH
Inspection Body Xa / Notified Body 0036

Klaus J. Meier

TÜV SÜD Industrie Service GmbH
Anlagensicherheit
Westendstraße 199
D-80686 München
Tel.: +49 (0) 89 / 5791-1891
Fax: +49 (0) 89 / 5791-1810
www.tuev-sued.de



Deutsche
Akkreditierungsstelle
D-IS-14153-02-03

ΠΙΣΤΟΠΟΙΗΤΙΚΟ

CERTIFICATE

ΕΓΚΡΙΣΗΣ ΣΥΣΤΗΜΑΤΟΣ ΠΟΙΟΤΗΤΑΣ (IS – ΑΡΧΙΚΟΣ ΕΛΕΓΧΟΣ ΚΑΙ ΔΟΚΙΜΕΣ) ΣΥΜΦΩΝΑ ΜΕ ΤΗΝ ΟΔΗΓΙΑ 2010/35/ΕΕ
QUALITY SYSTEM APPROVAL (IS – INITIAL INSPECTION AND TESTS)
ACCORDING TO DIRECTIVE 2010/35/EU

Αριθμός Πιστοποιητικού: MIRTEC1-00-1400CER12.2052001001
Certificate No:

Αριθμός Κοινοποίησης του Φορέα Πιστοποίησης: 0437
Notification No of the Certification body:

Ημερομηνία έκδοσης: 04.05.2020
Date of issue:

Αριθμός Έκθεσης: RPT12.2052001001
Test Report No: PE-R-2105/2/19

Όνομα και Διεύθυνση Κατασκευαστή
& Εργοστασίου Παραγωγής:
Name and Address of the Manufacturer
& Manufacturing Plant:

Zhejiang Winner Fire Fighting Equipment Co. Ltd./
Jiaxing Fuxin Fire Industrial Co. Ltd.
No 260, Zhendong Road, Xinhua Fengqiao Town,
Nanhu District, Jiaxing City, Zhejiang Province, 314008 P.R. China

Όνομα και Διεύθυνση Αιτούντος:
Name and Address of the Applicant:

MOBIAK A.E
Markou Mpotsari 96-98, 73136 Chania, GREECE

Τύπος Προϊόντος: Seamless Aluminum Gas Cylinders 1.5L, 2L, 3L and 5L, PT 300 bar, drawings No.:
Product Type: 1.5L: JX30-ML02/01-1Rev.2, 2L: JX30-ML02/02-1Rev.1 and JXML300G-1Rev.0,
3L: JX30-ML02/03-1Rev.2 & JXML300G-2Rev.0, 5L: JX30-ML04/05-1Rev.1 & JX30-ML04/05-1Rev.2

Εφαρμοζόμενα πρότυπα: EN 1975:1999 + A1:2003
Applied Standards:

Αριθμός Πιστοποιητικού Έγκρισης Τύπου: PE-C-128ARev.1/PED/B/01/11, PE-C-115ARev.2/PED/B/01/13
Type Approval Certificate No: PE-C-166ARev.2/PED/B/01/13, PE-C-141ARev.2/PED/B/01/12 MIRTEC S.A.

Η EBETAM A.E. πιστοποιεί ότι, στις εγκαταστάσεις δοκιμών του ως άνω Κατασκευαστή, έχει συσταθεί μια υπηρεσία επιθεώρησης εντός της επιχείρησης (IS), και εφαρμόζεται εγκεκριμένο Σύστημα Ποιότητας για την κατασκευή, την επίβλεψη της κατασκευής και τον αρχικό έλεγχο και τις δοκιμές των συγκεκριμένων τύπων προϊόντων που αναφέρονται παραπάνω, σύμφωνα με το μέρος 1.8.7.6 της συνθήκης ADR 2019.

MIRTEC S.A. certifies that, at the above stated testing facilities, the Manufacturer implements an in-house inspection service (IS) with an approved Quality System for the manufacture, supervision of manufacture and initial inspection and tests of the specific product types mentioned above, according to part 1.8.7.6 of ADR 2019 agreement.

Ημερομηνία Λήξης Πιστοποιητικού: 20.06.2021
Expiry Date of the Certificate:

ΔΗΛΩΣΗ- DECLARATION

Ο Κατασκευαστής ή ο αντιπρόσωπός του εξουσιοδοτείται –μετά την εξέταση των προαπαιτήσεων- να επιθέσει στα συγκεκριμένα προϊόντα τη σήμανση π συνοδευόμενη από τον αριθμό κοινοποίησης της EBETAM ΑΕ, ως κατωτέρω.
The Manufacturer or his agent is authorized-after examination of the prerequisites-to affix to the above mentioned products the π -mark accompanied by MIRTEC's notification number, as illustrated below.

π 0437

Το πιστοποιητικό ισχύει υπό τον όρο ότι το Σύστημα Ποιότητας του Κατασκευαστή υπόκειται σε επιτήρηση από τον Κοινοποιημένο Οργανισμό. Δύο επιθεωρήσεις κατ' ελάχιστον πρέπει να διενεργούνται σε διάρκεια 12 μηνών.
This certificate remains valid provided that the Manufacturer's Quality System is subjected to surveillance by the Notified Body. A minimum of two audits shall be carried out in a 12 month period.

Για την EBETAM ΑΕ
For MIRTEC SA

Για το Τμήμα Αξιολόγησης
For the Evaluation Dpt



A. Stamou
Technical Manager

K. Diamanti
Mechanical Engineer, PhD, IWE