

Traksiyon Atel Seti (ET-320)

Traction Splint Set

- Bacak kemiği kırıklarında, bacağı gergiye alarak kırığın birbirine temasını keserek sinir zedelenmesini önlemekle birlikte ağrıyı azaltmaktadır.
- Uzunluğu ayarlanabilmektedir.
- Her iki atelde 4 adet bacak bağlama aparatları bulunmaktadır.
- Metal malzemeden mamul gergi makarası bulunmaktadır.

- In cases of leg fractures, it prevents nerve trauma by cutting touch of the fractures one another by taking the leg into splint and reduces pain.
- Length can be adjusted.
- There are 4 fixing apparatus for leg. It has stretching roller made of metal material.
- Including tensioner pulley made of metal.



Şişme Atel Seti (ET-190)

Air Splint Set

- Özel taşıma çantası.
- Tüm kol, tüm bacak, yarım kol, yarım bacak, el - el bileği, ayak - ayak bileği olmak üzere 6 parçadan oluşur.
- Pompa veya nefes yardımı ile kolay ve hızlı uygulama imkanı sağlar.

- Special carriage bag.
- Consists of 6 parts as whole arm, whole leg, half arm, half leg, foot and ankle.
- Ensure fast application possibility by help of a pump or blowing.



DECLARATION OF CONFORMITY



EMS MOBİL SİSTEMLER A.Ş..

1. Organize Sanayi Bölgesi Kırım Hanlığı Caddesi No: 9

Sincan-Ankara /TURKEY

Tel : +90 312 267 18 67 - Faks : +90 312 267 18 66

Product : Traction Splint Set (Adult/Pediatric)

Product Code : ET-320

Classification : Class I - Rule I

We herewith declare with this document that compliance evaluation was applied upon the above mentioned products as per the Council Directive 93/42/EEC Medical Devices Regulation Annex 7 for Medical Device Directive

General Applicable Directives

Medical Device Directive council directive 93/42/EEC of June 1993 concerning medical devices. (MDD 93/42/EEC) Compliance Declaration evaluation was applied as per Annex 7.

Standarts

- | | |
|----------------------|---|
| TS EN 980:2011 | Symbols for use in the labeling of medical devices |
| TS EN 1041+A1:2015 | Information supplied by the manufacturer of medical devices |
| TS EN ISO 14971:2013 | Medical devices - Application of risk management to medical devices |

Signature

EMS MOBİL SİSTEMLER A.Ş.
1. Organize Sanayi Bölgesi
Kırımhanlığı Caddesi No: 9 Sincan/ANKARA
Tel: 0312 267 18 67 - Fax: 0312 267 18 66
Mersis No: 0475005500500012
Sincan Vergi Dairesi: 475 005 5005

Name : Mevlüt Şahin
Position : General Manager

Date : 15.05.2016