

EC Declaration of Conformity

We, TaiDoc Technology Corporation

B1-7F, No.127, Wugong 2nd Road, 24888 Wugu Dist, New Taipei City, TAIWAN declare under our sole responsibility that the product

Product Name

: Blood Glucose Monitoring System

Product Model

: TD-4116

Classification

: 98/79/EC (IVDD), Annex II, List B

Conformity Assessment Route

: 98/79/EC (IVD), Annex IV excluding section 4&6

EC Certificate Number

: V1 052126 0069 Rev.02

European Representative

: MedNet EC-REP GmbH

Borkstraße 10, 48163 Münster, Germany

Notified Body (CE0123)

: TÜV SÜD Product Service GmbH

Ridlerstraße 65, 80339 München, Germany

GMDN code

: 62537

to which this declaration relates is in conformity with the following standard(s) or other normative document(s):

ISO 13485:2016	Medical devices - Quality management systems - Requirements for regulatory purposes
EN ISO 14971:2012	Medical devices - Application of risk management to medical devices
EN ISO 15197:2015	In vitro diagnostic test systems —Requirements for blood-glucose monitoring systems for self-testing in managing diabetes mellitus
EN 15223-1:2016	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied. Part 1: General requirements
EN ISO 18113-1:2011	In vitro diagnostic medical devices. Information supplied by the manufacturer (labelling). In vitro diagnostic instruments for professional use
EN ISO 18113-2;2011	In vitro diagnostic medical Devices. Information supplied by the manufacturer (labelling). In vitro diagnostic reagents for professional use
EN ISO 18113-3:2011	In vitro diagnostic medical devices. Information supplied by the manufacturer (labelling). In vitro diagnostic instruments for professional use
EN ISO 18113-4:2011	In vitro diagnostic medical Devices. Information supplied by the manufacturer (labelling). In vitro diagnostic reagents for self-testing



泰博科技股份有限公司 新北市24888五股區五工二路127號81-7權 B1-7F., No. 127, Wugong 2nd Rd., Wugu Dist. New Taipei City 24888, Taiwan

Tel: +886-2-6625-8188 Fax: +886-2-6625-0288

www.taidoc.com	The state of the second st
EN ISO 18113-5:2011	In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 5: In vitro diagnostic instruments for self-testing
EN ISO 23640:2015	In vitro diagnostic medical devices. Evaluation of stability of in vitro diagnostic reagents
EN ISO 20417:2021	Medical device – Information to be supplied by the manufacturer.
EN 13532:2002	General requirements for in vitro diagnostic medical devices for self-testing
EN 13612:2002	Performance evaluation of in vitro diagnostic medical devices
EN 61010-1:2010	Safety requirements for electrical equipment for measurement, control and laboratory use. General requirements.
EN 61010-2-101:2015	Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment
EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use. EMC requirements. General requirements
EN 61326-2-6 :2013	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-6: Particular requirements - In vitro diagnostic (IVD) medical equipment
EN 62304:2006+A1:2015	Medical device software Software life cycle processes
ISO/IEC 12207:2008	Systems and software engineering- Software life cycle processes
EN 62366-1:2015	Medical devices Application of usability engineering to medical

2022, 1, 3.

Date of Issue

Im Jan

Jim Jan Management Representative