

**Honeywell**  
CERTIFICATE OF ANALYSIS

Honeywell Specialty Chemicals Seelze GmbH  
Wunstorferstrasse 40  
Seelze, 30926  
Telefon: +49 5137 999-0  
[lab.honeywell.com](http://lab.honeywell.com)

Brand: Honeywell Fluka  
Product: 35375  
Lot No: N2960  
Production date: 23.Oct.2023  
Expiry Date: 07.Oct.2026

Silver nitrate solution 0,1 mol/l, volumetric solution, Reag. Ph. Eur.

Parameter	Specification	Units	Result	Units
factor	0.999-1.001		1.000	
BAM/EMPA ref.-material	effektiv		71387	
BAM/EMPA Certificate no.	effektiv		U3-0319	
Traceable to NIST SRM	effektiv		919b	

Produced using quality grade "puriss. p.a." tested raw materials.

Test method: Argentometric titration related to  
EMPA/BAM referencematerial Sodium chloride

The factor (titer) was measured at 20°C.

BAM: Federal Institute for Materials Research and Testing  
Germany  
EMPA: Swiss Federal Laboratories for Materials Testing and Research  
NIST: National Institute of Standards and Technology  
USA

QC Release Date: 13.Nov.2023

The minimum shelf life is based on the current knowledge and holds only for proper storage conditions in the originally closed  
flasks/ packages.

We herewith confirm that the delivery is effected according to the technical delivery conditions agreed.  
Particular properties of the products or the suitability for a particular area of application are not assured.  
We guarantee a proper quality within our General Conditions of Sales.



Honeywell  
Quality Management

This document was produced electronically and is valid without a signature