

ACON BIOTECH (HANGZHOU) CO.,LTD

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Toxoplasma IgM EIA Test Kit

PURCHASE NO.: 37740068A

BATCH NO.: 2003699

EXPIRATION DATE: 2021/08/20

QUANTITY: 20kits

SPECIFICATION		STANDARD	CONCLUSION
Appearance		No obvious sediments or floccules in liquid component and color meets requirement	Complies
Blank		OD<0.050 (450nm/630nm) or OD<0.100 (450nm)	Complies
Sensitivity Standard	ToxoM-H	OD>0.500	Complies
	ToxoM-M	OD>0.300	Complies
	ToxoM-L	OD>0.200	Complies
Compliance of Negative Standard		$\geq 19/20$	Complies
Negative Control		OD<0.200	Complies
Cutoff Control		OD>0.200	Complies
Positive Control		OD>0.500	Complies
Precision		CV<15%,OD> 1.1* Cut-Off OD	Complies

Quality Department: Hedy zhang

Date: Sep. 23, 2020